

PIONEER 4.2 CARE

4 MOTORLUYOÇUNBAKIMHASTAKARYOLASI

MOTORSMODELINTENSIVECAREPATIENTBED



TEKNİK ÖZELLİKLER

- Elektronik motoryardımlı ile yükseklik, başucu ve ayakucu hareketi.
- Trendelenburg, terstrendelenburg, fowler ve vasküler pozisyonları.
- Hasta el kumandası.
- Hemşire kontrol ünitesi.
- Elektronik CPR tuşu.
- Otomatik kayma sistemine sahip başucu
- Auto contour.
- Hijyenik dizayna sahip plastik (PP), kilitlenebilen ve gizlenebilen yan korkuluklar.
- Kolaysökülmüş takılabilen ABS plastik muhafazalı, profil şase.
- Hijyenik dizayna sahip plastik (PP) kolaysökülmüş takılabilen başlıklar.
- Antistatiktelerler. (merkezi ve doğrusal kilit)
- Elektrostatik boyalı metal aksam.
- Elektrostatik termal sprey metal çerçeve (Opsiyonel).
- Yüksekliği ayarlanabilir paslanmaz çelik, serum askısı.
- Diz hareketlerini kolaylayabilme özelliği.
- Plastik çarpma tamponları.
- 24V akü ile kesintisiz çalışabilme.
- Sırt kısmında mekanik CPR.
- Başucu ve ayakucu köşelerinde aksesuar yuvası.
- Sırt aç göstergesi.
- Karyola aç göstergesi.
- Tek tuşla cardiac chair (Sandalye) pozisyonu.
- Tek tuşla shock pozisyonu.
- Tek tuşla semi fowler pozisyonu.
- Tek tuşla yataktan çıkış pozisyonu.
- Tek tuşla muayene pozisyonu.
- İstem dışı kullanımları önlemek için hemşire kumandasında aktivasyon tuşu.
- Yataklar platformuzatması 18cm. (Opsiyonel)
- Monitör tutucu. (Opsiyonel)
- Kompak laminant sırt ve X-Ray kaset yuvası (Opsiyonel).
- Oksijende posututucusu.
- Keten tutucu.
- IPX4 elektronik sistem.

TECHNICAL DETAILS

- Electrical operated backrest, height and grest adjustment.
- Electrical operated trendelenburg, reverse trendelenburg, fowler and vascular position.
- Patient hand set.
- Nurse control unit.
- Auto-CPR.
- Retracting backrest for advanced comfort, less pressure.
- Auto contour.
- Hygienic design (PP), lockable and dtuck away siderails.
- Profile frame with easy removable ABS plastic covers.
- Removable head and foot ends.
- Central and directional lockable castors.
- Electrostatic painted metal frame.
- Electronic thermal spray metal frame. (Optional)
- Height adjustable stainless steel IV pole.
- Easy adjustable knee break position.
- Plastic crash bumpers.
- Battery Backup.
- Dual sided manual CPR at backrest.
- Comprehensive range of accessories socket at corners.
- Angle gauge for back.
- Angle gauge for bed.
- Cardiac chair position with one button.
- Shock position with one button.
- Semi-Fowler position with one button.
- Bed exit position with one button.
- Examination position with one button.
- Activation key at remote control to prevent any undesired positions.
- Bed extension 18 cm (Optional)
- Monitor holder. (Optional)
- X-Ray translucent back rest and cassette carrier
- Oxygen tank holder.
- Linen holder.
- IPX4 standard electronic system.

HEMŞİRE KONTROL ÜNİTESİ
NURSE CONTROL UNIT



HASTA EL KUMANDASI
PATIENT HANDSET



KORKULUK MEKANİZMASI
SADE RAIL MECHANISM



MERKEZİ FREN SİSTEMİ
CENTRAL LOCK



EBAT / DIMENSIONS

- Uzunluk / Overall Length : 212 cm
- En / Overall Width : 102 cm
- Yükseklik (Şilteli) / Height Range : 40-82 cm
- Uyuyan yüzey uzunluğu / Sleeping surface Length : 195 cm
- Uyuyan yüzey genişliği / Sleeping surface Width : 86 cm
- Trendelenburg / Trendelenburg : 0° - 16°
- Ters Trendelenburg / Reverse Trendelenburg : 0° - 16°
- Ayak Ucu Kısmı (Maximum) / Legrest Angle (Max.) : 45°
- Sırt Kısmı (Maximum) / Backrest Angle (Max.) : 70°
- Teker Çapı / Castor Diameter : 15 cm
- Yatak Altı Boşluğu / Under Bed Clearance : 15 cm
- Güvenli Çalışma Kapasitesi / Safe Working Load : 230 kg
- (Aksesuarlar dahil) / (With Accessories)

ELEKTRONİKLER / ELECTRICAL DATA

- Güç Girişi / Power In : 1.5 A max. 230VAC 50-60 Hz
- Elektrik Çarpmasına Karşı Koruma / Electrical Shock Protection : Class I, Type B
- Su Sızmasına Karşı Koruma / Liquid Ingress Protection : IPX4 - IPX6
- Akü / Batteries : 2x 12V Series, Connected, Sealed, Rechargeable, Lead/Acid gel, 1.2 amp hrs.



Sırt Hareketi
Backrest



Ayak Hareketi
Legrest



Sırt-Ayak Hareketi
Auto Contour



Diz Kırma
Knee Break



Yükseklik
High/Low




Düz Trendelenburg
Trendelenburg



Ters Trendelenburg
Reverse Trendelenburg



Kardiyak
Cardiac Chair



PIONEER 4.2 CARE

4 MOTORLU YOÇUNBAKIM HASTA KARYOLASI
4 MOTORS MODEL INTENSIVE CARE PATIENT BED

PIONEER 4.2 CARE



Low Bed



Sirt Hareketi / Backrest



Trendelenburg / Trendelenburg



Kardiyak / Cardiac Chair



MEYSA TIBBİ CİHAZLAR SAN. ve TİC. A.Ş.

O. S. B. 16.Cad. No :73 Melikgazi / KAYSERİ / TÜRKİYE

+90 352 503 16 60

www.meissa.com.tr

+90 352 503 16 61

info@meissa.com.tr

DECLARATION OF CONFORMITY

As MEYSA TIBBİ CİHAZLAR SAN. VE TİC A.Ş., we declare that the following products are in compliance with the provisions of the Council Directive MDR- 2017/745 for medical devices. All supporting documents are stored within the manufacturer. We declare that we will fulfill the obligations imposed by the quality system and that we will maintain the quality system adequately and effectively.

Manufacturer: MEYSA TIBBİ CİHAZLAR SANAYİ VE TİCARET A.Ş.

SRN NUMBER: TR-MF-000024429, MEYSA TIBBİ CİHAZLAR SAN. VE TİC. A.Ş.

**Manufacturer Address: Tacettin veli mah. Halit Narin Cad. Bahadır Plaza apt. no:7/11
MELİKGAZI/KAYSERİ/TURKEY**

Product information: Check the Attached Product List.

Directives: MDR(EU) 2017/745- Medical Device Regulation

Classification: Class-I Non Sterile Products - Annex VIII

Conformity Assessment: Annex IV Declaration of conformity (Annex II & Annex III)

ŞANSEL SAĞIR


General Manager

30.05.2022

KAYSERİ/ TÜRKİYE



Primary Product Its number	Name of the product	TEMEL UDI-DI NUMBER
86816723878433	MIP 100 INFUSION HOLDER	868358388348R4
86816723878426	MIP 101 INFUSION HOLDER	868358388348R4
86816723878594	IMPULSE 01 PATIENT TRANSFER TROLLEY	868358388348R4
86816723878907	DONEE 01 BLOOD DONATION CHAIR	868358388348R4
86816723878020	NİCE 2 BABY COT	868358388348R4
86816723878068	MOTHERCARE 3 GYNECOLOGICAL EXAMINATION COUCH	868358388348R4
86816723878013	NİCE 1 BABY COT	868358388348R4
86816723878037	SUNRISE DELIVERY BED	868358388348R4
86816721878039	PIONEER 4.3 PLUS 4 MOTORS INTENSIVE CARE PATIENT BED	868358388348R4
86816721878336	PIONEER CLW 4 MOTORS COLUMN INTENSIVE CARE PATIENT BED WITH WEIGHT SCALE	868358388348R4
86816721878237	PIONEER 3 3 MOTORS ELECTRICAL PATIENT BED	868358388348R4
86816721878282	PIONEER CL 4 MOTORS COLUMN MODEL INTENSIVE CARE PATIENT BED	868358388348R4
86816723890077	GRANDE CL 4 MOTORS INTENSIVE CARE PATIENT BED	868358388348R4
86816723891074	GRANDE 4 4 MOTORS INTENSIVE CARE PATIENT BED	868358388348R4
86816723891876	GRANDE 3 3 MOTORS ELECTRICAL PATIENT BED	868358388348R4
86816723891371	GRANDE 3.1 Trend 3 MOTORS ELECTRICAL PATIENT BED	868358388348R4
86816723891173	PIONEER 4.1 Trend 4 MOTORS ELECTRICAL PATIENT BED	868358388348R4
86816723889170	PIONEER 4.2 Care 4 MOTORS INTENSIVE CARE PATIENT BED	868358388348R4
86816723989177	PIONEER 2 2 MOTORS ELECTRICAL PATIENT BED	868358388348R4
86816723989986	PIONEER 4 4 MOTORS ELECTRICAL PATIENT BED	868358388348R4
86816723981003	PEDI03 3 MOTORS PEDIATRIC ELECTRICAL PATIENT BED	868358388348R4
86816723981010	PRIMITIVE 2 2 ADJUSTMENTS MANUEL PATIENT CARE BED	868358388348R4
86816723981027	PRIMITIVE 3 3 ADJUSTMENTS MANUEL PATIENT CARE BED	868358388348R4
86816723981034	DUO 2 2 MOTORS ELECTRICAL PATIENT BED	868358388348R4
86816723981041	ELİTE 2 2 MOTORS ELECTRICAL PATIENT BED	868358388348R4
86816723981058	ELİTE CL 4 MOTORS INTENSIVE CARE PATIENT BED	868358388348R4
86816723981065	ELİT 3 3 MOTORS ELECTRICAL PATIENT BED	868358388348R4
86816723981072	ELİT 4 4 MOTORS INTENSIVE CARE PATIENT BED	868358388348R4
86816723981089	ELİT 4.3 Plus 4 MOTORS INTENSIVE CARE PATIENT BED	868358388348R4
86816723981119	ELİT JUNIOR 3 MOTORS ELECTRICAL PEDIATRİC PATIENT CARE BED	868358388348R4
86816723981225	JUNIOR PEDIATRİC PATIENT BED	868358388348R4
86816723981232	JUİÖR 3 3 MOTORS PEDIATRİC PATIENT CARE BED	868358388348R4
86816723981249	JUNIOR 4 4 MOTORS ELECTRICAL PEDIATRIC INTENSIVE PATIENT CARE BED	868358388348R4
86816723981256	DİAGNO 01 EXAMINATION CHOUCH	868358388348R4
86816723981263	DİAGNO 02 EXAMINATION CHOUCH	868358388348R4
86816723981287	DİAGNO 03 EXAMINATION CHOUCH	868358388348R4
86816726554327	MOTHER CARE 2 2 MOTORS GYNECOLOGICAL EXAMINATION COUCH	868358388348R4
86816726554358	DYNAMIC 2.0 GENERAL PURPOSE PATIENT STRETCHER	868358388348R4
86816726554907	MOVEE 2.0 PATIENT IMMEDIATE TREATMENT STRETCHER	868358388348R4
86816726554983	MOVEE 2.1 PATIENT IMMEDIATE TREATMENT STRETCHER	868358388348R4
86816726554976	MOVEE VERSUS OPERATION ROOM TRANSFER STRETCHER	868358388348R4
86816723890008	PIONEER 3.1 Trend 3 MOTORS PATIENT CARE BED	868358388348R4
86816723890701	IMPULSE STACKABLE WHEELCHAIR	868358388348R4



QUALITY SYSTEM CERTIFICATE

The Certification Body

B.E.C BELGELENDİRME VE MUAYENE HİZMETLERİ LTD.ŞTİ.

Certified that the

**MEYSA TIBBİ CİHAZLAR
SANAYİ VE TİCARET A.Ş.**

Organize sanayi bölgesi 16. Cd. No:73
MELİKGAZI/KAYSERİ-TÜRKİYE

has established and applies quality systems for

**"DESIGN AND PRODUCTION OF HOSPITAL FURNITURES;
MEDICAL AND PATIENT BEDS"**

has implemented and maintains a quality management system.

an audit, documented in 23.409.17/A report, has verified that this quality management system fulfills the requirements of the following standart.

TS EN ISO 9001:2015

Valid Date : 15.09.2023
Certificate number : 38.23.409.17/A
EA Code -NACE code : 23 NACE 32.50
Certificate Issuance Date /Rev. No: 15.09.2020 /0

MANAGER OF SYSTEM
CERTIFICATION



TÜRKAK BDS NO
YS-54B8-86CD



1.VISA 18.08.2020



2.VISA 17.08.2021

3.VISA 17.08.2022

This document is valid until the date determined above under the conditions that rules of certification are obeyed and it is fixed that periodical supervisions are undergone and appropriateness of the system is sustained. Confirm the validity of the document at www.bec.com.tr or on fax number + 90 262 641 16 68

B.E.C LTD. -555/2 SK. NO:8 GEBZE / KOCAELİ / TÜRKİYE - +90.262.642.04.02- www.becturkey.com



CERTIFICATE

This is to Certify that the
Medical Devices Quality Management System of

Meysa Tıbbi Cihazlar Sanayi ve Ticaret Anonim Şirketi

KayseriOsB Mah. 16 Cad. No: 73 Melikgazi Kayseri Turkey

has been independently assessed and is compliant
with the requirements of

ISO 13485:2016

This Certificate is applicable to the following product or service ranges:

Design, Production and After Sales Services of Hospital Furnitures, Medical Bed and Patient
Bed

Certificate No.: TR95717H

Date of initial registration	16 February 2022
Date of this Certificate	16 February 2022
Surveillance audit on or before	15 February 2023
Recertification Due / Certificate expiry	15 February 2025

This Certificate is property of LMS Certifications and remains valid
subject to satisfactory surveillance audits.

Director

For verification and updated information concerning the present certificate visit to www.lmscert.com

This Certificate is the property of LMS Certification Limited and shall be returned immediately when demanded.



KAB-MC-04

LMS Certification Limited
Labrynth Business Centre, 43 Middle Hill Gate, Stockport,
Great Manchester, England-SK1 3DG
Phone :+44 208 935 5094
Company No.: 11029176
Visit > www.lmscert.com
E-mail > info@lmscert.com



LMS/FM/001/Q/REV06