

Certificate of Compliance

Product Description	:	Cell Culture Slide Hybridwell™ (DLux)
Catalog Number	:	33201
Lot Number	:	-
Manufacture Date	:	-
Expiry Date	:	3years from date of manufactured
<hr/>		
Visual Attribute	:	PASS
Functional Test	:	PASS
Sterility	:	We hereby certify that the goods specified above have been duly sterilized by Ethylene-Oxide gas according to ISO 11135 standard<SAL 10 ⁻⁶ >
Cytotoxicity	:	Cytotoxicity test is conducted to qualify all material resins using ISO 10993-5 standards for cytotoxicity and have been shown to be non-toxic
Non-Pyrogenic	:	Products are tested non-pyrogenic(LAL test) to a limit of less than 0.01EU/mL(0.4EU/device) per USP <85>
RNase/DNase/DNA Free	:	RNase contamination <1x10 ⁻⁹ Kunitz unit/uL, DNase contamination <1x10 ⁻⁶ Kunitz unit/uL and Human DNA <2pg
BSE/TSE Free	:	All products raw materials are either free of animal derived ingredients or are processed according to EMA 410/01, current revision

SPL Lifesciences Co., Ltd. hereby certifies that the product identified above have been inspected in compliance with product quality specification and requirements as documented in our ISO 9001:2015 Quality Management System (K-QA-Q031478) in Korea

Date : 01. 24. 2022

Authorized Signature : H.R. LEE



Quality Control / Assistant Manager