


## VivaChek™ Ino Data Sheet

<b>Specification</b>	
<b>Technology</b>	Electrochemical biosensor
<b>Test range</b>	10~600 mg/dL (0.6~33.3 mmol/L)
<b>Enzyme</b>	Glucose Oxidase
<b>Sample size</b>	About 0.5uL
<b>Testing time</b>	About 5 seconds
<b>Coding method</b>	No coding
<b>Hematocrit range</b>	20-70%
<b>Result calibration</b>	Plasma-equivalent
<b>Measurement unit</b>	Pre-set with either mg/dL or mmol/L depends on the local standard
<b>Sample application</b>	Wicking action, end-fill
<b>Sample type</b>	Fresh capillary whole blood
<b>Primary site</b>	Fingertip
<b>AST (Alternate sites)</b>	Forearm, palm
<b>Memory</b>	900 results with time and date
<b>Test averaging</b>	7, 14 and 30-day average
<b>Operation temperature</b>	5°C - 45°C (41°F -113°F)
<b>Operation humidity</b>	R.H. ≅ 10-90%
<b>Test strip storage temperature</b>	2°C ~30°C (36°F - 86°F)
<b>Test strip unopened shelf life</b>	24 months
<b>Test strip opened shelf life</b>	6 months
<b>Battery</b>	2 x coin batteries
<b>Battery life</b>	Over 1,000 tests
<b>Data transfer port</b>	Micro USB
<b>Beep function</b>	Yes
<b>Automatic shutoff</b>	120 seconds after last action
<b>Strip ejector</b>	Yes
<b>Backlight</b>	No
<b>Hypo warning</b>	Yes
<b>Ketone warning</b>	Yes
<b>Test reminders (alarms)</b>	5
<b>Meal Markers</b>	Before and after meals