
	<p align="center">DECLARATION OF CONFORMITY According to EN ISO/IEC 17050-1 :2010</p>	
---	--	---

Manufacturer's Name: Agilent Technologies Australia (M) Pty Ltd.
Manufacturer's Address: 679 Springvale Rd, Mulgrave, Victoria, 3170, Australia

Declares under sole responsibility that the product as originally delivered

Product Name: Cary 60
Product Number: G6860A
Product Options: This declaration covers all options of the above product

complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

Low Voltage Directive	2014/35/EU
EMC Directive	2014/30/EU
RoHS Directive	2011/65/EU as amended by (EU) 2015/863

and conforms with the following product standards:

	Standard
EMC:	IEC 61326-1:2012 / EN 61326-1:2013 Canada: ICES-001 Issue 5 Australia/New Zealand: IEC 61326-1:2012
Safety:	IEC 61010-1 Ed. 3.1 b:2017 / EN 61010-1:2010+A1:2019 IEC 61010-2-081:2015 / EN 61010-2-081:2015 Canada: CAN/CSA C22.2 No. 61010-1-12 Canada: CAN/CSA C22.2 No. 61010-2-081-15 USA: UL 61010-1 (3 rd Edition) USA: UL 61010-2-081:2015
Other:	EN IEC 63000:2018

Supplementary Information:

The product was tested in a typical configuration with Agilent Technologies test systems.

This DoC applies to above-listed product placed on the EU market after:

February 14, 2022

Date



Cale Hunkin, Quality Manager

For further information, please contact your local Agilent Technologies sales office, agent or distributor, or Agilent Technologies Deutschland GmbH, Hewlett-Packard-Strasse 8, 76337 Waldbronn, Germany.

