

20, AVENUE APPIA - CH-1211 GENEVA 27 - SWITZERLAND - TEL CENTRAL +41 22 791 2111 - FAX CENTRAL +41 22 791 3111 - WWW.WHO.INT

Tel. direct: +41 22 791 5891 SD Biosensor, Inc

Fax direct: +41 22 791 4836 Attention: Mr Young-Gyun Kim

E-mail: diagnostics@who.int

C 4th and 5th, 16 Deogyeong-daero, 155

In reply please beon-gil

refer to: CC/vl Suwon-si, Geonggi-do

16690

Your reference: P17-370-9 République de Corée

11 June 2020

Dear Mr Kim.

Subject: WHO Prequalification of In Vitro Diagnostics – Final Public Report

Product name: STANDARD Q HIV 1/2 Ab 3-Line Test

Product codes: 09HIV30D and 09HIV30DM

Regulatory version: Rest-of-World regulatory version

Manufacturer: SD Biosensor Inc

PQDx Reference Number: PQDx 0383-117-00

We are pleased to inform you that the above-referenced product was prequalified on 10 June 2020 and listed on the World Health Organization (WHO) list of prequalified in vitro diagnostic products.

The following post-prequalification activities are required to maintain the prequalification status:

- 1. Notification to WHO of any planned changes to a prequalified product, in accordance with "WHO procedure for changes to a WHO prequalified in vitro diagnostic" (document number PQDx_121); and
- 2. Post-market surveillance activities, in accordance with "WHO guidance on post-market surveillance of in vitro diagnostics" (ISBN 978 92 4 150921 3).

You are also required to submit an annual report that details sales data and all categories of complaints in a summarized form. There are certain categories of complaints and changes to the product that must be notified immediately to WHO, as per the above-mentioned documents. The sales data will serve as denominator data to guide the frequency of re-inspection.

Failure to comply with any of the above-mentioned post-prequalification requirements may lead to remedial action by WHO, including but not limited to, de-listing from the WHO list of prequalified in vitro diagnostic products.

If you have any questions, please do not hesitate to contact us by email (diagnostics@who.int) or by telephone (+4122 791 5891).

Yours sincerely,

Mr Deus Mubangizi

Unit Head

Prequalification Unit

Regulation and Prequalification Department

Access to Medicines and other Health Products Division