

PDS-250 **DNA/RNA decontamination solution, spray**



1. Background The presence of contaminant DNA and RNA in molecular biology laboratories, espe-

plicon DNA and RNA contaminations.

contaminations are particularly sustainable. PDS-250 is a ready-to-use solution for the removal of nucleic acids from most surfaces at PCR workstations and/or lab devices and equipment. This cleansing solution contains a non-alkaline and non-carcinogenic agent -

water solution of phosphoric acid - and is highly active against plasmid, genomic and am-

PDS-250 is stable and heat resistant. PDS-250 should be stored at room tempera-

cially PCR workstations, can result in PCR artifacts, false positive results and inaccurate data. Removal of nucleic acid contaminations has proven to be no trivial matter, as DNA

ture. At lower temperatures a precipitate might occur which can be resolved easily at 37°C. Storage for up to 2 weeks at 65 °C does not reduce the quality of the product. 2. Protection and precaution information

Eye contact and prolonged skin contact with PDS-250 may cause irritation. There-

fore, safety glasses and disposable gloves should be worn while handling the reagent.

PDS-250 can be applied onto glass, ceramic, plastic, rubber, steel and precious

metal. PDS-250 should not be used for the cleaning of light metal or non-ferrous metals. Do not use PDS-250 spray on electronic devices, like powered dispenser, or pipettes (see below for details).

Instructions for use

3.

For the decontamination of smooth, non-porous surfaces spray PDS-250 directly on

the surface, let soak for 1 minute and dry with a paper towel. Then rinse thoroughly with

clean water and dry with a clean paper towel. For coated or sensitive surfaces, we recommend to spot test prior to use to avoid damage or discoloration.

tion, follow the manufacturer's instructions and remove the shaft from the pipette. Remove seals and gaskets from the shaft. Soak the shaft for 1 minute in PCR Clean™, rinse thoroughly with clean water, dry and reassemble.

Contamination of pipettes may occur even while using filtered tips. For decontamina-

Specifications 4.

Composition:<2% phosphoric acid, < 0,2% ethoxylated alcohol, purified water. Catalogue number:BS-040107-DK

Biosan SIA Ratsupites str. 7, k.2, Riga, LV-1067, Latvia Phone +371 67426137, fax +371 67428101

PDS-250

Edition 1.02 – August of 2020 http://www.biosan.lv/