

Subject: Notice of report GZHT02407734-S1

To: NEOMATRIX SRL

Address: #121, 31 August 1989 Street MD-2012 Chisinau, Republic of Moldova

From: Intertek Testing Services Shenzhen Ltd. Guangzhou Branch

Date: June 04, 2021

Tested samples of report GZHT02407734-S1 dated Apr 15, 2021 were found to comply with the Type IIR requirement of EN 14683:2019+AC:2019 with respect to microbial cleanliness, bacterial filtration efficiency, differential pressure and splash resistance pressure tests.

Based on the test results of In Vitro Cytotoxicity Test of report GZHT02407734-S1 dated Apr 15, 2021, it can be concluded that under the experimental conditions, the tested article Medical Face Masks of report GZHT02407734-S1 have no potential toxicity to L-929 in the MTT method.

If you have questions regarding the above information, please reach out to andreas.faisst@intertek.com

Intertek Testing Services Shenzhen Ltd. Guangzhou Branch

June 04, 2021



Apur

Jun 04. 2021

