

MEDIFIT

Mirandola, 27 ottobre 2016

STERILITY DECLARATION

We hereby declare that, based on the results of the tests at the laboratory authorized Crowned Consulting srl (4678-16 and 4679-16 Test reports), the range of products of the family "Veristeril" pouches and rolls, maintain the sterility of the product contained as follows:

STEAM STERILIZATION: 5 years

ETO STERILIZATION: 5 years

The maintenance of sterility requires to preserve the packaged products and sterile, away from direct sources of light and heat, in a dry place and at temperatures between 10 ° C and 40 ° C (recommended).


Medifit S.r.l.



Product Service

CERTIFICATE

No. Q1N 16 06 95628 001

Holder of Certificate: **Medifit S.r.l.**
Via Bruino, 72
41037 Mirandola (Mo)
ITALY

Facility(ies): Medifit S.r.l.
Via Bruino, 72, 41037 Mirandola (Mo), ITALY



Certification Mark:



Scope of Certificate: **Design and development, production and distribution of packaging materials for medical devices to be sterilized**

Applied Standard(s): EN ISO 13485:2012 + AC:2012
Medical devices - Quality management systems - Requirements for regulatory purposes (ISO 13485:2003 + Cor. 1:2009)
DIN EN ISO 13485:2012

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). See also notes overleaf.

Report No.: ITA273774

Valid from: 2016-08-30
Valid until: 2019-08-29



Date, 2016-08-30

S. Preiß
Stefan Preiß

Page 1 of 1



MEDIFIT

DECLARATION OF CONFORMITY

Manufacturer: Address:	MEDIFIT s.r.l. Via Bruino, 72 - 41037 Mirandola (MO) Italy
Medical Device:	POUCHES AND ROLLS "VeriSteril"
Classification Annex IX D. Lgs. 46/97	Class I not sterile

MEDIFIT s.r.l. declares that Medical Devices POUCHES AND ROLLS "VeriSteril" in each different models are conforming to the essential requirements described in annex I of the Medical Devices Directive 93/42/CEE, consolidated with the requirements to 2007/47/EC, and to the applicable standards.

MEDIFIT s.r.l. has developed a post sale surveillance procedure of its medical device according to MEDDEV 2.12/1. "guidelines on post sale surveillance of Medical Devices".

Applicable Directives:

- Medical Devices Directive 93/42/CEE, consolidated with the requirements to 2007/47/EC

Applicable Standards:

European standards	Title
EN ISO 11607-1:2009	Packaging for terminally sterilized medical devices - Part 1: Requirements for materials, sterile barrier systems and packaging systems (EN ISO 11607-1:2009/A1:2014)
EN ISO 13485:2012	Medical devices - Quality management systems - Requirements for regulatory purposes (ISO 13485:2012)
EN 868-5:2009	Packaging for terminally sterilized medical devices - Part 5: Sealable pouches and reels of porous materials and plastic film construction - Requirements and test methods
EN 868-3:2009	Packaging for terminally sterilized medical devices - Part 3: Paper for use in the manufacture of paper bags (specified in EN 868-4) and in the manufacture of pouches and reels (specified in EN 868-5) - Requirements and test methods

MEDIFIT s.r.l. put at the Authority disposal the Technical File with all the documentation required by Annex VII of the Directive 93/42/CEE for five years starting from the last production date of the device.

Mirandola, 08-01-2018



MEDIFIT s.r.l.
CEO
Mario Neri