CERTIFICATE OF ANALYSIS

Digitally signed by Bolea Petru
Date: 2026.04.03 14:15:55 EEST
Reason: MoldSign Signature
Location: Moldova
MOLDOVA EUROPEANĂ

Product : **SUP IRIS SALMONELLA (25G) 120 TABLETS**
Reference : BS07708 Expiry date : 2028.11.16
Batch N° : 0000114918

Analysis	Specifications	Results
Characters		
Appearance	Tablet	Conform
Color	White to pale blue	Conform
Appearance of prepared medium	Clear solution	Conform
Color of prepared medium	Blue green	Conform
Microbiology		
Cultural response after 18 hours of incubation at 41,5°C in Salmonella Enrichment Broth, followed by 24h of incubation at 37°C onto IRIS® Salmonella Agar		
Growth Salmonella Enteritidis CIP 82.97 (WDCM 00030) + Escherichia coli ATCC 25922 (WDCM 00013) + Pseudomonas aeruginosa ATCC 27853 (WDCM 00025)	> 10 magenta colonies	Conform
Growth Salmonella Typhimurium ATCC 14028 (WDCM 00031) + Escherichia coli ATCC 8739 (WDCM 00012) + Pseudomonas aeruginosa ATCC 27853 (WDCM 00025)	> 10 magenta colonies	Conform
Growth Escherichia coli ATCC25922 (WDCM 00013)	Partially inhibited, colorless colonies, score 0-1	Conform
Growth Enterococcus faecalis ATCC 29212 (WDCM 00087)	Inhibited, score 0	Conform

Beauvais November 04, 2025

CONFORM **Batch released by: Charlotte Naude (QC)**

This certificate is electronically generated and requires no signature
Biokar Diagnostics - Rue des Quarante Mines - ZAC de Ther - Allonne - BP10245 - F60002 BEAUVAIS CEDEX - FRANCE
Tel: +33(0)3 44 14 33 33 Fax: +33(0)3 44 14 33 34