

Test Report No : CPSA/210663482-CA64947
Company : ASAP International Sdn. Bhd.
No.1, Jalan Sitar 33/6, Seksyen 33,
40400 Shah Alam, Selangor,
Malaysia

TEST REPORT

Sample Description : Powder Free Nitrile Examination Gloves,
Device Basic UDI-DI: 955589090ASAPPFN4L
Brand Name : ASAP
Country of Origin : Malaysia
Country of Destination : All countries
Size : M
Quantity Tested : 200 pieces
Test Conducted : Freedom from holes
Test Method : EN455 Part 1:2020
Testing Period : 21 May 2021 – 04 June 2021

Based on submitted samples, the following results obtained:-

Acceptable Quality Limit (AQL) : 1.5 Accept : 7 Found : 3

Result : Within AQL

SIGNED FOR AND ON BEHALF OF
SGS (MALAYSIA) SDN BHD

**** End of Report ****



CHEE TUCK CHOON
SECTION HEAD
IKM No. M/3983/6401/12/14

This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <https://www.sgs.com/en/terms-and-conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <https://www.sgs.com/en/terms-and-conditions/terms-e-document>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 14 days (for perishable samples) or 30 days only.