



Certificate of Analysis

01/12/2025

AFLAPREP® M WIDE	
Product code	P124, P124B
Product size	10 and 50 columns
Batch number	NK 286
Date of manufacture	11 November 2025
Expiry	11 November 2027
Storage recommendations	2 - 8 °C. Do not freeze. Alternatively this product can be stored at room temperature (21 - 25 °C) for 12 months from the date of manufacture.

This document is electronically created and therefore valid without a signature

QC/CERTIFICATES/TEMPLATES/V9/11.2023





Certificate of Analysis


Tested to Specification: Quality Control Product Release Inspection and Test Specification			
Recovery testing method	4 ng of AFT M1 is applied in 25 ml of 10 % powdered milk (w/v).		
Recovery specifications	≥80 % to ≤110 % of AFT M1. This exceeds the specifications required by AOAC method 2000.08.		
Recovery results (n = 6)	AFT M1: 90 %	% RSD (n = 6)	AFT M1: 6.6 %
Capacity testing method	1500 ng of AFT M1 is applied in 25 ml of 10 % powdered milk (w/v).		
Capacity specification	Contains antibodies against AFT M1 with a capacity ≥100 ng, according to CEN specification (CEN Method No. NEN-EN14501).		
Capacity results	Recovery ≥70 % when challenged with 1500 ng of AFT M1. For optimum results, samples with greater amounts should be diluted.		

This document is electronically created and therefore valid without a signature

This is to certify that AFLAPREP® M WIDE batch NK 286 passes the required specifications and is released for sale.

QC/CERTIFICATES/TEMPLATES/V9/11.2023

R-Biopharm Rhône Ltd.
Quality Control Department



Lesley Eschenlauer
Quality Scientist

