



CERTIFICATE OF ANALYSIS

Product : **CONFIRM' SALMONELLA KIT 50 tests**
Reference : **BT01108**
Batch N° : **0109852**

Expiry date : **2025.09.22**

Analysis	Specifications	Results
N° batch R1	109852	
N° batch R2	109852	
N° batch R3	109852	

Characters

Appearance atex Salmonella (R1 reagent)	Milky suspension	Conform
Color latex Salmonella (R1 reagent)	White	Conform
Appearance positive control (R2 control)	Opalescente solution	Conform
Color positive control (R2 control)	White	Conform
Appearance negative control (R3 control)	Clear solution	Conform
Color négative control (R3 control)	Colorless	Conform

Microbiology

Agglutination Salmonella Typhimurium ATCC 14028 (WDCM 00031)	Positive	Conform
Agglutination Salmonella Enteritidis CIP 82.97 (WDCM 00030)	Positive	Conform
Agglutination Escherichia coli ATCC 25922 (WDCM 00013)	Negative	Conform
Agglutination positive control (R2)	Positive	Conform
Agglutination negative control (R3)	Negative	Conform

Beauvais January 27, 2025

CONFORM

Batch released by: Pascale GADONNA-WIDEHEM (QC)