

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Laboratory Equipment (Microscope)
Name and address of the applicant	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37 35578 Wetzlar GERMANY
Name and address of the manufacturer	Leica Microsystems CMS GmbH Ernst-Leitz-Straße 17-37, 35578 Wetzlar, GERMANY
Name and address of the factory	Leica Microsystems Ltd. Shanghai Floor 1, 2, 3A, 4A, and 6, Building T20-1, 258 Jinzang Road, China (Shanghai) Pilot Free Trade Zone, 201206 Shanghai, PEOPLE'S REPUBLIC OF CHINA Zhang Jia Gang Optical Instrument Co; Ltd Tangqiao Town, Luyuan Nanlin Development Zone, 215616 Zhangjiagang City, PEOPLE'S REPUBLIC OF CHINA
Ratings and principal characteristics	Rated voltage: 12 VDC Rated power: 18 W Protection class: III
Trade mark (if any)	Leica
Model/type Ref.	DM1000 LED
Additional information (if necessary)	Certificate DE 3 – 30741 issued 2018-06-26 is replaced by this version due to technical changes
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016 IEC 61010-2-101:2018
as shown in the Test Report Ref. No. which forms part of this certificate	713229324

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 056257 0277 Rev. 00

Date, 2022-01-28



(Erich Thurner)

