



EU Declaration of Conformity

as per Annex IV of the Regulation EU 2017/746 on in-vitro diagnostic medical devices

Manufacturer: Roche Diagnostics GmbH
Address: Sandhofer Strasse 116
68305 Mannheim
Germany

Single Registration Number: DE-MF-000006260

Roche Diagnostics GmbH declares, under the sole responsibility, that the product/the product line

Product Name	Cat. No.	Basic UDI-DI
Elecsys CEA	11731629322	7613336001349T

Intended Use:

Immunoassay for the in vitro quantitative determination of carcinoembryonic antigen in human serum and plasma. This assay is further indicated for serial measurement of CEA to aid in the management of cancer patients. The electrochemiluminescence immunoassay "ECLIA" is intended for use on Elecsys and cobas e immunoassay analyzers.

Product Name	Cat. No.	Basic UDI-DI
Elecsys CEA	04491777190	761333600279AN
Elecsys CEA	07027079190	761333600248AB
Elecsys CEA	07027079214	761333602051A4
Elecsys CEA	09755616190	761333602866BR

Intended Use:

Immunoassay for the in vitro quantitative determination of carcinoembryonic antigen in human serum and plasma. This assay is further indicated for serial measurement of CEA to aid in the management of cancer patients. The electrochemiluminescence immunoassay "ECLIA" is intended for use on cobas e immunoassay analyzers.

Product Name	Cat. No.	Basic UDI-DI
CEA CalSet	11731645322	7613336001359V

Intended Use:

CEA CalSet is used for calibrating the quantitative Elecsys CEA assay on cobas e immunoassay analyzers.

Risk Class: A B C D

Conformity Route: Self-Declaration of Conformity (Class A)

- Self-Declaration of Conformity after Notified Body involvement for sterile manufacturing conditions acc. Art. 48 (10) (Class A sterile)
- Technical Documentation Assessment Class B/C – Annex IX
- Technical Documentation Assessment Class D – Annex IX
- Technical Documentation Assessment Class B/C/D for Self-Testing – Annex IX
- Technical Documentation Assessment Class B/C/D for Near-Patient Testing – Annex IX
- Technical Documentation Assessment Class C/D for Companion Diagnostics – Annex IX

Certificates:

- EU QM Certificate No.: V12 010283 0639
- EU Technical Documentation Assessment Certificate No. (Class D, Near-Patient Testing, Self-Testing and Companion Diagnostics):

Other:

- Common Specifications:

Notified Body (NB) Name: TÜV Süd Product Service GmbH

NB Address: Ridlerstraße 65
80339 Munich
Germany


NB Ident. No.: 0123

to which this declaration relates fulfils the requirements of Regulation EU 2017/746 on in-vitro diagnostic medical devices.

Mannheim, 1 August 2024

Roche Diagnostics GmbH

i.V./on behalf of the company

DocuSigned by:

5E57330EEFE04C4...

Dr. Klaus Riebel
Site Quality Head / Network Lead Penzberg

ppa./on behalf of the company

DocuSigned by:

FC5EDEC1054B44C...

Dr. Stefan Scheib
Global Head of Regulatory Affairs, Core Lab

Contact address: Roche Diagnostics GmbH
Abt./Dept. Global Regulatory Affairs
Sandhofer Strasse 116
D-68305 Mannheim

EG-Konformitätserklärung/EC Declaration of Conformity

gemäß Anhang IV der Richtlinie 98/79/EG des Europäischen Parlaments und des Rates vom 27. Oktober 1998 mit TÜV SÜD Product Service GmbH (Ridlerstraße 65, 80339 München, Germany) als Notified Body (Nr. 0123)

as per Annex IV of Directive 98/79/EC of the European Parliament and Council of 27 October 1998 via TÜV SÜD Product Service GmbH (Ridlerstrasse 65, 80339 Munich, Germany) as the Notified Body (No. 0123)

Hersteller/Manufacturer: Roche Diagnostics GmbH

Adresse/Address: Sandhofer Strasse 116
D-68305 Mannheim

Die Roche Diagnostics GmbH erklärt, dass das Produkt/die Produktfamilie
Roche Diagnostics GmbH declares that the product/the product line

Produktname/Product name: **Elecsys Anti-HBc II**

Art.-Nr./Cat. No.: **09014926190**

Beschreibung/Description: Immunologischer In-vitro-Test zur qualitativen Bestimmung der IgG- und IgM-Antikörper gegen das Hepatitis B-Core-Antigen in Humanserum und -plasma.
Der ElektroChemilumineszenz ImmunoAssay "ECLIA" ist zur Durchführung an **cobas e** Immunoassay-Systemen vorgesehen.

*Immunoassay for the in vitro qualitative determination of IgG and IgM antibodies to the hepatitis B core antigen in human serum and plasma. The electrochemiluminescence immunoassay "ECLIA" is intended for use on **cobas e** immunoassay analyzers.*

auf das/die sich diese Erklärung bezieht, den Forderungen der Richtlinie 98/79/EG des Europäischen Parlaments und des Rates vom 27. Oktober 1998 über In-vitro-Diagnostica (bzw. seine Umsetzung in nationales Recht der Mitgliedsstaaten in welchen das Produkt vermarktet werden soll) entspricht.

to which this declaration relates fulfils the requirements of Directive 98/79/EC of the European Parliament and Council of 27 October 1998 on in-vitro diagnostic medical devices (and its relevant transposition into the national laws of the Member States in which the device is intended to be placed on the market).

Mannheim, 31 May 2021

Roche Diagnostics GmbH


ppa./on behalf of the company

DocuSigned by:

A45CC19E27A04F3...

Ralf Zielenski
Head of Quality
Centralised and Point of Care Solutions

ppa./on behalf of the company

DocuSigned by:

FC5EDEC1054B44C...

Dr. Stefan Scheib
Director Global Regulatory Affairs
Centralised and Point of Care Solutions

Kontaktadresse/Contact address: Roche Diagnostics GmbH
Abt./Dept. Global Regulatory Affairs
Sandhofer Strasse 116
D-68305 Mannheim

EG-Konformitätserklärung/EC Declaration of Conformity

gemäß Anhang IV der Richtlinie 98/79/EG des Europäischen Parlaments und des Rates vom 27. Oktober 1998 mit TÜV SÜD Product Service GmbH (Ridlerstraße 65, 80339 München, Germany) als Notified Body (Nr. 0123)

as per Annex IV of Directive 98/79/EC of the European Parliaments and Council of 27 October 1998 via TÜV SÜD Product Service GmbH (Ridlerstrasse 65, 80339 Munich, Germany) as the Notified Body (No. 0123)

Hersteller/Manufacturer: Roche Diagnostics GmbH

Adresse/Address: Sandhofer Strasse 116
D-68305 Mannheim

Die Roche Diagnostics GmbH erklärt, dass das Produkt/die Produktfamilie
Roche Diagnostics GmbH declares that the product/the product line

Produktname/Product name: **Elecsys Anti-HBe**

Art.-Nr./Cat. No.: **07026838190**

Beschreibung/Description: Immunologischer In-vitro-Test zur qualitativen Bestimmung von humanen Antikörpern gegen Hepatitis B e Antigen (HBeAg) in Humanserum und – plasma.
Der **ElektroChemiLumineszenz ImmunoAssay "ECLIA"** ist zur Durchführung an **cobas e** Immunoassay-Systemen vorgesehen.
Immunoassay for the in vitro qualitative determination of human antibodies to the hepatitis B e antigen (HBeAg) in human serum and plasma.
The electrochemiluminescence immunoassay "ECLIA" is intended for use on cobas e immunoassay analyzers.


auf das/die sich diese Erklärung bezieht, den Forderungen der Richtlinie 98/79/EG des Europäischen Parlaments und des Rates vom 27. Oktober 1998 über In-vitro-Diagnostica (bzw. seine Umsetzung in nationales Recht der Mitgliedsstaaten in welchen das Produkt vermarktet werden soll) entspricht.

to which this declaration relates fulfils the requirements of Directive 98/79/EC of the European Parliament and Council of 27 October 1998 on in-vitro diagnostic medical devices (and its relevant transposition into the national laws of the Member States in which the device is intended to be placed on the market).

Mannheim, 3 February 2021


Roche Diagnostics GmbH

ppa./on behalf of the company

DocuSigned by:

0AFDA3CC08B94C8...

Ralf Zielenski
Head of Quality
Centralised and Point of Care Solutions

ppa./on behalf of the company

DocuSigned by:

FC5EDEC1054B44C...

Dr. Stefan Scheib
Director Global Regulatory Affairs
Centralised and Point of Care Solutions

Kontaktadresse/Contact address: Roche Diagnostics GmbH
Abt./Dept. Global Regulatory Affairs
Sandhofer Strasse 116
D-68305 Mannheim

EG-Konformitätserklärung/EC Declaration of Conformity

gemäß Anhang IV der Richtlinie 98/79/EG des Europäischen Parlaments und des Rates vom 27. Oktober 1998 mit TÜV SÜD Product Service GmbH (Ridlerstraße 65, 80339 München, Germany) als Notified Body (Nr. 0123)

as per Annex IV of Directive 98/79/EC of the European Parliament and Council of 27 October 1998 via TÜV SÜD Product Service GmbH (Ridlerstrasse 65, 80339 Munich, Germany) as the Notified Body (No. 0123)

Hersteller/Manufacturer: Roche Diagnostics GmbH

Adresse/Address: Sandhofer Strasse 116
D-68305 Mannheim

Die Roche Diagnostics GmbH erklärt, dass das Produkt/die Produktfamilie
Roche Diagnostics GmbH declares that the product/the product line

Produktname/Product name: Elecsys Anti-HBs II

Art.-Nr./Cat. No.: 08498598190


Beschreibung/Description: Immunologischer In-vitro-Test zur quantitativen Bestimmung von Humanantikörpern gegen das Hepatitis-B-Oberflächenantigen (HBsAg) in Humanserum und -plasma.
Der ElektroChemiLumineszenz ImmunoAssay "ECLIA" ist zur Durchführung an **cobas e** Immunoassay-Systemen vorgesehen.
Immunoassay for the in vitro quantitative determination of human antibodies to the hepatitis B surface antigen (HBsAg) in human serum and plasma.
*The electrochemiluminescence immunoassay "ECLIA" is intended for use on **cobas e** immunoassay analyzers.*

auf das/die sich diese Erklärung bezieht, den Forderungen der Richtlinie 98/79/EG des Europäischen Parlaments und des Rates vom 27. Oktober 1998 über In-vitro-Diagnostica (bzw. seine Umsetzung in nationales Recht der Mitgliedsstaaten in welchen das Produkt vermarktet werden soll) entspricht.
to which this declaration relates fulfils the requirements of Directive 98/79/EC of the European Parliament and Council of 27 October 1998 on in-vitro diagnostic medical devices (and its relevant transposition into the national laws of the Member States in which the device is intended to be placed on the market).

Mannheim, 3 September 2019


Roche Diagnostics GmbH

ppa./on behalf of the company
ppa. Dr. Lydia Langen



Ralf Zielenski
Head of Quality
Centralised and Point of Care Solutions

ppa./on behalf of the company
i.V. Dr. Joachim Hoch



Dr. Stefan Scheib
Director Global Regulatory Affairs
Centralised and Point of Care Solutions

Kontaktadresse/Contact address: Roche Diagnostics GmbH
Abt./Dept. Global Regulatory Affairs
Sandhofer Strasse 116
D-68305 Mannheim

EG-Konformitätserklärung/EC Declaration of Conformity

gemäß Anhang IV der Richtlinie 98/79/EG des Europäischen Parlaments und des Rates vom 27. Oktober 1998 mit TÜV SÜD Product Service GmbH (Ridlerstraße 65, 80339 München, Germany) als Notified Body (Nr. 0123)

as per Annex IV of Directive 98/79/EC of the European Parliaments and Council of 27 October 1998 via TÜV SÜD Product Service GmbH (Ridlerstrasse 65, 80339 Munich, Germany) as the Notified Body (No. 0123)

Hersteller/Manufacturer: Roche Diagnostics GmbH

Adresse/Address: Sandhofer Strasse 116
D-68305 Mannheim

Die Roche Diagnostics GmbH erklärt, dass das Produkt/die Produktfamilie
Roche Diagnostics GmbH declares that the product/the product line

Produktname/Product name: **Elecsys Anti-HCV II**

Art.-Nr./Cat. No.: **08836981190**
08837031190
08837058190

Beschreibung/Description: Der Elecsys Anti-HCV II Test ist ein diagnostischer In-vitro-Test für den qualitativen Nachweis von Antikörpern gegen Hepatitis-C-Virus (HCV) in Humanserum und -plasma.

Der ElektroChemilumineszenz ImmunoAssay "ECLIA" ist zur Durchführung an **cobas e** Immunoassay-Systemen vorgesehen.

The Elecsys Anti-HCV II assay is an in vitro diagnostic test for the qualitative detection of antibodies to hepatitis C virus (HCV) in human serum and plasma.

*The electrochemiluminescence immunoassay "ECLIA" is intended for use on **cobas e** immunoassay analyzers.*

auf das/die sich diese Erklärung bezieht, den Forderungen der Richtlinie 98/79/EG des Europäischen Parlaments und des Rates vom 27. Oktober 1998 über In-vitro-Diagnostica (bzw. seine Umsetzung in nationales Recht der Mitgliedsstaaten in welchen das Produkt vermarktet werden soll) entspricht.

to which this declaration relates fulfils the requirements of Directive 98/79/EC of the European Parliament and Council of 27 October 1998 on in-vitro diagnostic medical devices (and its relevant transposition into the national laws of the Member States in which the device is intended to be placed on the market).

Mannheim, 9 February 2022


Roche Diagnostics GmbH

ppa./on behalf of the company

DocuSigned by:
 Ralf Zielenski
A7F0BA9FE91A46A...

Ralf Zielenski
Head Q&R Compliance, PRRC RDG
Centralised and Point of Care Solutions

ppa./on behalf of the company

DocuSigned by:
 Stefan Scheib
FC5EDEC1054B44C...

Dr. Stefan Scheib
Network Lead Core Lab, Global Regulatory Affairs
Centralised and Point of Care Solutions

Kontaktadresse/*Contact address*: Roche Diagnostics GmbH
Abt./*Dept.* Global Regulatory Affairs
Sandhofer Strasse 116
D-68305 Mannheim

EC Declaration of Conformity

as per Annex IV of the Regulation EU 2017/746 on in-vitro diagnostic medical devices

Manufacturer: Roche Diagnostics GmbH
Address: Sandhofer Strasse 116
 68305 Mannheim
 Germany

Single Registration Number: DE-MF-000006260

Roche Diagnostics GmbH declares, under the sole responsibility, that the product/the product line

Product Name	Cat. No.	Basic UDI-DI
Elecsys Anti-Tg	09004998190	7613336011419X
Elecsys Anti-Tg	09005021190	7613336011429Z
Anti-Tg CalSet	09005030190	761333601143A3

Risk Class: A B C D

Conformity Route:

- Self-Declaration of Conformity (Class A)
- Self-Declaration of Conformity after Notified Body involvement for sterile manufacturing conditions acc. Art. 48 (10) (Class A sterile)
- Technical Documentation Assessment Class B/C – Annex IX
- Technical Documentation Assessment Class D – Annex IX
- Technical Documentation Assessment Class B/C/D for Self-Testing – Annex IX
- Technical Documentation Assessment Class B/C/D for Near-Patient Testing – Annex IX
- Technical Documentation Assessment Class C/D for Companion Diagnostics – Annex IX

Certificates:

- EU QM Certificate No.: V12 010283 0639
- EU Technical Documentation Assessment Certificate No. (Class D, Near-Patient Testing, Self-Testing and Companion Diagnostics):

Other: Common Specifications:

Notified Body (NB) Name: TÜV Süd Product Service GmbH
NB Address: Ridlerstraße 65
 80339 Munich
 Germany
NB Ident. No.: 0123

to which this declaration relates fulfils the requirements of Regulation EU 2017/746 on in-vitro diagnostic medical devices.

Mannheim, 7 December 2021

Roche Diagnostics GmbH

ppa./on behalf of the company

DocuSigned by:

A7F0BA9FE91A46A...

Ralf Zielenski
Head Q&R Compliance, PRRC RDG
Centralised and Point of Care Solutions

ppa./on behalf of the company

DocuSigned by:

FC5EDEC1054B44C...

Dr. Stefan Scheib
Director Global Regulatory Affairs
Centralised and Point of Care Solutions

Contact address:

Roche Diagnostics GmbH
Abt./Dept. Global Regulatory Affairs
Sandhofer Strasse 116
D-68305 Mannheim

EG-Konformitätserklärung/EC Declaration of Conformity

gemäß Anhang III der Richtlinie 98/79/EG des Europäischen Parlaments und des Rates vom 27. Oktober 1998
as per Annex III of Directive 98/79/EC of the European Parliaments and Council of 27 October 1998

Hersteller/Manufacturer: Roche Diagnostics GmbH
Adresse/Address: Roche Professional Diagnostics
Sandhofer Straße 116
D-68305 Mannheim

Die Roche Diagnostics GmbH erklärt, dass das Produkt/die Produktfamilie (bei rezepturgleichen Produkten)
Roche Diagnostics GmbH declares that the product/the product line (in case of products manufactured by identical recipes)

Produktname/Product name: **Anti-TPO CalSet**
Art.-Nr./Id. No.: **06472931**
Beschreibung/Description: Anti-TPO CalSet wird zur Kalibration des quantitativen Elecsys Anti-TPO Tests auf Elecsys und **cobas e** Immunoassay-Systemen verwendet.
Anti-TPO CalSet is used for calibrating the quantitative Elecsys Anti-TPO assay on the Elecsys and cobas e immunoassay analyzers.

auf das/die sich diese Erklärung bezieht, den Forderungen der EG-Richtlinie 98/79/EG des Rates vom 27. Oktober 1998 (bzw. seine Umsetzung in nationales Recht der Mitgliedsstaaten in welchen das Produkt vermarktet werden soll) über In-vitro-Diagnostica entspricht.
to which this declaration relates fulfils the requirements of EC Directive 98/79/EC of the Council of 27 October 1998 (and its relevant transposition into the national laws of the Member States in which the device is intended to be placed on the market) concerning in-vitro diagnostic devices.

Mannheim, 19.09.2013

Roche Diagnostics GmbH
ppa./on behalf of the company



Dr. M. Thein
Head of Quality
Roche Professional Diagnostics

i. V./on behalf of the company



Dr. C. Fleischer
Head of Quality Control Penzberg
Roche Diagnostics Global Operations

Kontaktadresse/Contact address: Roche Professional Diagnostics
Abt./Dept. Global Regulatory Affairs
Sandhofer Straße 116
D-68305 Mannheim
Fax: +49 621/759 1448

06472931_Anti-TPO CalSet - la

Roche Diagnostics GmbH Diagnostics Division

Roche Diagnostics GmbH; Werk Penzberg; Nonnenwald 2; D 82377 Penzberg; Telefon +49 8856 60 0; Telefax +49 8856 60 3896

Sitz der Gesellschaft: Mannheim - Registergericht: AG Mannheim HRB 3962 - Geschäftsführung: Thomas Schmid, Sprecher: Edgar Vieth - Aufsichtsratsvorsitzender: Dr. Severin Schwan

EG-Konformitätserklärung/EC Declaration of Conformity

gemäß Anhang III der Richtlinie 98/79/EG des Europäischen Parlaments und des Rates vom 27. Oktober 1998
as per Annex III of Directive 98/79/EC of the European Parliament and Council of 27 October 1998

Hersteller/Manufacturer: Roche Diagnostics GmbH

Adresse/Address: Sandhofer Strasse 116
D-68305 Mannheim

Die Roche Diagnostics GmbH erklärt, dass das Produkt/die Produktfamilie
Roche Diagnostics GmbH declares that the product/the product line

Produktname/Product name: **Elecsys Anti-TPO**

Art.-Nr./Cat. No.: **07026935190**

Beschreibung/Description: Immunologischer In-vitro-Test zur quantitativen Bestimmung von Thyreoperoxidase-Antikörper in Humanserum und -plasma. Die Anti-TPO-Bestimmung dient als Hilfsmittel zur Diagnose von autoimmunen Schilddrüsenerkrankungen.

Der ElektroChemilumineszenz ImmunoAssay "ECLIA" ist zur Durchführung an **cobas e** Immunoassay-Systemen vorgesehen.

Immunoassay for the in vitro quantitative determination of antibodies to thyroid peroxidase in human serum and plasma. The anti-TPO determination is used as an aid in the diagnosis of autoimmune thyroid diseases.

*The electrochemiluminescence immunoassay "ECLIA" is intended for use on **cobas e** immunoassay analyzers.*

auf das/die sich diese Erklärung bezieht, den Forderungen der Richtlinie 98/79/EG des Europäischen Parlaments und des Rates vom 27. Oktober 1998 über In-vitro-Diagnostica (bzw. seine Umsetzung in nationales Recht der Mitgliedsstaaten in welchen das Produkt vermarktet werden soll) entspricht.

to which this declaration relates fulfils the requirements of Directive 98/79/EC of the European Parliament and Council of 27 October 1998 on in-vitro diagnostic medical devices (and its relevant transposition into the national laws of the Member States in which the device is intended to be placed on the market).

Mannheim, 24 February 2021

Roche Diagnostics GmbH


ppa./on behalf of the company

DocuSigned by:

0AFDA3CC08B94C8...

Ralf Zielenski
Head of Quality
Centralised and Point of Care Solutions

ppa./on behalf of the company

DocuSigned by:

FC5EDEC1054B44C...

Dr. Stefan Scheib
Director Global Regulatory Affairs
Centralised and Point of Care Solutions

Kontaktadresse/Contact address: Roche Diagnostics GmbH
Abt./Dept. Global Regulatory Affairs
Sandhofer Strasse 116
D-68305 Mannheim

EC Declaration of Conformity

as per Annex IV of the Regulation EU 2017/746 on in-vitro diagnostic medical devices

Manufacturer: Roche Diagnostics GmbH
Address: Sandhofer Strasse 116
68305 Mannheim
Germany

Single Registration Number: DE-MF-000006260

Roche Diagnostics GmbH declares, under the sole responsibility, that the product/the product line

Product name: AssayTip/AssayCup tray

Cat.-No.: 05694302001

Basic UDI-DI: 761333601957BN

Risk Class: A B C D

Conformity Route: Self-Declaration of Conformity (Class A)
 Self-Declaration of Conformity after Notified Body involvement for sterile manufacturing conditions acc. Art. 48 (10) (Class A sterile)
 Technical Documentation Assessment Class B/C – Annex IX
 Technical Documentation Assessment Class D – Annex IX
 Technical Documentation Assessment Class B/C/D for Self-Testing – Annex IX
 Technical Documentation Assessment Class B/C/D for Near-Patient Testing – Annex IX
 Technical Documentation Assessment Class C/D for Companion Diagnostics – Annex IX

Certificates: EU QM Certificate No.:
 EU Technical Documentation Assessment Certificate No. (Class D, Near-Patient Testing, Self-Testing and Companion Diagnostics):

Other: Common Specifications:

Notified Body (NB) Name: N/A
NB Address:


NB Ident. No.: N/A

to which this declaration relates fulfils the requirements of Regulation EU 2017/746 on in-vitro diagnostic medical devices.

Mannheim, 23 November 2021

Roche Diagnostics GmbH

ppa./on behalf of the company

DocuSigned by:

A7F0BA9FE91A46A...

Ralf Zielenski
Head Q&R Compliance, PRRC RDG
Centralised and Point of Care Solutions

ppa./on behalf of the company

DocuSigned by:

FC5EDEC1054B44C...

Dr. Stefan Scheib
Director Global Regulatory Affairs
Centralised and Point of Care Solutions

Contact address:

Roche Diagnostics GmbH
Abt./Dept. Global Regulatory Affairs
Sandhofer Strasse 116
D-68305 Mannheim



EU Declaration of Conformity

as per Annex IV of the Regulation EU 2017/746 on in-vitro diagnostic medical devices

Manufacturer: Roche Diagnostics GmbH
Address: Sandhofer Strasse 116
68305 Mannheim
Germany

Single Registration Number: DE-MF-000006260

Roche Diagnostics GmbH declares, under the sole responsibility, that the product/the product line

Product Name	Cat. No.	Basic UDI-DI
Elecsys CA 19-9	11776193122	761333600730AH
Elecsys CA 19-9	11776193214	761333602081AD
Elecsys CA 19-9	07027028190	761333600799BM
Elecsys CA 19-9	07027028214	761333602050A2

Intended Use:

Immunoassay for the in vitro quantitative determination of CA 19-9 in human serum and plasma. The electrochemiluminescence immunoassay "ECLIA" is intended for use on cobas e immunoassay analyzers.

Product Name	Cat. No.	Basic UDI-DI
CA 19-9 CalSet	11776215122	761333600732AM

Intended Use:

CA 19-9 CalSet is used for calibrating the quantitative Elecsys CA 19-9 assay on cobas e immunoassay analyzers.

Risk Class: A B C D

Conformity Route: Self-Declaration of Conformity (Class A)
 Self-Declaration of Conformity after Notified Body involvement for sterile manufacturing conditions acc. Art. 48 (10) (Class A sterile)
 Technical Documentation Assessment Class B/C – Annex IX
 Technical Documentation Assessment Class D – Annex IX
 Technical Documentation Assessment Class B/C/D for Self-Testing – Annex IX
 Technical Documentation Assessment Class B/C/D for Near-Patient Testing – Annex IX

Technical Documentation Assessment Class C/D for Companion Diagnostics – Annex IX

Certificates:

EU QM Certificate No.: V12 010283 0639

EU Technical Documentation Assessment Certificate No. (Class D, Near-Patient Testing, Self-Testing and Companion Diagnostics):

Other:

Common Specifications:

Notified Body (NB) Name: TÜV Süd Product Service GmbH

NB Address: Ridlerstraße 65
80339 Munich
Germany


NB Ident. No.: 0123

to which this declaration relates fulfils the requirements of Regulation EU 2017/746 on in-vitro diagnostic medical devices.

Mannheim, 12 August 2024

Roche Diagnostics GmbH

ppa./on behalf of the company

DocuSigned by:

485913ABEB04408...

Dr. Peer Lorenz
Site Quality Head / Network Lead, Mannheim

ppa./on behalf of the company

DocuSigned by:

FC5EDEC1054B44C...

Dr. Stefan Scheib
Global Head of Regulatory Affairs, Core Lab

Contact address: Roche Diagnostics GmbH
Abt./Dept. Global Regulatory Affairs
Sandhofer Strasse 116
D-68305 Mannheim



EU Declaration of Conformity

as per Annex IV of the Regulation EU 2017/746 on in-vitro diagnostic medical devices

Manufacturer: Roche Diagnostics GmbH
Address: Sandhofer Strasse 116
68305 Mannheim
Germany

Single Registration Number: DE-MF-000006260

Roche Diagnostics GmbH declares, under the sole responsibility, that the product/the product line

Product Name	Cat. No.	Basic UDI-DI
Elecsys CEA	11731629322	7613336001349T

Intended Use:

Immunoassay for the in vitro quantitative determination of carcinoembryonic antigen in human serum and plasma. This assay is further indicated for serial measurement of CEA to aid in the management of cancer patients. The electrochemiluminescence immunoassay "ECLIA" is intended for use on Elecsys and cobas e immunoassay analyzers.

Product Name	Cat. No.	Basic UDI-DI
Elecsys CEA	04491777190	761333600279AN
Elecsys CEA	07027079190	761333600248AB
Elecsys CEA	07027079214	761333602051A4
Elecsys CEA	09755616190	761333602866BR

Intended Use:

Immunoassay for the in vitro quantitative determination of carcinoembryonic antigen in human serum and plasma. This assay is further indicated for serial measurement of CEA to aid in the management of cancer patients. The electrochemiluminescence immunoassay "ECLIA" is intended for use on cobas e immunoassay analyzers.

Product Name	Cat. No.	Basic UDI-DI
CEA CalSet	11731645322	7613336001359V

Intended Use:

CEA CalSet is used for calibrating the quantitative Elecsys CEA assay on cobas e immunoassay analyzers.

Risk Class: A B C D

Conformity Route: Self-Declaration of Conformity (Class A)

- Self-Declaration of Conformity after Notified Body involvement for sterile manufacturing conditions acc. Art. 48 (10) (Class A sterile)
- Technical Documentation Assessment Class B/C – Annex IX
- Technical Documentation Assessment Class D – Annex IX
- Technical Documentation Assessment Class B/C/D for Self-Testing – Annex IX
- Technical Documentation Assessment Class B/C/D for Near-Patient Testing – Annex IX
- Technical Documentation Assessment Class C/D for Companion Diagnostics – Annex IX

Certificates:

- EU QM Certificate No.: V12 010283 0639
- EU Technical Documentation Assessment Certificate No. (Class D, Near-Patient Testing, Self-Testing and Companion Diagnostics):

Other:

- Common Specifications:

Notified Body (NB) Name: TÜV Süd Product Service GmbH

NB Address: Ridlerstraße 65
80339 Munich
Germany


NB Ident. No.: 0123

to which this declaration relates fulfils the requirements of Regulation EU 2017/746 on in-vitro diagnostic medical devices.

Mannheim, 1 August 2024

Roche Diagnostics GmbH

i.V./on behalf of the company

DocuSigned by:

5E57330EEFE04C4...

Dr. Klaus Riebel
Site Quality Head / Network Lead Penzberg

ppa./on behalf of the company

DocuSigned by:

FC5EDEC1054B44C...

Dr. Stefan Scheib
Global Head of Regulatory Affairs, Core Lab

Contact address: Roche Diagnostics GmbH
Abt./Dept. Global Regulatory Affairs
Sandhofer Strasse 116
D-68305 Mannheim



EC Declaration of Conformity

as per Annex IV of the Regulation EU 2017/746 on in-vitro diagnostic medical devices

Manufacturer: Roche Diagnostics GmbH
Address: Sandhofer Strasse 116
68305 Mannheim
Germany

Single Registration Number: DE-MF-000006260

Roche Diagnostics GmbH declares, under the sole responsibility, that the product/the product line

Product Name	Cat. No.	Basic UDI-DI
CleanCell M	04880293190	761333601330A4
CleanCell M	04880293214	761333602614AU

Risk Class: A B C D

Conformity Route: Self-Declaration of Conformity (Class A)
 Self-Declaration of Conformity after Notified Body involvement for sterile manufacturing conditions acc. Art. 48 (10) (Class A sterile)
 Technical Documentation Assessment Class B/C – Annex IX
 Technical Documentation Assessment Class D – Annex IX
 Technical Documentation Assessment Class B/C/D for Self-Testing – Annex IX
 Technical Documentation Assessment Class B/C/D for Near-Patient Testing – Annex IX
 Technical Documentation Assessment Class C/D for Companion Diagnostics – Annex IX

Certificates: EU QM Certificate No.:
 EU Technical Documentation Assessment Certificate No. (Class D, Near-Patient Testing, Self-Testing and Companion Diagnostics):

Other: Common Specifications:

Notified Body (NB) Name: N/A
NB Address:


NB Ident. No.: N/A

to which this declaration relates fulfils the requirements of Regulation EU 2017/746 on in-vitro diagnostic medical devices.

Mannheim, 30 May 2022


Roche Diagnostics GmbH

ppa./on behalf of the company

DocuSigned by:

A7F0BA9FE91A46A...

Ralf Zielenski
Head Q&R Compliance, PRRC RDG
Centralised and Point of Care Solutions

ppa./on behalf of the company

DocuSigned by:

FC5EDEC1054B44C...

Dr. Stefan Scheib
Network Lead Core Lab, Global Regulatory Affairs
Centralised and Point of Care Solutions

Contact address:

Roche Diagnostics GmbH
Abt./Dept. Global Regulatory Affairs
Sandhofer Strasse 116
D-68305 Mannheim

EU Declaration of Conformity

as per Annex IV of the Regulation EU 2017/746 on in-vitro diagnostic medical devices

Manufacturer: Roche Diagnostics GmbH
Address: Sandhofer Strasse 116
68305 Mannheim
Germany

Single Registration Number: DE-MF-000006260

Roche Diagnostics GmbH declares, under the sole responsibility, that the product/the product line

Product Name	Cat. No.	Basic UDI-DI
Elecsys C-Peptide	03184897190	761333600931AV
Elecsys C-Peptide	03184897214	761333602044A7

Intended Use:

Immunoassay for the in vitro quantitative determination of C-peptide in human serum, plasma and urine. The assay is intended for use as an aid in the diagnosis and treatment of patients with abnormal insulin secretion. The electrochemiluminescence immunoassay "ECLIA" is intended for use on Elecsys and cobas e immunoassay analyzers.

Product Name	Cat. No.	Basic UDI-DI
Elecsys C-Peptide	07027168190	761333600993BK
Elecsys C-Peptide	07027168214	761333602053A8

Intended Use:

Immunoassay for the in vitro quantitative determination of C-peptide in human serum, plasma and urine. The assay is intended for use as an aid in the diagnosis and treatment of patients with abnormal insulin secretion. The electrochemiluminescence immunoassay "ECLIA" is intended for use on cobas e immunoassay analyzers.

Product Name	Cat. No.	Basic UDI-DI
C-Peptide CalSet	03184919190	761333600932AX

Intended Use:

C-Peptide CalSet is used for calibrating the quantitative Elecsys C-Peptide assay on cobas e immunoassay analyzers.

Risk Class: A B C D

Conformity Route:

- Self-Declaration of Conformity (Class A)*
- Self-Declaration of Conformity after Notified Body involvement for sterile manufacturing conditions acc. Art. 48 (10) (Class A sterile)*
- Technical Documentation Assessment Class B/C – Annex IX*
- Technical Documentation Assessment Class D – Annex IX*
- Technical Documentation Assessment Class B/C/D for Self-Testing – Annex IX*
- Technical Documentation Assessment Class B/C/D for Near-Patient Testing – Annex IX*
- Technical Documentation Assessment Class C/D for Companion Diagnostics – Annex IX*

Certificates:

- EU QM Certificate No.: V12 010283 0639*
- EU Technical Documentation Assessment Certificate No. (Class D, Near-Patient Testing, Self-Testing and Companion Diagnostics):*

Other: *Common Specifications:*


Notified Body (NB) Name: TÜV Süd Product Service GmbH
NB Address: Ridlerstraße 65
80339 Munich
Germany
NB Ident. No.: 0123

to which this declaration relates fulfils the requirements of Regulation EU 2017/746 on in-vitro diagnostic medical devices.

Mannheim, 30 March 2023


Roche Diagnostics GmbH

i.V./on behalf of the company

DocuSigned by:

E3965E80F3E840E...

Dr. Christina Schmid
Head of Pre-Market Quality Core Lab

ppa./on behalf of the company

DocuSigned by:

FC5EDEC1054B44C...

Dr. Stefan Scheib
Global Head of Regulatory Affairs, Core Lab

Contact address: Roche Diagnostics GmbH
Abt./Dept. Global Regulatory Affairs
Sandhofer Strasse 116
D-68305 Mannheim