



The "LIFETIME" Surgical Aspirator has been designed to use in operation theaters during surgical operation to take off the accumulated liquids in the operated resin of the patient and for application in Gynecology and Dermatology (liposuction).

The electronically controlled foot paddle and selection of variable capacity accumulation jars give wide range of working facilities to the operator.

SPECIFICATION

PRODUCT MODEL	SA01HT	SA01HT MANUAL	SA01HT W/SENSOR
COMPRESSOR	OILLESS PISTON PUMP	OILLESS PISTON PUMP	OILLESS PISTON PUMP
VOLTAGE	230 VAC ± %10 , 50 Hz	230 VAC ± %10 , 50 Hz	230 VAC ± %10 , 50 Hz
MAX. FREE AIR FLOW	60 lt / min ± %10	60 lt / min ± %10	60 lt / min ± %10
MAX. VACUUM	-700 mmHg ± %5	-700 mmHg ± %5	-700 mmHg ± %5
VACUUM SETTING	YES	YES	YES
COLLECTION JAR	2 - 3 - 5 lt	2 - 3 - 5 lt	2 - 3 - 5 lt
PROBE CONTAINER	YES	YES	YES
MOBILE STAND	YES	YES	YES
STAND BASKET	YES	YES	YES
DEVICE BODY	PLASTIC	PLASTIC	PLASTIC
STAND BODY	ANTI-STATIC PAINTED METAL	ANTI-STATIC PAINTED METAL	ANTI-STATIC PAINTED METAL
VACUUM ENTRANCE	DOUBLE	SINGLE	DOUBLE
LEFT-RIGHT JAR CHANGE	WITH ELECTRONIC BUTTON	MANUEL	WITH ELECTRONIC BUTTON WITH SENSOR AUTOMATIC
FLOAT SYSTEM	YES	YES	YES
HYROPHIBIC FILTER	YES	YES	YES
FOOT PEDAL	OPTIONAL	NO	OPTIONAL
STANDARD	ISO 10079 - 1	ISO 10079 - 1	ISO 10079 - 1

