

Specificații tehnice (F4.1)

Numărul procedurii de achiziție: **MD-1549030142902**

Denumirea procedurii de achiziție: Piese pentru utilaj medical

		Denumirea bunurilor	Modelul articolului	Țara de origine	Producătorul	Specificarea tehnică deplină solicitată de către autoritatea contractantă	Specificarea tehnică deplină propusă de către ofertant	Standard e de referință
Cod CPV								
1		2	3	4	5	6	7	8
		Bunuri						
		Lotul 26 Cablul fibră optică compatibil cu sursa de lumină Karl Storz 250twin						
	26.1	Cablul fibră optică compatibil cu sursa de lumină Karl Storz 250twin	495NA	Germania	Karl Storz	Cablul fibră optică compatibil cu sursa de lumină Karl Storz 250twin	Cablul fibră optică compatibil cu sursa de lumină Karl Storz 250twin	



Semnat:

Numele, Prenumele: Cucerenco Olesea În calitate de: Jurisconsult

Ofertantul: SC "Denolga Medical" SRL Adresa: Mun. Chisinau, str. Banulescu Bodoni, 45, of. 316


Fiber Optic Light Cables

for Cold Light Fountains



Fiber Optic Light Cable with Straight Connector

- 495 NE **Fiber Optic Light Cable**, diameter 4.8 mm, length 300 cm
- 495 NCS **Fiber Optic Light Cable**, extremely heat-resistant, diameter 4.8 mm, length 250 cm
- NEW** 495 NCSC **Fiber Optic Light Cable**, extremely heat-resistant, with safety locking device, diameter 4.8 mm, length 250 cm
- 495 NB **Fiber Optic Light Cable**, diameter 4.8 mm, length 180 cm
- 495 ND **Fiber Optic Light Cable**, diameter 3.5 mm, length 300 cm
- 495 NA **Fiber Optic Light Cable**, diameter 3.5 mm, length 230 cm
- NEW** 495 NAC **Fiber Optic Light Cable**, with safety locking device, diameter 3.5 mm, length 230 cm
- 495 NL **Fiber Optic Light Cable**, diameter 3.5 mm, length 180 cm
- 495 NTA **Fiber Optic Light Cable**, diameter 2.5 mm, length 230 cm
- 495 NT **Fiber Optic Light Cable**, diameter 2.5 mm, length 180 cm
- NEW** 495 TIP **Fiber Optic Light Cable**, highly heat resistant, diameter 4.8 mm, length 300 cm, optimized for 3D TIPCAM®1 and TIPCAM®



Fiber Optic Light Cable with 90° Deflection to the Light Source

- 495 NWL **Fiber Optic Light Cable**, with 90° deflection to the cold light fountain on the fountain side, diameter 3.5 mm, length 300 cm
- 495 NWM **Fiber Optic Light Cable**, with 90° deflection to the cold light fountain on the fountain side, diameter 3.5 mm, length 230 cm
- 495 NW **Fiber Optic Light Cable**, with 90° deflection to the cold light fountain on the fountain side, diameter 3.5 mm, length 180 cm
- 495 NTX **Fiber Optic Light Cable**, with 90° deflection to the cold light fountain on the fountain side, diameter 2.5 mm, length 230 cm
- 495 NTW **Fiber Optic Light Cable**, with 90° deflection to the cold light fountain on the fountain side, diameter 2.5 mm, length 180 cm



495 NVC 495 NV/NVL/NVB

Fiber Optic Light Cable with 90° Deflection to the Instrument

- NEW** 495 NVC **Fiber Optic Light Cable**, with 90° deflection to the instrument, very narrow radius of curvature, diameter 4.8 mm, length 300 cm
- 495 NVB **Fiber Optic Light Cable**, with 90° deflection to the instrument, diameter 4.8 mm, length 300 cm
- 495 NVL **Fiber Optic Light Cable**, with 90° deflection to the instrument, diameter 3.5 mm, length 300 cm
- 495 NV **Fiber Optic Light Cable**, with 90° deflection to the instrument, diameter 3.5 mm, length 230 cm



Light Adaptor, angled 90°

- 495 EW **Light Adaptor**, angled 90°, diameter 4.8 mm, free rotatable, to connect with standard telescopes

Adaptor for connecting KARL STORZ fiber optic light cables with endoscopes and light sources from other manufacturers, see page TP 132





EC-DECLARATION OF CONFORMITY

EG-KONFORMITÄTSERKLÄRUNG

Device Name
Produkt Name Fiber Optic Light Cable, 230cm, Ø3.5mm

Model Number(s)
Modell Nummer(n) 495NA

Classification
Klassifizierung Class I per Annex IX, Rule 1 of Council Directive 93/42/EEC
Klasse I gemäß Anhang IX, Regel 1 der Richtlinie 93/42/EWG des Rates

We issue the present Declaration of Conformity on our sole responsibility and herewith declare self-dependent that the device mentioned above meets the Essential Requirements as defined in Annex VII.

Wir stellen die vorliegend Konformitätserklärung in Eigenverantwortung aus und erklären hiermit unter alleiniger Verantwortung, dass das oben genannte Produkt die Grundlegenden Anforderungen gemäß Anhang VII erfüllt.

This Declaration of Conformity is issued according to Annex VII Council Directive 93/42/EEC for Medical Devices (*for class I devices*).

Diese Konformitätserklärung ist erstellt gemäß Anhang VII Richtlinie 93/42/EWG des Rates über Medizinprodukte (*für Klasse I Produkte*).

Full list of applied standards, directives and laws (12-C2.3.F013-LOAS-CM006) on request.
Vollständige Liste angewandter Normen, Richtlinien und Gesetze (12-C2.3.F013-LOAS-CM006) auf Anfrage.

This Declaration is valid until: 2018-07-16
Diese Erklärung ist gültig bis: 2018-07-16



KARL STORZ SE & Co. KG
Dr.-Karl-Storz-Straße 34
78532 Tuttlingen
Germany

Tuttlingen, 07 Oktober 2013


i. V. Serkan Sezer
Vice President Global Quality
Management, Regulatory Affairs, RSB
& Service



This declaration loses all validity if KARL STORZ SE & Co. KG performs a product change which affects the Conformance to the Essential Requirements or an alteration of any kind not approved by KARL STORZ SE & Co. KG was made at the device mentioned above.
Diese Erklärung verliert ihre Gültigkeit sobald KARL STORZ SE & Co. KG Produktänderungen durchführt, welche die Konformität mit den Grundlegenden Anforderungen beeinflusst oder eine Änderung jeglicher Art ohne Freigabe durch die KARL STORZ SE & Co. KG am oben genannten Produkt durchgeführt wird.

