


DATE	03.01.2025
DOC.NO	FRM-KLT-BYM-069
PAGE NO	1
REV.NO	
REV.DATE	

### TECHNICAL DATA SHEET

<b>PRODUCT:</b>	Sterile Reinforced Surgical Gown - single packed
<b>Description of Product:</b>	Sterile Reinforced Surgical Gown, SMS 35 gsm + Reinforcement - Flat Pouch
<b>Raw Materials:</b>	SMS + Reinforcement
<b>Product Colour:</b>	Medical Blue
<b>Reference Code:</b>	SG-01202-02 SG-01202-03 SG-01202-04 SG-01202-05
<b>Weight in Grams:</b>	35 gsm(SMS) + Reinforcement
<b>Package:</b>	Flat Pouch
Product: Medical Device Directive MDR 2017/745 ECC ANNEX XI A ALL VERSIONS. / Products and materials used do not contain metal. Product Shelf Life - 5 Years	

Product Materials			Unit / Size		PROPERTIES																														
1	SMS + Reinforcement	AllSizes																																	
			<table border="1"> <thead> <tr> <th>X</th> <th>M</th> <th>L</th> <th>XL</th> <th>XXL</th> </tr> </thead> <tbody> <tr> <td>A</td> <td>125</td> <td>132</td> <td>140</td> <td>150</td> </tr> <tr> <td>B</td> <td>34</td> <td>36</td> <td>36</td> <td>37</td> </tr> <tr> <td>C</td> <td>58</td> <td>59</td> <td>60</td> <td>63</td> </tr> <tr> <td>D</td> <td>18</td> <td>19.5</td> <td>21</td> <td>22.5</td> </tr> <tr> <td>E</td> <td>146</td> <td>155</td> <td>160</td> <td>167</td> </tr> </tbody> </table>			X	M	L	XL	XXL	A	125	132	140	150	B	34	36	36	37	C	58	59	60	63	D	18	19.5	21	22.5	E	146	155	160	167
X	M	L	XL	XXL																															
A	125	132	140	150																															
B	34	36	36	37																															
C	58	59	60	63																															
D	18	19.5	21	22.5																															
E	146	155	160	167																															

<b>Tolerances:</b>	+/- 2% cm	<b>Package Information</b>
<b>Measurement:</b>	cm	The products in the sterilization bag are double-packaged to reduce all risks during transportation. Double packaged products are put into Baymed's standard sized carton; dimensions are as follows: Height = 44 cm; Length = 40 cm ve Width = 60 cm.
<b>Preparation Date</b>		<b>QUALITY CONTROL APPROVAL</b>