



# Certificate of Analysis

Product:

**Ficoll-Paque™ PLUS**

Code Number:

17144003

Lot No: 10291481

Test/Characteristic:	Limits:	Results:
<b>Sterility, F<sub>0</sub>-value</b> minutes	min. 15	Passed
<b>Density</b> g/ml, 20°C	1.076 - 1.078	1.077
<b>pH</b>	5.5 - 7.5	6.6
<b>UV / Visible spectra</b>	to pass test	Passed
<b>Sodium Diatrizoate</b> g/dl	8.1 - 9.9	9.2
<b>Colour; A = 405 nm</b> AU	max. 0.20	0.12
<b>Endotoxin activity</b> EU/ml	less than 0.12	<0.12

Note: Sterility is ensured by in-process control.

Manufactured in compliance with our ISO 9001 certified quality management system.

No animal derived material is used in raw materials or in the manufacture of this product.

Approval Date (Year-Month-Day): 2020-04-03

Expiry Date (Year-Month): 2023-03

Manufacturing Date (Year-Month): 2020-03

GE Healthcare Bio-Sciences AB  
Björkgatan 30  
SE-751 84 Uppsala, Sweden  
T + 46 (0)18 612 00 00  
www.gehealthcare.com  
Reg.No. SE 55 61 08 1919 01

Tests and limits according to AS 45-6006-81 Ed. AE

Issued (Year-Month-day): 2020-04-03 by Quality Assurance

Hjelmér, Anna

This document has been electronically produced and is valid without a signature.

DOC1103901 / 6