



**Agilent Technologies**

**DECLARATION OF CONFORMITY**

According to EN ISO/IEC 17050-1 :2010



**Manufacturer's Name:** Agilent Technologies, Incorporated  
**Manufacturer's Address:** 5301 Stevens Creek Blvd.  
Santa Clara, CA 95051, USA

**Declares under sole responsibility that the product as originally delivered**

**Product Name:** Cary 630 FTIR Instrument  
**Model Number:** G8043-64001, G8043-64002, G8043-64011  
G8043-64012  
**Product Options:** This declaration covers all options and sampling accessories of  
the above product(s)

**complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:**

Low Voltage Directive 2014/35/EU  
EMC Directive 2014/30/EU  
RoHS Directive 2011/65/EU

**and conforms with the following product standards:**

	<b>Standard</b>	<b>Limit</b>
<b>EMC</b>	IEC 61326-1:2012 / EN 61326-1:2013 CISPR 11:2011 Australia/New Zealand AS/NZS CISPR 11:2010 Canada: ICES-001:2014 / NMB-001:2004	Group 1 Class A
<b>Safety</b>	IEC 61010-1:2010 / EN 61010-1:2010	
<b>Other</b>	EN 50581:2012	

**Supplementary Information:**

FDA Accession Number 0321242

The products were tested in a typical configuration with Agilent Technologies test systems.

All desktop PCs, Notebook PCs, LCD Monitors, Printers, or Keyboards affiliated with the sale of this product system are exempt from this declaration of compliance.

**This DoC applies to above-listed products placed on the EU market after:**

December 19, 2017

Date

Cale Hunkin, Quality Manager

For further information, please contact your local Agilent Technologies sales office, agent or distributor, or Agilent Technologies Deutschland GmbH, Hewlett-Packard-Strasse 8, 76337 Waldbronn, Germany.

Part Number:



G8043-90000

Revision: F