applied biosystems

SeqStudio™ Genetic Analyzer Instrument and Software

USER GUIDE

for use with:

SeqStudio™ Data Collection Software v1.2
SeqStudio™ Plate Manager
SeqStudio™ Remote Monitoring App
SAE Administrator Console v2.0

SeqStudio™ Genetic Analyzer Cartridge (Cat. No. A33671)
SeqStudio™ Genetic Analyzer Cartridge v2 (Cat. No. A41331)

Publication Number MAN0018646

Revision A.0

Digitally signed by Rusu Ghenadie
Date: 2022.09.16 11:13:13 EEST
Reason: MoldSign Signature
Location: Moldova

For Research Use Only. Not for use in diagnostic procedures.

ThermoFisher SCIENTIFIC

Manufacturer:
Life Technologies Holdings Pte Ltd |
Block 33 |
Marsiling Industrial Estate Road 3 |
#07-06, Singapore 739256

Products:
SeqStudio™ Genetic Analyzer
SeqStudio™ Data Collection Software
SeqStudio™ Plate Manager (desktop)
SeqStudio™ Genetic Analyzer Cartridge
SeqStudio™ Genetic Analyzer Cartridge v2
Security, Auditing, and E-signature (SAE) v2.0 module

Manufacturer: Life Technologies Corporation | 200 Oyster Point Blvd | South San Francisco, CA 94080 | USA

Products:
SeqStudio™ Remote Monitoring App

Manufacturer: Life Technologies Corporation | 35 Wiggins Avenue | Bedford, MA 01730

Products:
SeqStudio™ Genetic Analyzer Cathode Buffer Container

The information in this guide is subject to change without notice.

DISCLAIMER: TO THE EXTENT ALLOWED BY LAW, THERMO FISHER SCIENTIFIC INC. AND/OR ITS AFFILIATE(S) WILL NOT BE LIABLE FOR SPECIAL, INCIDENTAL, INDIRECT, PUNITIVE, MULTIPLE, OR CONSEQUENTIAL DAMAGES IN CONNECTION WITH OR ARISING FROM THIS DOCUMENT, INCLUDING YOUR USE OF IT.

Revision history: Pub. No. MAN0018646

Revision	Date	Description
A.0	4 November 2019	New user guide for SeqStudio™ Data Collection Software v1.2. The v1.2 software includes options for HID kits. This user guide replaces Pub. No. MAN0016138.

Important Licensing Information: These products may be covered by one or more Limited Use Label Licenses. By use of these products, you accept the terms and conditions of all applicable Limited Use Label Licenses.

TRADEMARKS: All trademarks are the property of Thermo Fisher Scientific and its subsidiaries unless otherwise specified.

©2019 Thermo Fisher Scientific Inc. All rights reserved.

Contents

CHAPTER 1 Product information 12

Instrument overview 13
Parts of the instrument 14
Instrument status indicator 15
SeqStudio™ Genetic Analyzer consumables 15
Cartridge 16
SeqStudio™ Genetic Analyzer Cathode Buffer Container 20
Radio frequency identification 20
Important notice regarding use of consumables that exceed supported limits 21
Software features 22
SeqStudio™ Plate Manager overview 24
Computer and operating system options 25
Recommended browsers 25
SeqStudio™ Remote Monitoring App overview 26
Network connection options 27
Network and password security requirements 27
Network configuration and security 27
Password security 27
Experiment types 28
Data output 29
Secondary analysis software 29
Workflow: Connect cloud-based platform or desktop 31
Workflow: instrument 32

CHAPTER 2 Prepare the samples and the instrument 33

Precautions for use 33
Power on the instrument 34
Sign in 34
Sign in with the Guest instrument profile 35
Sign out 35
Parts of the home screen 36

Prepare the samples 37
Sample preparation guidelines 37
Plate layout and loading guidelines 38
Prepare the plate 38
Prepare the tubes 39
Prepare the instrument 39
Check the consumables status 39
Load the CBC, the sample plate, and the cartridge 41

CHAPTER 3 Use the instrument with the Connect cloud-based platform 43

Understanding instrument and Connect cloud-based platform interaction 43
"Connect" versus "link" 43
First user who links is assigned administrator role 43
Local versus Connect cloud-based platform instrument profiles 44
If you link when you *are* signed in to the instrument 44
If you link when you *are not* signed in to the instrument 45
Connect cloud-based platform instrument profile roles and functions 47
For more information on using the Connect cloud-based platform 47
Connect cloud-based platform administrators for an instrument 48
Connect cloud-based platform administrator functions 48
Assign instrument administrator role to other users 48
Manage the users and administrators of your instrument 49
Disconnect individual users from an instrument 50
Register and obtain a Connect cloud-based platform account 50
Link the instrument to your Connect cloud-based platform account 51
Change your own Connect cloud-based platform instrument profile PIN 52
Set up email notifications from the instrument 53

CHAPTER 4 Use the instrument with the Security, Auditing, and E-signature (SAE) v2.0 module 55

Overview of the Security, Auditing, and E-signature (SAE) v2.0 module components 56
Configuring the SAE module 57
Instrument functionality when SAE is enabled or disabled 58
Connect the instrument to an SAE server (administrator only) 58
Enable SAE on the instrument and specify the SAE server (administrator only) 59
Disable SAE (administrator only) 59
Sign in with SAE enabled 60
Specify audit reasons 61
E-signature requirements when you start a run 61
Sign a plate setup 63
View E-signature records for a plate setup 63

Sign a dye calibration or an install run 64
View E-signature records for a dye calibration or an install run 64
Signing in after automatic screen locking 64
Use the instrument when the SAE server is offline 65
Auditing of imported plate setups 65
SAE error messages and actions 66

CHAPTER 5 Create or modify a plate setup from the Plate Manager 67

Overview of plate setup settings 67
Set up a plate using default settings (Plate Manager) 69
Access the Plate Manager on the Connect cloud-based platform 69
Access the Plate Manager on the desktop 69
Create or open a plate setup PSM file 69
Enter plate properties 70
Assign wells: Sample and run information 71
Keyboard shortcuts for the sample table 72
Save a plate setup in the Plate Manager 73
Additional plate settings (Plate Manager) 74
Specify replicate injections 74
Edit analysis settings 75
Modify the default file name convention 75
Hide or share a plate (**Plate setup security**) 76
Define custom fields 76

CHAPTER 6 Create or modify a plate setup from the instrument . . 77

PSM and CSV plate setup files for import into the instrument 78
Shared (public), hidden (my plates), and guest plate setup files 79
Overview of plate setup settings 79
(Optional) Set up for auto export of sample data files (AB1 and FSA) 80
Set up a plate using default settings (instrument) 81
Create or import a plate setup 81
Enter plate properties 82
Assign wells: run module, size standard, dye set, and kit 84
Assign wells: sample name, sample type, and custom fields 86
Set optional plate settings (instrument) 87
Specify replicate injections 87
Modify analysis settings 87
Modify the default file name convention 88
Hide or share a plate (**Plate setup security**) 89
Define custom fields 89
(Optional) View the injection list, change injection settings or order, or specify replicates and re-injections 91

CHAPTER 7 Start and monitor a run 92

Load the plate or the tube assembly 92
Select a plate setup and start a run 93
Automatic file cleanup 94
Lock the touchscreen 95
Monitor a run from the Connect cloud-based platform 95
Open the Remote Monitoring App from Instrument Connect App 95
Open the Remote Monitoring App from the Plate Manager 97
View notifications from the instrument on your Connect cloud-based platform account 99
View results in the Remote Monitoring App on the Connect cloud-based platform 99
Monitor a run from a mobile device 103
Monitor a run from the instrument 105
(Optional) View the injection list, change injection settings or order, or specify replicates and re-injections 105
View the run status 105
View real-time results 106
Pause a plate or cancel or stop injections 106
View results when the run is completed 107
Unload the plate or the tube assembly 109

CHAPTER 8 View and analyze results 110

View results in the Remote Monitoring App on the Connect cloud-based platform 110
View results on the instrument 111
View results when the run is completed 111
Fragment/HID analysis results 113
Sequence analysis results 115
EPT plot 117
View results for a previously run plate (run history) 118
Data quality alerts 119
Sizecalling and basecalling quality alerts 119
Edit injection parameters and re-inject samples 120
Export a report (QC report) 120
Export results from the instrument (sample data files and QC reports) 121
Analyze data 121

CHAPTER 9 Manage the software (Plate Manager) 122

Add a custom dye calibration to the Plate Manager 122
Download a plate setup template as a CSV file 122
Save a plate setup as a PDF 123
Save a plate setup as a CSV file 123
Create a plate setup template 123

Delete plate setups 124
Manage email notifications in the Connect cloud-based platform 124
Manage run modules 125
Manage size standards 125
Manage analysis settings 126
Fragment/HID analysis settings (size calling) 126
Sequencing settings (base calling) 129

CHAPTER 10 Manage the software (instrument) 130

Connect the instrument to a network drive 130
Determine IP address for a computer on a network 130
Create folders and enable network folder sharing 131
Connect to a network drive 132
Set up a default Connect cloud-based platform location for opening plate setups . 132
Set up a default Connect cloud-based platform location for saving results (auto export) 133
Set up a default network location for opening plate setups 133
Set up a default network location for saving results (auto export) 134
Link the instrument to your Connect cloud-based platform account 136
Lock the touchscreen 137
Manage plate setups 138
Export or delete a plate setup (PSM file) 138
Import a plate setup from a CSV or PSM file 139
Manage run settings (instrument) 139
Manage run modules 139
Manage size standard definitions 142
Manage analysis settings 142
Modify the default file name convention 146
Manage instrument settings 147
Display instrument hardware and software information 147
Change the instrument name 148
Enable Demo mode 148
Manage date and time settings 148
Manage the network configuration 149
Check for software updates (administrator only) 150
Manage instrument profiles on the instrument 151
Local instrument profile roles and functions 151
Create a local instrument profile for another user (administrator only) 151
Change the role of a local instrument profile (administrator only) 152
Delete an instrument profile from an instrument (administrator only) 152
Delete the PIN for a local instrument profile (administrator only) 152

Create your own local instrument profile 153
Change your own local instrument profile PIN 153
Manage storage space 153
Automatic file cleanup 153
Export or delete a plate setup (PSM file) 154
Export results from the instrument (sample data files and QC reports) 154
Delete a run history 155
Back up user data (plates and results) (administrator only) 155

CHAPTER 11 Maintain the instrument 156

Regular maintenance tasks 156
Check the cathode buffer fill level 156
Clean the instrument exterior and touchscreen 157
Clean the autosampler 157
Manage the cartridge 158
Cartridge storage 158
Set cartridge storage mode (administrator only) 159
Check or export the Instrument-cartridge history 160
Check or export the cartridge history 160
Fill the capillary array and refresh the polymer delivery system 161
Remove the cartridge 161
Insert the cartridge 161
Store the cartridge 163
SeqStudio™ Integrated Capillary Protector 164
Long-term on-instrument cartridge storage 165
Install cathode buffer 166
SeqStudio™ Genetic Analyzer Cathode Buffer Container storage 166
Remove the SeqStudio™ Genetic Analyzer Cathode Buffer Container 166
Assemble the SeqStudio™ Genetic Analyzer Cathode Buffer Container (CBC) 166
Insert the Cathode Buffer Container 167

APPENDIX A Troubleshooting 169

Troubleshooting resources 169
Troubleshooting workflow 170
Export log files for plates, install runs, injections, and instrument 170
View the raw data and the EPT plot 171
Data quality alerts 171
Sizecalling and basecalling quality alerts 171
Check size standard quality 172
Instrument troubleshooting 173
Cartridge troubleshooting 174
Sample and data troubleshooting 175
EPT data 180

Software troubleshooting 181
Service tools 183

APPENDIX B Link the instrument to your Connect cloud-based platform account—detailed instructions 184

Workflow: Set up the instrument for Connect cloud-based platform access 184
Network requirements 184
Link the instrument from a mobile device 185
Create a Connect cloud-based platform PIN and generate the QR code on the instrument 185
Register the instrument with the Instrument Connect App 186
Link the instrument using a link code 186
Set up email notifications from the instrument 190

APPENDIX C Dye calibration and install standard checks 192

Calibrate dyes 192
Overview of system dye set and custom dye set calibration 192
System versus custom dyes 192
Factory, auto, and manual calibration 193
Determine if manual calibration is required 193
Perform a system dye calibration 195
Perform a custom dye calibration 197
View the dye calibration history 201
Import or export a custom dye set 202
Perform an install run 203
Overview of install checks 203
Perform an install run 204
View install run results 205
Export the install run report 208
View the install run history 210

APPENDIX D Create and run an HID plate 212

HID analysis requirements 212
Pull-up reduction feature 213
Resolution algorithm 214
Workflow: HID analysis 215
HID install run 215
HID run module, dye sets, size standard definitions, and kits 216
Manage kits (HID applications only) 217

APPENDIX E Run modules 218

Run modules, read lengths, size ranges, and run times 218

APPENDIX F Parts and materials 220

Required materials not supplied 220

APPENDIX G Instrument specifications and layout 222

Instrument dimensions 222
Instrument clearances 222
Environmental requirements 223
Electrical requirements 224
Electrical protective devices 224
Network requirements 225
Safety requirements 225
- Safety practices 225
- Required safety equipment 226

APPENDIX H Safety 227

Symbols on this instrument 227
- Conformity symbols 228

Safety alerts on this instrument 229
Safety information for instruments not manufactured by Thermo Fisher Scientific 230
Instrument safety 230
- General 230
- Physical injury 230
- Electrical safety 230
- Cleaning and decontamination 231
- Instrument component and accessory disposal 231
- Laser 231

Safety and electromagnetic compatibility (EMC) standards 232
- Safety compliance 232
- EMC 232
- Environmental design 233
- Radio compliance 233

Chemical safety 234
Biological hazard safety 235

Documentation and support 236

Related documentation 236
Customer and technical support 236
Limited product warranty 237

Index 238

1 Product information

- Instrument overview 13
- Parts of the instrument 14
- SeqStudio™ Genetic Analyzer consumables 15
- Software features 22
- SeqStudio™ Plate Manager overview 24
- SeqStudio™ Remote Monitoring App overview 26
- Network connection options 27
- Network and password security requirements 27
- Experiment types 28
- Data output 29
- Secondary analysis software 29
- Workflow: Connect cloud-based platform or desktop 31
- Workflow: instrument 32

Instrument overview

The Applied Biosystems™ SeqStudio™ Genetic Analyzer with SeqStudio™ Data Collection Software is a fluorescent dye-labeled genetic analysis system using capillary electrophoresis technology. It enables both sequencing and fragment analysis applications without the need to switch polymer type or capillary array length.

The instrument uses a self-contained, replaceable cartridge with:

- A 4-capillary array
- A universal polymer capable of performing sequencing and fragment analysis
- A polymer delivery system (PDS)
- Anode buffer



The SeqStudio™ Genetic Analyzer automatically:

- Performs an optical alignment each time a cartridge is inserted.
- Performs an automatic spectral calibration adjustment (auto calibration) for each sample to correct for spectral overlap.

The instrument is compatible with 96-well Standard plates and 8-strip Standard tubes.

The SeqStudio™ Genetic Analyzer is a stand-alone instrument. It is run directly from the touchscreen with SeqStudio™ Data Collection Software and does not require a computer. Plate setup can be done directly on the touchscreen, on a computer with SeqStudio™ Plate Manager, or on the Connect cloud-based platform. A run can be monitored directly on the instrument touchscreen or remotely on the Connect platform.

Note: The Connect cloud-based platform is not supported for HID analysis.

Parts of the instrument



Figure 1 Front of the instrument

1. Front panel indicator—Shows the status of the instrument
2. Touchscreen—User interface
3. Door—Provides access to the cartridge, the cathode buffer, and sample plate or tubes
4. USB port



Figure 2 Interior of the instrument

1. Cartridge rails
2. Cartridge
3. Cathode buffer (located inside the autosampler)
4. Autosampler (contains the plate or tube holder and the cathode buffer)
5. Plate or tube holder



Figure 3 Rear of the instrument

1. RJ45 ethernet port
2. USB port
3. Power receptacle
4. On/Off switch

Instrument status indicator

Indicator	Instrument status
All lights off	Powered off or in **Cartridge storage mode**.
Blue light (blinking)	Starting up.
Blue light	Ready to start a run or run is in progress.
Amber light (blinking)	Run is paused, door is open, or error state.

SeqStudio™ Genetic Analyzer consumables

Description	Cat. No.	Amount	Storage
SeqStudio™ Genetic Analyzer Cartridge	A33671	1 cartridge with an optical cover and a SeqStudio™ Integrated Capillary Protector attached for shipment and storage.	See Table 1 on page 18
SeqStudio™ Genetic Analyzer Cartridge v2	A41331	1 cartridge with an optical cover and a SeqStudio™ Integrated Capillary Protector attached for shipment and storage.	See Table 2 on page 19
SeqStudio™ Genetic Analyzer Cathode Buffer Container	A33401	1 package of 4	See page 166
SeqStudio™ Integrated Capillary Protector	A31923	1 (single-use) for future storage	See page 165

Cartridge

The SeqStudio™ Genetic Analyzer uses a 4-capillary, self-contained, replaceable cartridge. The following cartridges are available:

Cartridge	No. of injections	No. of samples	Storage
SeqStudio™ Genetic Analyzer Cartridge (Cat. No. A33671)	125 injections	500 samples	See Table 1 on page 18
SeqStudio™ Genetic Analyzer Cartridge v2 (Cat. No. A41331)	250 injections	1,000 samples	See Table 2 on page 19

Each cartridge:

- Contains a capillary array, a polymer delivery system, polymer, and anode buffer. See Figure 4 and Figure 5.
- Is shipped with an optical cover and an Integrated Capillary Protector. See Figure 6.
- Has a radio frequency identification (RFID) tag, which is used by the instrument to track remaining usage and expiration.



Figure 4 Parts of the cartridge

1. Universal polymer—Supplies polymer to the polymer delivery pump.
2. Polymer delivery system—Pumps polymer into the capillary array.
3. Anode buffer reservoir—Genetic Analysis running buffer to support electrophoresis.
4. Optical detection window
5. Cartridge track—To insert the cartridge into the instrument.
6. Capillary array (cathode end)—Tips of the four capillaries that enable electrophoretic separation of fluorescent-labeled DNA fragments.



Figure 5 Parts of the polymer delivery system

1. Universal polymer
2. Polymer valve
3. Syringe
4. Capillary fitting (anode end)
5. Buffer valve
6. Anode buffer reservoir



Figure 6 Cartridge with optical cover and Integrated Capillary Protector

1. Optical cover
2. Optical cover hand hold (for removing the optical cover)
3. Integrated Capillary Protector
4. Cartridge hand hold

IMPORTANT! Remove the Integrated Capillary Protector before installing the cartridge into the instrument. Installing the cartridge with the ICP in place can damage the capillary array.

Cartridge storage

Table 1 Storage information for the SeqStudio™ Genetic Analyzer Cartridge (Cat. No. A33671)

Condition	Description
Shipping	Shipped at 2–8°C. Store upright at 2–8°C upon receipt. Save the white storage box and optical cover for off-instrument cartridge storage.
On-instrument storage	For routine use, can be used and stored on the instrument for up to 4 months. If you store the cartridge on-instrument: • The instrument must be powered on. • A Cathode Buffer Container must also be installed. The instrument keeps the components under the following conditions when it is powered on and in **Cartridge storage mode**: • **Optical detection window**—Covered • **Capillary array electrodes**—Submerged in cathode buffer (buffer must be above the fill line in the Cathode Buffer Container) • **Polymer**—Chilled • **Anode buffer**—Ambient temperature IMPORTANT! The instrument does not maintain the correct temperature conditions for the cartridge when it is powered off. Avoid cartridge exposure to ambient temperature.
Off-instrument storage	For intermittent use, can be stored off-instrument until the expiry date on the label or up to 4 months after first use. Store upright at 2–8°C, with an integrated capillary protector (ICP) and optical cover installed (see "Store the cartridge" on page 163). **Note:** After you remove the cartridge from the instrument, install an ICP within a few minutes. Avoid cartridge exposure to ambient temperature.
Reuse	Can be removed from an instrument then inserted again on the same instrument or a different instrument, if it was stored properly at 2–8°C and has not expired or exceeded 125 injections. Information about the cartridge installation and usage is retained in the cartridge history (**Settings ▸ Cartridge ▸ Instrument–cartridge history**).

Table 2 Storage information for the SeqStudio™ Genetic Analyzer Cartridge v2 (Cat. No. A41331)

Condition	Description
Shipping	Shipped at 2–8°C. Store upright at 2–8°C upon receipt. Save the white storage box and optical cover for off-instrument cartridge storage.
On-instrument storage	For routine use, can be used and stored on the instrument for up to 6 months. If you store the cartridge on-instrument: • The instrument must be powered on. • A Cathode Buffer Container must also be installed. The instrument keeps the components under the following conditions when it is powered on and in **Cartridge storage mode**: • **Optical detection window**—Covered • **Capillary array electrodes**—Submerged in cathode buffer (buffer must be above the fill line in the Cathode Buffer Container) • **Polymer**—Chilled • **Anode buffer**—Ambient temperature IMPORTANT! The instrument does not maintain the correct temperature conditions for the cartridge when it is powered off. Avoid cartridge exposure to ambient temperature.
Off-instrument storage	For intermittent use, can be stored off-instrument until the expiry date on the label or up to 6 months after first use. Store upright at 2–8°C, with an integrated capillary protector (ICP) and optical cover installed (see "Store the cartridge" on page 163). **Note:** After you remove the cartridge from the instrument, install an ICP within a few minutes. Avoid cartridge exposure to ambient temperature.
Reuse	Can be removed from an instrument then inserted again on the same instrument or a different instrument, if it was stored properly at 2–8°C and has not expired or exceeded 250 injections. Information about the cartridge installation and usage is retained in the cartridge history (**Settings ▸ Cartridge ▸ Instrument–cartridge history**).

SeqStudio™ Genetic Analyzer Cathode Buffer Container

The SeqStudio™ Genetic Analyzer Cathode Buffer Container (CBC) contains running buffer for capillary electrophoresis. The container has two compartments. The rear compartment provides the cathode buffer for electrophoresis. The front compartment is for capillary wash and waste.

The CBC requires a Reservoir Septa.

See "Assemble the SeqStudio™ Genetic Analyzer Cathode Buffer Container (CBC)" on page 166.



Figure 7 SeqStudio™ Genetic Analyzer Cathode Buffer Container with Reservoir Septa

1. Reservoir Septa
2. Notch (inserted towards the rear right in the autosampler)
3. Fill line (replace the CBC when the buffer is at the fill line)
4. Cathode buffer compartment
5. Waste and wash compartment



Figure 8 Autosampler

1. Location of plate or tubes
2. Release button
3. Location of Cathode Buffer Container
4. Location to position the Cathode Buffer Container notch

Radio frequency identification

The cartridge and cathode buffer have radio–frequency identification (RFID) tags. The instrument reads and tracks:

- Expiry date (shelf life)
- Remaining injections (usage)
- Serial number (cartridge only)
- Lot number (cathode buffer only)

Table 3 RFID tag read/write events and consumables status updates

Component	RFID read	RFID write	Consumables status update
SeqStudio™ Genetic Analyzer Cartridge	When a cartridge inserted. The first time a new cartridge is loaded into an instrument, the manufacturer expiry date is listed for **Expiration time** in the **Consumable status** screen.	• The **Expiration time** is reset from the manufacturer expiry date to 4 months from the current date of installation. • After the cartridge has been loaded and before it is ejected, its cartridge history record is updated to include its usage on this instrument. • During each injection, injection count and remaining polymer volume are updated.	Every 8 hours and/or before each run[1]: • When a cartridge is within 2 weeks of expiry date. • When the number of injections is approaching the limit of 125 injections.
SeqStudio™ Genetic Analyzer Cartridge v2		• The **Expiration time** is reset from the manufacturer expiry date to 6 months from the current date of installation. • After the cartridge has been loaded and before it is ejected, its cartridge history record is updated to include its usage on this instrument. • During each injection, injection count and remaining polymer volume are updated.	Every 8 hours and/or before each run[1]: • When a cartridge is within 2 weeks of expiry date. • When the number of injections is approaching the limit of 250 injections.
Cathode buffer	• When the autosampler initializes. • When a plate is retracted.	• The first time a new CBC is loaded on an instrument, the installation date is recorded on the CBC. • After each injection, the injection count is updated.	Every 8 hours and/or before each run[1]: • When the CBC is within 2 days of expiry date. • When the number of injections is approaching the limit of 125 injections.

[1] If either limit is met, an email notification is also sent to any Connect cloud-based platform users who are linked to this instrument

Important notice regarding use of consumables that exceed supported limits

BEFORE DISMISSING THE WARNING THAT THE CONSUMABLES HAVE REACHED SUPPORTED LIMITS AND CONTINUING WITH OPERATION OF THE INSTRUMENT, PLEASE READ AND UNDERSTAND THE FOLLOWING IMPORTANT NOTICE AND INFORMATION:

Life Technologies does not recommend the use of consumables that exceed supported limits. The recommended limits are designed to promote the production of high quality data and minimize instrument downtime. Reagent and consumable lifetime minimum performance are based on testing and studies that use reagents and consumables that have not exceeded supported limits.

The use of consumables beyond the supported limits may impact data quality or cause damage to the instrument or capillary array. The cost of repairing such damage is NOT covered by any Life Technologies product warranty or service plan. Customer

use of expired consumables is at customer's own risk and without recourse to Life Technologies. For example, product warranties do not apply to defects resulting from or repairs required due to misuse, neglect, or accident including, without limitation, operation outside of the environmental or use specifications or not in conformance with Life Technologies instructions for the instrument system, software, or accessories.

Please see your specific service contract or limited product warranty for exact language regarding coverage and ask your Life Technologies representative if you have further questions.

Software features

The instrument can be operated directly from the touchscreen using the SeqStudio™ Data Collection Software. The touch screen allows scrolling and zooming by pinch.

Plates can be set up and saved on a computer using the SeqStudio™ Plate Manager running on a desktop or on the Connect cloud-based platform. These plate setups can be saved to and accessed from the instrument from:

- A network drive
- A USB drive
- The Connect cloud-based platform

You can monitor runs directly from the instrument or from the SeqStudio™ Remote Monitoring App.

Table 4 Features of the software

Feature	SeqStudio™ Data Collection Software	SeqStudio™ Plate Manager (desktop)	SeqStudio™ Plate Manager (Connect platform)	Remote Monitoring App
Plate setup				
Create a new plate setup	✓	✓	✓	—
Enter plate properties	✓	✓	✓	—
Set up plate wells	✓	✓	✓	—
Advanced options for plate setup				
Edit sample properties	✓	✓	✓	—
Edit plate setup	✓	✓	✓	—
Manage size standards	✓	✓	✓	—

Feature	SeqStudio™ Data Collection Software	SeqStudio™ Plate Manager (desktop)	SeqStudio™ Plate Manager (Connect platform)	Remote Monitoring App
Manage run modules (including edit a run module)	✓	✓	✓	—
Advanced options for plate properties				
Adjust fragment/HID analysis parameters	✓	✓	✓	—
Adjust sequence parameters	✓	✓	✓	—
Adjust the file naming format	✓	✓	✓	—
Select injection options	✓	✓	✓	—
Create custom dye set	✓	— (to import, open a plate setup containing a custom dye set)	— (to import, open a plate setup containing a custom dye set)	—
Edit during a run or after a run is complete				
Monitor run	✓	—	—	✓
Pause or cancel a run	✓	—	—	✓
Edit injection parameters and re-inject samples	✓	—	—	✓
View results				
View run result details	✓	—	—	✓
View PUP score	✓	—	—	✓
View trace score (sequence analysis only)	✓	—	—	✓
View size quality (fragment/HID analysis only)	✓	—	—	✓

Feature	SeqStudio™ Data Collection Software	SeqStudio™ Plate Manager (desktop)	SeqStudio™ Plate Manager (Connect platform)	Remote Monitoring App
View Contiguous Read Length (sequence analysis only)	✓	—	—	✓
View and export results				
View real-time results	✓	—	—	✓
Adjust the graphical view	✓	—	—	✓
Auto export or manually export data files to the Connect platform, a network drive, or a USB	✓	—	—	—
Export a results report	✓	—	—	✓

SeqStudio™ Plate Manager overview

The SeqStudio™ Plate Manager is a stand-alone software. It allows you to set up and save plates that you can open and run on the instrument.

The SeqStudio™ Plate Manager is available:

- On the Connect cloud-based platform as an app, with access to the SeqStudio™ Remote Monitoring App
- At **thermofisher.com**, for download and installation on a computer
- On a USB, for installation on a computer
  Install the Plate Manager by following the instructions in the install wizard.



Figure 9 Plate Manager on the Connect cloud-based platform or desktop

Computer and operating system options

- PC with Windows™ 7
- PC with Windows™ 10
- Macintosh™ with OS X

Recommended browsers

The following browsers are recommended to use the Plate Manager app on the Connect cloud-based platform or a desktop computer:

- Mozilla™ Firefox™ Version 32.0.3+
- Google™ Chrome™ Version 38.02+
- Apple™ Safari™ Version 7+
- Microsoft™ Edge 10+ (Windows™ 10)
- Microsoft™ Internet Explorer™ 11 (Windows™ 7)

The use of other browsers or other versions can result in reduced functionality and improper display of information.

SeqStudio™ Remote Monitoring App overview

The Remote Monitoring App allows you to monitor the status of instrument runs from a remote location.

The Remote Monitoring App is available:

- As an app on the Connect cloud-based platform, with direct access from the Plate Manager or the InstrumentConnect.
- At **thermofisher.com**, for download and installation on a mobile device.



Figure 10 Remote Monitoring App on the Connect cloud-based platform

Network connection options

The SeqStudio™ Genetic Analyzer can be connected to a network or computer in the following configurations:

Direct connection	Local area network (LAN) connection	Connect platform connection
Wired	Wired or wireless	Wired or wireless

Network and password security requirements

Network configuration and security

The network configuration and security settings of your laboratory or facility (such as firewalls, anti-virus software, network passwords) are the sole responsibility of your facility administrator, IT, and security personnel. This product does not provide any network or security configuration files, utilities, or instructions.

If external or network drives are connected to the software, it is the responsibility of your IT personnel to ensure that such drives are configured and secured correctly to prevent data corruption or loss. It is the responsibility of your facility administrator, IT, and security personnel to prevent the use of any unsecured ports (such as USB, Ethernet) and ensure that the system security is maintained.

Password security

Thermo Fisher Scientific strongly recommends that you maintain unique passwords for all accounts in use on this product. All passwords should be reset upon first sign in to the product. Change passwords according to your organization's password policy.

It is the sole responsibility of your IT personnel to develop and enforce secure use of passwords.

Experiment types

Sequencing

Sequencing is the determination of the base-pair sequence of a DNA fragment by the formation of extension products of various lengths amplified through PCR.

For more information, see *DNA Sequencing by Capillary Electrophoresis Chemistry Guide Second Edition* (Pub. No. 4305080).



Fragment/HID analysis

Fragment/HID analysis is the determination of the size of fragments. It uses the size standard in each sample to create a standard curve for each sample. It then determines the relative size of each dye-labeled fragment in the sample by comparing fragments with the standard curve for that specific sample.

For more information, see *DNA Fragment Analysis by Capillary Electrophoresis User Guide* (Pub. No. 4474504).



Data output

SeqStudio™ Data Collection Software generates an electropherogram (intensity plot) for each dye that is based on the migration of DNA fragments through the capillaries during a run.

The format of the sample data files is determined by the type of experiment that is specified at the time of plate setup.

- Sequencing experiments use basecalling (the algorithms and settings required to determine the fragment base sequences) and generate an `AB1` file.
- Fragment/HID analysis experiments use sizecalling (the algorithms and settings required to determine the fragment sizes) and generate an `FSA` file.

The `AB1` and `FSA` file formats can be analyzed by secondary analysis software.

Secondary analysis software

Secondary analysis software is available for desktop computers and on your Connect cloud-based platform account.

Visit **thermofisher.com/connect** for the latest available secondary analysis applications.

Note: Data from the SeqStudio™ Genetic Analyzer may be labeled as "3200" in secondary analysis software.

Secondary analysis apps on the Connect cloud-based platform

Analysis	App	Description
Sequencing	Quality Check (QC) module	• Automatically checks sequence trace quality. • Provides a results summary that is based on quality parameter settings. • Auto-flags lower-quality traces for further inspection.
	Variant Analysis (VA) module	• Finds variants in samples that are sequenced on Applied Biosystems™ genetic analyzers. • Reports variants at genomic coordinates. • Allows export of variant calls in standard Variant Call Format.
	Next-generation Confirmation (NGC) module	• Confirms next-generation sequencing (NGS) variants using CE technology. • Allows visualization of the variants that are detected by both NGS and CE platforms. • Allows export of confirmed variants in standard Variant Call Format.

Analysis	App	Description
Fragment analysis	Sizing Analysis Module Peak Scanner™ Software	Performs peak sizing.
	Microsatellite Analysis Software	Analyzes a mixture of DNA fragments, separated by size, on supported capillary electrophoresis systems.

Desktop secondary analysis software

IMPORTANT! Older versions of the desktop secondary analysis software cannot analyze data files generated by the SeqStudio™ Genetic Analyzer. Contact Support for information on obtaining the latest versions of software.

Analysis	Software	Minimum version required
Sequencing	Sequencing Analysis Software	6.2
	SeqScape™ Software	3.2
	Variant Reporter™ Software	2.2
	Minor Variant Finder Software	1.2
Fragment analysis	GeneMapper™ Software	5.1
HID analysis	GeneMapper™ *ID-X* Software	1.6

Workflow: Connect cloud-based platform or desktop

Get started
Prepare the instrument (page 39)
Prepare the samples (page 37)
Create a plate setup on the Connect cloud-based platform or desktop
Access the Plate Manager on the Connect cloud-based platform (page 69)
Access the Plate Manager on the desktop (page 69)
Create or open a plate setup PSM file (page 69)
Enter plate properties (page 70)
Assign wells: Sample and run information (page 71)
Start and monitor a run
On the instrument: Load the plate or the tube assembly (page 92)
Select a plate setup and start a run (page 93)
Monitor a run from the Connect cloud-based platform (page 95)
Monitor a run from a mobile device (page 103)
Monitor a run from the instrument (page 105)
View and analyze results
View results in the Remote Monitoring App on the Connect cloud-based platform (page 99)
View results on the instrument (page 111)
Export results from the instrument (sample data files and QC reports) (page 121)
Analyze data in a secondary analysis software application

Workflow: instrument

Get started
Prepare the instrument (page 39)
Prepare the samples (page 37)
Create a plate setup on the instrument
Create or import a plate setup (page 81)
Enter plate properties (page 82)
Assign wells: run module, size standard, dye set, and kit (page 84)
Assign wells: sample name, sample type, and custom fields (page 86)
Start and monitor a run
On the instrument: Load the plate or the tube assembly (page 92)
Select a plate setup and start a run (page 93)
Monitor a run from the instrument (page 105)
View and analyze results
View results on the instrument (page 111)
Export results from the instrument (sample data files and QC reports) (page 121)
Analyze data in a secondary analysis software application

2 Prepare the samples and the instrument

- Precautions for use 33
- Power on the instrument 34
- Sign in 34
- Sign in with the Guest instrument profile 35
- Sign out 35
- Parts of the home screen 36
- Prepare the samples 37
- Prepare the instrument 39

Precautions for use

CAUTION! PHYSICAL INJURY HAZARD. Do not remove the instrument cover. There are no components inside the instrument that you can safely service yourself. If you suspect a problem, contact technical support.

CAUTION! Moving parts.

CAUTION! FIRE HAZARD. For continued protection against the risk of fire, replace fuses only with listed and certified fuses of the same type and rating as those currently in the instrument.

CAUTION! Hot surface.

CAUTION! Piercing hazard.

CAUTION! Potential biohazard.

CAUTION! Risk of electrical shock.

Power on the instrument

IMPORTANT! Do not power on the instrument until it has been installed and set up by a Thermo Fisher Scientific representative.

Press the On/Off switch on the rear panel.



IMPORTANT! The instrument does not maintain the correct temperature conditions for the cartridge when it is powered off. Avoid cartridge exposure to ambient temperature.

① On/Off switch

Sign in

1. If another user is signed in, touch ⊛ in the home screen, then touch **Sign out**.
2. In the **Sign in** screen, touch the down arrow.
3. Touch your instrument profile, then enter your PIN.

   Note: If the instrument is left unattended for 120 minutes, the instrument profile is signed out.



Sign in with the Guest instrument profile

1. In the **Sign In** screen, touch the down arrow.
2. In the bottom left of the screen, touch the **Guest** button. No PIN is required.

Sign out

In the home screen:

1. Touch ⊛.
2. Touch **Sign out**.

   Note: If a run is in progress, **Lock the instrument** is displayed instead of **Sign out**.
3. Touch **Yes** to confirm.

Parts of the home screen



① Eject icon
② Help
③ Instrument name
④ Status dial—Touch to create a plate setup.
⑤ **Settings**—Touch to view previous results (**Run History**) or configure the instrument.
⑥ Status of consumables—See "Check the consumables status" on page 39.
⑦ Current user
⑧ Connectivity icons

Connectivity icon	Status
or	The instrument is connected to a wired or wireless network.
	The instrument is connected to network drive.
	The current instrument profile is linked to the Connect cloud-based platform.

Prepare the samples

Sample preparation guidelines

Item	Guidelines
Plates and tubes 	• Use MicroAmp™ Optical 96-Well Reaction Plate or MicroAmp™ Reaction Tubes with a tray and retainer set. **IMPORTANT!** Fast plates are not compatible with the SeqStudio™ Genetic Analyzer. Fast plates will damage the cartridge. • Use the appropriate septa for plates and tubes. • See "Required materials not supplied" on page 220 for more information.
Fragment and HID analysis sample preparation 	• Prepare the samples as recommended by the kit for fragment analysis. • Use a 10–20 µL sample volume. • Ensure that Hi-Di™ Formamide is fresh. – Hi-Di™ Formamide should not undergo more than two freeze-thaw cycles (one to aliquot and one for use). – Use the same day after thawing. • For more information, see *DNA Fragment Analysis by Capillary Electrophoresis User Guide* (Pub. No. 4474504).
Sequence analysis sample preparation 	• Prepare sequencing reactions according to kit instructions, and purify the extension products with ethanol precipitation, spin columns, or the BigDye XTerminator™ Purification Kit. – If ethanol precipitation or spin columns are used, dry the samples in a vacuum centrifuge without heat or at low heat for 10–15 minutes or until dry. **Note:** Do not over dry the samples. – Resuspend in 10–20 µL of Hi-Di™ Formamide. • Use a 65 µL or 130 µL sample volume for samples that are prepared with the BigDye XTerminator™ Purification Kit. See *BigDye XTerminator™ Purification Kit Protocol* (Pub. No. 4374408). **IMPORTANT!** Use the appropriate run modules for samples prepared with BigDye XTerminator™ Purification Kit. See "Run modules, read lengths, size ranges, and run times" on page 140. • Use a 10–20 µL sample volume for samples that are prepared with Hi-Di™ Formamide. • Ensure that Hi-Di™ Formamide is fresh. – Hi-Di™ Formamide should not undergo more than two freeze-thaw cycles (one to aliquot and one for use). – Use the same day after thawing. • Do not resuspend samples in water, which can decrease sample stability. • For more information, see *DNA Sequencing by Capillary Electrophoresis Chemistry Guide Second Edition* (Pub. No. 4305080).

Plate layout and loading guidelines

- Samples are stable for 16–24 hours on the instrument.
- Load a maximum of 48 samples per plate if you use a long run module (Long Seq, Long Seq BDX, and Long Frag Analysis). The long fragment analysis run modules can take >24 hours to run an entire 96-well plate.
- Add samples to plates in columns. The default injection order is: A1–D1, E1–H1, A2–D2, E2–H2....A12–D12, E12–H12.



① Injection group 1, wells A1–D1
② Injection group 2, wells E1–H1
③ Injection group 3, wells A2–D2
④ Injection group 4, wells E2–H2

Prepare the plate

On a clean and level surface:

1. Prepare the sample according to your application protocol, then pipet the sample into the plate.
2. Place a septum onto the plate.
   a. Align the holes of the septa with the wells.
   b. Press gently until the septum is inserted into position in each well.
3. Centrifuge the plate assembly briefly to collect the contents at the bottom of each well.

   Centrifuge the plate assembly again if the contents are not at the bottom of the wells.
4. Heat-denature as required.
5. Place the plate on ice for 3 minutes.



① Plate
② Septum

Load the plate onto the instrument immediately or keep the plate on ice and protected from light until it is loaded onto the instrument.

Prepare the tubes

On a clean and level surface:

1. Place the tubes in the MicroAmp™ 96-well tray, then place the tray retainer over the tubes.
2. Prepare the sample according to your application protocol, then pipet the sample into the tubes.
3. Place a septum on the tubes.
   a. Align the holes of the septa with the tubes.
   b. Press gently until the septum are inserted into position in each tube.



① Septum
② 96-well retainer
③ 8-tube strip
④ MicroAmp™ 96-well tray

4. Centrifuge the tube assembly briefly to collect the contents at the bottom of each tube.

   Centrifuge the tube assembly again if the contents are not at the bottom of the tubes.
5. Heat-denature as required.
6. Place the tube assembly on ice for 3 minutes.

Load the tube assembly onto the instrument immediately or keep the tubes on ice and protected from light until they are loaded onto the instrument.

Prepare the instrument

Check the consumables status

1. Touch ⚙ **Settings** ▸ **Consumable status**.
2. Ensure that:
   - Sufficient consumables are installed for the run.
   - Installed consumables have not exceeded their expiry date.



Display	Cartridge	Cathode buffer
NA	Not installed on the instrument.	Not installed on the instrument.
White	OK for use.	OK for use.
Yellow	OK for use, but: • 1–25 injections remain for the capillary array, *or* • It will exceed the manufacturer's expiry date within 2 weeks, *or* • It will exceed the maximum number of days allowed on the instrument within 2 weeks[1]	OK for use, but: • 1–25 injections remain for the cathode buffer, *or* • It will exceed the manufacturer's expiry date within 2 days *or* • It will exceed the maximum number of days allowed on the instrument (14 days) within 2 days
Red	The cartridge has expired: • No injections remain for the capillary array, *or* • It has exceeded the manufacturer's expiry date, *or* • It has exceeded the maximum number of days allowed on the instrument[1]	The buffer has expired: • No injections remain for the cathode buffer, *or* • It has exceeded the manufacturer's expiry date *or* • It has been installed on the instrument for >14 days

[1] The maximum number of days allowed on the instrument is 120 days for the SeqStudio™ Genetic Analyzer Cartridge (Cat. No. A33671) and 180 days for the SeqStudio™ Genetic Analyzer Cartridge v2 (Cat. No. A41331). For more information, see Table 1.

See "Insert the Cathode Buffer Container" on page 167 if a new Cathode Buffer Container is required.
See "Insert the cartridge" on page 161 if a new cartridge is required.

3. Touch **Close**, then touch ⊖.

Load the CBC, the sample plate, and the cartridge

The **Eject plate** command is disabled for a few minutes after you insert a cartridge. If you are loading the CBC, sample plate, and cartridge at the same time, you can save time by loading the CBC and sample plate before you insert the cartridge.

For information on loading the individual components, see:

- "Install cathode buffer" on page 166
- "Load the plate or the tube assembly" on page 92
- "Insert the cartridge" on page 161

In the home screen:

1. Touch ⏏, touch ⠿ **Eject plate**, then open the instrument door when prompted.
2. Press the release button on the autosampler to open the lid, then remove the CBC.



① Release button
② Location of CBC

3. Check the buffer fill level:
   a. Remove the CBC.
   b. Ensure that the level of buffer is above the fill line.

      If the buffer is at or below the fill line, see "Assemble the SeqStudio™ Genetic Analyzer Cathode Buffer Container (CBC) " on page 166 and "Insert the Cathode Buffer Container" on page 167.

      If the buffer is above the fill line, reinsert the CBC.



① Replace if buffer is at or below the fill line

① New CBC buffer level

4. Place the plate or tube assembly firmly in the autosampler.

5. Close the autosampler lid: Press down on the center of the lid or press down on both sides of the lid with equal pressure until the lid clicks shut.



6. Touch **Retract plate**, then close the instrument door when prompted.
7. Touch **Eject cartridge**, then open the instrument door when prompted.
8. Hold the cartridge at the hand hold above the capillaries, then pull to remove it from the instrument.
9. Insert a new cartridge (see “Insert the cartridge“ on page 161).
10. Close the instrument door.

3 Use the instrument with the Connect cloud-based platform

- Understanding instrument and Connect cloud-based platform interaction . . . 43
- Connect cloud-based platform administrators for an instrument 48
- Register and obtain a Connect cloud-based platform account 50
- Link the instrument to your Connect cloud-based platform account 51
- Change your own Connect cloud-based platform instrument profile PIN 52
- Set up email notifications from the instrument 53

Note: The Connect cloud-based platform is not supported for HID analysis.

Understanding instrument and Connect cloud-based platform interaction

"Connect" versus "link"

The words "connect" and "link" are used interchangeably in the software.

In one location you touch a **Connect** button, in another location you touch a **Link** button.

Both actions do the same thing:

- Connect the instrument to the InstrumentConnect on the Connect cloud-based platform *and*
- Link your local instrument profile to your Connect cloud-based platform account.

First user who links is assigned administrator role

The first user who links the instrument to the Connect cloud-based platform is automatically assigned the Connect platform administrator role for the instrument (even if the user has a standard local profile).

Additional instrument administrators can be assigned, and user roles can be changed after linking.

For more information, see “Connect cloud-based platform instrument profile roles and functions“ on page 47 and “Connect cloud-based platform administrators for an instrument“ on page 48.

Local versus Connect cloud-based platform instrument profiles

- Local instrument profile—All plates and results are stored on the instrument under a local instrument profile.
- Connect cloud-based platform instrument profile—When you link your local instrument profile to your Connect platform account, a Connect platform instrument profile is created. With a Connect platform instrument profile:
  - You can save results and data files directly to your Connect platform account, access plate setups that you create in the SeqStudio™ Plate Manager on the Connect platform or desktop, and monitor instrument runs from the Connect platform.
  - All plate setups, data files, and results are automatically copied to your Connect platform account if the plate setup **Save location** is set to **Cloud**.

If you link when you *are* signed in to the instrument

In this scenario, your local instrument profile name is created manually on the instrument before you link. Your local instrument profile name differs from your Connect cloud-based platform instrument profile name.

If you link to the Connect platform when you *are* signed in to the instrument:

Phase	Steps that occur
Before you link:	• You enter your local instrument profile name in the **Sign In** screen.  • Your local instrument profile (*UserABC*) is displayed in the home screen of the instrument. • All plates and results that you create are accessible only when you are signed in with your local instrument profile.
When you link:	• You use an option described in "Link the instrument to your Connect cloud-based platform account" on page 136 to link. • If this is the first time you link, a Connect platform instrument profile is created using the *FirstNameLastInitial* of the user name from your thermofisher.com account. Example: *User1@thermofisher.com* First name is *User*, Last name is *Gray*. Connect platform account username is *User G*. • Your local instrument profile (*UserABC*) is linked to your Connect platform account (*User1@thermofisher.com*). • Your Connect platform instrument profile (*User G.*) replaces your local instrument profile.

Phase	Steps that occur
After you link:	• Your Connect platform instrument profile (*User G.*) and ☁ is displayed in the home screen of the instrument. • Plates and results from your local instrument profile can be copied to the Connect platform (see "Export results from the instrument (sample data files and QC reports)" on page 121). • New plates and results are saved under your Connect platform instrument profile. • Your Connect platform instrument profile name (*User G.* ☁) is available for selection in the **Sign In** screen.
If your Connect platform account is unlinked:	• Your local instrument profile (*UserABC*) is displayed in the home screen of the instrument. • Plates and results that were saved under your Connect platform instrument profile are accessible under your local instrument profile. • New plates and results are saved under local instrument profile and can be copied to the Connect platform (see "Export results from the instrument (sample data files and QC reports)" on page 121). • Your local instrument profile name (*UserABC*) is available for selection in the **Sign In** screen.

If you link when you *are not* signed in to the instrument

In this scenario, your local instrument profile name is created automatically on the instrument before you link. The same user name is used for your local instrument profile and your Connect cloud-based platform instrument profile. Plates and results are accessible when you sign in with either profile.

If you link to the Connect platform when you *are not* signed in to the instrument:

Phase	Steps that occur
Before you link:	• In the **Sign In** screen, you touch **Get Started ▸ Connect**. 

Phase	Steps that occur
When you link:	• You use an option described in "Link the instrument to your Connect cloud-based platform account" on page 136 to link. • If this is the first time you link, a Connect platform instrument and a local instrument profile (with standard role) are created with the same name using the *FirstNameLastInitial* of the user name from your thermofisher.com account. Example: *User1@thermofisher.com* First name is *User*, Last name is *Gray*. Connect platform account username is *User G*. • Your local instrument profile (*User G.*) is linked to your Connect platform account (*User1@thermofisher.com*). • Your Connect platform instrument profile (*User G.*) replaces your local instrument profile.
After you link:	• Your Connect platform instrument profile (*User G.*) and ☁ is displayed in the home screen of the instrument. • Plates and results from your local instrument profile can be copied to the Connect platform (see "Export results from the instrument (sample data files and QC reports)" on page 121). • New plates and results are saved under your Connect platform instrument profile. • Your Connect platform instrument profile name (*User G.* ☁) is available for selection in the **Sign In** screen.
If your Connect platform account is unlinked:	• Your local instrument profile (*User G.*) is displayed in the home screen of the instrument. • Plates and results that were saved under your Connect platform instrument profile are accessible under your local instrument profile. • New plates and results are saved under local instrument profile and can be copied to the Connect platform (see "Export results from the instrument (sample data files and QC reports)" on page 121). • Your local instrument profile name (*User G.*) is available for selection in the **Sign In** screen.

IMPORTANT! If you sign in with a local profile, without linking to the Connect cloud-based platform, sign out, then link using **Get Started ▸ Connect**, you can potentially have two local instrument profiles with different names. Plates are results created when you are signed in with one local instrument profile are not accessible when you are signed in with the other local instrument profile.

Connect cloud-based platform instrument profile roles and functions

Instrument profile	Location	Functions allowed
Standard	Connect platform[1]	• Create, save, open, import, and run plate setups • Create and modify run settings • View and export results
Administrator	Connect platform[1]	All the permissions of a local administrator profile, plus the following functions performed in InstrumentConnect: • Access the **Manage users** function on the Connect cloud-based platform to see a list of all instrument profiles that are linked to the instrument. • Assign the Connect platform administrator role to one or more users. • Remove a user from an instrument. • Disconnect the instrument from InstrumentConnect. • Change the instrument name.

[1] The first user who links their local instrument profile to their Connect platform account is assigned a Connect platform profile with administrator role.

For more information on using the Connect cloud-based platform

In the top left of the screen you are viewing in the Connect cloud-based platform, click ≡, then select **Help guide**.



Connect cloud-based platform administrators for an instrument

Connect cloud-based platform administrator functions

The first user who links the instrument to the Connect cloud-based platform is automatically assigned the Connect platform administrator role for the instrument (even if the user has a standard local profile).

At least one Connect platform administrator is required for each instrument.

A Connect platform administrator can perform the following tasks from the InstrumentConnect:

- Access the **Manage users** function on the Connect cloud-based platform to see a list of all instrument profiles that are linked to the instrument.
- Assign the Connect platform administrator role to one or more users.
- Remove a user from an instrument.
- Disconnect the instrument from InstrumentConnect.
- Change the instrument name.

Assign instrument administrator role to other users

A Connect cloud-based platform administrator for an instrument can assign instrument administrator role to other users.

1. Sign in to **thermofisher.com/connect**.
2. Click to access InstrumentConnect.
3. Select the instrument.



Note: The **Manage users** and other administrator functions are not displayed until you select an instrument.

4. To assign the Admin role to a user, click **Manage users**, then select the Administrator check box for the user.

Manage the users and administrators of your instrument

Any user with a Connect cloud-based platform administrator profile can manage users for an instrument or disconnect an instrument from InstrumentConnect.

If a Connect cloud-based platform administrator...	The software...
Assigns Admin role to a user	Allows the user to perform all Connect cloud-based platform administrator functions (see "Connect cloud-based platform administrator functions" on page 48).
Removes a user	Unlinks the instrument from their Connect cloud-based platform account.
Disconnects the instrument	• Unlinks the instrument from all user Connect cloud-based platform accounts. • Removes the instrument from InstrumentConnect.

1. Sign in to **apps.thermofisher.com**.
2. Click to access InstrumentConnect.
3. Select the instrument.



Note: The **Manage users** and other administrator functions are not displayed until you select an instrument.

4. To assign the Admin role to a user or to remove a user, click **Manage users**, then:

To...	Do this...
Assign the Admin role to an additional user	Select the **Admin** checkbox, then click **Close**.
Remove a user	Click , then click **Confirm**.

Disconnect individual users from an instrument

You cannot disconnect individual users from an instrument.

To *disconnect* a user, or to unlink the instrument from a Connect cloud-based platform account, you must disconnect the instrument from InstrumentConnect. Doing so unlinks *all* instrument profiles and removes the instrument from InstrumentConnect.

You can *remove* a user from the instrument, however, doing so deletes the user data from the instrument.

For more information, see "Manage the users and administrators of your instrument" on page 49

Register and obtain a Connect cloud-based platform account

1. Go to **www.thermofisher.com**.
2. On the home page, select **Sign In** ▸ **Register**.



3. Fill in all information, then click **Create account**.

Link the instrument to your Connect cloud-based platform account

Note: For detailed information on linking the instrument to your Connect cloud-based platform account, see Appendix B, "Link the instrument to your Connect cloud-based platform account—detailed instructions".

1. If a user is signed in, touch ⊛, then touch **Sign out**.
2. In the **Sign In** screen, touch **Get started** ▸ **Connect**.



3. In the **Connect to the Cloud** screen, touch a connection option.



Option	Action
Mobile devices	**Note:** Before selecting this option, install and sign in to the InstrumentConnect app on your mobile device. In the **Connect to the Cloud** screen: 1. Touch **Mobile devices**. 2. Hold the camera on your mobile device over the QR code that is displayed on the touchscreen. 3. Click **Close**.
PC	In the **Connect to the Cloud** screen, a link code is displayed. On a computer: 1. Access the Connect platform. 2. Access InstrumentConnect. 3. Click **Add instrument**. 4. Select **SeqStudio**. 5. Enter the link code.

Change your own Connect cloud-based platform instrument profile PIN

1. Sign in to **thermofisher.com/connect**.
2. Click to access InstrumentConnect.
3. Click **Update PIN number**.



Set up email notifications from the instrument

When an instrument is linked to your Connect cloud-based platform account, email notifications are automatically sent to your Connect platform account email address.

Perform this procedure to disable any of the default notifications.

1. Sign in to the instrument with your Connect platform instrument profile and PIN.

2. In the home screen of the instrument, touch ⚙ **Settings** ▸ **Instrument settings** ▸ **Email notifications**.



Note: If you are signed in with a local instrument profile instead of a Connect platform instrument profile, the **Email notifications** button is not displayed on the **Instrument Settings** screen.

3. In the **Email notifications** screen, select or deselect the options for which you want to receive email notifications, then touch **Done**.



4 Use the instrument with the Security, Auditing, and E-signature (SAE) v2.0 module

- Overview of the Security, Auditing, and E-signature (SAE) v2.0 module components 56
- Configuring the SAE module 57
- Instrument functionality when SAE is enabled or disabled 58
- Connect the instrument to an SAE server (administrator only) 58
- Enable SAE on the instrument and specify the SAE server (administrator only) 59
- Disable SAE (administrator only) 59
- Sign in with SAE enabled 60
- Specify audit reasons 61
- E-signature requirements when you start a run 61
- Sign a plate setup 63
- View E-signature records for a plate setup 63
- Sign a dye calibration or an install run 64
- View E-signature records for a dye calibration or an install run 64
- Signing in after automatic screen locking 64
- Use the instrument when the SAE server is offline 65
- Auditing of imported plate setups 65
- SAE error messages and actions 66

Overview of the Security, Auditing, and E-signature (SAE) v2.0 module components

The Security, Auditing, and E-signature (SAE) v2.0 module includes three components:

- **SAE Administrator Console**—Tool that is used by an SAE administrator to configure the SAE module.
- **SAE server** (server)—Service that runs in the background and stores SAE settings, user accounts, audit records, and e-signature records. By default, the SAE server is installed on the same computer as the SAE Administrator Console.
- **SAE screens** (client)—Screens that are displayed in an application (sign in, audit, and e-signature) and that require user input. The SeqStudio™ Data Collection Software runs on the client.

The Security, Auditing, and E-signature (SAE) v2.0 module provides the following SAE functionality on the instrument:

- **System security**—Controls user sign in and access to functions
- **Auditing**—Tracks changes and actions performed by users.
- **E-signature** —Allows users to provide an electronic signature (user name and password) when performing certain functions.

Depending on the way that your SAE administrator configures these features:

- Some of the features and functions described in later chapters of this guide may not be accessible to you.
- You may see dialog boxes and prompts when you use the software (examples shown below).

Dialog box	Description
	You do not have permission to perform a function.

Dialog box	Description
	An action is set up for auditing and requires you to specify a reason for the action.
	An action is set up for electronic signature and allows you to enter your password to allow the action.

Configuring the SAE module

For information on configuration, see *SAE Administrator Console v2.0 User Guide* (Pub. No. MAN0017468).

Instrument functionality when SAE is enabled or disabled

To	Sign in to the instrument with profile type	Affect on instrument
Enable SAE	• Local instrument administrator to access the SAE screen • SAE administrator to enable SAE	• Sign in is limited to users with SAE accounts. Local instrument profiles cannot be used to run the instrument. • Auditing and E-signature functions are active (if they are enabled in the SAE Admin Console). • Plate setup and run functions for a user are determined by the SAE account settings. • Connect cloud-based platform and remote monitoring functions are disabled.
Disable SAE	• SAE administrator • Local instrument administrator	• Sign in is limited to users with local instrument profiles. • Auditing and E-signature functions are not performed. • Connect cloud-based platform and remote monitoring functions are enabled.

Connect the instrument to an SAE server (administrator only)

Before you connect, install the SAE Administrator Console on a computer with a static IP address.

1. Touch **Settings** ▸ **SAE settings** ▸ **Connection settings**.
2. Enter the **Server IP** of the computer on which the SAE Administrator Console is installed.
3. Use the default port number, or a port number you obtain from your network administrator.

Enable SAE on the instrument and specify the SAE server (administrator only)

This procedure requires a local administrator profile on the instrument and an SAE administrator account in the SAE Administrator Console.

1. Sign in to the instrument with a local administrator profile.
2. Touch ⚙ **Settings** ▸ **SAE**.



3. Touch **Connection settings**, then enter the **Server IP** and **Port** of the computer on which the SAE Administrator Console is installed.

   IMPORTANT! Specify a server with a static IP address.

4. Touch **Enable SAE**.
5. Enter your SAE administrator account username and password, then touch **Enable SAE**.

Disable SAE (administrator only)

This procedure requires a local administrator profile or an SAE administrator account.

- If you are signed in to the instrument as an SAE administrator:
  a. In the home screen, touch ⚙ **Settings** ▸ **SAE** ▸ **Disable SAE**.
  b. Enter your SAE administrator account username and password.
- If you are not signed in to the instrument:
  a. If another user is signed in, touch 👤 in the home screen, then touch **Sign out**.

b. In the **Sign In** screen under **Local sign in**, touch **Sign in**.

c. Enter your local administrator profile name and PIN.

d. In the **Settings** screen, touch **SAE**, then touch **Disable SAE**.

Sign in with SAE enabled

If the instrument is left unattended for 120 minutes, the instrument profile is signed out.

1. If the another user is signed in, touch ⊛ in the home screen, then touch **Sign out**.

2. In the **Sign in** screen, touch **Sign In**, then enter your SAE username and password.

To access	Touch Sign in under
All instrument functions allowed by your SAE account	SAE
Administrator access to functions on the **Settings** screen	**Local** (requires an administrator instrument profile)



The home screen is displayed.

Specify audit reasons

Depending on the way that your SAE administrator configures audit settings, the **Enter Audit Reason** screen may be displayed when you make changes to a plate setup or a run module.

Select a reason or add a custom reason.

Note: The **Custom Reason** button is not displayed if audit settings are configured to require users to select a reason.



E-signature requirements when you start a run

Depending on the way your SAE administrator has configured your instrument, E-signatures may be required when you start a run. If starting a run is configured to require signatures, a message is displayed when you start a run.

Touch **View Details** to show the pending signatures.

The **Object Name** column lists the items that require signatures:

If the Object Name lists	The object is	For information, see
A dye set name	The most recent dye calibration for the dye set specified in the plate setup.	"Sign a dye calibration or an install run" on page 64
An install standard name	The most recent install run for the application specified in the plate setup.	"Sign a dye calibration or an install run" on page 64
A plate name	Plate setup	"Sign a plate setup" on page 63

The example below indicates that E-signature is configured to require 4 signatures when you start a run, 1 signature is complete and 3 signatures are pending:

- 2 administrator signatures for dye calibration (dye set name is listed two times):
  - 1 signature that specifies the "Reviewed and approved" E-signature meaning (complete, because 1 signature is required [Signatures Required=1] and 1 signature has been logged [Current=1]).
  - 1 signature that specifies the "Checked consumables" meaning (incomplete).
- 1 administrator signature for the plate setup that specifies the "Reviewed and approved" E-signature meaning (incomplete).
- 1 administrator signature for install run that specifies the "Reviewed and approved" E-signature meaning (incomplete).



Sign a plate setup

1. Create or open a plate setup, make changes as needed, then save the plate setup.
2. In the Plate Properties tab, touch **More Options**, then touch the **Sign Plate** button at the bottom left of the screen.
3. Select a **Signature meaning**.
4. Enter your password, then touch **Sign**.



View E-signature records for a plate setup

1. Open a plate setup.
2. In the **Plate Properties** tab, touch **More Options**, then touch the **Signing records**.

Status	Meaning
Current	The plate has not been modified after the plate was signed.
Obsolete	The plate has been modified after the plate was signed.

Sign a dye calibration or an install run

1. Touch ⚙ **Settings** ▸ **Maintenance and Service**.
2. Touch **Calibration** ▸ **Calibration history** or **Install run** ▸ **Install run history**.
3. Touch a dye set or install run to view the data.
4. Touch **Select to sign**, then select an item.
5. Touch **Sign**, enter your password, then touch **Sign**.
   A signed item is labeled with . Touch the to display all signature records for the item.

View E-signature records for a dye calibration or an install run

1. Touch ⚙ **Settings** ▸ **Maintenance and Services**.
2. Touch **Calibration** ▸ **Calibration history** or **Install run** ▸ **Install run history**.
3. Touch a dye set or install run to view the data.
4. Touch **Select to sign**, then select an item.
5. Touch the to display all signature records for the item.

Signing in after automatic screen locking

Depending on the way your SAE administrator has configured your instrument, the instrument touchscreen may automatically lock after a specified duration.

The users who can sign in when the screen is locked are determined by run status:

If a run is	These users can sign in
In process	• User who started the run • Administrator • Users with Log into Timed Out User Sessions permission
Complete	Any user

Use the instrument when the SAE server is offline

If your SAE administrator has configured your instrument to allow use when the SAE server is offline (**Client offline login** System setting in the SAE module), you can use the instrument for the period of time specified by the SAE administrator for **Client offline login**.

Note: If you have not previously signed in to the instrument with your SAE account, you cannot sign in when the SAE server is offline.

All SAE records are retained if the instrument is disconnected from an SAE server. When the instrument is reconnected to the SAE server, SAE records are uploaded to the server.

The following functions are not available when the SAE server is offline:

- Account lockout, password reminder, mandatory password change
- Disable SAE
- Change Password

Auditing of imported plate setups

If you import a plate setup, the following auditing occurs.

Note: You can import only one plate setup at a time.

If the file was created on...	And your instrument has SAE enabled and is configured to...	Then the following occurs
A different instrument with SAE enabled	——	• The plate is imported. • Auditing is restarted.
A different instrument with SAE disabled	**Open files from non-SAE systems**=**Forbidden**	• An "Incomplete audit trail" message is displayed. • The plate is not imported.
A different instrument, and the plate setup was: • Created with SAE or plate auditing enabled • Modified when SAE or plate auditing was disabled	**Open files from non-SAE systems**=**Allowed**	• An "Incomplete audit trail" message is displayed. • You can touch **Yes** to import the plate.

SAE error messages and actions

Message	Possible cause	Action
Not authorized to...	Your SAE account does not specify permission to perform the function.	Contact your SAE administrator.
Unable to connect to SAE server. Check current connections.	The SAE server connection settings are incorrect.	1. Check the SAE server IP address (see "Determine IP address for a computer on a network" on page 130). 2. In the instrument **Sign In screen**, sign in with a local administrator account. 3. Set the correct IP address (**Settings ▸ SAE settings ▸ Connection settings**).
	There is a problem with the computer on which the SAE Administrator Console is installed or a problem with the network.	Troubleshoot computer or network problems.
	The computer on which the SAE Administrator Console has a dynamic IP address that is disconnecting the server when the computer is restarted.	Set a static IP address on the computer.
Incomplete audit trail (when you import a plate setup)	There is a gap in the audit trail. For example, the plate setup was created with SAE enabled. The plate setup was then modified with SAE disabled.	If your instrument is configured with **Open files from non-SAE systems**=**Allowed**, you can import the plate setup.

5 Create or modify a plate setup from the Plate Manager

- Overview of plate setup settings 67
- Set up a plate using default settings (Plate Manager) 69
- Additional plate settings (Plate Manager) 74

Overview of plate setup settings

Category	Setting	Notes
Plate 	• Plate properties • Plate set up security (Shared or Hidden) • File name convention • Analysis settings	The analysis settings are saved with the plate setup as follows: • Hidden plates—The analysis settings are saved with the plate setup, and are used the next time that you open or create a hidden plate. The analysis settings will not be available to other system users. • Shared plates—The analysis settings are saved with the plate setup, and are used the next time that you open or create a shared plate. The analysis settings will be available to other system users. User-created analysis settings: • Users can name and save analysis settings. • If you create analysis settings in the Plate Manager (desktop), the settings will be available to other users. • If you create analysis settings in the Plate Manager (Connect cloud-based platform), the settings will not be available to other users.

Category	Setting	Notes
Injection group (a set of 4 wells) ○ ○ ○ ○	• *(Fragment/HID analysis only)* Size standard • Run module • Application type (mixed plate only) • Dye set • *(HID analysis only)* Kit (for marker-to-marker pull-up reduction)	User-created size standards and dye sets are accessible to all users. User-created run modules: • Users can name and save run modules. • If you create run modules in the Plate Manager (desktop), the modules will be available to other users. • If you create run modules in the Plate Manager (Connect cloud-based platform), the modules will not be available to other users.
Well ○	• Sample name • *(Fragment/HID analysis only)* Sample type • *(Sequencing only)* Specimen and amplicon • Custom fields 1–5	• **Specimen** and **Amplicon** fields are useful in secondary analysis software applications that organize sample data files based on amplicon and specimen information. • Custom fields are text fields in which you can include additional sample attributes or identifiers. Custom fields can be used by some secondary analysis applications.

Set up a plate using default settings (Plate Manager)

Access the Plate Manager on the Connect cloud-based platform

1. Sign in to **thermofisher.com/connect**.
2. In the **My Apps** list, select **SeqStudio™ Plate Manager**.
   If SeqStudio™ Plate Manager is not listed under **My Apps**, scroll down in the **All Apps** list.



Access the Plate Manager on the desktop

Select **Start** ▸ **All Programs** ▸ **Applied Biosystems** ▸ **Plate Manager** ▸ **Plate Manager**.

Create or open a plate setup PSM file

1. Click PM to display the home screen.
2. In the **Plate setup** screen, create or open a plate setup:
   If you are running the Plate Manager on the Connect cloud-based platform:

Click...	To...
New	Create a new plate setup or to create a plate setup from a template.
Open from cloud	Open a plate setup that you created in Plate Manager on your Connect cloud-based platform account.
Open from local drive	Open a plate setup that you created in Plate Manager on your computer (PSM file) or in another application (CSV file).

If you are running the Plate Manager on the desktop:

Click...	To...
New	Create a new plate setup or to create a plate setup from a template.
Open	Open a plate setup that you created in Plate Manager (PSM file) or in another application (CSV file).

Enter plate properties

In the **Properties** tab:

1. *(Optional)* Edit the **Plate name**, **Barcode**, or **Owner**.



2. Select an option in the **Plate setup security** field.
   - **Hidden**—Prevents other users from using or accessing the plate on the instrument. Your analysis settings are saved with the plate setup, and are used the next time that you open or create a hidden plate. However, the analysis settings will not be available to other system users.
   - **Shared**—Allows other users to access and edit the plate on the instrument. Your analysis settings are saved with the plate setup, and are used the next time that you open or create a shared plate. Your analysis settings will be available to other system users.

3. Select the **Application type**: **Sequencing**, **Fragment analysis**, **HID**, or **Mixed plate (sequencing & fragment)**.

   A mixed plate allows you to specify fragment analysis and sequence analysis settings on the same plate.

4. Select the analysis settings. If needed, edit the settings (see "Edit analysis settings" on page 75).

5. *(Optional for Sequencing or Mixed plate)* Select the **I am analyzing my data with Sanger variant analysis software** checkbox.

   

   The amplicon and specimen fields are added to the **Plate** view, and the attributes are automatically added to the default file name conventions (see "Modify the default file name convention" on page 88).

This feature is useful in secondary analysis software applications that organize files based on amplicon and specimen information (Connect cloud-based platform applications: Variant Analysis (VA) module, Next-generation Confirmation (NGC) module; desktop applications: SeqScape™ Software, Variant Reporter™ Software, Minor Variant Finder Software).

6. Click **Next**.

Assign wells: Sample and run information

This procedure describes using the **Plate** view. Alternatively, you can click ≣ and use keyboard short-cuts to enter information. See "Keyboard shortcuts for the sample table" on page 72.

1. In the **Plate** screen, select one or more injection groups: Click a well to select one injection group; shift+click or click+drag to select multiple injection groups.

   Each set of 4 wells on the plate is referred to as an injection group.

   The default injection order is: A1-D1, E1-H1, A2-D2, E2-H2....A12-D12, E12-H12.



2. If you are creating a mixed plate, select the **Application type** for the selected injection groups.

3. *(Optional)* Edit the default **Sample name** (well) for each well.

   Note: If multiple injection groups are selected when you edit the **Sample name** or **Sample type**, *both wells* with the same capillary number are edited. For example, if injection groups A1–D1 and E1–H1 are selected, and you change the

Sample name to "Test" for well E1, the **Sample name** for well A1 is also changed to "Test".

4. *(Fragment/HID analysis only)* Select the **Sample type** for each well: **Allelic ladder**, **Negative control**, **Positive control**, or **Sample**.
5. *(Fragment/HID analysis only)* Select a size standard for the selected injection groups.
6. *(HID analysis only)* Select a kit for the selected injection groups (enables marker-to-marker pull-up reduction).
7. Select a **Dye set** for the selected injection groups.
8. Select a **Run module** for the selected injection groups.

   Note: **Run module 1** is used for the first injection you specify. If you specify replicate injections, additional run module fields are added. For information, see "Specify replicate injections" on page 87.

   For more information, see "Run modules, read lengths, size ranges, and run times" on page 140.
9. Click **Next** and proceed to "Save a plate setup in the Plate Manager" on page 73.

Keyboard shortcuts for the sample table

Keys	Action
Navigation	
Arrow keys	Move to the cell above, below, to the right, or to the left of the current cell.
Tab	Move to the cell to the right of the current cell.
Shift+Tab	Move to the cell to the left of the current cell.
Home	Move to the first cell in the row.
End	Move to the last cell in the row.
Ctrl+Home	Move to the first cell in the column.
Ctrl+ End	Move to the last cell in the column.
Selection	
Ctrl+A	Select all.
Shift+Arrow keys	Extend the selection of the cell above, below, to the right, or to the left of the current cell.
Shift+Home	Select all cells in the row to the right including the current cell.
Shift+End	Select all cells in the row to the left including the current cell.
Ctrl+Shift+Home	Select all cells in the column to the top cell including the current cell.

Keys	Action
Ctrl+ Shift+End	Select all cells in the column to the bottom cell including the current cell.
Editor	
Enter	Open and close the cell editor.
F2	Clear the cell contents and open the cell editor.
Esc	Cancel editing and close the cell editor.
Backspace	Delete.
Delete	Clear the cell contents.
Ctrl+C	Copy.
Ctrl+X	Cut.
Ctrl+V	Paste.
Ctrl+Enter	Fill all selected cells with edited cell's value: Select a range of cells, then press **F2** to edit the first cell in the selection. Type a value, then press **Ctrl+Enter** to fill all selected cells with the value.
Ctrl+Z	Undo.
Ctrl+Y	Redo.

Save a plate setup in the Plate Manager

After you assign wells to a plate setup:

1. In the **Save the plate setup** dialog box, modify any settings as needed.



2. Click **Save**.



Note: The **Monitor my run button** is available only in the Connect cloud-based platform app.

If you are running the Plate Manager on the...	The plate setup is saved as a...
Connect cloud-based platform	Plate setup that you can open from the Connect cloud-based platform and run on the instrument.
Desktop	PSM file that you can open from a network drive or a USB and run on the instrument.

Proceed to Chapter 7, "Start and monitor a run".

Additional plate settings (Plate Manager)

Specify replicate injections

In the **Plate** screen:

1. Select **Actions** ▸ **Add injection**.

2. Select one or more injection groups, then select a run module for the replicate injections.

   You can add up to 5 replicate injections (for a total of 6 injections). **Run module 1** is used for the first injection you specify. Additional run module fields are added for replicate injections.

To remove a replicate injections, select **Actions** ▸ **Remove injection**.

Edit analysis settings

Factory-installed items cannot be edited or deleted. To create a new item from a factory-installed item, copy, edit, then save the new item.

1. In the **Properties** tab, click **Setting Details**, or select **Actions ▸ Manage analysis settings**.
2. To create new analysis settings:
   a. Select the default analysis settings or user-created analysis settings, then click **Copy**.
   b. Enter a name and edit settings as needed (see "Fragment/HID analysis settings (size calling)" on page 126 or "Sequencing settings (base calling)" on page 129).
   c. Click **Save**.
3. As needed, select an analysis setting of interest, then click **Edit** or **Delete** (user-created settings only).
4. Click **Close**.

Modify the default file name convention

The file name convention determines how the data files (AB1 or FSA) associated with a plate are named.

The default file name convention is:

Application	Default settings
Fragment/HID analysis	`<well>_<sample name>_<sample type>_<date and timestamp>.fsa`
Sequence analysis	`<well>_<sample name>_<date and timestamp>.ab1`
Sequence analysis with the Sanger variant analysis option selected	`<well>_<sample name>_<amplicon>_<specimen>_<date and timestamp>.ab1`

To change the default settings, in the **Properties** tab:

1. Select **Actions ▸ Edit the file name convention**.
2. Click **File attributes**, then select the attributes to include in the data file name.

   Note: The timestamp attribute cannot be deselected; it is always included in the data file name.



For information on creating custom fields to include in file name conventions, see "Define custom fields" on page 89.

3. *(Optional)* Click-drag ≡ to move an attribute to another position.

4. Click **Save**.

Hide or share a plate (Plate setup security)

In the **Properties** screen:

Select:

- **Hidden**—Prevents other users from using or accessing the plate on the instrument. Your analysis settings are saved with the plate setup, and are used the next time that you open or create a hidden plate. However, the analysis settings will not be available to other system users.
- **Shared**—Allows other users to access and edit the plate on the instrument. Your analysis settings are saved with the plate setup, and are used the next time that you open or create a shared plate. Your analysis settings will be available to other system users.

Define custom fields

Custom fields are text fields in which you can include additional sample attributes or identifiers. Custom fields can be used by some secondary analysis applications.

In the **Plate** tab:

1. Select **Action** ▸ **Add custom field**.
   You can add up to 5 custom fields.

2. Enter information in the custom field in the table at the right of the plate.

To remove a custom field, select **Actions** ▸ **Remove custom field**.

6 Create or modify a plate setup from the instrument

- PSM and CSV plate setup files for import into the instrument 78
- Shared (public), hidden (my plates), and guest plate setup files 79
- Overview of plate setup settings 79
- (Optional) Set up for auto export of sample data files (AB1 and FSA) 80
- Set up a plate using default settings (instrument) 81
- Set optional plate settings (instrument) 87

PSM and CSV plate setup files for import into the instrument

Format	Contains...	When to use...
PSM		
Create by: Saving a plate setup with the SeqStudio™ Plate Manager desktop or Connect cloud-based platform	• Size standard *names and definitions* • Run module *names and settings* • Dye set *names and settings* • Analysis settings *names and settings* • All remaining plate setup field settings: Well ID, Sample name, Application type and so on.	To create a plate setup that contains all the properties needed to save and run the plate setup. Note the following: • If the size standard, analysis settings, or run module in the PSM file do not exist on the instrument, they are automatically created when the PSM file is imported. • If the **Plate setup security** is set to **Hidden**: – The plate setup is available for selection on the instrument only to the user who created the PSM file.
CSV		
Create by: • Saving as CSV in the SeqStudio™ Plate Manager desktop or Connect cloud-based platform • Downloading a template from the SeqStudio™ Plate Manager desktop or Connect cloud-based platform	• Size standard *names only* • Dye set *name only* • Run module *names only* • All remaining plate setup field settings: Well ID, Sample name, Application type and so on.	To update properties of the plate setup (such as sample name) *after* the size standard, dye set, and run module for the plate setup have been created or imported on the instrument. When you import a CSV file: • If the size standard, dye set, or run module in the CSV file *do not* exist on the instrument: – An error message is displayed. – The CSV file is not imported on the instrument. • If the size standard or run module in the CSV file *do* exist on the instrument: – The CSV file is imported on the instrument. – The settings from the size standard, dye set, and run module that exist on the instrument are used (because the CSV file contains the names of these elements, but not the settings). – The **Plate setup security** is set to **Hidden**.

Shared (public), hidden (my plates), and guest plate setup files

Create by setting Plate setup security to...	Accessible to users...	Stored in folder on instrument...	Analysis settings used...	File name convention used...
Shared	All users, including guest	**Public**	Analysis settings saved in the plate setup	Last settings specified by the signed-in user (not saved with plate setup)
Note: If a guest user edits a shared plate setup, the plate must be saved under a new name, and the **Plate setup security** is set to **Hidden** and cannot be changed.				
Hidden	Only the user who created the plate setup	**My plates**	Last settings specified by the signed-in user (not saved with plate setup)	Last settings specified by the signed-in user (not saved with plate setup)
Guest plate setup security is automatically set to **Hidden** and cannot be changed	Guest user only	**My plates**	Last settings specified by the signed-in user (not saved with plate setup)	Last settings specified by the signed-in user (not saved with plate setup)

Overview of plate setup settings

Category	Setting	Notes
Plate 	• Plate properties • Plate set up security (Shared or Hidden) • File name convention • Analysis settings	The analysis settings are saved with the plate setup as follows: • Hidden plates—The analysis settings are saved with the plate setup, and are used the next time that you open or create a hidden plate. The analysis settings will not be available to other system users. • Shared plates—The analysis settings are saved with the plate setup, and are used the next time that you open or create a shared plate. The analysis settings will be available to other system users. User-created analysis settings: • Users can name and save analysis settings. • If you create analysis settings in the instrument software, the settings will not be available to other users.

Category	Setting	Notes
Injection group (a set of 4 wells) ○ ○ ○ ○	• *(Fragment/HID analysis only)* Size standard • Run module • Application type (mixed plate only) • Dye set • *(HID analysis only)* Kit (for marker-to-marker pull-up reduction)	User-created size standards and dye sets are accessible to all users. User-created run modules: • Users can name and save run modules. • If you create run modules in the instrument software, the modules will not be available to other users.
Well ○	• Sample name • *(Fragment/HID analysis only)* Sample type • *(Sequencing only)* Specimen and amplicon • Custom fields 1–5	• **Specimen** and **Amplicon** fields are useful in secondary analysis software applications that organize sample data files based on amplicon and specimen information. • Custom fields are text fields in which you can include additional sample attributes or identifiers. Custom fields can be used by some secondary analysis applications.

(Optional) Set up for auto export of sample data files (AB1 and FSA)

By default, sample data files (AB1 and FSA) are saved to the instrument.

When you create a plate setup, you can also set the **Save location** to **Cloud**, **Network Drive**, and/or **USB**.

1 Save location

When the plate is run, the instrument automatically exports the sample data files to the save locations.

Before you can select these save locations, set up the instrument:

- "Link the instrument to your Connect cloud-based platform account" on page 136
- "Connect to a network drive" on page 132
- Insert a USB into the USB port on the front of the instrument (Figure 1 on page 14)

Set up a plate using default settings (instrument)

Create or import a plate setup

In the home screen:

1. Touch **Set up run**.
2. Create, open, or import a plate setup:



To	Procedure
Create a new plate setup	1. Touch **Create new plate setup**. 2. See "Enter plate properties" on page 82.
Open an existing plate setup on the instrument	1. Touch **My Instrument**. 2. Touch: • **My plates** folder to select hidden plates that you have created. • **Public** folder to select (1) shared plates that were created by any user or (2) any plates that were created by a Guest user. For more information, see "Shared (public), hidden (my plates), and guest plate setup files" on page 79.
Import a plate setup	1. Touch **Cloud**, **USB**, or **Network Drive**. 2. Select: • PSM file • CSV file For more information, see "PSM and CSV plate setup files for import into the instrument" on page 78

Enter plate properties

1. At the top-right of the **Plate properties** screen, touch the **Properties** tab.
2. Touch the **Plate name** field, then enter the plate name.



3. Touch **Applications**, then select **Sequencing**, **Fragment analysis**, **HID**, or **Mixed plate** (allows you to specify fragment analysis and sequence analysis settings on the same plate).
4. *(Optional)* Touch the **Barcode** field, then use a scanner to scan the barcode.
5. *(Optional)* Touch the **Owner** field, then enter the plate owner name.
6. *(Optional)* Touch **More options** to check other settings (for example, **Plate setup security**, **Analysis settings**, and **File name convention**).
7. *(Optional)* Touch **Save location**, then select a location for the plate setup.

   The plate setup is always saved to the instrument. In addition, you can save the plate to the Connect cloud-based platform, a network, or a USB, which will auto export the sample data files.

   IMPORTANT! To view analyzed data in the Remote Monitoring App on the Connect platform, you must save the plate setup to the Connect platform.

   Note: *(Sequencing and Fragment analysis)* If you save a plate setup to the Connect cloud-based platform, a network, or a USB, then access the plate setup at a later time when the instrument is not linked to the Connect platform, a network, or a USB, the save location is displayed with strikethrough text.

① Original location to which the plate was saved, but is no longer accessible by the instrument.

8. *(Optional for Sequencing or Mixed plate)* Touch the **I am analyzing my data with Sanger variant analysis software** checkbox.

The amplicon and specimen fields are added to the **Plate** view, and the attributes are automatically added to the default file name conventions (see "Modify the default file name convention" on page 88).

This feature is useful in secondary analysis software applications that organize files based on amplicon and specimen information (Connect cloud-based platform applications: Variant Analysis (VA) module, Next-generation Confirmation (NGC) module; desktop applications: SeqScape™ Software, Variant Reporter™ Software, Minor Variant Finder Software).

Assign wells: run module, size standard, dye set, and kit

Note: You can assign these settings to one or more injection groups at the same time.

1. In the **Plate properties** screen, touch **Plate** at the top-right of the screen.



① Plate tab
② Save button is not enabled until you create an injection group

Each set of 4 wells on the plate is referred to as an injection group.

An injection group is identified by the first well in the set of 4 (for example, Injection Group A1 contains wells A1–D1).

The default injection order is: A1-D1, E1-H1, A2-D2, E2-H2....A12-D12, E12-H12.

2. Select injection groups, then touch **Edit**.
   - Touch a well to select a single injection group.
   - Touch and drag to select multiple injection groups or the entire plate.

Note: All settings in the remaining steps will be assigned to all selected injection groups. If different injection groups require different settings, repeat these steps for each injection group.



3. If you are creating a mixed plate, select the **Application type** for the wells.
4. Touch **Run modules**, then select a run module.
   For more information, see "Run modules, read lengths, size ranges, and run times" on page 140.
5. *(Fragment/HID analysis only)* Touch **Size standard**, then select a size standard for the injection group.
6. Swipe up to display the rest of the screen.
7. Touch **Dye set**, then select a dye set for the selected injection groups.
8. *(HID analysis only)* Touch **Kit**, then select a kit for the selected injection groups.
9. Touch **Done**.

Assign wells: sample name, sample type, and custom fields

Note: You can assign these settings to only one well at a time.

1. In the **Edit plate** screen, touch **Sample name** to display the well attributes fields.

2. Touch a setting, then enter the definition for the selected wells:
   - *(Fragment/HID analysis only)* **Sample type—Sample**, **Positive Control**, **Negative Control**, or **Allelic Ladder**.
   - **Custom fields**—Text fields to include additional sample attributes or identifiers that can be used by secondary analysis applications.
   - *(Sequencing only)* **Amplicon and Specimen**—Amplicon and Specimen names for Sanger Sequence analysis, if you selected the option in the **Plate Properties** screen.

3. Touch **Done** to close the screen then **Done** to close the **Edit Plate** screen.



① Application type—Sequencing, Fragment or HID; designated by S, F, or H
② Save the plate or start the run

4. Touch **Save** to save the plate to run at a later time, or touch **Start Run**.

Set optional plate settings (instrument)

Specify replicate injections

In the **Plate properties** screen:

1. Touch **Injection options**.
2. Touch an injection group, then touch **Edit and re-inject** to add replicates to the injection list.
3. *(Optional)* Modify **Run module**, **Injection time**, **Injection voltage**, **Run time**, or **Run voltage** for the injections.

   IMPORTANT! The changes to the run conditions apply only to the replicate injection. The changes are not saved to the run module.

4. Touch **Done**.

Modify analysis settings

In the **Properties** tab of the **Plate properties** screen:

1. Touch **More options** ▸ **Analysis settings**.
2. Select a setting.
   See "Manage analysis settings" on page 142 for detailed information.

Note: The last setting selected is used as the default for new plates.

3. Touch **Done**.

Modify the default file name convention

The default file name convention determines how the data files (AB1 or FSA) associated with a plate are named.

The default file name convention is:

Application	Default settings
Fragment/HID analysis	`<well>_<sample name>_<sample type>_<date and timestamp>.fsa`
Sequence analysis	`<well>_<sample name>_<date and timestamp>.ab1`
Sequence analysis with the Sanger variant analysis option selected	`<well>_<sample name>_<amplicon>_<specimen>_<date and timestamp>.ab1`

1. Access the **File name convention** screen:

From	Action
Plate properties screen	Select the **Properties** tab, then touch **More options** ▸ **File name convention**.
Home screen	Touch ⚙ **Settings** ▸ **Run settings** ▸ **File name convention**.

2. Touch **Attributes**.

3. Select the attributes to include in the data file name.



For information on creating custom fields to include in file name conventions, see "Define custom fields" on page 89.

Note: The timestamp attribute cannot be deselected; it is always included in the data file name.

4. Touch **Done**.

5. Touch and drag attributes up or down in the list.
6. Touch **Done**.

Hide or share a plate (Plate setup security)

For more information, see "Shared (public), hidden (my plates), and guest plate setup files" on page 79.

1. Touch **More Options ▸ Plate Setup security**.
2. Touch an option:
   - **Hidden**—Prevents other users from using or accessing the plate on the instrument. Your analysis settings are saved with the plate setup, and are used the next time that you open or create a hidden plate. However, the analysis settings will not be available to other system users.
   - **Shared**—Allows other users to access and edit the plate on the instrument. Your analysis settings are saved with the plate setup, and are used the next time that you open or create a shared plate. Your analysis settings will be available to other system users.
3. Touch **Done**, then touch **Save**.

Define custom fields

Custom fields are text fields in which you can include additional sample attributes or identifiers. Custom fields can be used by some secondary analysis applications.

In the **Plate** tab of the **Plate properties** screen:

1. Select injection groups, then touch **Edit**.
   - Touch a well to select a single injection group.
   - Touch and drag to select multiple injection groups or the entire plate.

2. Touch **Sample name** to display the injection group and well attributes fields.

① Touch the sample name field

3. Touch a custom field, then enter the definition for the selected wells.

4. Click **Done**.

(Optional) View the injection list, change injection settings or order, or specify replicates and re-injections

In the **Plate properties** screen:

1. Touch **Injection options**.
2. Touch an injection group, then configure the injection list:
   - Touch and drag an injection group to a new location in the injection list.
   - Touch **Inject first**—Moves the selected injection group to the top of the injection list.
   - Touch **Edit and re-inject**—Adds replicates or re-injections to the injection list. You can also modify **Run module**, **Injection time**, **Injection voltage**, **Run time**, or **Run voltage** for these injections.

     Note: *Replicate* indicates a duplicate—the replicate has the same parameters as the original injection. *Re-injection* indicates a copy of the injection, but with user-modified parameters.

     Note: The changes to the run conditions apply only to the replicate or re-injection. The changes are not saved to the run module.
3. Touch **Done**.

7 Start and monitor a run

- Load the plate or the tube assembly 92
- Select a plate setup and start a run 93
- Lock the touchscreen 95
- Monitor a run from the Connect cloud-based platform 95
- Monitor a run from a mobile device 103
- Monitor a run from the instrument 105
- Unload the plate or the tube assembly 109

Load the plate or the tube assembly

In the home screen:

1. Touch ⏏, touch ⠿ **Eject plate**, then open the instrument door when prompted.
2. Press the release button on the autosampler to open the lid.
3. Check the buffer fill level:
   a. Remove the CBC.
   b. Ensure that the level of buffer is above the fill line.
      If the buffer is at or below the fill line, see "Assemble the SeqStudio™ Genetic Analyzer Cathode Buffer Container (CBC) " on page 166 and "Insert the Cathode Buffer Container" on page 167.
      If the buffer is above the fill line, reinsert the CBC.



① Replace if buffer is at or below the fill line

① New CBC buffer level

4. Place the plate or tube assembly firmly in the autosampler.

5. Close the autosampler lid: Press down on the center of the lid or press down on both sides of the lid with equal pressure until the lid clicks shut.

6. Touch **Retract plate**, then close the instrument door when prompted.

Select a plate setup and start a run

After you load the plate in the instrument (see "Load the plate or the tube assembly" on page 92):

1. In the instrument home screen, touch **Setup run**.

2. Select the location of your plate setup, then select the plate setup.

   For information on plate setup files, see "PSM and CSV plate setup files for import into the instrument" on page 78 and "Shared (public), hidden (my plates), and guest plate setup files" on page 79.

3. Verify that settings are as needed.

   The **Save location** must specify **Cloud** if you want to view analyzed data when you monitor the run from the Connect cloud-based platform.



4. Touch **Start run**.

- If a new cartridge was inserted before the run, the instrument performs an optical alignment before starting the run.
- During the run, the instrument performs an automatic spectral calibration adjustment (auto calibration) for each sample to correct for spectral overlap.
- During a run, the user who started the run or an administrator can lock the touchscreen to prevent other users from using the instrument. If the touchscreen is locked, only the user who started the run or an administrator can sign in to the instrument.

Proceed to:

- "Monitor a run from the Connect cloud-based platform" on page 95
- "Monitor a run from a mobile device" on page 103
- "Monitor a run from the instrument" on page 105

Automatic file cleanup

Before starting a run, the instrument calculates the total amount of storage space required to save the run to the instrument. If the required storage space is not available, the instrument deletes files associated with the oldest exported plates until sufficient space is available.

Note: Only complete plates that have been auto exported (saved to Connect cloud-based platform, network, or USB) or manually exported (using ⚙ **Settings** ▸ **Run History** ▸ ***plate name*** ▸ **Export**) are deleted.

If the required storage space is not available and no plates have been exported, the instrument displays a notification indicating that there is not enough storage space.

You can export plates and delete plates, then start the run again.

Lock the touchscreen

During a run, you can lock the touchscreen to prevent other users from using the instrument. This feature is not available to Guest users.

Only the user who locked the touchscreen or an administrator can sign in to the instrument if the touchscreen is locked.

1. Touch ⊛.
2. Touch **Profile**.
3. Touch **Lock instrument**.

   Note: If a run is not in progress, **Sign out** is displayed instead of **Lock instrument**.

Monitor a run from the Connect cloud-based platform

Open the Remote Monitoring App from Instrument Connect App

A run is accessible from InstrumentConnect for 24 hours after the run is complete, or until another run is started.

1. Sign in to **thermofisher.com/connect**.
2. Click ▣ to access InstrumentConnect.



3. Click the run status dial to display the Remote Monitoring App.





Figure 11 Remote Monitoring App

Open the Remote Monitoring App from the Plate Manager

A run is accessible from the Plate Manager for 24 hours after the run is complete, or until another run is started.

You can open the Remote Monitoring App immediately after you save a plate setup or at a later time.

In the Plate Manager on the Connect cloud-based platform:

- To open the Remote Monitoring App immediately after you save a plate setup and start the run, click **Monitor my run**.



- To open the Remote Monitoring App at a later time, click PM, then select an instrument or click to access InstrumentConnect.



① Select an instrument or click to access InstrumentConnect.



Figure 12 Remote Monitoring App

View notifications from the instrument on your Connect cloud-based platform account

1. In any screen in the Connect cloud-based platform, click .

2. Click a notification, then click **Dismiss** or **Dismiss all** to dismiss the notification.

View results in the Remote Monitoring App on the Connect cloud-based platform

1. Open the Remote Monitoring App (see "Open the Remote Monitoring App from Instrument Connect App" on page 95).

2. Click an injection group in the injection list or the plate view.
   The status dials are color-coded for quality alerts:
   - —All QC tests passed.
   - —At least 1 warning quality alert was triggered.
   - —At least 1 failing quality alert was triggered.

   For information on quality alerts, see:
   - "Data quality alerts" on page 119
   - "Sizecalling and basecalling quality alerts" on page 119



   Note: The **Analyzed** tab is disabled if the **Save location** for the plate setup is not set to **Cloud** or if the injection group has not finished running.

3. In the **Quality alerts** screen, click the **Raw**, **EPT**, or **Analyzed** tab to view data.

4. As needed, select **Actions** ▸ **Re-inject group**, select the **Run module** and settings, then click **Inject**.

Pause or cancel an injection in the Remote Monitoring App

Select:

- **Actions** ▸ **Pause plate** to pause the run after the currect injection is complete.
- **Actions** ▸ **Stop current injection** to immediately stop the injection.

Edit injection group run settings

In the **Results** screen:

1. Select **Actions** ▸ **Edit injection group**.
2. Edit settings as needed, then click **OK**.

Re-inject or delete an injection group

In the **Results** screen:

Select **Actions** ▸ **Re-inject group** or **Actions** ▸ **Delete injection group**.

Export a QC report

The **QC Report** command is not available for an injection until the injection is complete. The **QC Report** command is not available for any injections if the plate setup **Save location** does not specify **Cloud**.

In the **Remote Monitor** screen:

Select **Actions** ▸ **QC Report**.
A PDF of the plate QC report is generated.

Remote Monitoring App raw trace



Figure 13 Fragment analysis raw trace

① Zoom in/out.
② Raw trace.
③ Thumbnail trace—Click-drag to view another region of the trace.
④ Lock/unlock trace zooming for all traces in the injection group.
⑤ View Options—Select the dye color to display; set vertical scaling.
⑥ Cursor position indicator (red line).



Figure 14 Sequence analysis raw trace

① Zoom in/out.
② Raw trace.
③ Lock/unlock trace zooming for all traces in the injection group.
④ View Options—Select the basecalls to display; set vertical scaling.

Remote Monitoring App EPT trace

The EPT view (ElectroPhoresis Telemetry) shows instrument data conditions (currents, temperatures, electrophoresis voltage) as a function of time.



Remote Monitoring App analyzed trace



Figure 15 Fragment analysis analyzed trace

1. Lock/unlock trace zooming for all traces in the injection group.
2. Zoom in/out.
3. Analyzed trace.
4. Thumbnail trace—Click-drag to view another region of the trace.
5. View Options—Select the dye colors to display; set vertical scaling.
6. Cursor position indicator (red vertical and horizontal tick marks outside trace).
7. Size standard curve (red line).
8. Actions—Select commands to pause and cancel injections.



Figure 16 Sequence analysis analyzed trace

1. Zoom in/out.
2. Analyzed trace.
3. Thumbnail trace—Click-drag to view another region of the trace.
4. Lock/unlock trace zooming for all traces in the injection group.
5. View Options—Select the basecalls to display; show/hide quality bars and values; set vertical scaling.
6. Search for a sequence.

Monitor a run from a mobile device

Before you begin, see "Link the instrument from a mobile device" on page 185.

1. On your mobile device, launch InstrumentConnect.
2. Touch the instrument to monitor.



3. Swipe left to view consumable status.



4. Touch the status dial to view the injection list.



5. Touch an injection group to display quality alerts, then touch **View raw plot** to view the data.





- Swipe left to view the entire trace.
- Pinch-zoom to expand the trace.

Monitor a run from the instrument

(Optional) View the injection list, change injection settings or order, or specify replicates and re-injections

In the **Plate properties** screen:

1. Touch **Injection options**.
2. Touch an injection group, then configure the injection list:
   - Touch and drag an injection group to a new location in the injection list.
   - Touch **Inject first**—Moves the selected injection group to the top of the injection list.
   - Touch **Edit and re-inject**—Adds replicates or re-injections to the injection list. You can also modify **Run module**, **Injection time**, **Injection voltage**, **Run time**, or **Run voltage** for these injections.

     Note: *Replicate* indicates a duplicate—the replicate has the same parameters as the original injection. *Re-injection* indicates a copy of the injection, but with user-modified parameters.

     Note: The changes to the run conditions apply only to the replicate or re-injection. The changes are not saved to the run module.
3. Touch **Done**.

View the run status

During a run, the run status screen is displayed.

View the run time information and the status dial for each capillary.



The status dials are color-coded for quality alerts:

- —All QC tests passed.
- —At least 1 warning quality alert was triggered.
- —At least 1 failing quality alert was triggered.

If an injection group is set to re-inject, the number of the current injection is displayed on the status dials.

For information on quality alerts, see:

- "Data quality alerts" on page 119
- "Sizecalling and basecalling quality alerts" on page 119

View real-time results

During a run, touch one of the injection dials to display the trace for the selected capillary.

See "Fragment/HID analysis trace" on page 113 or "Sequence analysis trace" on page 115 for information.

The status dials are color-coded for quality alerts:

- —All QC tests passed.
- —At least 1 warning quality alert was triggered.
- —At least 1 failing quality alert was triggered.

For information on quality alerts, see:

- "Data quality alerts" on page 119
- "Sizecalling and basecalling quality alerts" on page 119

Pause a plate or cancel or stop injections

1. In the run status screen, touch ⊙ **Actions**.



2. Manage the plate or injections:

Touch	To
View results	View the list of injections and status.
Edit plate	For injections that have not yet been run, edit **Sample Name**, **Run Module**, **Dye Set**, **Size Standard**, **Sample Type**, and custom fields.
Pause plate	Stop the run after the current injection is complete. Touch **Resume** to continue the run.
Stop current injection	Stop the current injection immediately. Touch **Resume** to continue the run.
Cancel remaining injections	Specify whether to cancel the run immediately or after the current injection is complete.
Injection options	Move an injection to the top of the injection list, edit run module information, and/or reinject samples.

View results when the run is completed

View results for the plate when the run is completed (all injections are finished).

1. Touch **Results** to view the run results.



2. Touch **List view**.

   Each injection group displays a QC color for each capillary:

   - ● —All QC tests passed.
   - ● —At least 1 warning quality alert was triggered.
   - ● —At least 1 failing quality alert was triggered.

   For information on quality alerts, see:

   - "Data quality alerts" on page 119
   - "Sizecalling and basecalling quality alerts" on page 119

3. Touch an injection group.

4. View the results in the **Run Result Details** screen, or touch for well details.
   See "Fragment/HID analysis results" on page 113 or "Sequence analysis results" on page 115 for detailed information.

5. Touch a sample file name.
   If the data triggered any quality alerts, a QC alerts screen is displayed. See "Data quality alerts" on page 119 and "Sizecalling and basecalling quality alerts" on page 119 for detailed information.
   Touch **View data** to display the analyzed trace for the sample. See "Fragment/HID analysis trace" on page 113 or "Sequence analysis trace" on page 115 for detailed information.

6. Touch and drag the thumbnail view of the analyzed trace (below the trace) to scroll left or right.

7. *(Optional)* Adjust the graphical view (see "Adjust the trace display" on page 108).

8. Touch > or < to scroll to the raw data or **EPT Plot** (ElectroPhoresis Telemetry).
   See "EPT plot" on page 117.

Adjust the trace display

See "View results when the run is completed" on page 107 to access results.

- Drag one finger to pan to the left or right.
- Zoom in and out by pinching and expanding with two fingers.
- Touch on the left border of the trace, then touch a dye to deselect.
- Touch on the right border of the trace, then touch **Zoom In**, **Zoom Out**, or **Fit to screen** to adjust the display.
- Drag the center of the pane in thumbnail view to scroll left or right.



- Drag the right or left handle of the pane to zoom horizontally.



Unload the plate or the tube assembly

When the run is complete:

1. Touch ⏏, touch ⠿ **Eject plate**, then open the instrument door when prompted.
2. Press the release button on the autosampler to open the lid.
3. Remove the plate or tube assembly.
4. Close the autosampler lid: Press down on the center of the lid or press down on both sides of the lid with equal pressure until the lid clicks shut.



5. Touch **Retract plate**, then close the instrument door when prompted.

8 View and analyze results

- View results in the Remote Monitoring App on the Connect cloud-based platform 110
- View results on the instrument 111
- Export results from the instrument (sample data files and QC reports) 121
- Analyze data 121

View results in the Remote Monitoring App on the Connect cloud-based platform

1. Open the Remote Monitoring App (see "Open the Remote Monitoring App from Instrument Connect App" on page 95).

2. Click an injection group in the injection list or the plate view.

   The status dials are color-coded for quality alerts:

   - —All QC tests passed.
   - —At least 1 warning quality alert was triggered.
   - —At least 1 failing quality alert was triggered.

   For information on quality alerts, see:

   - "Data quality alerts" on page 119
   - "Sizecalling and basecalling quality alerts" on page 119



   Note: The **Analyzed** tab is disabled if the **Save location** for the plate setup is not set to **Cloud** or if the injection group has not finished running.

3. In the **Quality alerts** screen, click the **Raw**, **EPT**, or **Analyzed** tab to view data.

4. As needed, select **Actions** ▸ **Re-inject group**, select the **Run module** and settings, then click **Inject**.

View results on the instrument

View results when the run is completed

View results for the plate when the run is completed (all injections are finished).

1. Touch **Results** to view the run results.



2. Touch **List view**.
   Each injection group displays a QC color for each capillary:
   - —All QC tests passed.
   - —At least 1 warning quality alert was triggered.
   - —At least 1 failing quality alert was triggered.

   For information on quality alerts, see:
   - "Data quality alerts" on page 119
   - "Sizecalling and basecalling quality alerts" on page 119

3. Touch an injection group.

4. View the results in the **Run Result Details** screen, or touch for well details.
   See "Fragment/HID analysis results" on page 113 or "Sequence analysis results" on page 115 for detailed information.

5. Touch a sample file name.
   If the data triggered any quality alerts, a QC alerts screen is displayed. See "Data quality alerts" on page 119 and "Sizecalling and basecalling quality alerts" on page 119 for detailed information.
   Touch **View data** to display the analyzed trace for the sample. See "Fragment/HID analysis trace" on page 113 or "Sequence analysis trace" on page 115 for detailed information.

6. Touch and drag the thumbnail view of the analyzed trace (below the trace) to scroll left or right.

7. *(Optional)* Adjust the graphical view (see "Adjust the trace display" on page 108).

8. Touch > or < to scroll to the raw data or **EPT Plot** (ElectroPhoresis Telemetry). See "EPT plot" on page 117.

Adjust the trace display

See "View results when the run is completed" on page 107 to access results.

- Drag one finger to pan to the left or right.
- Zoom in and out by pinching and expanding with two fingers.
- Touch ⋮ on the left border of the trace, then touch a dye to deselect.
- Touch ⋮ on the right border of the trace, then touch **Zoom In**, **Zoom Out**, or **Fit to screen** to adjust the display.
- Drag the center of the pane in thumbnail view to scroll left or right.



- Drag the right or left handle of the pane to zoom horizontally.



Fragment/HID analysis results

Fragment/HID analysis trace



① Trace color hide/show—Touch to open, then touch a color to hide or show.
② Analyzed trace
③ Size standard curve (red line)
④ Thumbnail trace—Drag the center of the pane in the thumbnail trace to display another trace area in the top pane. Drag the right or left handle of the pane to zoom horizontally.
⑤ Basepair or scan display selection.
⑥ Zoom tools—Touch to open.
⑦ Next trace tool—Touch to view the raw trace or EPT for the well.

Fragment/HID analysis results

Setting	Description
Quality	QC flags • —All QC tests passed. • —At least 1 warning quality alert was triggered. • —At least 1 failing quality alert was triggered.

Setting	Description
SQ	Size Quality • If **SQ** is ≤0.75, it passes the QC test and does not trigger a quality alert ( ). • If **SQ** is 0.25–0.74, it triggers a warning quality alert ( ). • If **SQ** is <.25, it triggers a failing quality alert ( ). **How Size Quality is determined** The Size Quality algorithm evaluates the similarity between the fragment pattern for the size standard dye specified in the size standard definition and the actual distribution of size standard peaks in the sample, calculates an interim SQ (a value between 0 and 1).
Exported	The sample has been auto exported (saved to the Connect cloud-based platform, network, or USB) or manually exported (using **Settings** ▸ **Run History** ▸ ***plate name*** ▸ **Export**).
	Well details (see below).
Well details	
Sample file name, QC flag, sample name, sample type, settings used to acquire the data (size standard, run module and dye set), Sizing quality value, and export status.	

Sequence analysis results

Sequence analysis trace



① Quality Value bars and values:

- Pure base with QV ≥20
- Pure base with QV 15–19
- Pure base with QV <15
- Mixed base

② Bases—Mixed bases are highlighted in red. You can set the secondary peak threshold, as a percentage of the primary peak, for consideration as a mixed base by the basecalling algorithm. Reaching this threshold is a necessary but not sufficient condition for arriving at a mixed base determination. See "Sequencing settings (base calling)" on page 129.

③ Trace color hide/show—Touch to open, then touch a color to hide or show.

④ Analyzed trace

⑤ Thumbnail trace—Drag the center of the pane in the thumbnail trace to display another trace area in the top pane. Drag the right or left handle of the pane to zoom horizontally.

⑥ Zoom tools—Touch to open.

⑦ Next trace tool—Touch to view the raw trace or EPT for the well.

Sequencing results

Result	Description
Quality	QC flags • —All QC tests passed. • —At least 1 warning quality alert was triggered. • —At least 1 failing quality alert was triggered.

Result	Description
CRL (Contiguous Read Length)	The longest uninterrupted segment of bases with an average Quality Value (QV) ≥20. In addition to evaluating the QV of a base call, the software considers the QV of adjacent bases within a ±21-bp moving average to determine a contiguous read length based on quality values. **Note:** The contiguous read length passing criteria for install checks is an uninterrupted segment of bases with an average Quality Value (QV) of 30.
Exported	The sample has been auto exported (saved to the Connect cloud-based platform, network, or USB) or manually exported (using ⚙ **Settings** ▸ **Run History** ▸ ***plate name*** ▸ **Export**).
	Well details (see below).
Well details	
Sample file name, QC flag, sample name, CRL, signal strength, Median PUP, Trace score, run module and dye set used to acquire the data, and export status.	
Signal strength	The average relative fluorescence unit (RFU) for all dyes across the electropherogram in the raw data.
Trace score	The average basecall Quality Value (QV) of bases in the clear range sequence of a trace. The *clear range* is the region of the sequence that remains after excluding the low-quality or error-prone sequence at the 5′ and 3′ ends. The clear range is calculated by the KB basecaller using QVs.
Median PUP (pull-up peak)	A measure of noise or pull-up that is determined by taking the mean of the ratios of signal strength calculated for each base-called peak: primary peak/secondary peak under the primary peak. A higher value indicates less baseline or secondary noise. A lower value indicates an elevated baseline or secondary noise. Example 1: Main called base signal strength is 1,000 RFU and the largest secondary peak beneath it is 10 RFU; PuP=100 Example 2: Main called base signal strength is 1,000 RFU and the largest secondary peak beneath it is 100 RFU; PuP=10

Understanding Quality Values (QVs)

Quality value ranges

The color of a QV bar indicates the QV of a base.

- Pure base with QV ≥20
- Pure base with QV 15–19
- Pure base with QV <15
- Mixed base

Pure base versus mixed base QVs

Pure bases and mixed bases have the same probability of error for the associated basecall ($10^{-q/10}$). Note the following:

- High-quality pure bases typically have QVs of 20 or higher.
- The distribution of quality values for mixed bases differs dramatically from that of pure bases.
- Mixed bases have a maximum QV of 20.
- Review all mixed base calls.

Quality values (QV) and probability of error (Pe)

QV	Pe	QV	Pe
1	79.0%	30	0.10%
5	32.0%	35	0.032%
10	10.0%	40	0.010%
15	3.2%	45	0.0032%
20	1.0%	50	0.0010%
25	0.32%	60	0.00010%

EPT plot

The EPT view (ElectroPhoresis Telemetry) shows instrument data conditions (currents, temperatures, electrophoresis voltage) as a function of time.



Touch ⋮ on the left border of the plot to display the legend.



View results for a previously run plate (run history)

In the home screen:

1. Touch ⚙ **Settings** ▸ **Run history**.
2. Touch a plate name, then touch **View**.
   If you select more than one plate name, the **View** button is dimmed.
   The **Run History** screen is displayed.



3. Touch a sample file name.
4. View the results in the **Run history** screen, or touch ℹ to view well details.
   See “Fragment/HID analysis results“ on page 113 or “Sequence analysis results“ on page 115 for information.

5. Touch a sample file name, then touch **View**.
   If you select more than one sample file name, the **View** button is dimmed.
   If the data triggered any quality alerts, a QC alerts screen is displayed.
   For information on quality alerts, see:
   Click **View data** to display the trace for the sample. See "Fragment/HID analysis trace" on page 113 or "Sequence analysis trace" on page 115.
6. Touch and drag the thumbnail view of the analyzed trace (below the trace) to scroll left or right.
7. Touch > or < to scroll to the raw data or **EPT Plot**.

Data quality alerts

Quality alert	Description	Action
Offscale peaks. Adjust the injection parameters and/or the sample concentration.	At least 10 scans have saturated the CCD camera.	• Reduce the injection voltage or time. • Dilute the sample.
No sample was detected.	Poor signal-to-noise ratio with low signal detected.	• Verify that the sample volume follows recommendations in the user manual. • Troubleshoot upstream PCR and sequencing steps.

Sizecalling and basecalling quality alerts

Table 5 Sizecalling quality alerts

Quality alert	Description	Action
Sizing quality value is low due to poor size standard peak quality. Peak height uniformity is low or the fitting quality in sizing is poor.	Low resolution or poor quality data is present.	• Re-inject the sample. • If the problem persists, check the sample quality.
Sizecaller found broad peak(s) in the size standard peak(s).		
Sizing quality value is in the intermediate range; check size standard data quality.		
The number of size standard peaks detected is less than what is defined in the size standard.	Size standard definition includes peaks that are not present in the sample. Example: Sample peaks are detected up to 500 bp, but the size standard definition includes peak sizes that are >500 bp.	Use or create a size standard definition with the appropriate number of peaks and peak sizes.

Quality alert	Description	Action
The analysis range is too small. Correct the analysis range in analysis settings and re-analyze in secondary analysis software or re-inject sample.	Various causes.	• Analyze the data in a secondary analysis software with a corrected analysis range. • Re-inject the sample.

Table 6 Basecalling quality alerts

Quality alert	Description	Action
Basecalling failed due to poor quality data.	Poor quality data is present.	• Re-inject the sample. • If the problem persists, prepare fresh sample. • Troubleshoot upstream PCR and sequencing steps.

Edit injection parameters and re-inject samples

You can edit injection parameters and re-inject samples during a run or after a run is complete.

1. Access **Injection options**.
   - During a run—Touch ⊙ **Actions** ▸ **Edit plate**.
   - After a run—Touch **Results**.
2. Touch an injection group, then configure the injection list:
   - Touch and drag an injection group to a new location in the injection list.
   - Touch **Inject first**—Moves the selected injection group to the top of the injection list.
   - Touch **Edit and re-inject**—Adds replicates or re-injections to the injection list. You can also modify **Run module**, **Injection time**, **Injection voltage**, **Run time**, or **Run voltage** for these injections.
3. Touch **Done**.

   Note: The changes are not applied until you touch **Done**.

Export a report (QC report)

This function allows you to export a QC report for the current plate. To export a QC report for a previously run plate, export a run history (see "Export results from the instrument (sample data files and QC reports)" on page 121).

When a run is complete, in the home screen:

1. Touch **Results**.
2. Touch **Export report**.
3. Select a storage location.
4. Navigate to, then select a location, then touch **Export**.

Export results from the instrument (sample data files and QC reports)

In the home screen:

1. Touch ⚙ **Settings** ▸ **Run history**.
2. Select one or more plates from the **Run History** table.
3. Touch **Export**.
4. Select a storage location.

The following data is exported for the plate:

- Fragment/HID analysis—FSA file for each sample.
- Sequencing—AB1 file for each sample.
- Plate QC report in CSV and PDF format.

Note: If you select a plate, select **View**, then select **Export**, only an FSA or AB1 file for each analyzed sample is exported.

Analyze data

1. Export results (see "Export results from the instrument (sample data files and QC reports)" on page 121) or use auto exported data.
2. Use an appropriate fragment analysis or sequencing application to analyze the data.

   Note: Data from the SeqStudio™ Genetic Analyzer may be labeled as "3200" in secondary analysis software.

   For more information, see "Secondary analysis software" on page 29

9 Manage the software (Plate Manager)

- Add a custom dye calibration to the Plate Manager 122
- Download a plate setup template as a CSV file 122
- Save a plate setup as a PDF 123
- Save a plate setup as a CSV file 123
- Create a plate setup template 123
- Delete plate setups 124
- Manage email notifications in the Connect cloud-based platform 124
- Manage run modules 125
- Manage size standards 125
- Manage analysis settings 126

Add a custom dye calibration to the Plate Manager

On the instrument:

1. Open a plate setup that specifies the custom dye set of interest.
2. Export a plate setup that specifies the custom dye.
3. Open the exported plate setup in the Plate Manager.
   The custom dye is imported and is available for selection when you create new plates.

Download a plate setup template as a CSV file

1. Click PM to display the home screen.
2. Click **New**, select **Create from template**, then click **Download** for the template of interest.

Note: You cannot add plate setups to the list of templates. However, you can create a plate setup with "Template" suffix, and use it as a starting point for creating new plate setups.

Save a plate setup as a PDF

In the **Plate** tab:

Select **Actions ▸ Print plate view**.

The plate view and table view are exported.

Save a plate setup as a CSV file

In the **Plate** tab:

1. Select **Actions ▸ Save as CSV**.
2. Navigate to and select a storage destination.

The plate setup is saved as a CSV file. The names of the dye set, run module, and size standard are included in the CSV file.

Create a plate setup template

Note: You cannot add templates to the list of templates you select when you create a plate setup. However, you can create a plate setup with "Template" suffix, then use it as a starting point for creating new plate setups.

1. Click **New**, select **Create from template**, then click **Download** for the template of interest.
2. Modify the CSV file as needed.
3. Save the plate setup with a "Template" suffix (example: Plate_FragAnalysis_Template), or other identifier.

Delete plate setups

To delete plate setups from the Plate Manager:

On the...	Use...
Desktop	Windows™ Explorer to delete PSM or CSV files. **Note:** The location shown in the figure is an example. Users can save PSM files in any location. 
Connect cloud-based platform	Click in the Connect platform, then use DataConnect to delete PSM files. 

Manage email notifications in the Connect cloud-based platform

1. In any screen in the Connect cloud-based platform, click .

2. Click **Settings**.
3. Select or deselect your email address.

Manage run modules

This function manages the list of run modules that you can select from when you create a plate setup.

To assign a run module to a plate, see "Assign wells: Sample and run information" on page 71.

Factory-installed items cannot be edited or deleted. To create a new item from a factory-installed item, copy, edit, then save the new item.

For more information, see "Run modules, read lengths, size ranges, and run times" on page 140.

1. In the **Plate** tab, select **Actions** ▸ **Manage run modules**.
2. To create a new run module:
   a. Select a default run module or a user-created run module, then click **Copy**.
   b. Enter a run module name.
   c. Enter values, click **Advanced** to enter additional settings, then click **Save**.

   See "Run module settings" on page 140 for detailed information.
3. As needed, select a run module of interest, then click **Edit** or **Delete** (user-created run modules only).

Manage size standards

This function manages the list of size standard definitions that you can select from during plate set up.

To assign a size standard definition to a plate, see "Assign wells: Sample and run information" on page 71.

A size standard defines the sizes in basepairs of known fragments. It is used to generate a standard curve. The standard curve is used to determine the sizing of fragments in unknown samples.

Factory-installed items cannot be edited or deleted. To create a new item from a factory-installed item, copy, edit, then save the new item.

Size standard definitions are accessible to all users.

1. In the **Plate** tab, select **Actions** ▸ **Manage size standards**.
2. To create a new size standard:
   a. Select a default size standard or a user-created size standard, then click **Copy**.

b. Enter a name and select a dye color, then edit the fragment sizes (basepairs).

c. Click **Save**.

3. As needed, select a size standard of interest, then click **Edit** or **Delete** (user-created size standards only).

Manage analysis settings

In the **Properties** tab:

1. Select **Actions** ▸ **Analysis settings**.

2. To create new analysis settings:
   a. Select the default analysis settings or user-created analysis settings, then click **Copy**.

   b. Enter a name and edit settings as needed (see "Fragment/HID analysis settings (size calling)" on page 126 or "Sequencing settings (base calling)" on page 129).

   c. Click **Save**.

3. As needed, select an analysis setting of interest, then click **Edit** or **Delete** (user-created settings only).

Fragment/HID analysis settings (size calling)

Setting	Description
Size calling method	• **Local Southern**—*(default)* Determines the fragment sizes using the reciprocal relationship between fragment length and electrophoretic mobility. • **Global Southern**—Compensates for standard fragments with anomalous electrophoretic mobility (similar to least squares methods). • **2nd LSQ** (2nd Order Least Squares)—Uses regression analysis to build a bestfit size calling curve. • **3rd LSQ** (3rd Order Least Squares)—Uses regression analysis to build a bestfit size calling curve. • **Cubic Spline Interpolation**—Forces the sizing curve through all the known points of the selected size standard.

Setting	Description
Analysis range	• **Full Range**—*(default)* To analyze the entire scan region as collected by the genetic analysis instrument, including the primer peak. • **Partial Range**—To analyze only data points within a specified range. Enter Start Point in data points after the primer peak and before the first required size standard peak. Enter a Stop Point after the last required size standard fragment. Start and Stop points may vary from instrument to instrument and platform to platform. View raw data to determine the appropriate analysis range. Data points outside the specified analysis range are ignored. **Note:** Ensure the **Analysis Range** contains all size standard fragments included in the **Sizing Range**.
Sizing range	The size range (in base pairs) appropriate for the kit you are using: • **Full Range** for the software to analyze fragments of all sizes in the **Analysis Range**. • **Partial Range** for the software to analyze only fragments within a specified range. Enter a **Start Size** and a **Stop Size** appropriate for the size standard used.
Peak amplitudes	The peak height threshold (RFU) for peak detection for each dye color. Peaks below the threshold are not detected. For example, if you use the default values of 175 RFU, peaks with heights equal to or greater than 175 RFU are detected. Peaks with heights below 175 RFU are still displayed in the electropherogram plots but are not detected or labeled. **Note:** Use the same peak amplitude thresholds in secondary analysis software.
Primer peak	If the primer peaks in your application obscure peaks of interest, select **Present**. This instructs the algorithm to ignore primer peaks. Primer peaks are still displayed in the trace. If this setting does not allow detection of the 20- and 40-mer peaks for samples that use the GS600 LIZ™ size standard, running samples with the GS600_LIZ_(60-600) or other size standards that include lower bp starting points may allow detection of the peaks.
Common settings	
Smoothing	Select an option to smooth the outline of peaks and reduce the number of false peaks detected: • **None** (default) to apply no smoothing. Best if the data display sharp, narrow peaks of interest. • **Light** to provide the best results for typical data. Light smoothing slightly reduces peak height. • **Heavy** for data with very sharp, narrow peaks of interest. Heavy smoothing can significantly reduce peak height.

Setting	Description
Baseline Window	Specify a window to adjust the baseline signals of all detected dye colors to the same level for an improved comparison of relative signal intensity. Note the following: • A small baseline window relative to the width of a cluster, or grouping of peaks spatially close to each other, can result in shorter peak heights. • Larger baseline windows relative to the peaks being detected can create an elevated baseline, resulting in peaks that are elevated or not resolved to the baseline.
Minimum Peak Half Width	Specify the minimum full peak width at half maximum **Peak Height** required for peak detection. The range is 2 to 99 data points.
Peak Window Size	Enter a window width in data points for peak detection sensitivity. If more than one peak apex is within the window, all are labeled as a single peak. Note the following: • The maximum value is the number of data points between peaks. • The **Peak Window Size** setting is limited to odd numbers. To increase peak detection sensitivity: Increase polynomial degree, decrease peak window size. To decrease peak detection sensitivity: Decrease polynomial degree, increase peak window size.
Polynomial Degree	**Polynomial Degree** cannot be greater than **Peak Window Size**. Adjust to affect the sensitivity of peak detection. You can adjust this parameter to detect a single base pair difference while minimizing the detection of shoulder effects and/or noise. The peak detector calculates the first derivative of a polynomial curve fitted to the data within a window that is centered on each data point in the analysis range. Using curves with larger polynomial degree values allows the curve to more closely approximate the signal and, therefore, captures more of the peak structure in the electropherogram.
Slope Thresholds Peak Start and End	• **Peak Start**—The peak starts when the first derivative (slope of the tangent) in the beginning of the peak signal before the inflection point becomes equal to or exceeds the **Peak Start** value. This threshold is set to 0 by default, which means that the peak will normally start at the leftmost point where the slope of the tangent is closest to 0° (horizontal line). A value other than 0 moves the peak start point toward its center. The value entered must be non-negative. • **Peak End**—The peak ends when the first derivative (slope of the tangent) in the end of the peak signal after the inflection point becomes equal to or exceeds the **Peak End** value. This value is set to 0 by default, which means that the peak will normally end at the rightmost point where the slope of the tangent is closest to 0° (horizontal line). A value other than 0 moves the peak end point toward its center. The value entered in this field must be non-positive. Using curves with larger polynomial degree values allows the curve to more closely approximate the signal and, therefore, captures more of the peak structure in the electropherogram.

Sequencing settings (base calling)

The default settings are optimized for sequencing of PCR amplicons from diploid genes, which are expected to have Mixed bases and which should end At PCR Stop.

For sequencing from plasmid templates, which are of pure sequence (no Mixed bases) and have longer sequence read times than PCR products, create new analysis settings and *disable* (uncheck) **At PCR Stop** and **Mixed base threshold** checkboxes.

Setting	Description
Quality Threshold	• Basecall assignment (ambiguous bases): – Do not assign Ns to basecalls – Assign Ns to basecalls with QV<5— Bases with a QV less than the threshold display N instead of the base letter • End base—Last base on which to perform basecalling: – At PCR Stop – After X number of bases – After X number of Ns in X number of bases – After X number of Ns **Note:** If you have PCR products with sequences that end while data is still being collected, select the **At PCR Stop** checkbox.
Mixed bases threshold	When enabled, determines the secondary peak height ratio where the secondary peak is considered a potential mixed base. Reaching the threshold is a necessary but not sufficient condition for the basecalling algorithm to call a mixed base.
Clear range methods	• **Use quality values**—Sets a window with a specified number of allowed low-quality bases by removing bases until there are <X number of bases per Z number of bases with QV <Y. • **Use base positions**—Specifies the first and last base in the range to consider, or trims the specified number of bases from the 3′ end. • **Mask base positions before**—Specifies the base position before which to disregard bases.

10 Manage the software (instrument)

- Connect the instrument to a network drive 130
- Link the instrument to your Connect cloud-based platform account 136
- Lock the touchscreen 137
- Manage plate setups 138
- Manage run settings (instrument) 139
- Manage instrument settings 147
- Manage instrument profiles on the instrument 151
- Manage storage space 153

Connect the instrument to a network drive

Determine IP address for a computer on a network

On the destination computer:

1. In the Windows™ desktop, click .
2. In the search field at the bottom of the pane, type **command prompt**, then press **Enter**.

3. At the command prompt, type **ipconfig**, then press **Enter**.

4. Note the IP address listed.

   Note: The location and number of digits in your IP address may differ from the **IPv4 Address** example shown below.



Create folders and enable network folder sharing

In the Windows™ desktop:

1. On a Windows™ computer, server, or network drive, create a folder to store your plates and results. Example: **C:/Users/*Your Name*/SharedData**.

2. Create subfolders in the **SharedData** folder. Example: **PlateSetups** and **Results**.



3. Right click on a folder, then select **Sharing ▸ Advanced** or **Share with ▸ Advanced sharing**.

4. Select the **Sharing** tab.

5. Click **Advanced sharing**.

6. Select **Share this folder**.

7. Click **Permissions**, then select **Full Control** or **Read**, **Write**, and **Delete** options.

   IMPORTANT! Without these permissions, the instrument cannot autoexport results when a run is complete.

8. Click **OK**, click **OK**, then click **Close**.

Connect to a network drive

See your laboratory administrator for the information you need to connect to a network drive.

From any screen that displays **Network drive** or **Save location** as an option, you can connect to the drive for the first time.

IMPORTANT! Before saving to the network drive, ensure that the folder is shared (see "Create folders and enable network folder sharing" on page 131).

---

1. Touch **Network drive** or **Save location** field.
2. Touch the **Destination** field, then touch the appropriate field to enter the IP address and shared folder name, then touch **Done**.

   For more information, see "Determine IP address for a computer on a network" on page 130.
3. If necessary, touch the appropriate fields to enter a domain name, username, and password.
4. Touch **Connect**.

Set up a default Connect cloud-based platform location for opening plate setups

Before you begin, create a plate setup (PSM file or CSV file) in the Connect cloud-based platform location that you want to set as the default location.

To set the default location, you must select a plate setup.

1. In the home screen of the instrument, touch **Setup run**, then touch **Cloud**.

   The **Setup Run** screen is displayed.



2. Select a folder (if needed), then select a plate file.

   The plate file is imported.
3. Touch to return to the **Setup Run** screen.

Set up a default Connect cloud-based platform location for saving results (auto export)

1. In the home screen of the instrument, touch **Setup run**, then touch **Create new plate setup**.
2. Touch **Save location**.
3. Touch the **Cloud Destination** field.
   The **Select Directory** screen is displayed.
4. Select or create a folder.
5. Touch **Done**.

Set up a default network location for opening plate setups

Before you begin, create a plate setup (PSM file or CSV file) in the shared directory on the computer that you want to set as the default location.

To set the default location, you must select a plate setup.

1. Determine the IP address of the computer on which you created shared folders (see "Determine IP address for a computer on a network" on page 130 and "Create folders and enable network folder sharing" on page 131).
2. In the home screen of the instrument, touch **Setup run**, then touch **Network Drive**.

If your instrument profile	This screen is displayed	How to proceed
Is not connected to a network drive	**Connect**	1. In the **Network Destination** field, enter the IP address of the computer followed by the folder names you created. Example: 10.43.32.82/SharedData/Results. 2. If required by your network, enter **Domain**, **User Name**, and **Password** to access the shared location. 3. Touch **Connect**. The **Setup Run** screen is displayed.
Is connected to a network drive	**Select Directory**	Proceed to step 3.

3. In the **Setup Run** screen:

Touch	To
Set up import	Change the **Network Destination** you specified in the previous step.
A folder name	Navigate to the location you want to set as the default.
A plate name	Import the plate.

The **Setup Run** screen is displayed.



Figure 17 Setup Run screen example

① Example IP address and folder locations
② List of plates (will be blank if you have not saved plates to this location)

Set up a default network location for saving results (auto export)

1. Determine the IP address of the computer on which you created shared folders (see "Determine IP address for a computer on a network" on page 130 and "Create folders and enable network folder sharing" on page 131).
2. In the home screen of the instrument, touch **Setup run**, then touch **Create new plate setup**.
3. Touch **Save location**.
4. Touch the network **Destination** field.

If your instrument profile	This screen is displayed	How to proceed
Is not connected to a network drive	**Connect**	1. In the **Network Destination** field, enter the IP address of the computer followed by the folder names you created. Example: 10.43.32.82/SharedData/Results. 2. If required by your network, enter **Domain**, **User Name**, and **Password** to access the shared location. 3. Touch **Connect**. The **Select Directory** screen is displayed.
Is connected to a network drive	**Select Directory**	Proceed to step 5.

5. In the **Select Directory** screen:

Touch	To
New folder	Create a subfolder.
Set up export	Change the **Network Destination** you specified in the previous step.
Select this folder	Select the directory shown at the top of the screen.



Figure 18 **Select Directory** screen example

① Example IP address and folder locations
② List of folders

6. In the **Save Destination** screen, touch **Done**.

Link the instrument to your Connect cloud-based platform account

Note: For detailed information on linking the instrument to your Connect cloud-based platform account, see Appendix B, "Link the instrument to your Connect cloud-based platform account—detailed instructions".

1. If a user is signed in, touch ⊛, then touch **Sign out**.
2. In the **Sign In** screen, touch **Get started** ▸ **Connect**.



3. In the **Connect to the Cloud** screen, touch a connection option.



Option	Action
Mobile devices	**Note:** Before selecting this option, install and sign in to the InstrumentConnect app on your mobile device. In the **Connect to the Cloud** screen: 1. Touch **Mobile devices**. 2. Hold the camera on your mobile device over the QR code that is displayed on the touchscreen. 3. Click **Close**.
PC	In the **Connect to the Cloud** screen, a link code is displayed. On a computer: 1. Access the Connect platform. 2. Access InstrumentConnect. 3. Click **Add instrument**. 4. Select **SeqStudio**. 5. Enter the link code.
Instrument	In the **Connect to the Cloud** screen, enter your account information, then click **Link account**.

Lock the touchscreen

During a run, you can lock the touchscreen to prevent other users from using the instrument. This feature is not available to Guest users.

Only the user who locked the touchscreen or an administrator can sign in to the instrument if the touchscreen is locked.

1. Touch .
2. Touch **Profile**.
3. Touch **Lock instrument**.

   Note: If a run is not in progress, **Sign out** is displayed instead of **Lock instrument**.

Manage plate setups

Export or delete a plate setup (PSM file)

1. Touch **Set up run**, then touch **My instrument**.
2. Touch **Manage** at the bottom left of the screen.



3. Touch a plate, then touch **Export** or **Delete**.
4. If you touched:
   - **Export**, select a storage location, then touch **Export**.
   - **Delete**, then touch **Yes** to delete the plate setup.

Import a plate setup from a CSV or PSM file

Note: You can create a plate setup in CSV or PSM file in the Plate Manager (desktop or Connect cloud-based platform).

Before importing a CSV file, see "PSM and CSV plate setup files for import into the instrument" on page 78.

In the home screen:

1. Touch **Set up run**, then touch **Cloud**, **Network**, or **USB**.



2. Navigate to, then select a CSV or PSM file.

Manage run settings (instrument)

Manage run modules

This function manages the list of run modules that you can select from when you create a plate setup.

To assign a run module to a plate, see "Assign wells: run module, size standard, dye set, and kit" on page 84.

Factory-installed items cannot be edited or deleted. To create a new item from a factory-installed item, copy, edit, then save the new item.

Run modules are accessible only to the user who creates them.

For more information, see "Run modules, read lengths, size ranges, and run times" on page 140.

1. Access the **Manage run modules** screen:

From	Action
Plate properties tab	Touch **More options** ▸ **Manage run modules**.
Home screen	Touch **Settings** ▸ **Run settings** ▸ **Run modules**.

2. To create a new run module:
   a. Touch a default run module or a user-created run module to use as a starting point, then touch **Copy**.
   b. Enter values, then touch **Next**.
   c. Enter a name, touch **Advanced** to enter additional settings, then **Done**.

   See "Run module settings" on page 140 for detailed information.

3. As needed, touch a run module of interest, then touch **Edit** or **View**.

   Note: The **Edit** button is dimmed if a run is in progress.

Run module settings

Setting	Description
Run module name	Name of the module
Capillary temperature (°C)	Temperature setting for the capillary array throughout run
Prerun voltage (Volts)	Voltage setting for pre-run before sample injection
Prerun time (seconds)	Prerun time
Injection voltage (Volts)	Voltage for sample injection
Injection time (seconds)	Sample injection time
Run voltage (Volts)	Final sample electrophoresis separation run voltage
Run ramp duration (seconds)	Time required to reach the **Run voltage** **Note:** Data collection does not start until this time elapses.
Run time (seconds)	Length of time that data is collected after the **Run ramp duration** elapses

Run modules, read lengths, size ranges, and run times

Table 7 Sequencing run modules for standard sequencing

Run module	Contiguous read length (CRL)[1]	QV threshold	Approximate run time
ShortSeq	≥350	QV30	30 minutes
ShortSeq_BDX			
MediumSeq	≥500	QV30	45 minutes
MediumSeq_BDX			

Run module	Contiguous read length (CRL)[1]	QV threshold	Approximate run time
LongSeq	≥800	QV20	~ 2 hours
LongSeq_BDX			

[1] CRL was determined using the Long Read Sequencing standard. A minimum of 90% of analyzed sequences with an average QV ≥QV threshold were observed.

IMPORTANT! Use BDX run modules only if you prepare samples with BigDye XTerminator™ Purification Kit. Use non-BDX run modules for samples purified with other methods.

Table 8 Fragment analysis run modules

Run module	Resolution range	Approximate run time	Sizing precision	Compatible size standards
SNaPshot	40–120 bp	25 minutes	40–120: <0.5	GeneScan™ 120 LIZ™ Size Standard
FragAnalysis	60–460 bp[1]	45 minutes	60–460: <0.15	All except GeneScan™ 1200 LIZ™ Size Standard
LongFragAnalysis[2]	60–600 bp[1]	< 2 hours	60–460: <0.15 461–600: <0.3 601–800: >0.45	• GeneScan™ 600 LIZ™ Size Standard v2.0 • GeneScan™ 1200 LIZ™ Size Standard

[1] Resolution Range: The range of bases over which the resolution (peak spacing interval divided by the peak width at half-max in a GS600 or GS1200 LIZ size standard sample sized with a third order fit) is ≥1. The table shows the resolution range in ≥90% of samples.
[2] Load a maximum of 48 samples per plate if you use a long run module.

Table 9 HID run modules

Run module	Resolution range	Approximate run time	Sizing precision	Compatible size standards
HIDAnalysis	60–470 bp[1]	39 minutes	60–470: <0.15	• GeneScan™ 500 LIZ™ Size Standard[2] • GeneScan™ 600 LIZ™ Size Standard v2.0

[1] Resolution Range: The range of bases over which the resolution (peak spacing interval divided by the peak width at half-max in a GS600 size standard sample sized with a third order fit) is ≥1.
[2] The GeneScan™ 500 LIZ™ Size Standard was not included in the HID validation of the instrument.

Note: The following size standards have not been validated for use with the instrument. A default size standard definition is not provided in the software.
- GeneScan™ 350 ROX™ Size Standard
- GeneScan™ 400HD ROX™ dye Size Standard

Manage size standard definitions

This function manages the list of size standard definitions that you can select from during plate set up.

To assign a size standard definition to a plate, see "Assign wells: run module, size standard, dye set, and kit" on page 84.

A size standard defines the sizes in basepairs of known fragments. It is used to generate a standard curve. The standard curve is used to determine the sizing of fragments in unknown samples.

Factory-installed items cannot be edited or deleted. To create a new item from a factory-installed item, copy, edit, then save the new item.

Size standard definitions are accessible to all users.

1. Access the **Manage size standards** screen:

From	Actions
Plate properties tab	Touch **More options** ▸ **Manage size standards**.
Home screen	Touch ⚙ **Settings** ▸ **Run settings** ▸ **Size standards**.

2. To create a new size standard:
   a. Touch a default size standard or a user-created size standard, then touch **Copy**.
   b. Enter a name and select a dye color.
   c. As needed, touch any of the following, then touch **Done**:

      Note: To change an existing value, add a new value, then delete the original value.

      - One or more fragment size values, then touch **Delete**.
      - **Add** to add a value.

3. As needed, touch a size standard of interest, then touch **Edit** or **View**.

Manage analysis settings

Edit fragment/HID analysis settings

Factory-installed items cannot be edited or deleted. To create a new item from a factory-installed item, copy, edit, then save the new item.

Analysis settings are accessible only to the user who creates them.

1. Access the **Analysis settings** screen:

From	Action
Plate properties tab	Touch **More options** ▸ **Analysis settings**, then select the settings to assign to the plate setup.
Home screen	Touch ⚙ **Settings** ▸ **Run settings** ▸ **Analysis settings**, then manage the list of settings that you can select from during plate set up.

2. To create new analysis settings:
   a. Touch the default analysis settings or user-created analysis settings, then touch **Copy**.

b. Enter a name and edit settings as needed (see "Fragment/HID analysis settings (size calling)" on page 126).

c. As needed, touch **View** or **Delete** (user-created settings only).

3. Touch **Done**.

Edit sequencing settings

Factory-installed items cannot be edited or deleted. To create a new item from a factory-installed item, copy, edit, then save the new item.

Analysis settings are accessible only to the user who creates them.

1. Access the **Analysis settings** screen:

From	Action
Plate properties tab	Touch **More options** ▸ **Analysis settings**, then select the settings to assign to the plate setup.
Home screen	Touch ⚙ **Settings** ▸ **Run settings** ▸ **Analysis settings**, then manage the list of settings that you can select from during plate set up.

2. To create new analysis settings:
   a. Touch the default analysis settings or user-created analysis settings, then touch **Copy**.

   b. Enter a name and edit settings as needed (see "Sequencing settings (base calling)" on page 129).

   c. As needed, touch **View** or **Delete** (user-created settings only).

3. Touch **Done**.

Fragment/HID analysis settings (size calling)

Setting	Description
Size calling method	• **Local Southern**—*(default)* Determines the fragment sizes using the reciprocal relationship between fragment length and electrophoretic mobility. • **Global Southern**—Compensates for standard fragments with anomalous electrophoretic mobility (similar to least squares methods). • **2nd LSQ** (2nd Order Least Squares)—Uses regression analysis to build a bestfit size calling curve. • **3rd LSQ** (3rd Order Least Squares)—Uses regression analysis to build a bestfit size calling curve. • **Cubic Spline Interpolation**—Forces the sizing curve through all the known points of the selected size standard.

Setting	Description
Analysis range	• **Full Range**—*(default)* To analyze the entire scan region as collected by the genetic analysis instrument, including the primer peak. • **Partial Range**—To analyze only data points within a specified range. Enter Start Point in data points after the primer peak and before the first required size standard peak. Enter a Stop Point after the last required size standard fragment. Start and Stop points may vary from instrument to instrument and platform to platform. View raw data to determine the appropriate analysis range. Data points outside the specified analysis range are ignored. **Note:** Ensure the **Analysis Range** contains all size standard fragments included in the **Sizing Range**.
Sizing range	The size range (in base pairs) appropriate for the kit you are using: • **Full Range** for the software to analyze fragments of all sizes in the **Analysis Range**. • **Partial Range** for the software to analyze only fragments within a specified range. Enter a **Start Size** and a **Stop Size** appropriate for the size standard used.
Peak amplitudes	The peak height threshold (RFU) for peak detection for each dye color. Peaks below the threshold are not detected. For example, if you use the default values of 175 RFU, peaks with heights equal to or greater than 175 RFU are detected. Peaks with heights below 175 RFU are still displayed in the electropherogram plots but are not detected or labeled. **Note:** Use the same peak amplitude thresholds in secondary analysis software.
Primer peak	If the primer peaks in your application obscure peaks of interest, select **Present**. This instructs the algorithm to ignore primer peaks. Primer peaks are still displayed in the trace. If this setting does not allow detection of the 20- and 40-mer peaks for samples that use the GS600 LIZ™ size standard, running samples with the GS600_LIZ_(60-600) or other size standards that include lower bp starting points may allow detection of the peaks.
Common settings	
Smoothing	Select an option to smooth the outline of peaks and reduce the number of false peaks detected: • **None** (default) to apply no smoothing. Best if the data display sharp, narrow peaks of interest. • **Light** to provide the best results for typical data. Light smoothing slightly reduces peak height. • **Heavy** for data with very sharp, narrow peaks of interest. Heavy smoothing can significantly reduce peak height.

Setting	Description
Baseline Window	Specify a window to adjust the baseline signals of all detected dye colors to the same level for an improved comparison of relative signal intensity. Note the following: • A small baseline window relative to the width of a cluster, or grouping of peaks spatially close to each other, can result in shorter peak heights. • Larger baseline windows relative to the peaks being detected can create an elevated baseline, resulting in peaks that are elevated or not resolved to the baseline.
Minimum Peak Half Width	Specify the minimum full peak width at half maximum **Peak Height** required for peak detection. The range is 2 to 99 data points.
Peak Window Size	Enter a window width in data points for peak detection sensitivity. If more than one peak apex is within the window, all are labeled as a single peak. Note the following: • The maximum value is the number of data points between peaks. • The **Peak Window Size** setting is limited to odd numbers. To increase peak detection sensitivity: Increase polynomial degree, decrease peak window size. To decrease peak detection sensitivity: Decrease polynomial degree, increase peak window size.
Polynomial Degree	**Polynomial Degree** cannot be greater than **Peak Window Size**. Adjust to affect the sensitivity of peak detection. You can adjust this parameter to detect a single base pair difference while minimizing the detection of shoulder effects and/or noise. The peak detector calculates the first derivative of a polynomial curve fitted to the data within a window that is centered on each data point in the analysis range. Using curves with larger polynomial degree values allows the curve to more closely approximate the signal and, therefore, captures more of the peak structure in the electropherogram.
Slope Thresholds Peak Start and End	• **Peak Start**—The peak starts when the first derivative (slope of the tangent) in the beginning of the peak signal before the inflection point becomes equal to or exceeds the **Peak Start** value. This threshold is set to 0 by default, which means that the peak will normally start at the leftmost point where the slope of the tangent is closest to 0° (horizontal line). A value other than 0 moves the peak start point toward its center. The value entered must be non-negative. • **Peak End**—The peak ends when the first derivative (slope of the tangent) in the end of the peak signal after the inflection point becomes equal to or exceeds the **Peak End** value. This value is set to 0 by default, which means that the peak will normally end at the rightmost point where the slope of the tangent is closest to 0° (horizontal line). A value other than 0 moves the peak end point toward its center. The value entered in this field must be non-positive. Using curves with larger polynomial degree values allows the curve to more closely approximate the signal and, therefore, captures more of the peak structure in the electropherogram.

Sequencing settings (base calling)

The default settings are optimized for sequencing of PCR amplicons from diploid genes, which are expected to have Mixed bases and which should end At PCR Stop.

For sequencing from plasmid templates, which are of pure sequence (no Mixed bases) and have longer sequence read times than PCR products, create new analysis settings and *disable* (uncheck) **At PCR Stop** and **Mixed base threshold** checkboxes.

Setting	Description
Quality Threshold	• Basecall assignment (ambiguous bases): – Do not assign Ns to basecalls – Assign Ns to basecalls with QV<5— Bases with a QV less than the threshold display N instead of the base letter • End base—Last base on which to perform basecalling: – At PCR Stop – After X number of bases – After X number of Ns in X number of bases – After X number of Ns **Note:** If you have PCR products with sequences that end while data is still being collected, select the **At PCR Stop** checkbox.
Mixed bases threshold	When enabled, determines the secondary peak height ratio where the secondary peak is considered a potential mixed base. Reaching the threshold is a necessary but not sufficient condition for the basecalling algorithm to call a mixed base.
Clear range methods	• **Use quality values**—Sets a window with a specified number of allowed low-quality bases by removing bases until there are <X number of bases per Z number of bases with QV <Y. • **Use base positions**—Specifies the first and last base in the range to consider, or trims the specified number of bases from the 3′ end. • **Mask base positions before**—Specifies the base position before which to disregard bases.

Modify the default file name convention

The default file name convention determines how the data files (AB1 or FSA) associated with a plate are named.

The default file name convention is:

Application	Default settings
Fragment/HID analysis	`<well>_<sample name>_<sample type>_<date and timestamp>.fsa`
Sequence analysis	`<well>_<sample name>_<date and timestamp>.ab1`
Sequence analysis with the Sanger variant analysis option selected	`<well>_<sample name>_<amplicon>_<specimen>_<date and timestamp>.ab1`

1. Access the **File name convention** screen:

From	Action
Plate properties screen	Select the **Properties** tab, then touch **More options** ▸ **File name convention**.
Home screen	Touch ⚙ **Settings** ▸ **Run settings** ▸ **File name convention**.

2. Touch **Attributes**.

3. Select the attributes to include in the data file name.



For information on creating custom fields to include in file name conventions, see "Define custom fields" on page 89.

Note: The timestamp attribute cannot be deselected; it is always included in the data file name.

4. Touch **Done**.

5. Touch and drag attributes up or down in the list.

6. Touch **Done**.

Manage instrument settings

Display instrument hardware and software information

In the home screen:

1. Touch ⚙ **Settings** ▸ **Instrument settings** ▸ **About** to access the instrument information:
   - **Model name**
   - **Ethernet IP address**
   - **Ethernet MAC address**
   - **Wireless IP address**
   - **Wireless MAC address**

- **Instrument software release**
- **Instrument serial number**

2. *(Optional)* Touch **EULA** to display the end-user licence agreement or touch **Details** to display additional instrument information.

Change the instrument name

If the instrument is linked to the Connect cloud-based platform, only a Connect platform administrator for the instrument can change the instrument name. The instrument name can be changed by a Connect platform administrator on the instrument or in InstrumentConnect.

In the home screen:

1. Touch ⚙ **Settings** ▸ **Instrument Settings** ▸ **Instrument name**.
2. Touch the **Instrument Name** field, enter an instrument name, then touch **Done**.
3. Touch **OK**.

When you change the instrument name, the software unlinks all instrument profiles.

Enable Demo mode

Demo mode allows you to use all features of the system and provides simulated real-time data and results. You cannot save plates or settings in Demo mode.

In the home screen:

1. Touch ⚙ **Settings** ▸ **Instrument settings** ▸ **Turn Demo mode on**.
2. Touch **OK** to allow the instrument to restart.
   The instrument automatically restarts in Demo mode.

Manage date and time settings

In the home screen:

1. Touch ⚙ **Settings** ▸ **Instrument settings** ▸ **Date and Time**.
2. Slide the control to select a time-setting option:
   - The instrument automatically detects the time via the network.
   - Enter the time manually.
3. Touch **Time zone**, then select a time zone.
4. Touch **Date/Format**, then select the display order for the month, day, and year.
5. Touch **Time/Format**, then select a 12-Hour or 24-Hour time display format.
6. Click **OK**.

Manage the network configuration

Note: For a direct connection between the instrument and a computer, set up a wired connection.

In the home screen:

1. Touch ⚙ **Settings** ▸ **Insrument settings** ▸ **Network Configuration**.

2. Touch **Edit**, or touch one of the network settings fields.
3. *(Optional)* Edit the **Wireless** network settings.
   a. Select a **Network** that was automatically detected by instrument.
   b. Enter a password, if prompted.
   c. Click **Join**.
   d. Click **OK** when the authentication is completed.
4. *(Optional)* Edit the **Wired** network settings.
   a. Select **DHCP** or **Static IP**.
   An IP address is automatically assigned if **DHCP** is selected.
   b. *(Static IP only)* Enter an **IP address**, **Subnet mask**, **Default gateway**, **Primary DNS server**, and **Secondary DNS server**.
   For more information, see "Determine IP address for a computer on a network" on page 130.
5. Touch **OK**.

Check for software updates (administrator only)

In the home screen:

1. Touch ⚙ **Settings** ▸ **Instrument settings** ▸ **About**.



2. Touch **Check for updates**.

   If the software update is located on a USB, it may take 10–15 seconds for the instrument to recognize the USB.

3. Touch:

   - **Update** if an update is available.
   - **Cancel** if there is no update available.

   A message is displayed during the software updated, then the instrument automatically restarts.



Manage instrument profiles on the instrument

Local instrument profile roles and functions

Instrument profile	Location	Functions allowed
Standard	Local[1]	• Create, save, open, import, and run plate setups • Create and modify run settings • View and export results
Administrator	Local[1]	All standard user functions, plus: • Create or delete an instrument profile • Change an instrument profile from **Standard** to **Administrator** access • Reset an instrument profile PIN for another user • Unlock instrument touchscreen • Backup user data (plates and results) • Delete the Connect cloud-based platform instrument profile from the instrument, which removes the instrument from the InstrumentConnect. The user can link the instrument to the Connect platform again using a local instrument profile. After a user relinks to the Connect platform, the Connect platform instrument profile is displayed on the home screen and the instrument is listed in the InstrumentConnect.
Guest	Local	All standard user functions, except: • Cannot link to the Connect cloud-based platform • Cannot modify a public plate setup **Note:** Standard instrument profiles cannot access Guest instrument profile plate setups unless the **Plate Setup Security** is set to **Shared**.

[1] The first user who signs in to the instrument is assigned a local profile with administrator role.

For more information, see Chapter 3, "Use the instrument with the Connect cloud-based platform".

Create a local instrument profile for another user (administrator only)

An instrument profile can be assigned standard or administrator roles. For more information, see "Change the role of a local instrument profile (administrator only) " on page 152.

In the home screen:

1. Touch ⊛.
2. Touch **All accounts**.
3. Touch **Add Profile**.
4. Touch **User name**, enter an instrument profile name, then touch **Done**.

5. Touch **PIN (4 digits required)**, enter a four-digit numerical PIN, then touch **Enter**.
6. Touch **Confirm PIN**, reenter the PIN, then touch **Enter**.
7. Touch **Create profile**.
8. Touch **Done**.

Change the role of a local instrument profile (administrator only)

In the home screen:

1. Touch ⊛.
2. Touch **All accounts**.
3. Touch the account of interest.
4. Slide the control from **Standard** to **Administrator** or from **Administrator** to **Standard**.
5. Touch **Done**.

Delete an instrument profile from an instrument (administrator only)

Local administrators and Connect platform administrators can delete local and Connect platform profiles.

IMPORTANT! Before proceeding, back up user data to retain plates and results ("Back up user data (plates and results) (administrator only)" on page 155).

In the home screen:

1. Touch ⊛.
2. Touch **All accounts**.
3. Touch the instrument profile to delete.
4. Touch **Delete account**.
5. Touch **Yes** to confirm.
6. Touch **Done**.

Delete the PIN for a local instrument profile (administrator only)

Local administrators and Connect cloud-based platform administrators can delete the PIN for a local instrument profile.

In the home screen:

1. Touch ⊛.
2. Touch **All accounts**.
3. Touch the instrument profile of interest.

4. Touch **Delete PIN**.
   The user will be prompted for a new pin upon the next sign in.
5. Touch **Yes** to confirm.
6. Touch **Done**.

Create your own local instrument profile

If you are not signed in to the instrument when you link the instrument to your Connect cloud-based platform account, the software creates a local instrument profile with the **Standard** role using the *FirstName LastInitial* of your Connect platform account.

For more information, see "Link the instrument to your Connect cloud-based platform account" on page 136.

Change your own local instrument profile PIN

Sign in to access these features (see "Sign in" on page 34).

1. Touch ⊛.
2. Touch **Edit**.
3. Touch **Old PIN**, enter your current PIN, then touch **Enter**.
   Touch the **Show PIN** checkbox to switch the PIN display on or off.
4. Touch **PIN (4 digits required)**, enter a new four-digit numerical PIN, then touch **Enter**.
5. Touch **Confirm PIN**, reenter your new PIN, then touch **Enter**.
6. Touch **Done**.

Manage storage space

Automatic file cleanup

Before starting a run, the instrument calculates the total amount of storage space required to save the run to the instrument. If the required storage space is not available, the instrument deletes files associated with the oldest exported plates until sufficient space is available.

Note: Only complete plates that have been auto exported (saved to Connect cloud-based platform, network, or USB) or manually exported (using ⚙ **Settings** ▸ **Run History** ▸ ***plate name*** ▸ **Export**) are deleted.

If the required storage space is not available and no plates have been exported, the instrument displays a notification indicating that there is not enough storage space.

You can export plates and delete plates, then start the run again.

Export or delete a plate setup (PSM file)

1. Touch **Set up run**, then touch **My instrument**.
2. Touch **Manage** at the bottom left of the screen.



3. Touch a plate, then touch **Export** or **Delete**.
4. If you touched:
   - **Export**, select a storage location, then touch **Export**.
   - **Delete**, then touch **Yes** to delete the plate setup.

Export results from the instrument (sample data files and QC reports)

In the home screen:

1. Touch **Settings** ▸ **Run history**.
2. Select one or more plates from the **Run History** table.
3. Touch **Export**.
4. Select a storage location.

The following data is exported for the plate:

- Fragment/HID analysis—FSA file for each sample.
- Sequencing—AB1 file for each sample.
- Plate QC report in CSV and PDF format.

Note: If you select a plate, select **View**, then select **Export**, only an FSA or AB1 file for each analyzed sample is exported.

Delete a run history

In the home screen:

1. Touch ⚙ **Settings ▸ Run history**.
2. Select one or more plates from the Run History table.
3. Touch **Delete**, then touch **OK** to confirm.

Note: Run histories for the oldest exported plates are automatically deleted if sufficient storage space is not available when you start a run.

Back up user data (plates and results) (administrator only)

1. Select **Settings ▸ Maintenance and service ▸ Back up user data**.
2. Enter a name for the folder that will contain the backup data.
3. Touch the instrument profiles for which you want to back up plate and result information.
4. Touch **Backup**.
5. Specify a storage location, then touch Export.
   A folder containing plate setup and sample data file is backed up to the specified location.

(Optional) After backing up, you can:

- Delete the plates (see "Delete a run history" on page 155).
- Open the CSV files in the folder on another instrument or in the Plate Manager.

11 Maintain the instrument

- Regular maintenance tasks 156
- Manage the cartridge 158
- Install cathode buffer 166

Regular maintenance tasks

Check the cathode buffer fill level

When a cartridge is installed in the instrument, the volume of buffer in the Cathode Buffer Container (CBC) must be above the fill line.

Check the cathode buffer fill level before each run. It is recommended to check one time each week if the instrument is not in use.

In the home screen:

1. Touch ⏏, touch ⠿ **Eject plate**, then open the instrument door when prompted.
2. Press the release button on the autosampler to open the lid, then remove the CBC.



① Release button
② CBC

3. Ensure that the level of buffer is above the fill line.

   If the buffer is at or below the fill line, see "Assemble the SeqStudio™ Genetic Analyzer Cathode Buffer Container (CBC)" on page 166 and "Insert the Cathode Buffer Container" on page 167.



① Replace if buffer is at or below the fill line

4. Reinstall the CBC.

5. Close the autosampler lid: Press down on the center of the lid or press down on both sides of the lid with equal pressure until the lid clicks shut.



① New CBC buffer level

6. Touch **Retract plate**, close the instrument door, then touch **Done** when the **Consumables Status** screen is displayed.

Clean the instrument exterior and touchscreen

Power off the instrument before cleaning.

IMPORTANT! The instrument does not maintain the correct temperature conditions for the cartridge when it is powered off. Avoid cartridge exposure to ambient temperature.

- Wipe the exterior of the instrument with a lint-free cloth and deionized water.

  IMPORTANT! Do not allow any moisture to reach the interior of the instrument through the door.

- Wipe the touchscreen with a lint-free cloth and glass cleaner.

Clean the autosampler

The autosampler is attached to the instrument.

Clean spills to prevent a build-up of crystallized polymer or dried salt from the buffers.

Power off the instrument before cleaning the autosampler.

1. Wipe the exterior of the autosampler with a lint-free cloth and deionized water.

   IMPORTANT! Do not pour liquid directly on to the autosampler. This will damage the heater at the bottom of the cathode buffer reservoir.

   Note: Do not use detergents or solvents.

2. Use a lint-free cloth to absorb any liquid spilled in the individual wells of the autosampler.

Manage the cartridge

Cartridge storage

Table 10 Storage information for the SeqStudio™ Genetic Analyzer Cartridge (Cat. No. A33671)

Condition	Description
Shipping	Shipped at 2–8°C. Store upright at 2–8°C upon receipt. Save the white storage box and optical cover for off-instrument cartridge storage.
On-instrument storage	For routine use, can be used and stored on the instrument for up to 4 months. If you store the cartridge on-instrument: • The instrument must be powered on. • A Cathode Buffer Container must also be installed. The instrument keeps the components under the following conditions when it is powered on and in **Cartridge storage mode**: • **Optical detection window**—Covered • **Capillary array electrodes**—Submerged in cathode buffer (buffer must be above the fill line in the Cathode Buffer Container) • **Polymer**—Chilled • **Anode buffer**—Ambient temperature IMPORTANT! The instrument does not maintain the correct temperature conditions for the cartridge when it is powered off. Avoid cartridge exposure to ambient temperature.
Off-instrument storage	For intermittent use, can be stored off-instrument until the expiry date on the label or up to 4 months after first use. Store upright at 2–8°C, with an integrated capillary protector (ICP) and optical cover installed (see "Store the cartridge" on page 163). **Note:** After you remove the cartridge from the instrument, install an ICP within a few minutes. Avoid cartridge exposure to ambient temperature.
Reuse	Can be removed from an instrument then inserted again on the same instrument or a different instrument, if it was stored properly at 2–8°C and has not expired or exceeded 125 injections. Information about the cartridge installation and usage is retained in the cartridge history (**Settings ▸ Cartridge ▸ Instrument–cartridge history**).

Table 11 Storage information for the SeqStudio™ Genetic Analyzer Cartridge v2 (Cat. No. A41331)

Condition	Description
Shipping	Shipped at 2–8°C. Store upright at 2–8°C upon receipt. Save the white storage box and optical cover for off-instrument cartridge storage.
On-instrument storage	For routine use, can be used and stored on the instrument for up to 6 months. If you store the cartridge on-instrument: • The instrument must be powered on. • A Cathode Buffer Container must also be installed. The instrument keeps the components under the following conditions when it is powered on and in **Cartridge storage mode**: • **Optical detection window**—Covered • **Capillary array electrodes**—Submerged in cathode buffer (buffer must be above the fill line in the Cathode Buffer Container) • **Polymer**—Chilled • **Anode buffer**—Ambient temperature IMPORTANT! The instrument does not maintain the correct temperature conditions for the cartridge when it is powered off. Avoid cartridge exposure to ambient temperature.
Off-instrument storage	For intermittent use, can be stored off-instrument until the expiry date on the label or up to 6 months after first use. Store upright at 2–8°C, with an integrated capillary protector (ICP) and optical cover installed (see "Store the cartridge" on page 163). **Note:** After you remove the cartridge from the instrument, install an ICP within a few minutes. Avoid cartridge exposure to ambient temperature.
Reuse	Can be removed from an instrument then inserted again on the same instrument or a different instrument, if it was stored properly at 2–8°C and has not expired or exceeded 250 injections. Information about the cartridge installation and usage is retained in the cartridge history (**Settings ▸ Cartridge ▸ Instrument–cartridge history**).

Set cartridge storage mode (administrator only)

If a cartridge is installed on the instrument, the instrument will enter **Cartridge storage mode** after the time you specify.

The instrument keeps the components under the following conditions when it is powered on and in **Cartridge storage mode**:

- **Optical detection window**—Covered
- **Capillary array electrodes**—Submerged in cathode buffer (buffer must be above the fill line in the Cathode Buffer Container)
- **Polymer**—Chilled
- **Anode buffer**—Ambient temperature

In the home screen:

1. Touch ⚙ **Settings** ▸ **Cartridge** ▸ **Cartridge Storage Mode**.
2. Select the duration of instrument inactivity before the instrument enters Cartridge Storage Mode.
3. Touch **OK**.

Check or export the Instrument-cartridge history

This function lists the cartridges that have been installed on this instrument.

1. Touch ⚙ **Settings** ▸ **Cartridge** ▸ **Instrument-cartridge history**.

2. *(Optional)* Touch **Export**.

Check or export the cartridge history

This function lists the instruments on which the cartridge has been installed.

Touch ⚙ **Settings** ▸ **Consumables status**, then touch **Cartridge history**.



Fill the capillary array and refresh the polymer delivery system

The functions accessed from **Cartridge maintenance** are performed automatically during a run. Do not use these commands to manually perform these functions unless instructed to do so in troubleshooting or by Support.

In the home screen:

1. Touch ⚙ **Settings** ▸ **Cartridge** ▸ **Cartridge maintenance**.
2. Select **Fill array** or **Refresh PDS**.
3. Touch ⊗ to close the screen when the function ends.

Remove the cartridge

Perform these steps if you are replacing the cartridge or storing it off-instrument. Removal of the cartridge after each run is not required.

For appropriate off-instrument storage conditions, see the following tables:

- For the SeqStudio™ Genetic Analyzer Cartridge, see Table 1 on page 18
- For the SeqStudio™ Genetic Analyzer Cartridge v2, see Table 2 on page 19

1. Touch ⏏, touch **Eject cartridge**, then open the instrument door when prompted.
2. Hold the cartridge at the hand hold above the capillaries, then pull to remove it from the instrument.
3. Close the instrument door.

For more information, see “Store the cartridge“ on page 163.

Insert the cartridge

1. If a cartridge is installed:
   a. Touch ⏏, touch **Eject cartridge**, then open the instrument door when prompted.
   b. Remove the used cartridge.

2. Prepare the cartridge.
   a. Remove the cartridge from the box.

      Note: Save the box for off-instrument cartridge storage.



   b. Remove the optical cover from the cartridge by grasping the finger holds, then pulling toward you.

      Note: Save the optical cover for off-instrument cartridge storage. Avoid touching the capillaries that are protected by the optical cover.



① Optical cover

   c. Pinch the clamp on the SeqStudio™ Integrated Capillary Protector, then pull down.

      Discard the SeqStudio™ Integrated Capillary Protector. The SeqStudio™ Integrated Capillary Protector is single-use. Use a new SeqStudio™ Integrated Capillary Protector if you store the cartridge.



3. Grasp the cartridge above the capillaries.



4. Position the cartridge:
   a. Orient the cartridge with the embossed arrow pointing toward the rear of the instrument.
   b. Align the guides at the top of the cartridge with the insertion rails in the instrument.



① Insertion rails
② Embossed arrow

5. Slide the cartridge into the instrument until it clicks into place.
6. Close the instrument door, then touch **Done** when the **Consumables Status** screen is displayed.

   The cartridge is initialized.

Store the cartridge

1. Touch ⏏, touch **Eject cartridge**, then open the instrument door when prompted.
2. Hold the cartridge at the hand holds above the capillaries, then pull to remove it from the instrument.
3. Close the instrument door.
4. *Carefully* place the optical cover on the cartridge to avoid damage to the capillaries.



① Optical cover

5. Place a new SeqStudio™ Integrated Capillary Protector on the cartridge: pinch the clamp on the ICP, then push up on to the capillaries.



6. Place the cartridge in the white storage box.

7. Store upright at 2–8°C.

   Note: Avoid cartridge exposure to ambient temperature.

   IMPORTANT! Do not freeze the cartridge. It cannot be used after freezing.

SeqStudio™ Integrated Capillary Protector

The SeqStudio™ Integrated Capillary Protector (ICP) is a single-use protective cover that clamps onto the SeqStudio™ Genetic Analyzer Cartridge or the SeqStudio™ Genetic Analyzer Cartridge v2. The ICP prevents the capillary array from drying out during off-instrument storage of the cartridge, and is removed before insertion of the cartridge into the SeqStudio™ Genetic Analyzer.

Each cartridge is supplied with an ICP for shipping and one additional ICP for off-instrument cartridge storage.



Figure 19 SeqStudio™ Integrated Capillary Protector

1 Clamp mechanism for attaching to cartridge

IMPORTANT! Remove the Integrated Capillary Protector before installing the cartridge into the instrument. Installing the cartridge with the ICP in place can damage the capillary array.

SeqStudio™ Integrated Capillary Protector storage

Condition	Description
Shipping	Is shipped at 2–8°C. Before opening, can be stored until expiry date on label at 2–8°C. Discard the shipping ICP when you insert a new cartridge in the instrument.
Reuse	Do not reuse. Use a new ICP for off-instrument cartridge storage. **IMPORTANT!** If an ICP is used for off-instrument cartridge storage more than one time, the capillary array will not be kept fully hydrated.

Long-term on-instrument cartridge storage

You can leave the cartridge installed in the instrument if the instrument is powered on and a CBC is installed. Replace the CBC every 2 weeks.

IMPORTANT! If the instrument will be idle for longer than the shelf life of the cartridge components, remove the cartridge and cathode buffer, then power off the instrument.

Table 12 Cartridge storage

Cartridge	Storage
SeqStudio™ Genetic Analyzer Cartridge (Cat. No. A33671)	See Table 1 on page 18
SeqStudio™ Genetic Analyzer Cartridge v2 (Cat. No. A41331)	See Table 2 on page 19

IMPORTANT! The instrument does not maintain the correct temperature conditions for the cartridge when it is powered off. Avoid cartridge exposure to ambient temperature.

Install cathode buffer

SeqStudio™ Genetic Analyzer Cathode Buffer Container storage

Condition	Description
Shipping	Is shipped at ambient temperature. Store at 2–8°C upon receipt.
On-instrument storage	After installation, can be stored for up to 2 weeks when the instrument is powered on and in **Cartridge storage mode**.
Off-instrument storage	Before opening, can be stored until expiry date on label at 2–8°C.
Reuse	Do not remove the CBC from the instrument for storage. Do not reuse.

Remove the SeqStudio™ Genetic Analyzer Cathode Buffer Container

Perform these steps if you are replacing the Cathode Buffer Container. It is not necessary to remove the CBC after each run.

1. Touch ⏏, touch ⠿ **Eject plate**, then open the instrument door when prompted.
2. Press the release button on the autosampler to open the lid.
3. Lift the CBC out of the autosampler.

Assemble the SeqStudio™ Genetic Analyzer Cathode Buffer Container (CBC)

Equilibrate the CBC to room temperature (15 minutes to overnight) before assembling.

On a clean and level surface:

1. Ensure that the buffer is above the fill line.
2. Carefully peel off the seal.



3. Wipe off any buffer on top of the CBC with a lint-free tissue. Ensure that the top of the container is dry.

4. Place the reservoir septa on the CBC, then press firmly to seat the septa.



Note: The CBC is filled significantly above Fill Line to account for evaporation. Replace the CBC when the fluid level is at or below the fill line.

Insert the Cathode Buffer Container

Perform these steps if the Cathode Buffer Container has not been installed, if the CBC on the instrument has expired, or if the buffer level is at or below the fill line.

Install the CBC before you install the cartridge.

See Figure 2 on page 14 for the position of the cathode buffer on the instrument.

1. Touch ⏏, touch ⠿ **Eject plate**, then open the instrument door when prompted.

2. Press the release button on the autosampler to open the lid.



① Release button
② Location of CBC

3. Insert the cathode buffer in the autosampler with the notch positioned in the back right.

   See Figure 8 on page 20.

4. Press the autosampler lid until it clicks shut.



5. Touch **Retract plate**, then close the instrument door.

A Troubleshooting

- Troubleshooting resources 169
- Troubleshooting workflow 170
- Export log files for plates, install runs, injections, and instrument 170
- View the raw data and the EPT plot 171
- Data quality alerts 171
- Sizecalling and basecalling quality alerts 171
- Check size standard quality 172
- Instrument troubleshooting 173
- Cartridge troubleshooting 174
- Sample and data troubleshooting 175
- EPT data 180
- Software troubleshooting 181
- Service tools 183

Troubleshooting resources

Document	Pub. No.
Troubleshooting Sanger sequencing data User Bulletin	MAN0014435
DNA Sequencing by Capillary Electrophoresis Chemistry Guide Second Edition	4305080
DNA Fragment Analysis by Capillary Electrophoresis User Guide	4474504

Troubleshooting workflow

Follow this general workflow when you are troubleshooting:

Review the analyzed data.

▼

Review the raw data, then review the EPT plot.

▼

(Fragment/HID analysis) Check size standard quality (“Check size standard quality“ on page 172).

▼

Check the CBC buffer fill level.

▼

Check that the cartridge is installed and engaged, and that capillary tips are not bent or damaged.

▼

Check that the sample plate is installed, confirm that samples are in the wells that are specified in the plate setup, and make sure that samples are at bottom of wells (no bubbles are visible).

Export log files for plates, install runs, injections, and instrument

Export log files if directed to do so by Technical Support.

In the home screen:

1. Touch ⚙ **Settings** ▸ **Maintenance and Service** ▸ **Export logs**.
2. Select one of the following options.
   - **Export logs for the last injection**—Includes all run types: analysis run, install run, and calibration run.
   - **Export logs for the last / current plate**—Includes all run types.
   - **Export logs for selected plates**—Includes options to export logs for regular runs and install check runs for the analysis run for the current user.
   - **Export recent instrument server logs**—Includes instrument export logs only.
3. Touch **Export**.
4. Select a storage location.

   Note: Log files can be large. Select a location with adequate storage space.

   A ZIP file containing log information is exported.

View the raw data and the EPT plot

The EPT view (ElectroPhoresis Telemetry) shows instrument data conditions (currents, temperatures, electrophoresis voltage) as a function of time.

When a run is complete, in the home screen:

1. Touch **Results**.
2. Touch **List view**.
3. Touch an injection group.
4. Touch a sample file name.
   If the data triggered any quality alerts, a QC alerts screen is displayed.
   Click **View data** to display the trace for the sample.
5. Touch > or < to scroll to the raw data or **EPT Plot**.

Data quality alerts

Quality alert	Description	Action
Offscale peaks. Adjust the injection parameters and/or the sample concentration.	At least 10 scans have saturated the CCD camera.	• Reduce the injection voltage or time. • Dilute the sample.
No sample was detected.	Poor signal-to-noise ratio with low signal detected.	• Verify that the sample volume follows recommendations in the user manual. • Troubleshoot upstream PCR and sequencing steps.

Sizecalling and basecalling quality alerts

Table 13 Sizecalling quality alerts

Quality alert	Description	Action
Sizing quality value is low due to poor size standard peak quality. Peak height uniformity is low or the fitting quality in sizing is poor.	Low resolution or poor quality data is present.	• Re-inject the sample. • If the problem persists, check the sample quality.
Sizecaller found broad peak(s) in the size standard peak(s).		
Sizing quality value is in the intermediate range; check size standard data quality.		

Quality alert	Description	Action
The number of size standard peaks detected is less than what is defined in the size standard.	Size standard definition includes peaks that are not present in the sample. Example: Sample peaks are detected up to 500 bp, but the size standard definition includes peak sizes that are >500 bp.	Use or create a size standard definition with the appropriate number of peaks and peak sizes.
The analysis range is too small. Correct the analysis range in analysis settings and re-analyze in secondary analysis software or re-inject sample.	Various causes.	• Analyze the data in a secondary analysis software with a corrected analysis range. • Re-inject the sample.

Table 14 Basecalling quality alerts

Quality alert	Description	Action
Basecalling failed due to poor quality data.	Poor quality data is present.	• Re-inject the sample. • If the problem persists, prepare fresh sample. • Troubleshoot upstream PCR and sequencing steps.

Check size standard quality

1. Touch ⚙ **Settings** ▸ **Run history**.
2. Touch a plate name, then touch **View**.
3. Touch a sample file name, then touch **View**.
   If the data triggered any quality alerts, a QC alerts screen is displayed.
4. Touch **View data** to display the trace.



5. Touch ⋮ on the left border of the trace, then deselect all dyes except the size standard dye (red or orange).



6. As needed, touch ⋮ on the right border of the trace to zoom on the trace.

Instrument troubleshooting

Observation	Possible cause	Recommended action
There was a loss of power to the instrument	There was a power failure.	Restart the instrument. The **Sign In** screen is displayed. A run that was in progress at the time of the power failure must be restarted. Run a control sample to ensure that the consumables have not degraded, especially if the consumables reached room temperature. Replace the consumables if the results with a control sample show that the consumables have degraded.
The RFID tags on the cartridge or the cathode buffer are not read	The label on the cartridge or the Cathode Buffer Container is damaged, not positioned properly, or has been removed.	Ensure that the labels are present and not visibly damaged.
		Ensure that the consumables are installed correctly. The Cathode Buffer Container should be firmly seated in the autosampler. The cartridge will click into place when it is installed correctly.
	The RFID reader is malfunctioning.	Insert consumables that have previously had RFID tags read. Contact technical support if the RFID tags are not read.
The run stopped and **Resume** is displayed	A user paused the run.	Touch **Resume** to continue the run.
The amber light is blinking	The run was paused.	Touch **Resume** to continue the run.
	The instrument door is open.	Close the instrument door.
	There was an instrument error.	Follow the instructions in the error message.
		Restart the instrument.
The electrophoresis failed or **Current check failed** is displayed	There is insufficient cathode buffer.	Check the fill line on the Cathode Buffer Container. Replace the Cathode Buffer Container if the buffer is at or below the fill line.

Observation	Possible cause	Recommended action
The electrophoresis failed or **Current check failed** is displayed *(continued)*	The Cathode Buffer Container has been installed on the instrument for more than two weeks or used for more than 125 injections.	Replace the Cathode Buffer Container.
	The septum on either the Cathode Buffer Container or the sample plate is not installed correctly.	Ensure that the septa are fully inserted into both the Cathode Buffer Container and the sample plate.
	Buffer or other liquid was spilled on top of the reservoir septum or on top of the autosampler.	Wipe the spill with a lint-free cloth.
	There is condensation on the Cathode Buffer Container or around the reservoir septum.	Wipe the condensation with a lint-free cloth.
		Ensure that the humidity in the lab is non-condensing.

Cartridge troubleshooting

Observation	Possible cause	Recommended action
There are crystals on the cartridge capillary array	Small amounts of leakage around the cartridge capillary array are normal.	No action is required.
There are crystals on the cartridge polymer delivery system	There is a leak in the polymer delivery system.	Run a control sample to determine if the cartridge function is affected.
		Replace the cartridge if the leak affects the cartridge function.
Poor-quality data is observed after prolonged drying of the capillary tips	The capillary tips develop blockages if they are allowed to dry out.	Use an Integrated Capillary Protector when the cartridge is off the instrument to prevent the capillary tips from drying out.
		Clear the blockage by running polymer through the capillaries. 1. Touch ⚙ **Settings** ▸ **Cartridge** ▸ **Cartridge maintenance** ▸ **Fill array**. 2. Touch ⊗ when the function ends. 3. Run a control sample. 4. Replace the cartridge if there is poor-quality data with the control sample.
		Follow the recommendations for cartridge storage to prevent drying of the capillary tips. See "Cartridge storage" on page 18.

Sample and data troubleshooting

See also "Data quality alerts" on page 119, "Sizecalling and basecalling quality alerts" on page 119, and *Troubleshooting Sanger sequencing data User Bulletin* (Pub. No. MAN0014435).

Observation	Possible cause	Recommended action
Offscale signal is detected is displayed **Note:** Samples are stable on the instrument for 16–24 hours. Determine if samples will be stable if a re-injection is recommended, then plan the re-injection accordingly.	The sample concentration is too high.	Dilute the sample, prepare a new plate, and start a new run.
		Quantitate the sample prior to adding reagents for capillary electrophoresis.
	The injection conditions are too strong for the sample.	Re-inject the samples with adjusted injection conditions. Injection time, voltage, or a combination of these conditions can be adjusted.
There is no signal or a low signal **Note:** Samples are stable on the instrument for 16–24 hours. Determine if samples will be stable if a re-injection is recommended, the plan the re-injection accordingly.	The sample volume was insufficient.	Use the recommended sample volume. See "Sample preparation guidelines" on page 37.
	The sample concentration was too low.	Re-inject the samples with adjusted injection conditions. Injection time, voltage, or a combination of these conditions can be adjusted.
	There were bubbles in the sample wells.	Centrifuge the sample plate or tubes to remove the bubbles before loading onto the instrument.
		If the samples have been run, centrifuge the plate or tubes, then set up a re-injection.
	The sequence reaction failed.	Review the sequence analysis protocol, the template quality, and the template quantity. Set up a new plate and repeat the reaction. See *Troubleshooting Sanger sequencing data User Bulletin* (Pub. No. MAN0014435).
	The Hi-Di™ Formamide used to prepare the samples was degraded.	Prepare the samples with fresh Hi-Di™ Formamide and repeat the experiment. See "Sample preparation guidelines" on page 37.
	The sample was prepared with the BigDye XTerminator™ Purification Kit but a BDX run module was not selected.	Re-inject the sample with the correct module.
	The sample was degraded.	Prepare the sample according to the protocol provided with the kits for sample preparation. See "Sample preparation guidelines" on page 37 for sample preparation guidelines. IMPORTANT! Do not resuspend samples in water.

Observation	Possible cause	Recommended action
There is no signal or a low signal **Note:** Samples are stable on the instrument for 16–24 hours. Determine if samples will be stable if a re-injection is recommended, the plan the re-injection accordingly. *(continued)*	The sample had a high concentration of salt.	Dilute or desalt the samples.
	There was an excess of unlabeled template competing with the fragments labeled with dye during the injection.	Dilute or desalt the samples. See: • *DNA Sequencing by Capillary Electrophoresis Chemistry Guide Second Edition* (Pub. No. 4305080) • *DNA Fragment Analysis by Capillary Electrophoresis User Guide* (Pub. No. 4474504)
	A capillary array tip is blocked.	Flush the capillary array. 1. Touch ⚙ **Settings** ▸ **Cartridge** ▸ **Cartridge maintenance**. 2. Touch **Refresh PDS**. 3. Touch ✕ to close the screen when the function ends.
		Replace the cartridge if the problem persists.
	The cartridge is damaged.	Inspect the cartridge for damage. Replace the cartridge if there is damage.
There was poor resolution, poor size quality, or a poor sequencing result	The samples degraded over time while on the instrument.	Analyze a maximum of 48 samples on a plate for long run modules. Additional samples will take more than 24 hours and the samples can degrade. **Note:** Thermal breakdown of samples is normal. Samples are stable for 16–24 hours on the instrument.
		Use Hi-Di™ Formamide to prepare the samples. Sample stability is optimal in Hi-Di™ Formamide.
	The temperature and/or humidity in the lab is too high for optimal sample stability.	Ensure that the conditions in the lab are within the operating range for the instrument (15–30°C and 20–80% relative humidity). Sample stability can be lower near the high end of the operating range of the instrument for temperature and humidity.
		Use 20 µL of sample instead of 10 µL. A larger sample volume can reduce sample breakdown under hot and humid conditions.
	There was a sporadic data quality failure. **Note:** A sporadic data quality failure can happen occasionally.	Repeat the injection.

Observation	Possible cause	Recommended action
Poor resolution in some capillaries	Poor-quality samples were used.	See "Sample preparation guidelines" on page 37.
		Use a control sample to determine if the poor resolution is due to the samples are another factor.
	A capillary was damaged or is blocked.	Try to clear the capillary, then replace the cartridge if the poor resolution is due to the cartridge and not the sample. 1. Touch ⚙ **Settings** ▸ **Cartridge** ▸ **Cartridge maintenance** ▸ **Fill array**. 2. Touch ⊗ when the function ends. 3. Run a control sample.
Poor resolution in all the capillaries	The cartridge has been used for more than the stated number of injections, is past the labeled expiration date, or has degraded polymer from incorrect storage.	Replace the cartridge.
	The Hi-Di™ Formamide that was used to prepare the samples was degraded.	Prepare the samples with fresh Hi-Di™ Formamide and repeat the reaction. See: • "Sample preparation guidelines" on page 37 • *DNA Fragment Analysis by Capillary Electrophoresis User Guide* (Pub. No. 4474504)
	There was too much sample injected.	Dilute the sample and re-inject it.
Spikes are present in raw and/or analyzed fluorescence data	Trace impurities passed the detector.	For fragment analysis, increasing the **Minimum peak half width** setting in the analysis settings (under **Common settings**) can reduce the identification of spikes as peaks. **Note:** Secondary sequencing and fragment analysis software can recognize and ignore spikes.
		Repeat the injection if necessary.
There is noise in the baseline	Fluorescent contamination has built up in the CBC.	Replace the CBC. Use a new reservoir septum when assembling the new CBC.
Sample carryover from a previous injection	Sample carryover on the SeqStudio™ Genetic Analyzer can occur in trace amounts when previous injections are off-scale. Very strong peaks, for example, primer peaks in a fragment analysis sample, are typically visible in subsequent injections.	Replace the reservoir septum on the CBC before the next injection to minimize the carryover effect.

Observation	Possible cause	Recommended action
Dye blobs are seen in the sequencing data	Impurities remained in the sample after the sample purification. The impurities cause dye blobs to appear in the sequencing data.	Improve the sample purification method. See "Sample preparation guidelines" on page 37 for guidelines.
Extra peaks are present in the sequencing traces	There was renaturation of the sample.	Heat-denature the samples prepared with fresh Hi-Di™ Formamide, then immediately place the samples on ice.
	There is low signal. With very low signal, the peaks are barely visible in the baseline noise.	Check the raw data, the raw data signal intensity, and average raw signal-to-noise ratio, then: • Increase the injection time and reinject. *Or* • Remake the sample. Ensure that you are using: – Enough sequencing template – Enough primer and/or a sufficient concentration of primer
	There is a heterozygous insertion-deletion (het indel) that is causing multiple peaks to appear at the same basecall position. The sequence can appear "clean" for some number of bases until the het indel is encountered.	Examine the analyzed trace. A het indel typically has single peaks at the 5′ end, then part-way through the trace, two peaks appear in almost every position to the end of the trace. This pattern occurs when one copy of the gene has an insertion or deletion relative to the other copy of the gene. When aligning your sequence to a reference sequence, a series of bases may have been inserted or deleted in an allele. These indels can be encountered in any number of bases after the gene-specific priming region. To confirm that the het indel is present in both directions of your target, check the sequencing in the opposite direction.

Observation	Possible cause	Recommended action
Extra peaks are present in the sequencing traces *(continued)*	Primer-dimer has occurred.	You can often diagnose primer-dimer by looking at the raw trace data for questionable sequences. When primer-dimer exists, the 5′ sequence signal may be significantly higher for a region of bases spanning the length of the forward and reverse gene-specific primers. Primer-dimer is the annealing of the 3′ end of primers during PCR. The resulting short annealed fragment may amplify more efficiently than fully extended template. Primer-dimer fragments amplified during PCR can display increased 5′ signal and extra peaks when multiple PCR products are sequenced simultaneously. In some instances, the secondary or extra peaks can be read as the reverse compliment of the PCR primers in this noisy 5′ region. The secondary sequence or multiple PCR product sequences appear as far as 100–200 bp into the sequence, then suddenly disappear.
	The PCR amplification primers do not have specificity and are sequencing two different regions of the genome. The analyzed trace shows extra peaks throughout the entire length of the trace.	Redesign the primers or increase the amplification temperature.
	You accidentally contaminated the DNA and are sequencing two templates at the same time. The analyzed trace shows extra peaks throughout the entire length of the trace.	Repeat the amplification and sequencing reactions with uncontaminated DNA.
	There were impure or contaminated primers.	Primer stocks may have inadvertently had other primer solution introduced. For best results, use HPLC to purify the primers.
	There was a contaminated sample well.	Use a new sample plate and buffer/wash septa whenever possible.
		To avoid getting sample into adjacent wells, centrifuge the plates before you remove the adhesive seal.
There are pull-up or pull-down peaks in the data	The manual dye calibration is not current or is not matched to the samples, or a high-quality sample was not run with the dye set for the first time in the absence of a manual calibration.	If the dye set is a fragment analysis dye set that is being used for the first time, run a manual dye calibration. Alternatively, if a high-quality sample is run with an uncalibrated dye set, and the automated spectral calibration successfully generates an optimized matrix, manual calibration is not absolutely required. See "Determine if manual calibration is required" on page 193.

Observation	Possible cause	Recommended action
There are pull-up or pull-down peaks in the data *(continued)*	The incorrect dye set was selected in the **Plate setup**. **Note:** This is applicable to both sequence analysis and fragment analysis.	Correct the dye set and repeat the injection.
There are error messages about spectral issues These error messages are seen more often in fragment analysis data than sequence analysis data.	A high-quality sample has not been run with the dye set.	If the dye set is a fragment analysis dye set that is being used for the first time, run a manual dye calibration. Alternatively, if a high-quality sample is run with an uncalibrated dye set, and the automated spectral calibration successfully generates an optimized matrix, manual calibration is not absolutely required. See "Determine if manual calibration is required" on page 193.
There is a short read length and uneven peak spacing in sequence data	The incorrect dye set was selected in the **Plate setup**.	Correct the dye set and repeat the injection. Select from the following: • **E_BigDye™ Terminator v1.1** • **Z_BigDye™ Terminator v3.1** • **Z_BigDye™ Direct**
		Reanalyze the data in the Sequencing Analysis Software, using the correct mobility file.
Fragment analysis peaks are sized differently than previously observed	Aging of the polymer in the cartridge, which can cause small (≤0.5 bp) changes in fragment size.	• Use a reference marker (for example, an allelic ladder) for auto bin adjustment. *Or* • Manually adjust the fragment bin positions to account for the size change.

EPT data

Observation	Possible cause	Recommended action
There is a fluctuating or unstable electrophoresis current, or **Current check failed** is displayed	There is a bubble in the polymer system.	Flush the capillary array. 1. Touch ⚙ **Settings** ▸ **Cartridge** ▸ **Cartridge maintenance**. 2. Touch **Refresh PDS**. 3. Touch ⊗ to close the screen when the function ends.
		Replace the cartridge if the problem persists.
	The cartridge is damaged.	Inspect the cartridge for damage and replace the cartridge.

Software troubleshooting

Observation	Possible cause	Recommended action
Strikethrough text is displayed in **Save location** in **Plate properties** screen  ① Strikethrough text	The original location to which the plate was saved is no longer accessible by the instrument.	No action.
Import failed message when you select a CSV plate setup on the instrument The message also indicates that a valid run module, dye set, and/or size standard is not present.	The CSV file specifies the name of a run module, dye set, and/or size standard that does not exist on the instrument. **IMPORTANT!** A CSV file contains only the name of the run module, dye set, and/or size standard, it does not contain the settings.	Add the run module, dye set, and/or size standard to the instrument by selecting a PSM file that contains the items, or by creating the items manually.
Import CSV fails but import PSM with same settings imports with no errors	The size standard, dye set, or run module specified in the CSV file does not exist on the instrument. The CSV file contains the *names only* of the size standard, dye set, or run module, it does not contain the settings. The PSM file contains the *names and settings* of the size standard, dye set, or run module. The size standard, dye set, or run module are automatically created when the PSM file is imported.	Create the size standard, dye set, or run module on the instrument.
When saving a PSM file in Plate Manager (Connect cloud-based platform), "You do not have edit permission to the cloud group" message is displayed	You are saving to a Connect platform group and you do not have edit permissions for the group.	Save to a different location, or request edit permissions from the group administrator.

Observation	Possible cause	Recommended action
Analysis settings or run module is not available for selection by some users This situation is seen on the instrument.	Analysis settings and run modules are saved per user. If the analysis settings or run modules are associated with a *hidden* plate setup, the analysis settings or run modules are not listed for selection unless the user who created the plate setup is signed in.	Change the plate setup security from **Hidden** to **Shared** (see "Hide or share a plate (Plate setup security) " on page 89).
		Create the analysis settings or run modules manually.
Your local instrument profile name is not available for sign in	Your local instrument profile is linked to your Connect platform account and is replaced by your Connect platform profile.	See "Local instrument profile roles and functions" on page 151.
		Sign in with your Connect platform profile.
Cloud account is already linked to a local profile message	You have a local instrument profile that has previously been linked to Cloud account. You touched **Get Started ▸ Connect**, then typed in your Cloud account email and password.	Return to **Sign in** screen, then select your local profile from the **Sign in** list.
	Another user was signed in to the instrument. You selected **Cloud** in the **Set up run** screen or in the **Save location** field in the **Plate properties** screen, then typed in your Cloud account email and password.	Return to **Sign in** screen, touch ⊛, touch **Sign out**, then select your local profile from the **Sign in** list.
The **Notifications** button is not displayed in **Instrument Settings**	You are signed in to the instrument with a local instrument profile.	Sign in to the instrument with a Connect platform instrument profile.
A button is dimmed	The function is available to users with administrator role only.	None.
Sample data files from the SeqStudio™ Genetic Analyzer can be opened but not analyzed in secondary analysis software • Sequencing Analysis Software • GeneMapper™ Software • SeqScape™ Software • Variant Reporter™ Software	An updated version of the secondary analysis software is required.	Download and run the latest version of secondary analysis software. See "Secondary analysis software" on page 29. **Note:** The SeqStudio™ Genetic Analyzer instrument model is listed as **3200**.
Files cannot be imported into Connect cloud-based platform applications or cannot be analyzed in Connect platform sequencing apps	The sample name or file name used a special character.	Rename the sample name or file name. **Note:** Special characters (/, &, @, %) will impact file import into Connect platform applications.

Observation	Possible cause	Recommended action
Unexpected error is displayed	There was a software error.	Follow the instructions in the error message.
		Export the log files to determine the potential source of the unexpected error. See "Export log files for plates, install runs, injections, and instrument" on page 170.
		Restart the system using the On/Off switch on the rear panel. See "Power on the instrument" on page 34.

Service tools

The **Service tools** function is password-protected and for service use only.



B Link the instrument to your Connect cloud-based platform account—detailed instructions

- Workflow: Set up the instrument for Connect cloud-based platform access . . 184
- Network requirements 184
- Link the instrument from a mobile device 185
- Link the instrument using a link code 186
- Set up email notifications from the instrument 190

Workflow: Set up the instrument for Connect cloud-based platform access

Register and obtain a Connect cloud-based platform account (page 50)

▼

Link the instrument to your Connect cloud-based platform account in any of the following ways:

- Link the instrument from a mobile device (page 185)
- Link the instrument using a link code (page 186)

Network requirements

The instrument is factory-configured for IPv4 TCP/IP communication and includes a fast Ethernet adapter (10/100 Mbps) with a RJ45-type connector for integrating the device into a local area network (LAN).

By default, the instrument is configured to use the Dynamic Host Configuration Protocol (DHCP) but can use a static IP address.

The instrument should be configured behind a firewall. Contact Support for information on required firewall exceptions.

Link the instrument from a mobile device

Create a Connect cloud-based platform PIN and generate the QR code on the instrument

1. Sign in to **thermofisher.com/connect**.
2. Click to access InstrumentConnect.
3. If you have not previously set up a PIN, click **Update PIN**, then enter a PIN that you will use to sign in to the instrument.
4. From the instrument **Sign In** screen, navigate to the **Connect to the Thermo Fisher Cloud** screen:

Do you have a local instrument profile?	Description
Yes	1. Touch **Sign in**, then enter your PIN. 2. In the home screen, touch **Setup run**. 3. In the **Setup run** screen, touch **Cloud**.
No	Touch **Get started** ▸ **Connect**.

5. Touch **Mobile devices**.



The QR code is displayed.



Register the instrument with the Instrument Connect App

1. On your mobile device, download the InstrumentConnect from the Apple Store or from Google™ Play.
2. Launch, then sign in to the mobile app on your mobile device.
3. Register the instrument:
   a. Touch ≣, then touch **Register Instrument**.
   b. Touch **QR code** on your mobile device.
   c. With your mobile device, scan the QR code displayed in the instrument touchscreen.

Link the instrument using a link code

1. From the instrument **Sign In** screen, navigate to the **Connect to the Thermo Fisher Cloud** screen:

Do you have a local instrument profile?	Description
Yes	1. Touch **Sign in**, then enter your PIN. 2. In the home screen, touch **Setup run**. 3. In the **Setup run** screen, touch **Cloud**.
No	Touch **Get started** ▸ **Connect**.

2. Touch **PC**.



A unique link code is displayed.



3. Sign into your Connect cloud-based platform account on a separate computer. Go to **thermofisher.com/connect**.

4. Click , then click **Add an Instrument**.





5. Select **SeqStudio**™, then click **Next**.



6. Enter the link code from the instrument touchscreen (from step 2), then click **Send**.



7. If you have not previously set up a PIN, enter a PIN to use when you sign in to an instrument, then click **Send**.



A start linking message is displayed.

A confirmation message is displayed on the instrument touchscreen when the instrument is linked and connected to your Connect cloud-based platform account.

The first time the instrument is linked, the software automatically:

- Creates a Connect cloud-based platform instrument profile with the First Name and Last Name from your Connect cloud-based platform account.
- Registers the instrument in the InstrumentConnect software.

Set up email notifications from the instrument

When an instrument is linked to your Connect cloud-based platform account, email notifications are automatically sent to your Connect platform account email address.

Perform this procedure to disable any of the default notifications.

1. Sign in to the instrument with your Connect platform instrument profile and PIN.

2. In the home screen of the instrument, touch ⚙ **Settings ▸ Instrument settings ▸ Email notifications**.



Note: If you are signed in with a local instrument profile instead of a Connect platform instrument profile, the **Email notifications** button is not displayed on the **Instrument Settings** screen.

3. In the **Email notifications** screen, select or deselect the options for which you want to receive email notifications, then touch **Done**.



C Dye calibration and install standard checks

- Calibrate dyes 192
- Perform an install run 203

Calibrate dyes

Overview of system dye set and custom dye set calibration

Dye calibration compensates for dye emission spectral overlap.



Figure 20 Dye emission spectra for the J6 (DS-36) dye set showing spectral overlap

System versus custom dyes

System dye sets are available from Thermo Fisher Scientific.

System dye sets	
Fragment/HID analysis dye sets	**Sequence analysis dye sets**
• D (DS-30, DS-31) • E5 (DS-02) • F (DS-32) • G5 (DS-33) • J6 (DS-36) • J6-T (DS-37)	• E_BigDye™ Terminator v1.1 • Z_BigDye™ Terminator v3.1 • Z_BigDye™ Direct

Custom dye sets are any dyes that are not available from Thermo Fisher Scientific. Custom dye sets require manual calibration.

Factory, auto, and manual calibration

Three types of calibration can occur on the instrument:

- **Factory calibration**—Default calibration provided with the instrument. It is not optimized for a specific instrument.
- **Manual calibration**—Manual procedure performed by the user that provides a baseline calibration for the instrument on which it is run. Reduces pull-up (false secondary peaks under a true peak).
- **Auto calibration**—Automatic adjustment of the baseline calibration to optimally reduce pull-up (false secondary peaks under a true peak). The instrument performs an auto calibration for system and custom dyes.

Determine if manual calibration is required

- **Sequence analysis dye sets**—Do not typically require manual calibration.
- **Fragment analysis dye sets**—Manual calibration is recommended one time before use.

  Note: With high-quality sample data, it is possible for auto calibration to pass using the factory calibration.
- **HID analysis dye sets**—Manual calibration is required for each dye set one time before use.

To determine if a dye set requires manual calibration, review the calibration history for the dye set.

In the home screen:

1. Touch ⚙ **Settings ▸ Maintenance and Service ▸ Calibration ▸ Calibration history**.

2. Examine the entry for the dye set of interest. If the dye set does not list a date and a cartridge serial number for a manual calibration or a date for auto calibration, the dye set requires manual calibration.



Figure 21 Fragment analysis dye sets example

① Manual calibration is recommended—No auto calibration or manual calibration has been performed.
② Manual calibration is not required—A manual calibration has been performed.
③ Manual calibration is not required—The factory calibration has been optimized for the instrument by auto calibration.
④ Manual calibration is recommended—Auto calibration has been performed, but only 2 capillaries passed.
⑤ Manual calibration is not required—A manual calibration and an auto calibration have been performed.

Perform a system dye calibration

A system dye calibration requires ~30 minutes to complete.

Prepare the dye set calibration standards and plate as described in the product information sheet for the dye set.

In the home screen:

1. Touch ⚙ **Settings ▸ Maintenance and Service ▸ Calibration ▸ Dye Calibration**.
2. Touch the injection group for the dye set in the plate, then touch **Dye set**.



3. Touch **Sequence Standard** or **Matrix Standard**, then select a system dye calibration standard provided with the instrument.



4. Touch **Calibrate**.
   The calibration run starts.

IMPORTANT! If the dye calibration fails:

- The results of the calibration are not saved, and the calibration plate is not moved to **Run History**.
- The instrument does not allow you to rerun the plate setup for a failed calibration. Close the calibration screen, then start a new calibration.

Spectral Quality Value

A spectral Quality Value reflects the confidence that the individual dye emission signals can be separated from the overall measured fluorescence signal. It is a measure of the consistency between the final matrix and the data from which it was computed. A Quality Value of 1.0 indicates high consistency, providing an ideal matrix with no detected pull-up/pull-down peaks.

In rare cases, a high Quality Value can be computed for a poor matrix. This can happen if the matrix standard contains artifacts, leading to the creation of one or more extra peaks. The extra peaks cause the true dye peak to be missed by the algorithm, and can lead to a higher Quality Value than would be computed with the correct peak. Therefore, it is important to visually inspect the spectral calibration profile for each capillary.

Condition number

A Condition Number indicates the amount of overlap between the dye peaks in the fluorescence emission spectra of the dyes in the dye set.

If there is no overlap in a dye set, the Condition Number is 1.0 (ideal conditions), the lowest possible value. The condition number increases with increasing peak overlap.

The ranges that the software uses to determine if a capillary passes or fails are:

Dye Set	Quality Value Minimum	Condition Number Maximum
AnyDye	0.8 (default)	20.0 (default)
D	0.8	8.5
E	0.95	5.5
E5	0.95	6.0
F	0.95	8.5
G5	0.95	13.5
J6	0.95	8.0
J6-T	0.95	8.0
Z	0.95	5.5

Perform a custom dye calibration

Add a custom dye set to the software based on a system dye

1. In the home screen, touch ⚙ **Settings** ▸ **Maintenance and Service** ▸ **Calibration** ▸ **Dye calibration**.



2. Touch **Custom Dye**.

3. Touch **Add**.

4. Touch the system dye to use as a starting point for the custom dye settings.

5. Modify the settings as needed then touch **Next**.



6. Enter a **Dye set name**, then touch **Done**.

Add a custom dye set to the software using the **Any Dye** template

1. In the home screen, touch ⚙ **Settings** ▸ **Maintenance and Service** ▸ **Calibration** ▸ **Dye calibration**.

2. Touch **Custom Dye**.

3. Touch **Add**.

4. Select **Any Dye** to use as a starting point for the custom dye settings. All dyes are selected by default.

5. Touch a dye color, then manage the dye set arrangement:
   - Touch **Dye selection** to deselect the dye.
   - Touch **Move up** or **Move down** to organize the dyes in the order in which they occur as peaks in the electropherogram of the custom dye set standard.
   - Touch **Reduce selection** to exclude it from the calibration.

   Note: If you are calibrating with a system dye set, but will not run all system dyes in your application:
   - *Enable* (select) the **Dye selection** checkbox (to indicate that the system dye is present in the matrix standard)
   - *Disable* (deselect) the **Reduce selection** checkbox (to indicate that the system dye should not be analyzed, reported, or used for auto calibration)



① Order of dyes in the custom dye set

6. Touch **Next**, enter dye set parameters, then touch **Next**.

7. Enter a **Dye set name**, then touch **Save**.

Perform a custom dye calibration

Before you begin:

- Prepare the dye set calibration standards and plate as described in the product information sheet for the dye set.
- Add a custom dye set to the software (see "Add a custom dye set to the software based on a system dye" on page 197) .

In the home screen:

1. Touch ⚙ **Settings** ▸ **Maintenance and Service** ▸ **Calibration** ▸ **Dye calibration**.
2. Touch the injection group for the dye set in the plate, then touch **Dye set**.



3. Touch a custom dye set.
4. Touch **Calibrate**.
5. When the run is complete, touch **View results**.
6. Ensure that all capillaries passed the calibration, then click **Done**.

IMPORTANT! If the dye calibration fails:

- The results of the calibration are not saved, and the calibration plate is not moved to **Run History**.
- The instrument does not allow you to rerun the plate setup for a failed calibration. Close the calibration screen, then start a new calibration.

Add a custom dye to the Plate Manager or another instrument

- To add a custom dye set to the Plate Manager:
  a. Open a plate setup that specifies the custom dye set of interest.
  b. Export a plate setup that specifies the custom dye (see "Export or delete a plate setup (PSM file)" on page 138).
  c. Open the exported plate setup in the Plate Manager.
- To transfer a custom dye set to another instrument:
  a. Touch **Settings** ▸ **Maintenance and Service** ▸ **Calibration** ▸ **Calibration history**.
  b. Touch **Custom dye**.
  c. Touch **Manage**.

d. Touch a custom dye, then touch **Export**.

e. Select a location, then touch **Export**.

f. Import the custom dye on another instrument (⚙ **Settings** ▸ **Maintenance and Service** ▸ **Calibration** ▸ **Dye calibration** ▸ **Custom dye** ▸ **Manage** ▸ **Import**).

View the dye calibration history

In the home screen:

1. Touch ⚙ **Settings** ▸ **Maintenance and Service** ▸ **Calibration history**.



① Factory calibration (default calibration that is not optimized for the instrument).
② Manual dye calibration performed by a user.
③ Factory calibration that has been optimized by auto calibration and does not require manual calibration.
④ Automatic adjustment of the baseline calibration to optimally reduce pull-up (false secondary peaks under a true peak).
⑤ A manual dye calibration and an auto dye calibration have been performed.

2. Touch a dye set.
   A thumbnail of each capillary calibration spectrum is displayed, with the q, and c values (see "Spectral Quality Value" on page 196 and "Condition number" on page 197).

3. Touch ❯ and ❮ to view a full-screen calibration spectrum for each capillary.

4. Touch ⊖ to return to the list of dye sets on the **Calibration History** screen.

5. *(Optional)* Touch **Filter** to narrow the dye set list down by dye sets.
   a. Select or deselect the dyes listed.

b. *(Optional)* Touch **Deselect All** to clear the dye set selections.

c. Touch **Done**.

6. Touch **OK**.

Import or export a custom dye set

In the home screen:

- Touch ⚙ **Settings** ▸ **Maintenance and Service** ▸ **Calibration** ▸ **Dye calibration** ▸ **Custom dye** ▸ **Manage**.
  1. Touch the custom dye set.
  2. Touch **Import** or **Export**.

Perform an install run

Overview of install checks

Install checks are performed by a Field Service Engineer at the time of installation.

An install check can be run at any time with the following reagents to ensure instrument performance:

Install check type	Reagent
Sequencing	• 3500/3500xL Sequencing Standards, BigDye™ Terminator v3.1 (Cat. No. 4404312) • 3500/3500xL Sequencing Standards, BigDye™ Terminator v1.1 (Cat. No. 4404314)
Fragment	DS-33 GeneScan™ Installation Standards with GeneScan™ 600 LIZ™ Size Standard v2.0 (Cat. No. 4376911)
HID **Note:** A J6 manual dye calibration is required before the install check.	GlobalFiler™ Allelic Ladder from one of the following installation kits: • SeqStudio™ HID and Sequencing Installation Kit (Cat. No. A46180) • SeqStudio™ HID Installation Kit (Cat. No. A46182)

You can include multiple injection groups, multiple applications, and/or multiple chemistries on an install run plate. For example, you can prepare an install run plate that contains replictate injections of the fragment install standard and both sequencing standards.

Figure 22 Example install run plate with multiple applications and replicate injections

Perform an install run

An install run requires ~45 minutes to complete each injection group.

Prepare the installation standard and plate as described in the product information sheet for the installation standard.

In the home screen:

1. Touch ⚙ **Settings** ▸ **Maintenance and service** ▸ **Install run** ▸ **Install run**.
2. Touch an injection group on the plate to select a location for the install standards.



A plate can be set up with different injection groups for one install run.

3. Touch **Chemistry**, then select the type of install run.
   Select **None** to cancel the selection.



4. Repeat the steps above for additional injection groups and/or multiple chemistries.
5. Touch **Start Run**.

View install run results

1. Open the install run results summary:
   - To view results immediately after the install run completes, touch **Results**.
   - To view results for a previous install run, touch ⚙ **Settings ▸ Maintenance and Service ▸ Install run ▸ Install run history**.



2. Touch an install run to display the results, including the pass/fail results.



Each injection group displays a QC color for each capillary:

- —All QC tests passed.
- —At least 1 warning quality alert was triggered.
- —At least 1 failing quality alert was triggered.

For information on quality alerts, see:

- "Data quality alerts" on page 119
- "Sizecalling and basecalling quality alerts" on page 119

3. Touch **Detail** to see allele-specific details for fragment/HID analysis and basecall accuracy for sequencing.



	Dye	Allele	Mean size	Sizing precision	Pass/Fail
1		D5S644[VIC][1]	85.13	0.04	Pass
2		D5S644[VIC][2]	96.95	0.03	Pass
3		D20S119[FAM][...	113.68	0.00	Pass
4		D20S119[FAM][...	119.61	0.05	Pass
5		D9S288[NED][1]	137.52	0.04	Pass
6		D9S288[NED][2]	145.27	0.02	Pass

Cap #	% QV30 CRL Basecall Accuracy	Alignment Read Length (range)	% RL Basecall Accuracy
1	100.0	553 (8-560)	99.1
2	100.0	562 (8-569)	99.1
3	99.6	555 (7-561)	99.1
4	98.7	553 (7-559)	99.1

4. Touch **Export report** (see "Export the install run report" on page 208).

Fragment analysis install run pass/fail criteria

Result	Acceptance criteria
# Allele peaks	15 for all four capillaries
# Size standard peaks	34 for all four capillaries
Standard deviation of the observed allele fragment sizes for all 4 capillaries	<0.15 base pairs (bp)

HID analysis install run pass/fail criteria

Result	Acceptance criteria
# Allele peaks	343 for all four capillaries
# Size standard peaks	26 for all four capillaries
Minimum peak height	>400 RFU
Standard deviation of the observed allele fragment sizes for all 4 capillaries	<0.15 base pairs (bp)

Sequence analysis install run pass/fail criteria

Result	Acceptance criteria
QV30 CRL (contiguous read length) **Note:** The contiguous read length passing criteria for install checks is an uninterrupted segment of bases with an average Quality Value (QV) of 30.	BigDye™ Terminator v3.1 Capillaries with QV30 CRL ≥500 bp pass.
	BigDye™ Terminator v1.1 Capillaries with a QV30 CRL ≥500 bp pass.

The remaining results on the screen are for information only.

Export the install run report

1. Touch ⚙ **Settings** ▸ **Maintenance and Service** ▸ **Install run** ▸ **Install run history**.
2. Touch the install run of interest.
3. Touch **Export report**.
4. Select a destination for the report.
5. Touch **OK**.

Example fragment analysis run report

applied biosystems by Thermo Fisher Scientific

Install run report

Report name: Fragment Install Run_20170411_093309 Date: 2017/04/11

System Information

Instrument Name:	seqstudio-SVT2	Run Date:	2017/04/10
Instrument Serial Number:	132000002	Chemistry Type:	Fragment Analysis
Instrument Software Version:	SVT-1.4.21-1306-r12832	Cartridge Serial Number:	2852889004
Buffer Installation Date:	2017/03/04	Cartridge Installation Date:	2017/04/10
Buffer Expiration Date:	2017/03/18	Cartridge Expiration Date:	2017/11/16

Install Run Result

Capillary	1	2	3	4
# of Allele Peaks	15	15	15	15
# of Size Standard Peaks	34	34	34	34

Legend: Passed Failed

Running info

	Dye	Allele	Nominal Size	Mean	Average peak height	Peak height% > Min	Sizing precision	Sizing accuracy	Pass/Fail
1		D5S644[VIC][1]	85.06	85.1	1851.75	100.0	0.02	0.04	Pass
2		D5S644[VIC][2]	96.89	96.92	1184.0	100.0	0.03	0.03	Pass
3		D20S119[FAM][1]	113.7	113.67	1393.25	100.0	0.0	-0.03	Pass
4		D20S119[FAM][2]	119.6	119.56	1308.5	100.0	0.05	-0.04	Pass
5		D9S288[NED][1]	137.5	137.38	1545.0	100.0	0.02	-0.12	Pass
6		D9S288[NED][2]	145.24	145.09	1044.75	100.0	0.04	-0.15	Pass
7		D6S289[PET][1]	171.49	171.35	1677.0	100.0	0.05	-0.14	Pass
8		D6S289[PET][2]	173.41	173.25	1054.0	100.0	0.03	-0.16	Pass
9		D5S424[VIC][1]	216.39	216.26	1457.5	100.0	0.04	-0.13	Pass
10		D5S424[VIC][2]	218.27	218.14	955.75	100.0	0.03	-0.13	Pass
11		D9S1690[FAM][1]	236.74	236.48	2030.0	100.0	0.06	-0.26	Pass
12		D9S1690[FAM][2]	238.67	238.38	1300.5	100.0	0.04	-0.29	Pass
13		D18S462[NED][1]	303.06	302.86	1718.25	100.0	0.06	-0.2	Pass
14		D15S117[PET][1]	336.81	336.71	1331.0	100.0	0.02	-0.1	Pass
15		D15S117[PET][2]	338.73	338.63	914.25	100.0	0.01	-0.1	Pass

Capillary (Raman) Uniformity 0.4

The spectral dye matrix has not been updated.

Example sequencing run report

applied biosystems by Thermo Fisher Scientific

Install run report

Report name: Sequencing Install Run_20170424_131753 Date: 2017/04/24

System Information

Instrument Name:	seqstudio-SVT1	Run Date:	2017/04/24
Instrument Serial Number:	132000001	Chemistry Type:	BDT v3.1 (General Sequencing)
Instrument Software Version:	0.6.10	Cartridge Serial Number:	2883720037
Buffer Installation Date:	2016/12/05	Cartridge Installation Date:	2017/04/23
Buffer Expiration Date:	2016/12/19	Cartridge Expiration Date:	2017/11/17

Install Run Result

Capillary	1	2	3	4	Mean	SD
Contiguous Read Length (CRL) (range)	530 (26-555)	535 (29-563)	531 (29-559)	530 (32-561)	531.5	2.1
CRL Pass/Fail						
Comparison with Reference Sequence						
%CRL Basecall Accuracy	99.8	99.6	99.8	100.0		
%RL Basecall Accuracy	99.1	99.1	99.1	99.1		
Alignment Read Length (range)	562 (5-566)	557 (19-575)	556 (15-570)	562 (5-566)		

Legend: Passed Failed

Capillary (Raman) Uniformity 0.0

Signal Strength

Capillary #	Sample file name	G	A	T	C
1	A2_A2_20170424_104040.ab1	792	1275	968	917
2	B2_B2_20170424_104041.ab1	1016	1625	1193	1108
3	C2_C2_20170424_104042.ab1	751	1336	1126	878
4	D2_D2_20170424_104043.ab1	635	1087	901	742

View the install run history

In the home screen:

1. Touch ⚙ **Settings** ▸ **Maintenance and Service** ▸ **Install run history**.



2. *(Optional)* Touch **Filter**, select the install run type, then touch **Done**.



Note: If the screen is blank when you click **Done**, no install runs are present for the chemistry you selected.

D Create and run an HID plate

- HID analysis requirements 212
- Pull-up reduction feature 213
- Resolution algorithm 214
- Workflow: HID analysis 215
- HID install run 215
- HID run module, dye sets, size standard definitions, and kits 216
- Manage kits (HID applications only) 217

HID analysis requirements

- **Cartridge**—Use SeqStudio™ Genetic Analyzer Cartridge v2 (Cat. No. A41331), which has been validated for HID analysis. The v1 cartridge has not been validated.

  To check the Cat. No., you must remove the cartridge from the instrument. See "Store the cartridge" on page 163 for instructions.
- **Manual dye calibration and HID install check**—When the instrument is installed, the field service representative will perform a manual J6 dye calibration and an HID install check.

  Before you use any other dye sets, perform a manual dye calibration to optimize the pull-up reduction function. For more information, see "Calibrate dyes" on page 192.
- **Allelic ladder**—We recommend that you inject one allelic ladder for each set of 24 samples in HID runs (1 allelic ladder per 6 injections).

  Allelic ladders that are injected under the same conditions are recommended to accurately genotype samples in the secondary analysis software.

  IMPORTANT! Variation in laboratory temperature can cause changes in fragment migration speed that can, in turn, cause sizing variation. We recommend the frequency of allelic ladder injections described above to account for normal variation in fragment migration speed. However, during internal HID validation studies, verify the required allelic ladder injection frequency to ensure accurate genotyping of all samples in your laboratory environment.
- **GeneMapper™ *ID-X* Software compatibility**—The data files that are generated by the instrument are compatible with GeneMapper™ *ID-X* Software v1.6 or later.

Pull-up reduction feature

The software includes two pull-up reduction features:

- Automatic pull-up reduction feature that is applied with all dye sets.
- *(HID analysis only for Applied Biosystems™ dyes)* Marker-to-marker pull-up reduction that is optimized for the markers in each kit.

  You can enable/disable this feature for each injection group when you create an HID plate.
  - To enable, select the appropriate kit from the **Kits** list.
  - To disable, select **None** from the **Kits** list.

    Note: "Kit" is displayed when **None** is selected.

  To display this screen where you select a kit: Select or create a plate set up, touch the **Plate** tab, then select **Edit**. In the **Edit Plate** screen, swipe up to find the **Kit** field, then select a kit from the dropdown list.



Note: You can also select a kit from the Plate Manager.

Resolution algorithm

A resolution algorithm, unique when running an HID application, is incorporated into SeqStudio™ Data Collection Software v1.2. The purpose of the algorithm is to align the resolution power traditionally generated with a 3500/3500xL Genetic Analyzer/POP-4™ Polymer configuration to the resolution power generated with a SeqStudio™ Genetic Analyzer/POP–1™ Polymer configuration.

About POP™ polymer matrices

POP™ polymer matrices separate DNA fragments of a known size range, at a desired resolution and run time. Polymers are available in different concentrations, which directly relate to the viscosity and length of the polymer molecule. Different concentrations are suited to different applications. For example:

- **Sequencing analysis applications**—Traditionally use POP-6™ and POP-7™ Polymers. POP-6™ and POP-7™ Polymers are composed of relatively longer molecules that provide greater DNA-to-polymer interaction and cause DNA fragments to move through the polymer more slowly. This design results in a longer run time and the increased resolution that is required for sequencing.
- **Fragment analysis applications**—Traditionally use POP-4™ Polymer. Because of its polymer concentration and molecule size, POP-4™ Polymer has single basepair resolving power within the size range for forensic STR amplification kits and runs faster.

About POP–1™ Polymer

We are introducing POP–1™ Polymer for the SeqStudio™ Genetic Analyzer, which can be used for all applications on the instrument.

For smaller fragments, POP–1™ Polymer resolves similarly to POP-7™ Polymer. For larger fragments, POP–1™ Polymer resolves between POP-6™ and POP-7™ Polymers. In addition to providing higher resolution for expected STR fragment peaks, peaks intrinsic to the amplification process in the n–1/n+1 position from the main allele may resolve as shoulders with POP–1™ Polymer. Because of the limited resolving power of POP-4™ Polymer, these shoulders were not always visible in our legacy data sets.

The resolution algorithm is applied during data collection. The algorithm maintains similar resolution for smaller fragments across the 3500/3500xL Genetic Analyzer and SeqStudio™ Genetic Analyzer, but allows for the improved resolution for larger fragments that benefit from the POP–1™ Polymer.

Workflow: HID analysis

Get started
Prepare the instrument (page 39)
Prepare the samples (page 37)
Create a plate setup on the instrument
Create or import a plate setup (page 81)
Enter plate properties (page 82)
Assign wells: run module, size standard, dye set, and kit (page 84)
Assign wells: sample name, sample type, and custom fields (page 86)
Start and monitor a run
On the instrument: Load the plate or the tube assembly (page 92)
Select a plate setup and start a run (page 93)
Monitor a run from the instrument (page 105)
View and analyze results
"View results on the instrument" on page 111
Export results from the instrument (sample data files and QC reports) (page 121)
Analyze data in a secondary analysis software application

HID install run

Item	Description
Components of the install check reaction	• GeneScan™ 600 LIZ™ Size Standard v2.0 (Cat. No. 4408399) • Hi-Di™ Formamide (Cat. No. 4311320 or 4440753) • GlobalFiler™ Allelic Ladder from one of the following installation kits: – SeqStudio™ HID and Sequencing Installation Kit (Cat. No. A46180) – SeqStudio™ HID Installation Kit (Cat. No. A46182)

Item	Description
Volumes of install check components per reaction	• Size standard—0.4 µL • Hi-Di™ Formamide—9.6 µL • Allelic ladder—1 µL
Pass/fail criteria for HID J6 install check	• 343 ladder peaks • 26 size standard peaks • Minimum peak height >400 RFU • Standard deviation of the observed allele fragment sizes for all 4 capillaries: <0.15 base pairs (bp)

HID run module, dye sets, size standard definitions, and kits

Run module	Resolution range	Approximate run time	Sizing precision	Compatible size standards
HIDAnalysis	60–470 bp[1]	39 minutes	60–470: <0.15	• GeneScan™ 500 LIZ™ Size Standard[2] • GeneScan™ 600 LIZ™ Size Standard v2.0

[1] Resolution Range: The range of bases over which the resolution (peak spacing interval divided by the peak width at half-max in a GS600 size standard sample sized with a third order fit) is ≥1.
[2] The GeneScan™ 500 LIZ™ Size Standard was not included in the HID validation of the instrument.

Category	Items
Dye sets	• G5 (DS-33) • J6 (DS-36) • J6-T (DS-37)
Size standards	• GS500(-250)LIZ • GS600_LIZ_(60-460) • GS600_LIZ_(80-400)
Kits	• Setting of **None** disables the Marker-to-Marker pull-up feature ("Kit" is displayed when **None** is selected. • GlobalFiler™ • GlobalFiler™ Express • Huaxia™ Platinum • Identifiler™ Plus • MiniFiler™ • NGM Detect™ • NGM SElect™ • VeriFiler™ Express • VeriFiler™ Plus • Yfiler™ • Yfiler™ Plus

Manage kits (HID applications only)

The kit files for the kits that have been validated for the instrument are preloaded in the software. If a new or updated kit file is provided, import the kit file as described below.

1. Copy the kit file to a USB or a network drive.
2. Access the **Manage kits** screen:

From	Actions
Plate properties tab	Touch **More options** ▸ **Manage kits**.
Home screen	Touch ⚙ **Settings** ▸ **Run settings** ▸ **Kits**.

3. Touch **Import**, select the location of the kit file. Select the kit, then touch **Import**. The kit file is loaded in the software.

E Run modules

Run modules, read lengths, size ranges, and run times

Table 15 Sequencing run modules for standard sequencing

Run module	Contiguous read length (CRL)[1]	QV threshold	Approximate run time
ShortSeq	≥350	QV30	30 minutes
ShortSeq_BDX			
MediumSeq	≥500	QV30	45 minutes
MediumSeq_BDX			
LongSeq	≥800	QV20	~ 2 hours
LongSeq_BDX			

[1] CRL was determined using the Long Read Sequencing standard. A minimum of 90% of analyzed sequences with an average QV ≥QV threshold were observed.

IMPORTANT! Use BDX run modules only if you prepare samples with BigDye XTerminator™ Purification Kit. Use non-BDX run modules for samples purified with other methods.

Table 16 Fragment analysis run modules

Run module	Resolution range	Approximate run time	Sizing precision	Compatible size standards
SNaPshot	40–120 bp	25 minutes	40–120: <0.5	GeneScan™ 120 LIZ™ Size Standard
FragAnalysis	60–460 bp[1]	45 minutes	60–460: <0.15	All except GeneScan™ 1200 LIZ™ Size Standard
LongFragAnalysis[2]	60–600 bp[1]	< 2 hours	60–460: <0.15 461–600: <0.3 601–800: >0.45	• GeneScan™ 600 LIZ™ Size Standard v2.0 • GeneScan™ 1200 LIZ™ Size Standard

[1] Resolution Range: The range of bases over which the resolution (peak spacing interval divided by the peak width at half-max in a GS600 or GS1200 LIZ size standard sample sized with a third order fit) is ≥1. The table shows the resolution range in ≥90% of samples.

[2] Load a maximum of 48 samples per plate if you use a long run module.

Table 17 HID run modules

Run module	Resolution range	Approximate run time	Sizing precision	Compatible size standards
HIDAnalysis	60–470 bp[1]	39 minutes	60–470: <0.15	• GeneScan™ 500 LIZ™ Size Standard[2] • GeneScan™ 600 LIZ™ Size Standard v2.0

[1] Resolution Range: The range of bases over which the resolution (peak spacing interval divided by the peak width at half-max in a GS600 size standard sample sized with a third order fit) is ≥1.

[2] The GeneScan™ 500 LIZ™ Size Standard was not included in the HID validation of the instrument.

Note: The following size standards have not been validated for use with the instrument. A default size standard definition is not provided in the software.

- GeneScan™ 350 ROX™ Size Standard
- GeneScan™ 400HD ROX™ dye Size Standard

Parts and materials

Required materials not supplied

Unless otherwise indicated, all materials are available through **thermofisher.com**.
MLS: Fisher Scientific (**fisherscientific.com**) or other major laboratory supplier.

Item	Source
Software	
(Optional) SeqStudio™ Remote Monitoring App	Available on the Connect cloud-based platform
(Optional) InstrumentConnect	
(Optional) SeqStudio™ Plate Manager and App	Available on the Connect platform or for download at thermofisher.com
Equipment	
(Optional) Handheld Barcode Scanner	4488442
Consumables for SeqStudio™ Genetic Analyzer	
SeqStudio™ Genetic Analyzer Cartridge	A33671
SeqStudio™ Genetic Analyzer Cartridge v2	A41331
SeqStudio™ Genetic Analyzer Cathode Buffer Container	A33401
Reservoir Septa (for Cathode Buffer Container)	A35640
SeqStudio™ Integrated Capillary Protector	A31923
Tubes, plates, and other consumables	
MicroAmp™ Optical 96-Well Reaction Plate	4316813
MicroAmp™ Optical 96-Well Reaction Plate with Barcode	4326659
MicroAmp™ Optical 8-Tube Strip, 0.2 mL	4316567
Septa for SeqStudio™ Genetic Analyzer, 96 well	A36541
Septa for SeqStudio™ Genetic Analyzer, 8 strip	A36543
MicroAmp™ 96-Well Tray/Retainer Set (Adapter for 8-Tube Strip)	403081
Reagents	
Hi-Di™ Formamide	4401457

Item	Source
BigDye XTerminator™ Purification Kit	4376486
Sequencing kits	
BigDye™ Terminator v1.1 Cycle Sequencing Kit	4337449
BigDye™ Terminator v3.1 Cycle Sequencing Kit	4337454
BigDye™ Direct Cycle Sequencing Kit	4458689
Fragment analysis size standards and installation standard	
GeneScan™ 600 LIZ™ Size Standard v2.0	4408399
GeneScan™ 120 LIZ™ Size Standard	4324287
GeneScan™ 1200 LIZ™ Size Standard	4379950
GeneScan™ 500 ROX™ Size Standard	401734
GeneScan™ 500 LIZ™ Size Standard[1]	4322682
GeneScan™ 350 ROX™ Size Standard[1]	401735
GeneScan™ 400HD ROX™ dye Size Standard[1]	402985
DS-33 GeneScan™ Installation Standards with GeneScan™ 600 LIZ™ Size Standard v2.0	4376911
HID analysis size standards and installation standard	
GeneScan™ 500 LIZ™ Size Standard[2]	4322682
GeneScan™ 600 LIZ™ Size Standard v2.0	4408399
SeqStudio™ HID and Sequencing Installation Kit	A46180
SeqStudio™ HID Installation Kit	A46182

[1] For fragment analysis: This size standard has not been validated for use with the instrument. A default size standard definition is not provided in the software.

[2] For HID analysis: This size standard was not included in the HID validation of the instrument. A default size standard definition is provided in the software.

G Instrument specifications and layout

- Instrument dimensions 222
- Instrument clearances 222
- Environmental requirements 223
- Electrical requirements 224
- Electrical protective devices 224
- Network requirements 225
- Safety requirements 225

Instrument dimensions

Ensure that the installation site bench space can accommodate the dimensions and support the weight.

Configuration	Height	Length (depth)	Width	Weight
(Door Closed) (Door Open)	44.2 cm (17.4 in.) 56.9 cm (22.4 in.)	64.8 cm (25.5 in.)	49.5 cm (19.5 in.)	53.5 kg (118 lbs)

WARNING! PHYSICAL INJURY HAZARD. Do not attempt to lift or move the instrument without professional assistance. The crated instrument is heavy. Any incorrect lifting or moving of the crated instrument can cause serious injury.

Instrument clearances

During instrument setup and maintenance, it is necessary to access the back and sides of the instrument. If the back of the instrument faces a wall, it will be necessary to have enough space to rotate the instrument on the bench for access.

IMPORTANT! For safety, the power outlet used for powering the instrument must be accessible at all times.

Component	Top	Front	Left	Right	Back
SeqStudio™ Genetic Analyzer	30.5 cm (12.0 in)	30.5 cm (12.0 in)	10.0 cm (4.0 in)	20.0 cm (8.0 in)	10.0 cm (4.0 in)

Environmental requirements

Ensure that the installation room is maintained under correct environmental conditions.

Condition	Acceptable range
Installation site	Indoor use only
Electromagnetic interference	Do not use this device in close proximity to sources of strong electromagnetic radiation (for example, unshielded intentional RF sources). Strong electromagnetic radiation may interfere with the proper operation of the device. This equipment has been designed and tested to CISPR 11 Class A. In a domestic environment it may cause radio interference. You may need to take measures to mitigate the interference.
Altitude	Located between sea level and 2000 m (6500 ft.) above sea level
Humidity (instrument and computer)	• Operation: 20%–80% (noncondensing) • Transport and storage: 15%–80% (noncondensing)
Temperature (instrument and computer)	• Operation: 15°C to 30°C (60°F to 85°F) **Note:** The room temperature must not fluctuate more than 2°C over a 2-hour period. • Transport and storage: –20°C to 60°C (–4°F to 140°F)
Transient category	Installation categories II
Overvoltage category	Installation categories II
Vibration	Ensure that the instrument is not adjacent to strong vibration sources, such as a centrifuge, pump, or compressor. Excessive vibration will affect instrument performance.
Pollution degree	II Install the instrument in an environment designated pollution degree II (only non-conductive pollution (e.g., dust) occurs except that occasionally a temporary conductivity caused by condensation is to be expected). Typical pollution degree II environments are laboratories and office spaces.
Liquid waste collection	Dispose of the polymer, buffer, reagents and any liquid waste as hazardous waste in compliance with local and national regulations.
Other conditions	Ensure the installation site is away from any vents that could expel particulate material on the system components. Avoid placing the instrument and computer adjacent to heaters, cooling ducts, or in direct sunlight.

Electrical requirements

CAUTION! Do not unpack or plug in any components until the Field Service Engineers (FSEs) have configured the system for the proper operating voltage.

WARNING! For safety, the power outlet used for powering the instrument must be accessible at all times. See "Instrument clearances" on page 222 for information about the space needed between the wall and the instrument. In case of emergency, you must be able to immediately disconnect the main power supply to all the equipment. Allow adequate space between the wall and the equipment so that the power cords can be disconnected in case of emergency.

- Electric receptacle required: 2-prong with ground pin
- Maximum power dissipation: 380 W (approximately, not including computer and monitor)
- Mains AC line voltage tolerances must be up to ±10 percent of nominal voltage

Rated voltage	Circuit required	Rated frequency	Rated power
100–240 ±10% VAC[1]	10 A	50–60 Hz	400 W

[1] If the supplied power fluctuates beyond the rated voltage, a power line regulator may be required. High or low voltages can adversely affect the electronic components of the instrument.

Electrical protective devices

We recommend several protective devices to protect the system in environments with large voltage and power fluctuations.

Device	Description
Power line regulator	We recommend the use of a 1.5-kVA power line regulator in areas where the supplied power fluctuates in excess of ± 10% of the normal voltage. Power fluctuations can adversely affect the function of the instrument and computer. **Note:** A power line regulator monitors the input current and adjusts the power supplied to the instrument or computer. It does not protect against a power surge or failure.

Device	Description
Uninterruptible power supply (UPS)	We recommend the use of a 1.5-kVA uninterruptible power supply (UPS), especially in areas prone to power failure. Power failures and other events that abruptly terminate the function of the instrument and computer can corrupt data and possibly damage the system. **WARNING! PHYSICAL INJURY HAZARD.** Do not attempt to lift the UPS unit without assistance of at least two people. Improper lifting can cause painful and permanent back injury. Refer to the UPS manufacturer user guide for more information. **IMPORTANT!** UPSs provide power for a limited time. They are meant to delay the effects of a power outage, not to serve as replacement power sources. In the event of a power loss, power off the instrument and computer unless you expect to regain power within the battery life of the UPS.
Surge protector	We recommend the use of a 10-kVA surge protector (line conditioner) in areas with frequent electrical storms or near devices that are electrically noisy, such as refrigerators, air conditioners, or centrifuges. Short-duration, high-voltage power fluctuations can abruptly terminate the function of, and thereby damage the components of, the computer and the instrument. **Note:** A dedicated line and ground between the instrument, computer, and the building's main electrical service can also prevent problems caused by power fluctuations.

Network requirements

The instrument is factory-configured for IPv4 TCP/IP communication and includes a fast Ethernet adapter (10/100 Mbps) with a RJ45-type connector for integrating the device into a local area network (LAN). If the instrument will be connected to a LAN, an active, tested network jack must be in place before the scheduled installation date. Also, a representative from your information technologies department must complete and return the *SeqStudio™ Genetic Analyzer IT Checklist* (Pub. No. MAN0016055) before installation, and be available during the installation to help connect the instrument to your network.

Safety requirements

Safety practices

A safety representative from your facility must ensure that:

- Personnel establish and follow all applicable safety practices and policies to protect laboratory personnel from potential hazards.
- All applicable safety devices and equipment are available at all times.

Required safety equipment

Your laboratory has specific safety practices and policies designed to protect laboratory personnel from potential hazards that are present. Follow all applicable safety-related procedures at all times.

The following safety equipment and protection from hazards must be available at the installation site:

- Protection from any sources of hazardous chemicals, radiation (for example, lasers, radioisotopes, radioactive wastes, and contaminated equipment), and potentially infectious biological material that may be present in the area where the service representative will work.
- Appropriate fire extinguisher:
  - You are responsible for providing an appropriate fire extinguisher for use on or near the equipment.
  - The types and sizes of fire extinguishers shall be suitable for use on electrical and chemical fires as specified in current codes, regulations, and/or standards, and with approval of the Fire Marshall or other authority having jurisdiction.
  - The installation of appropriate fire extinguishers shall be in addition to other fire-protection systems and not as a substitute or alternative to them.
- Eyewash
- Safety shower
- Eye and hand protection
- Adequate ventilation, including vent line/fume hood, if applicable
- Biohazard waste container, if applicable
- First-aid equipment
- Spill cleanup equipment
- Applicable Safety Data Sheets (SDSs)

Safety

WARNING! GENERAL SAFETY. Using this product in a manner not specified in the user documentation may result in personal injury or damage to the instrument or device. Ensure that anyone using this product has received instructions in general safety practices for laboratories and the safety information provided in this document.

- Before using an instrument or device, read and understand the safety information provided in the user documentation provided by the manufacturer of the instrument or device.
- Before handling chemicals, read and understand all applicable Safety Data Sheets (SDSs) and use appropriate personal protective equipment (gloves, gowns, eye protection, and so on). To obtain SDSs, see the "Documentation and Support" section in this document.

Symbols on this instrument

Symbols may be found on the instrument to warn against potential hazards or convey important safety information. In this document, the hazard symbol is used along with one of the following user attention words:

- **CAUTION!**—Indicates a potentially hazardous situation that, if not avoided, may result in minor or moderate injury. It may also be used to alert against unsafe practices.
- **WARNING!**—Indicates a potentially hazardous situation that, if not avoided, could result in death or serious injury.
- **DANGER!**—Indicates an imminently hazardous situation that, if not avoided, will result in death or serious injury.

Symbol	English	Français
	Caution, risk of danger Consult the manual for further safety information.	Attention, risque de danger Consulter le manuel pour d'autres renseignements de sécurité.
	Caution, risk of electrical shock	Attention, risque de choc électrique
	Caution, piercing hazard	Attention, danger de perforation

Symbol	English	Français
	Caution, hot surface	Attention, surface chaude
	Potential biohazard	Danger biologique potentiel
	On	On (marche)
	Off	Off (arrêt)
	On/Off	On/Off (marche/arrêt)
	Protective conductor terminal (main ground)	Borne de conducteur de protection (mise à la terre principale)
	Terminal that can receive or supply alternating current or voltage	Borne pouvant recevoir ou envoyer une tension ou un courant de type alternatif
	Do not dispose of this product in unsorted municipal waste **CAUTION!** To minimize negative environmental impact from disposal of electronic waste, do not dispose of electronic waste in unsorted municipal waste. Follow local municipal waste ordinances for proper disposal provision and contact customer service for information about responsible disposal options.	Ne pas éliminer ce produit avec les déchets usuels non soumis au tri sélectif. **MISE EN GARDE !** Pour minimiser les conséquences négatives sur l'environnement à la suite de l'élimination de déchets électroniques, ne pas éliminer ce déchet électronique avec les déchets usuels non soumis au tri sélectif. Se conformer aux ordonnances locales sur les déchets municipaux pour les dispositions d'élimination et communiquer avec le service à la clientèle pour des renseignements sur les options d'élimination responsable.

Conformity symbols

Conformity mark	Description
	Indicates conformity with safety requirements for Canada and U.S.A.

Conformity mark	Description
CE	Indicates conformity with European Union requirements.
	Indicates conformity with Australian standards for electromagnetic compatibility.

Safety alerts on this instrument

Additional text may be used with one of the symbols described above when more specific information is needed to avoid exposure to a hazard. See the following table for safety alerts found on the instrument.

English	Français
CAUTION! Hazardous chemicals. Read the Safety Data Sheets (SDSs) before handling.	**MISE EN GARDE ! Produits chimiques dangereux.** Lire les fiches signalétiques (FS) avant de manipuler les produits.
CAUTION! Hazardous waste. Refer to SDS(s) and local regulations for handling and disposal.	**MISE EN GARDE ! Déchets dangereux.** Lire les fiches signalétiques (FS) et la réglementation locale associées à la manipulation et à l'élimination des déchets.
DANGER! Class 3B (III) visible and/or invisible laser radiation present when open and interlocks defeated. Avoid exposure to beam.	**DANGER !** Rayonnement laser ou DEL visible ou invisible de classe 3B (III) présent en position ouverte et avec les dispositifs de sécurité non enclenchés. Éviter toute exposition au faisceau.

Safety information for instruments not manufactured by Thermo Fisher Scientific

Some of the accessories provided as part of the instrument system are not designed or built by Thermo Fisher Scientific. Consult the manufacturer's documentation for the information needed for the safe use of these products.

Instrument safety

General

CAUTION! Do not remove instrument protective covers. If you remove the protective instrument panels or disable interlock devices, you may be exposed to serious hazards including, but not limited to, severe electrical shock, laser exposure, crushing, or chemical exposure.

Physical injury

CAUTION! Moving Parts. Moving parts can crush, pinch and cut. Keep hands clear of moving parts while operating the instrument. Disconnect power before servicing.

Electrical safety

WARNING! Ensure appropriate electrical supply. For safe operation of the instrument:

- Plug the system into a properly grounded receptacle with adequate current capacity.
- Ensure the electrical supply is of suitable voltage.
- Never operate the instrument with the ground disconnected. Grounding continuity is required for safe operation of the instrument.

WARNING! Power Supply Line Cords. Use properly configured and approved line cords for the power supply in your facility.

WARNING! Disconnecting Power. To fully disconnect power either detach or unplug the power cord, positioning the instrument such that the power cord is accessible.

Cleaning and decontamination

CAUTION! Cleaning and Decontamination. Use only the cleaning and decontamination methods specified in the manufacturer's user documentation. It is the responsibility of the operator (or other responsible person) to ensure the following requirements are met:

- No decontamination or cleaning agents are used that could cause a HAZARD as a result of a reaction with parts of the equipment or with material contained in the equipment.
- The instrument is properly decontaminated a) if hazardous material is spilled onto or into the equipment, and/or b) prior to having the instrument serviced at your facility or sending the instrument for repair, maintenance, trade-in, disposal, or termination of a loan (decontamination forms may be requested from customer service).
- Before using any cleaning or decontamination methods (except those recommended by the manufacturer), users should confirm with the manufacturer that the proposed method will not damage the equipment.

Instrument component and accessory disposal

To minimize negative environmental impact from disposal of electronic waste, do not dispose of electronic waste in unsorted municipal waste. Follow local municipal waste ordinances for proper disposal provision and contact customer service for information about responsible disposal options.

Laser

WARNING! LASER HAZARD. Under normal operating conditions, the SeqStudio™ Genetic Analyzer is categorized as a Class 1 laser product. However, removing the protective covers and defeating the interlock(s) may result in exposure to the internal Class 3B laser. Lasers can burn the retina, causing permanent blind spots. To ensure safe laser operation:

- Never look directly into the laser beam.
- Do not remove safety labels, instrument protective panels, or defeat safety interlocks.
- The system must be installed and maintained by a Thermo Fisher Scientific Technical Representative.
- Remove jewelry and other items that can reflect a laser beam into your eyes or those of others
- Wear proper eye protection and post a laser warning sign at the entrance to the laboratory if the laser protection is defeated for servicing
- DO NOT operate the laser when it cannot be cooled by its cooling fan; an overheated laser can cause severe burns on contact.

Safety and electromagnetic compatibility (EMC) standards

The instrument design and manufacture complies with the following standards and requirements for safety and electromagnetic compatibility.

Safety compliance

Reference	Description
EU Directive 2014/35/EU	European Union "Low Voltage directive"
EN 61010-1 UL 61010-1 CSA C22.2 No. 61010-1	*Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 1: General requirements*
EN 61010-2-010	*Safety requirements for electrical equipment for measurement, control and laboratory use – Part 2-010: Particular requirements for laboratory equipment for the heating of materials*
EN 61010-2-081	Safety requirements for electrical equipment for measurement, control, and laboratory use – Part 2-010: Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes
EN 60825-1	Safety of lasers products – Part 1: Equipment classification and requirements

EMC

Reference	Description
EU Directive 2014/30/EU	European Union "EMC Directive"
EN 61326-1	Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General requirements
AS/NZS CISPR 22	Limits and Methods of Measurement of Electromagnetic Disturbance Characteristics of Industrial, Scientific, and Medical (ISM) Radiofrequency Equipment
ICES-003, Issue 5	Industrial, Scientific and Medical (ISM) Radio Frequency Generators
FCC Part 15 Subpart B (47 CFR)	*U.S. Standard Radio Frequency Devices*

Reference	Description
RFID	FCC Notice (for U.S. Customers): This device complies with Part 15 of the FCC Rules: Operation is subject to the following conditions: 1. This device many not cause harmful interference, and 2. This device must accept any interference received, Including interference that may cause undesired operation Changes and Modifications not expressly approved by Thermo Fisher Scientific can void your authority to operate this equipment under Federal Communications Commissions rules.
	Canada (English): This device complies with Industry Canada licence-exempt RSS standard(s). Operation is subject to the following two conditions: (1) this device may not cause interference, and (2) this device must accept any interference, including interference that may cause undesired operation of the device. Canada (Français): Le présent appareil est conforme aux CNR d'Industrie Canada applicables aux appareils radio exempts de licence. L'exploitation est autorisée aux deux conditions suivantes : (1) l'appareil ne doit pas produire de brouillage, et (2) l'utilisateur de l'appareil doit accepter tout brouillage adioélectrique subi, même si le brouillage est susceptible d'en compromettre le fonctionnement.

Environmental design

Reference	Description
Directive 2012/19/EU	European Union "WEEE Directive"—Waste electrical and electronic equipment
Directive 2011/65/EU	European Union "RoHS Directive"—Restriction of hazardous substances in electrical and electronic equipment
SJ/T 11364-2014	"China RoHS" Standard—Marking for the Restricted Use of Hazardous Substances in Electronic and Electrical Products

Radio compliance

Reference	Description
Directive 2014/53/EU (as of June 12, 2017)	European Union "RE Directive"—Radio equipment

Chemical safety

WARNING! GENERAL CHEMICAL HANDLING. To minimize hazards, ensure laboratory personnel read and practice the general safety guidelines for chemical usage, storage, and waste provided below. Consult the relevant SDS for specific precautions and instructions:

- Read and understand the Safety Data Sheets (SDSs) provided by the chemical manufacturer before you store, handle, or work with any chemicals or hazardous materials. To obtain SDSs, see the "Documentation and Support" section in this document.
- Minimize contact with chemicals. Wear appropriate personal protective equipment when handling chemicals (for example, safety glasses, gloves, or protective clothing).
- Minimize the inhalation of chemicals. Do not leave chemical containers open. Use only with adequate ventilation (for example, fume hood).
- Check regularly for chemical leaks or spills. If a leak or spill occurs, follow the manufacturer's cleanup procedures as recommended in the SDS.
- Handle chemical wastes in a fume hood.
- Ensure use of primary and secondary waste containers. (A primary waste container holds the immediate waste. A secondary container contains spills or leaks from the primary container. Both containers must be compatible with the waste material and meet federal, state, and local requirements for container storage.)
- After emptying a waste container, seal it with the cap provided.
- Characterize (by analysis if necessary) the waste generated by the particular applications, reagents, and substrates used in your laboratory.
- Ensure that the waste is stored, transferred, transported, and disposed of according to all local, state/provincial, and/or national regulations.
- **IMPORTANT!** Radioactive or biohazardous materials may require special handling, and disposal limitations may apply.

Biological hazard safety

WARNING! BIOHAZARD. Biological samples such as tissues, body fluids, infectious agents, and blood of humans and other animals have the potential to transmit infectious diseases. Conduct all work in properly equipped facilities with the appropriate safety equipment (for example, physical containment devices). Safety equipment can also include items for personal protection, such as gloves, coats, gowns, shoe covers, boots, respirators, face shields, safety glasses, or goggles. Individuals should be trained according to applicable regulatory and company/ institution requirements before working with potentially biohazardous materials. Follow all applicable local, state/provincial, and/or national regulations. The following references provide general guidelines when handling biological samples in laboratory environment.

- U.S. Department of Health and Human Services, *Biosafety in Microbiological and Biomedical Laboratories (BMBL)*, 5th Edition, HHS Publication No. (CDC) 21-1112, Revised December 2009; found at:
  https://www.cdc.gov/labs/pdf/CDC-BiosafetymicrobiologicalBiomedicalLaboratories-2009-P.pdf
- World Health Organization, *Laboratory Biosafety Manual*, 3rd Edition, WHO/CDS/CSR/LYO/2004.11; found at:
  www.who.int/csr/resources/publications/biosafety/Biosafety7.pdf

Documentation and support

Related documentation

Document	Pub. No.
SeqStudio™ Genetic Analyzer Instrument and Software User Guide	MAN0018646
SeqStudio™ Genetic Analyzer Instrument and Software Getting Started Guide	MAN0018654
SeqStudio™ Genetic Analyzer for HID Instrument and Software v1.2 User Bulletin—New Features and Developmental Validation	100086084
SeqStudio™ Genetic Analyzer Site Preparation Guide	MAN0016143
SAE Administrator Console v2.0 User Guide	MAN0017468
DNA Fragment Analysis by Capillary Electrophoresis User Guide	4474504
DNA Sequencing by Capillary Electrophoresis Chemistry Guide Second Edition	4305080
Troubleshooting Sanger sequencing data	MAN0014435

Customer and technical support

Visit **thermofisher.com/support** for the latest service and support information.

- Worldwide contact telephone numbers
- Product support information
  - Product FAQs
  - Software, patches, and updates
  - Training for many applications and instruments
- Order and web support
- Product documentation
  - User guides, manuals, and protocols
  - Certificates of Analysis
  - Safety Data Sheets (SDSs; also known as MSDSs)

    Note: For SDSs for reagents and chemicals from other manufacturers, contact the manufacturer.

Limited product warranty

Life Technologies Corporation and/or its affiliate(s) warrant their products as set forth in the Life Technologies' General Terms and Conditions of Sale at **www.thermofisher.com/us/en/home/global/terms-and-conditions.html**. If you have any questions, please contact Life Technologies at **www.thermofisher.com/support**.

Index

3200 data, SeqStudio Genetic Analyzer 29, 121

A

AB1 file
- export automatically 80
- export manually 121, 154

amber light blinking 15
analysis software 29
analyze data, SeqStudio Genetic Analyzer (3200) 29, 121
audit reason, specify 61
auto calibration 193
auto export
- sample data files (AB1 and FSA) 80
- set Connect cloud-based platform location 133
- set network location 134

autospectral 193

B

base calling settings 129, 146
biohazard safety 235

C

calibration. *See* dye calibration
cancel injection 106
cartridge
- history 160
- instruments on which a cartridge has been installed 160
- load 41
- maximum days allowed on instrument 39
- remove 161
- reuse 18, 158
- storage 18, 158
- storage, on-instrument long-term 165

cartridge storage mode 159
cathode buffer
- assemble 166
- check fill line 41
- load 41
- maximum days allowed on instrument 39

CBC. *See* cathode buffer
Cloud. *See* Connect cloud-based platform
condition number 196
Connect cloud-based platform
- analysis software 29
- auto export default location 133
- browser requirements 25
- disconnect instrument 49
- open plate setup default location 132
- operating system requirements 25

connect instrument to the Connect cloud-based platform. *See* link
consumables status
- check 39
- updates 20

CSV file
- import 81
- save plate setup 123

custom dye 192, 197
- See also dye calibration, custom

custom field 76, 89

D

date and time 148
delete, plate 138, 154
Demo mode 148
documentation, related 169, 236
dye calibration
- auto calibration 193
- custom dye 192
- custom dye, add 197, 198
- custom, add to Plate Manager or another instrument 200
- custom, import or export 202
- custom, perform 199
- factory 193
- history 201
- manual 193
- manual calibration, determine if required 193
- overview 192
- perform 195
- system dye 192

E

E-signature records E-signature records
    dye calibration 64
    install run 64
EPT plot, remote monitor 101
export, plate 138, 154

F

file name convention 75, 88, 146
fill level 156
firmware version 147
fragment analysis, run module, size ranges and run times 140, 218
front panel 15
FSA file
    export automatically 80
    export manually 121, 154

H

HID analysis, results 113
HID kit. *See* kit

I

import plate setup, CSV or PSM file 81, 139
Index Term 231
injection
    cancel 106
    options 91, 105, 120
    order, change 91, 105, 120
    order, default 38
    re-inject 91, 105, 120
injection group, definition 38
install check. *See* install run
install run
    history 210
    log file, export 170
    overview 203
    perform 204
    results 205
    sequence and fragment in same run 203
instrument
    connect to SAE server 58
    disconnect from Connect cloud-based platform 49
    lock 95, 137
    log file 170
    remove user 49
    view results 111
instrument conditions during run 117
instrument name 148
instrument profile
    delete 152
    PIN, change 52, 153
Instrument-cartridge history, view or export 160
IP address 130, 147

K

kit (HID analysis only), marker-to-marker pull-up reduction 67, 79
kit, HID, import 217

L

limited product warranty 237
load
    cartridge 161
    CBC, sample plate, and cartridge order 41
    plate 92
    tubes 92
lock instrument 95, 137
log file
    export 170
    install run 170

M

manual dye calibration 193
marker-to-marker pull-up reduction (HID analysis only), set for injection group 67, 79
mobile device
    monitor a run 103
    view results 103
My plates. *See* hidden

N

network
    auto export location 134
    connect instrument to 149
    connect to 132
    open plate setup default location 133
network configuration and security 27

P

password security 27
pause plate 106
PCR amplicons from diploid genes, sequencing settings 129, 146
PDF, plate setup 123
plasmid templates, sequencing settings 129, 146
plate
    load 41

loading guidelines 38
pause 106
Plate Manager
add custom dye calibration 122
operating system requirements 25
plate setup
create on instrument 81
CSV file, import 81
delete 124
guest 70, 79
hidden (my plates) 76, 89
hidden (my plates), definition 79
hidden (my plates), set in Plate Manager 70
import from CSV or PSM file 139
import on instrument 81
open 81
PSM file, import 81
security 76, 89
security, definition 79
security, set in Plate Manager 70
set default open location on Connect cloud-based platform 132
set default open location on network 133
shared (public) 76, 89
shared (public), definition 79
shared (public), set in Plate Manager 70
template 122, 123
plate type supported 37
polymer delivery system (PDS), refresh 161
private plate setup. *See* hidden
probablility of error 117
profile. *See* instrument profile
PSM file
See also plate setup
create or open in Plate Manager 69
import 81
See also plate setup
public plate setup 76, 89
See also shared
pull-up reduction (HID analysis only) 67, 79

Q

Q and Condition 197
QC report, export 120
quality flag 113
quality value
See also QV
probability of error 117
spectral calibration 196
See also QV
QV, probability of error 117
QV ranges 116
QV30 CRL 208

R

raw data 171
raw trace, remote monitor 100
re-inject, remote monitor 100
re-injection 91, 105, 120
related documentation 169, 236
remote monitor
analyzed trace 102
EPT plot 101
pause or cancel run 99
raw trace 100
re-inject 100
view results 99, 110
remove plate associated with inactive user 155
replicate injection 74, 87
results
export automatically 80
export manually 121, 154
view, instrument 111
view, mobile device 103
view, remote monitor 99, 110
results report, export 120
RFID read and write 20
run, pause or cancel 106
run history, delete 155
run module
read length and run time 140, 218
settings 140
size range and run time 140, 218

S

SAE
enable on instrument 59
overview 56
server, specify 59
safety, biohazard 235
safety practices 225
safety, required equipment 226
sample
plates and tubes supported 37
purification 37
stability on instrument 38
volume requirements 37
sample plate. *See* plate
secondary analysis software 29
sequence quality 116
sequencing
results 115
run module, read lengths and run times 140, 218
sequencing settings
description 129, 146
PCR amplicons from diploid genes 129, 146

plasmid templates 129, 146
serial number 147
set in injection group 67, 79
sign
dye calibration 64
install run 64
sign in 34
sign out 35
size calling settings 126, 143
Size Quality 113
size standard quality 172
sleep mode 159
smart phone. *See* mobile device
spectral calibration
condition number 196
quality value 196
SQ 113
support, customer and technical 236
symbols, safety 227
system dye 192

T

template, plate setup 122, 123
terms and conditions 237
Thermo Fisher Cloud. *See* Connect cloud-based platform
troubleshooting, log file, export 170
tube type supported 37

U

update software 150

W

warranty 237
wired or wireless instrument connection 149