

## 2,20 mm NONGUARDED SLIT KNIFE TECHNICAL SPECIFICATION

1. It is disposable, cutting edge is 2.20 mm. It has a 45 degree angle and a full handle.
2. Due to the quality of the incision, it is double-bevel.
3. It is suitable for making a lateral incision in the cornea.
4. The blade does not shine under the microscope.
5. The brand and model is written on the handle of the knife. (label is tyvek, not sticking label)
6. It has a transparent protection guard made of polycarbonate that protects the tip of the blade (to see the model of the blade). The handle of the knife is made of sterilization-resistant polycarbonate.
7. The blade is in a hard blister container to withstand impacts that may occur until the end of use.)
8. It has a special coating system that protects the surface of the blade. This feature preserves the sharpness of the blade and prevents deformation on its surface.
9. The blister container is sealed with Medical Grade Tyvek to complete sterilization..
- 10- Product information should not be sticking labels, must be printed by the original manufacturer.
- 11-The blades are produced in different colors to avoid mixing with different models (Slit, Stab, Mvr, Crescent) for ease of use during surgery.
- 12- Products are FDA certified.