

EC CERTIFICATE OF CONFORMITY

NO. N8 04 03 45542 715



Product Service

HOLDER OF
CERTIFICATE:Zhanjiang Star Enterprise
Co., Ltd.No. 49 Jinchuan Road
524094 Zhanjiang
PEOPLE'S REPUBLIC OF CHINAZhanjiang Star Enterprise Co., Ltd.
No. 49 Jinchuan Road, 524094 Zhanjiang, PEOPLE'S REPUBLIC
OF CHINA

FACILITY(IES):

Zhanjiang Star Enterprise Co., Ltd.
No. 49 Jinchuan Road, 524094 Zhanjiang, PEOPLE'S REPUBLIC
OF CHINA

CERTIFICATION MARK:



SCOPE OF CERTIFICATE:

Manufacturer of transducers for
ultrasound device.

TESTED ACCORDIN TO:

EN 60334-3/B17:2017

This EC-Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 73/23/EEC relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

TEST REPORT:

CPSA87540912

VALID UNTIL:

2020-10-30

DATE,

2017-11-01



PAGE 1 OF 1



After preparation of the necessary technical documentation as well as the conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.

CERTIFICATE

NO. Q1N 04 03 45342 060



Product Service

**HOLDER OF
CERTIFICATE:**

**Zhanjiang Star Enterprise
Co., Ltd.**

No. 49 Jinchuan Road
524094 Zhanjiang
PEOPLE'S REPUBLIC OF CHINA

Zhanjiang Star Enterprise Co., Ltd.
No. 49 Jinchuan Road, 524094 Zhanjiang, PEOPLE'S REPUBLIC
OF CHINA

FACILITY(IES):

Zhanjiang Star Enterprise Co., Ltd.
No. 49 Jinchuan Road, 524094 Zhanjiang, PEOPLE'S REPUBLIC
OF CHINA

CERTIFICATION MARK:



SCOPE OF CERTIFICATE: Manufacturer of transducers for
ultrasound device.

**APPLIED
STANDART(S):**

INCLUDING THE IMPLEMENTATION MEETS REQUIREMENTS OF THE
STANDARD: ISO 9001:2015

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned
above has established and is maintaining a quality system which meets the requirements of the
listed standard(s). See also notes overleaf.

REPORT NO: 95853426098

VALID UNTIL: 2020-09-30


DATE, 2017-10-01

Reiner Krumme



PAGE 1 OF 1


Датчик C362



- Характеристики

- Датчик SonoScape C362 - это конвексный датчик высокой плотности для обследований взрослых, который имеет минимальную частоту в 3,5 MHz, максимальную в 7 MHz. Применяется в различных направлениях. Совместимость: SonoScape SSI-8000, S8, S8Exp, S9, S9Pro, S20, S22, S30, S40.

Датчик L752



- Характеристики

- Центральная частота: 8,0 МГц;
- Частота отображения: 4,0-16,0 МГц;
- FOV: 52 мм;
- Элементы: 256;
- Радиус: не доступно;
- Глубина изображения: 2-100 мм;
- Руководство по биопсии: доступный;
- Footprint: 55x10 мм;
- Совместимость: SonoScape SSI-8000, S8, S8Exp, S9, S9Pro, S20, S22, S30, S40.