

**Plaster of Paris Bandage Analysis
Report**

DATE

20/05/2021

PRODUCT NAME: Plaster of Paris Bandage

SIZE : 20CM X 2,70M

LOT NUMBER :140521106

TESTS PERFORMED	TS 4434	CONCLUSION	CONDITIONS OF CONFORMITY
OF THE WRAP SHEET MASS PER UNIT AREA(g/m ²)	MIN 26 g	30,62 gr	A
THE WRAP AND WEFT WIRE WINDING NUMBER OF CLOTH	Cm2 WIRE MUST BE AT LEAST 18 PER	18 wire	A
FEATURE OF THE WRAP SHEET	100% COTTON AND SHOULD BE HYDROPHILE	100% COTTON AND HYDROPHIL	A
PLASTERY MASS PER UNIT AREA OF THE WRAP SHEET	AT LEAST 340 g	440 gr	A
WETTABILITY	WHEN THROWN IN TO THE WATER SHOULD NOT DRY SURFACE	FULL WETTING OCCURRED AFTER 6 SECONDS 30°C WATER	A
SETTING TIME	1-8 MINUTES	2,45 MINUTES	A

CONDITIONS OF CONFORMITY : A:Available NA: None/Available

CONCLUSION: Test results conform to the relevant standard

QUALITY CONTROL OFFICER

GÖZDE DURGİT

MEDİTEKS SAĞLIK HİZMETLERİ
TIBBİ MALZEME TEKSTİL ÜRÜNLERİ
SAN. DİŞ. TİC. A.Ş. EDİRNE SÜLOĞLU ŞUBESİ
Organize San. Bölgesi 4.Cad. No:5 Süloğlu-Edirne
Davutpaşa V. d. No: 13 012 4314
Mersis No: 0612 0024 314 00016