

TERMENII ȘI CONDIȚIILE DE LIVRARE

Denumirea **“Consolidarea controlului Tuberculozei și reducerea mortalității cauzate de SIDA în Republica Moldova”**
Cumpărător: **Instituția Publică „Unitatea de Coordonare, Implementare și Monitorizare a Proiectelor în Domeniul Sănătății”**
Destinatar: **IMSP IFP „Ch. Draganiuc”, Chișinău.**
Pachet nr.: **293/GD/T**
Furnizor: **MedPunkt SRL**

1. Prețuri și Termeni de livrare:

Furnizorul recunoaște că va fi responsabil de:

- a) Livrarea mărfurilor la adresa destinatarului, IMSP IFP „Chiril Draganiuc”.
- b) Instalarea echipamentului.
- c) Instruirea personalului medical pentru întreținerea de rutină a echipamentului.

LOT: “ Congelator pentru laborator (Ultra Low-freezer), vertical, 680-750 L, - 50-86 °C”:

Nr	Articol	<u>Cantitate totală, unit</u>	Preț per unitate Chișinău, <u>MDL</u>	Preț total Chișinău, <u>MDL</u>	<u>Termenul de livrare</u>	<u>Destinatar</u>
1.	Congelator pentru laborator (Ultra Low-freezer), vertical, 680-750 L, - 50-86 °C	2	266611	533222	90 zile calendaristice	IFP, Chișinău
	<i>TOTAL:</i>			533222		

Notă: În caz de discrepanță între prețul unitar și totalul calculat din prețul unitar, corectarea va fi efectuată în conformitate cu prevederile punctului 10 din Invitația pentru prezentarea ofertei de preț.

- 2. Preț Fix: Prețurile menționate mai sus sunt fixe și nu vor suferi ajustări pe durata îndeplinirii contractului. Se va confirma că aceste prețuri nu includ taxele vamale, accizele, taxele pentru procedurile vamale și taxa pe valoarea adăugată (TVA) din Republica Moldova prin semnarea Termenilor și Condițiilor de Livrare.
- 3. Țara de origine: Produsele oferite vor proveni din țările membre ale Băncii Mondiale. Ca dovadă, Cumpărătorul ar putea solicita furnizarea unui certificat de origine pentru fiecare articol.
- 4. Programul de livrare: Livrarea va fi efectuată conform Termenilor de livrare stipulați în Termenii și Condițiile de Livrare.
- 5. Legislația aplicabilă: Contractul va fi interpretat în conformitate cu legile țării Cumpărătorului.

6. Soluționarea litigiilor: Cumpărătorul și Furnizorul vor depune toate eforturile pentru a rezolva orice neînțelegere sau litigiu între ei, în temeiul sau în legătură cu Contractul, pe cale amiabilă, prin negocieri directe informale. În cazul unui litigiu între Cumpărător și Furnizor, litigiul va fi soluționat în conformitate cu prevederile legilor din țara Cumpărătorului.
7. Livrarea și Documentele: La livrarea bunurilor, Furnizorul va notifica Cumpărătorul prin email sau fax cu privire la detaliile complete ale livrării, inclusiv numărul contractului, descrierea bunurilor, cantitatea etc.
8. Plata facturii se va efectua în totalitate (100%), după livrarea bunurilor, prin transfer bancar către contul Băncii Furnizorului, în decurs de zece (10) zile bancare de la data livrării bunurilor și a serviciilor conexe conform cerințelor din punctul 1 al acestor Termeni și Condiții de Livrare. Aceasta va fi confirmat prin semnarea unui Act de Recepție finală pentru bunuri și servicii conexe, eliberat de Cumpărător și avizat de Destinatar/Beneficiar final (denumit în continuare „Actul de Recepție Finală”).
9. Garanție: Bunurile oferite trebuie să fie acoperite de garanția producătorului (în ceea ce privește perioada de valabilitate), așa cum este specificat în Anexa A a Termenilor și Condițiilor de Livrare, începând de la data Actului de Recepție Finală.
10. Autorizația producătorului. Cumpărătorul poate solicita Furnizorului înainte de atribuirea contractului, să furnizeze autorizația de la producător pentru bunurile incluse în contract.
11. Instrucțiuni privind ambalarea și marcarea: Furnizorul va asigura ambalarea standard a bunurilor pentru a preveni deteriorarea acestora în timpul transportării către destinația finală, conform specificațiilor din Contract.
12. Defecte: Furnizorul va corecta toate defectele identificate fără costuri suplimentare pentru Cumpărător în termen de 30 de zile de la data notificării din partea Cumpărătorului.
13. Forță majoră: Furnizorului nu i se vor fi aplica penalități sau reziliere a contractului în cazul în care întârzierea sau neîndeplinirea obligațiilor sale conform Contractului rezultă dintr-un eveniment de forță majoră.

În sensul acestei clauze, „Forță Majoră” se referă la un eveniment care nu poate fi controlat de Furnizor, nu implică vina sau neglijența Furnizorului, și nu este previzibil. Astfel de evenimente pot include, dar fără a se limita la, acțiuni suverane ale Cumpărătorului, războaie sau revoluții, incendii, inundații, epidemii, restricții de carantină și embargouri de mărfuri.

În cazul apariției unei situații de forță majoră, Furnizorul va notifica imediat Cumpărătorul în scris despre această situație și cauza acesteia. Cu excepția cazului în care Cumpărătorul solicită altfel în scris, Furnizorul va continua să își îndeplinească obligațiile în temeiul Contractului, în măsura în care este rezonabil din punct de vedere practic, și va căuta toate mijloacele alternative rezonabile de executare care nu sunt împiedicate de evenimentul de forță majoră.

14. Specificațiile tehnice necesare
 - a. Toate bunurile trebuie să fie noi, nefolosite, să fie cele mai recente și actuale modele, incorporând toate îmbunătățirile recente ale designului și materialelor, cu excepția cazului în care se prevede altfel în aceste specificații.
 - b. Detalii specifice și standarde tehnice - conform Anexei A la Termenii și condițiile de Livrare
 - c. Furnizorul va confirma că bunurile corespund specificațiilor de mai sus. În cazul în care există abateri, Furnizorul va enumera toate aceste abateri.

15. Neexecutarea: În cazul în care Furnizorul nu respectă termenii și condițiile menționate mai sus, Cumpărătorul poate anula Contractul, după o notificare prealabilă de 10 zile, fără a-și asuma nicio răspundere față de Furnizor.
16. Întârzieri: În cazul în care Furnizorul nu reușește să livreze oricare sau toate bunurile până la data de livrare sau să efectueze Serviciile conexe în perioada specificată în Contract (așa cum este prevăzut în Termenii de livrare de mai sus), Cumpărătorul poate, fără a afecta celelalte clauze ale Contractului, să deducă din Prețul Contractului, drept despăgubiri forfetare, o sumă echivalentă cu 0,1% din prețul oferit al bunurilor întârziate sau al serviciilor neexecutate pentru fiecare zi lucrătoare sau parte din acestea până la o deducere maximă de zece (10 %) din Prețul Contractului.
17. Frauda și corupția: Politica Fondului Global impune ca toți ofertanții, furnizorii, contractanții și agenții acestora (declarați sau nu), precum și personalul, subcontractanții, sub-consultanții, furnizorii de servicii în cadrul contractelor finanțate de Fondul Global să respecte cele mai înalte standarde etice în timpul achizițiilor și execuției acestor contracte.¹ În acest context, Fondul Global a elaborat un **Cod de Conduită pentru Furnizori** (<https://www.theglobalfund.org/en/governance-policies/>), care are ca scop asigurarea că Furnizorii și Reprezentanții Furnizorilor participă la procesul de achiziții într-o manieră transparentă, echitabilă, responsabilă și onestă. Acest lucru include respectarea tuturor legilor și reglementărilor aplicabile privind concurența loială, precum și a standardelor recunoscute de bune practici în achiziții.
- Pentru confirmarea asumării acestor responsabilități, fiecare ofertant va semna adițional Codul de Conduită pentru Furnizori (Code of Conduct for Suppliers/Codul de Conduită pentru Furnizori), Anexa B.
18. În calitate de ofertant, confirmăm prin semnătura noastră de mai jos, că ne asumăm responsabilitatea pentru informațiile prezentate în cadrul acestei achiziții.

NUMELE FURNIZORULUI: MedPunkt SRL

Semnătura și ștampila autorizate _____

Numele și funcția semnatarului: Administrator Antonova Anna

Locul: mun. Chisinau

Data: 08.05.2024

¹ În acest context, orice acțiune întreprinsă de un ofertant, furnizor, contractant sau de orice al lor personal, agenți, subcontractanți, subconsulanți, furnizori de servicii, furnizori și/sau angajații acestora în vederea influențării asupra procesului de achiziție sau a executării contractului pentru un avantaj necuvenit este improprie.

Detalii specifice și standarde tehnice

LOT: „Congelator pentru laborator (Ultra Low-freezer), vertical, 680-750 L, - 50-86 °C”:

	Specificatii tehnice solicitate		Specificatii tehnice oferite
	Model		BDF-86V728
	Producător		Jinan Biobase Biotech Co., Ltd.
	Cod produs		
I.	Description of function and use		Description of function and use
	Main function	<u>Congelator pentru laborator (Ultra Low-freezer), vertical, 680-750 L, - 50-86 °C// Vertical Deep Freezer - 50-86 °C, 680-750L, with 3 racks</u> suitable for use in scientific research, cryogenic test on special materials, blood plasma cryopreservation, low temperature resistance test on biological materials, biological products, etc. Suitable for use in research institutions, hospitals, the health & disease prevention system, laboratories, etc.	<u>Congelator pentru laborator (Ultra Low-freezer), vertical, 728 L, - 40-86 °C// Vertical Deep Freezer - 40-86 °C, 728L, with 3 racks</u> suitable for use in scientific research, cryogenic test on special materials, blood plasma cryopreservation, low temperature resistance test on biological materials, biological products, etc. Suitable for use in research institutions, hospitals, the health & disease prevention system, laboratories, etc.
II.	Main specifications		Main specifications
	Temperature range:	- -50 °C - 86 °C.	- -40 °C - 86 °C.
	Configuration:	- Vertical.	- Vertical.
	Capacity:	- 680 -750 L.	- 728 L
	Externa Door:	- 1 external door. - Lockable door, with key lock.	- 1 external door. - Lockable door
	Shelves:	- 3 pieces, closed with a freezing flap	- 3 pieces, closed with a freezing flap. The flaps are inner additional doors. The inner doors are removable for cleaning or defrosting. See page 10 p.15 of the attached User Manual .

	Wheels:	- Equipped with 4 wheels, to be easily moved. - Adjustable feet for levelling.	- Equipped with 4 wheels, to be easily moved. - Adjustable feet for levelling.
	Interior Material:	- Stainless steels.	- Stainless steel SS 304
	External Material:	- Cold-rolled steel with anti-corrosive coating.	- Cold-rolled steel with anti-corrosive coating.
	Temperature display:	- Digital.	- Digital.
	Screen:	- LED. - Graphic display.	- LCD touch screen
	Alarm System:	- Acoustic. - Visual - High and low temperature. - Door open. - Power failure. - Voltage abnormal. - Ambient temperature abnormal.	- Acoustic. - Visual - High and low temperature. - Door open. - Power failure. - Voltage abnormal. - Ambient temperature abnormal.
	Alarm setting options:	- Maximum temperature limit. - Minimum temperature limit.	- Maximum temperature limit. - Minimum temperature limit.
	Internal battery for alarm and alarm data retention:	- Yes, minimum 48 hours.	- Yes, 72 hours.
	Cooling system:	- Ventilated or static.	- Ventilated or static.
	Refrigerant:	- Without CFC / HCFC.	- Refrigerant: High: R290; Low: R170
	Noise level:	- ≤ 60 db.	- 55 dB See page 26 of the attached User Manual .
III.	Electricity requirements		Electricity requirements
	Supply voltage:	- 230 ± 10 V - Voltage and plugs type adapted to meet the requirement of electrical system in Moldova.	- 220V - Voltage and plugs type adapted to meet the requirement of electrical system in Moldova.
	Frequency:	- 50/60 Hz.	- 50 Hz

IV.	Documentation		Documentation
	Manufacturer's certificate	- ISO 13485:2016 - Quality Management Systems for Medical Devices – mandatory. - CE Marking - Indicating compliance with European Union regulations and directives related to medical devices.	- ISO 13485:2016 - Quality Management Systems for Medical Devices – mandatory. - CE Marking - Indicating compliance with European Union regulations and directives related to medical devices.
	Quality and safety standards	EU Electrical Safety Standards Compliance: - EN 60601-1: Medical Electrical Equipment - General Requirements for Basic Safety and Essential Performance - IEC 61010-1: Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - EN 55014 - Electromagnetic compatibility. - Protection class (in accordance with EN 60529).	EU Electrical Safety Standards Compliance: - EN 60601-1: Medical Electrical Equipment - General Requirements for Basic Safety and Essential Performance - IEC 61010-1: Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - EN 55014 - Electromagnetic compatibility. Protection class (in accordance with EN 60529).
V.	Accessories		Accessories
	Freezer racks		Freezer racks
	Quantity	- 12	- 12 pcs
	External dimensions	- To be compatible with the internal size of the freezer	- Are compatible with the internal size of the freezer
	Internal dimensions	- To be compatible with freezers box 133*133*52 mm	- Model LD5X4 - 5 layers and 4 columns - Made of 304 stainless steel - Size: 560*140*295mm - Thickness 1.2mm - Are compatible with freezers box 133*133*52 mm Check please the link: https://www.biobase.com/product/freezer-racks-and-freezer-boxes
	Material	- Made of corrosion-resistant stainless steel, it can be frozen and thawed repeatedly and stored for a long time.	- Made of corrosion-resistant 304 stainless steel, it can be frozen and thawed repeatedly and stored for a long time. - The product can withstand temperature range: -196°C to 121C

VI.	Operation, maintenance and installation		Operation, maintenance and installation
	Operation and maintenance manual:	- At least one set of operation, maintenance and service manuals for each equipment. The manuals to include instructions for: - o setting up the equipment - o routine cleaning and maintenance - o planning for periodic maintenance.	- One set of operation, maintenance and service manuals for each equipment. The manuals to include instructions for: - o setting up the equipment - o routine cleaning and maintenance - o planning for periodic maintenance.
	Language of documentation:	- Written in United Nations languages: - o at least in English and - o mandatory in Romanian.	- Written in United Nations languages: - o in English and - mandatory in Romanian.
	Installation and maintenance:	- The bidder must arrange for the equipment to be installed by certified or qualified personnel. Any prerequisites for installation will be communicated to the beneficiary in advance, described in detail in written form. - The bidder will provide a comprehensive maintenance plan. The cost of the maintenance plan to be defined and guaranteed over the period of warranty. - The supplier will provide an after-sale service. - The service to have competent staff, adequate infrastructure and sufficient spare parts to be able to respond to any complaints and to repair or replace within 14 days.	- The bidder will arrange for the equipment to be installed by certified or qualified personnel. Any prerequisites for installation will be communicated to the beneficiary in advance, described in detail in written form. - The bidder will provide a comprehensive maintenance plan. The cost of the maintenance plan to be defined and guaranteed over the period of warranty. - The supplier will provide an after-sale service. - The service to have competent staff, adequate infrastructure and sufficient spare parts to be able to respond to any complaints and to repair or replace within 14 days.
	Training:	- The bidder will provide user training (including how to use and maintain the equipment).	- The bidder will provide user training (including how to use and maintain the equipment).
	Standard maintenance tools:	- All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication material, to be included in the offer. - o Bidders to specify the quantity of every item included in their offer (including items not specified above).	- All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication material, to be included in the offer. - Bidders to specify the quantity of every item included in their offer (including items not specified above).
	Spare parts:	- Each assembled equipment to be accompanied by an authorized list of spare parts.	- Each assembled equipment to be accompanied by an authorized list of spare parts.

	Warranty:	- At least 3 years .	- Yes, extended warranty 3 years
	Remarks above:	- The equipment offered, including its power supply, to be designed and constructed to operate properly and continuously in the conditions of the Republic of Moldova. - The equipment may need to tolerate high humidity (as high as 90% at 35 °C), ambient temperatures of 5–40 °C, fungi, and spikes in the electricity supply.	- The equipment offered, including its power supply, to be designed and constructed to operate properly and continuously in the conditions of the Republic of Moldova. - The equipment may need to tolerate high humidity (as high as 90% at 35 °C), ambient temperatures of 5–40 °C, fungi, and spikes in the electricity supply.

Toate bunurile vor fi testate de beneficiar. Acceptarea bunurilor și semnarea actelor de recepție va avea loc după testarea pozitivă efectuată de beneficiar.

În calitate de ofertant, confirmăm prin semnătura noastră de mai jos, că ne asumăm responsabilitatea pentru informațiile prezentate în cadrul acestei achiziții.

NUMELE FURNIZORULUI : MedPunkt SRL

Semnătura și ștampila autorizate _____

Locul: mun.Chisinau

Data: 08.05.2024