



Standard Sterile Surgical Set Specifications











OPS ADVANCED	STANDARD STERILE SURGICAL SET
Reference(s)	Z10-001-01
Description	 Field For Instrument Table With Reinforced Zone From SMS 150x190 -1 pc. Field For Mayo Table With SMS Reinforcement 80x40 - 1 pcs. Napkin For Absorption - Not to leave lint - 40x40 2 pcs. Lower Part Field 190x195 - 1 pcs. Positioning tape - on the entire surface of the incision area, impregnated in the fabric, 5cm wide Absorbent area - on the entire incision surface; liquid penetration resistance of min. 150cm of H20/cm2 SMS material on the entire surface of the field with liquid resistance 50cm H20/cm2 Upper Part Field 150x270 - 1 pcs. Positioning tape - on the entire surface of the incision area, impregnated in the fabric, width of 5cm Absorbent area - on the entire incision surface; resistance to liquid penetration of min. 150cm of H20/cm2
	SMMS material on the entire surface of the field with liquid resistance min. 50cm H20/cm2 . OP-Tape 10x50 - 1 pcs.
Main Materials	SMS (40 GR, %100 PP), BIFLEX SPUNBOND (56 GR, SB+PE FILM)
Packaging	Sterilization Pack 42x60
Manufacturer	ZENMED TIBBİ CİHAZLAR VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD. ŞTİ. ISO 13485:2016
Regulation	In compliance with Medical Devices Directive 93/42/EEC HP: High performance according to EN 13795 standard MDD class: I sterile

Latex Free





