



181119340120



China
检测
Accreditation
TESTING CNAS L2865

Test report

Report No. DNW-20200800046

Product Name Gas Roots Flow Meter

Client: Zhejiang Cangnan Instrument Group Co., Ltd.

Inspected Unit: Zhejiang Cangnan Instrument Group Co., Ltd.

Inspection Category Commissioned Inspection

Zhejiang Institute of Metrology

August 28, 2020



Zhejiang
浙江省

protection

Zhejiang Institute of Metrology

Test report

Sample name	Gas roots flow meter	Detection category	Conventional inspection
Model specifications	RM-80-G100	trademark	east star
Production date or batch number	2020. 8	Factory number	/
entrusting unit	Zhejiang Cangnan Instrument Group Co., Ltd.	Inspected unit	Zhejiang Cangnan Instrument Group Co., Ltd.
Entrusting unit address	Lingai Town Industrial Demonstration Park, Cangnan County, Zhejiang Province	production unit	Zhejiang Cangnan Instrument Group Co., Ltd.
Sample accuracy grade/measurement uncertainty/maximum permissible error (when applicable)	Level 1.0		
Sample delivery unit	Zhejiang Cangnan Instrument Group Co., Ltd.	Sample grade	/
Sample quantity	2 units	Sample arrival date	2020. 8. 24
sampler	/	Sampling date	/
Sampling quantity	/	Sampling location	/
Sampling base	/	Testing location	Environmental and Reliability Laboratory of the Institute of Electric Power
Test basis	GB/T 4208-2017	Test date	2020. 8. 25
Test items	Housing protection level		
Sample status description	Appearance normal, suitable for testing		
Test	<p>According to GB/T 4208-2017 "Enclosure Protection Grade (IP Code)", only the test results are provided for the tested items.</p> <p>Request for all 送检项目进行检 Province (Special seal for testing)</p> <p>Issue date: August 28, 2020</p> <p>3301180049820</p>		
Remark	The sample source information is provided to customers, and this organization is not responsible for its authenticity.		

approve:

周红国

Review:

Zhang Honghe

prepared by:

Hiroshi

Position:

Professor level senior engineer

Test items	Standard or technical requirements	Sample number	Test results
1 Enclosure protection grade test	IP67	IP6X	93020254429-01 conform to
		IPX7	93020254429-02 conform to

Blank below

science
dedicated
0049820

statement

1. The report is invalid without the "special test stamp".
2. This organization is only responsible for the complete report stamped with the "special testing seal".
3. Partial copying of this report will be invalid without the written approval of this institution.
4. Any report that is altered will be invalid.
5. If you have any objection to this report, please raise it in time.

6. The test results are only valid for the tested samples.

Address: No. 300 Xiasha Road, Jianggan District, Hangzhou City, Zhejiang Province

Consultation hotline: 0571-85027145

IN MY AUNT I'M WINTERING

Postal Code:310018

www.zjim.cn

ZJIM

