



EuroProxima


Streptomycin

CERTIFICATE OF ANALYSIS

Doc. No:	5111STREP-CA V11
Batch No:	250504
Code:	5111STREP
Exp. Date:	03-2027

This certificate relates to the above specified product batch and number only. The product has been tested according to our standard quality control program. The product is guaranteed when stored according to the instruction in the manual till the indicated expiry date.

Components:	Lot No:	Exp. Date:
Microtiter plate	250300	04-2027
Conjugate	250438	06-2027
Standards 20 - 10 - 2 - 1 - 0,5 - 0,25 - 0 ng/ml	250439	06-2027
Standard 100 ng/ml	250440	06-2027
Antibody	250441	06-2027
Dilution buffer	250196	03-2027
Rinsing buffer	250333	04-2027
TMB substrate	020373	03-2028
Stop solution	250185	03-2027
Protocol	Code 5111STREP [18]02.24	



Validation criteria

Bmax > 0.8 OD.

Blank < 0.2 OD.

Concentration (ng/ml)	Accepted Range %Bmax	Mean %Bmax
STD 10	12 - 22	16
STD 2	25 - 47	38
STD 1	46 - 60	51
STD 0,5	60 - 72	67

Performed by:

BB

date: 10-06-2025

Accepted by:

QA/QC Manager

ISO 9001
BUREAU VERITAS
Certification



This product is produced under the control of a management system approved by Bureau Veritas certification in accordance with the standard ISO 9001: 2015 Bureau Veritas Certification N°3089368

r-biopharm®

