



# CERTIFICATE OF REGISTRATION

## Lorne Laboratories Ltd

Unit 1 Cutbush Park Industrial Estate  
Danehill  
Lower Earley  
Berkshire RG6 4UT UNITED KINGDOM

UL LLC®(UL) issues this certificate to the Firm named above, after assessing the Firm's quality system and finding it in compliance with:

**ISO 13485:2016**

**EN ISO 13485:2016**

The manufacture of in vitro diagnostic blood grouping reagents. The purchase for resale of in vitro diagnostic serology test kit.

Authorized by



**Michael J. Windler, P.E.**

**Manager of Global Regulatory Service**  
Distinguished Member of the Technical Staff  
Life and Health Sciences, UL LLC




Check Certificate  
Status: [here](#)

File Number	A12241	Cycle Start	May 23, 2020
Certificate Number	1458.200523	Effective Date	May 23, 2020
Initial Issue Date	June 26, 2018	Expiry Date	May 22, 2023

This quality system registration is included in UL's Directory of Registered Firms and applies to the provision of goods and/or services as specified in the scope of registration from the address(es) shown above. By issuance of this certificate the firm represents that it will maintain its registration in accordance with the applicable requirements. This certificate is not transferable and remains the property of UL LLC.



UL LLC  
333 Pfingsten Road  
Northbrook, IL 60062-2096 USA



# LORNE LABORATORIES LTD



Unit 1 Cutbush Park Industrial Estate  
Danehill, Lower Earley, Berkshire, RG6 4UT  
United Kingdom

Phone: +44 (0) 118 921 2264  
Fax: +44 (0) 118 986 4518  
Email: info@lornelabs.com

## CERTIFICATE OF ANALYSIS

DESCRIPTION	LOT NO.	EXPIRY	PRODUCT CODE
RPR Carbon Kit	LO16261	2023-03	044150A

**STORAGE:** Refrigerated at 2 – 8°C. Protect from light. Do not freeze.

**SHIPPING:** This product has data supporting stability tolerance during fluctuations in ambient shipping temperature.

This product is in compliance with **MEDICAL DEVICES REGULATIONS 2002** for Annex 3 and Self-Certification. This product was manufactured, packaged and tested in accordance with **LORNE QUALITY SYSTEMS** and meets all product specifications.

REAGENT	SPECIFICATIONS			RESULT
	Appearance	Colour	Functionality	
RPR Carbon Kit	Liquid suspension free of macroscopic particles	Grey	Tested against kit (+) and (-) controls	PASS
Positive Control Negative Control	Liquid solution	Clear and transparent		PASS

If applicable:

- Components from human origin have been tested and found negative for the presence of antibody to HIV as well as HBsAg and HCV. Handle with caution as potentially infectious.
- This product was tested by methods described in the manufacturers package insert.
- This product is intended for *In Vitro* Diagnostic use only.

---

**We certify that this product has been released as meeting our acceptance criteria**

---

APPROVED BY:

**Eddy Velthuis**  
Technical Director

DATE: 14 April 2021

---



# CERTIFICATE OF REGISTRATION

## Lorne Laboratories Ltd

Unit 1 Cutbush Park Industrial Estate  
Danehill  
Lower Earley  
Berkshire RG6 4UT UNITED KINGDOM

UL LLC®(UL) issues this certificate to the Firm named above, after assessing the Firm's quality system and finding it in compliance with:

**ISO 13485:2016**

**EN ISO 13485:2016**

The manufacture of in vitro diagnostic blood grouping reagents. The purchase for resale of in vitro diagnostic serology test kit.

Authorized by



**Michael J. Windler, P.E.**

**Manager of Global Regulatory Service**  
Distinguished Member of the Technical Staff  
Life and Health Sciences, UL LLC




Check Certificate  
Status: [here](#)

File Number	A12241	Cycle Start	May 23, 2020
Certificate Number	1458.200523	Effective Date	May 23, 2020
Initial Issue Date	June 26, 2018	Expiry Date	May 22, 2023

This quality system registration is included in UL's Directory of Registered Firms and applies to the provision of goods and/or services as specified in the scope of registration from the address(es) shown above. By issuance of this certificate the firm represents that it will maintain its registration in accordance with the applicable requirements. This certificate is not transferable and remains the property of UL LLC.



UL LLC  
333 Pfingsten Road  
Northbrook, IL 60062-2096 USA



# LORNE LABORATORIES LTD



Unit 1 Cutbush Park Industrial Estate  
Danehill, Lower Earley, Berkshire, RG6 4UT  
United Kingdom

Phone: +44 (0) 118 921 2264  
Fax: +44 (0) 118 986 4518  
Email: info@lornelabs.com

## CERTIFICATE OF ANALYSIS

DESCRIPTION	LOT NO.	EXPIRY	PRODUCT CODE
RPR Carbon Kit	LO16261	2023-03	044150A

**STORAGE:** Refrigerated at 2 – 8°C. Protect from light. Do not freeze.

**SHIPPING:** This product has data supporting stability tolerance during fluctuations in ambient shipping temperature.

This product is in compliance with **MEDICAL DEVICES REGULATIONS 2002** for Annex 3 and Self-Certification. This product was manufactured, packaged and tested in accordance with **LORNE QUALITY SYSTEMS** and meets all product specifications.

REAGENT	SPECIFICATIONS			RESULT
	Appearance	Colour	Functionality	
RPR Carbon Kit	Liquid suspension free of macroscopic particles	Grey	Tested against kit (+) and (-) controls	PASS
Positive Control Negative Control	Liquid solution	Clear and transparent		PASS

If applicable:

- Components from human origin have been tested and found negative for the presence of antibody to HIV as well as HBsAg and HCV. Handle with caution as potentially infectious.
- This product was tested by methods described in the manufacturers package insert.
- This product is intended for *In Vitro* Diagnostic use only.

---

**We certify that this product has been released as meeting our acceptance criteria**

---

APPROVED BY:

Eddy Velthuis  
Technical Director

DATE: 14 April 2021

---