


Certificate of Analysis

Product Name:

Tris-EDTA buffer solution - 100X concentrate for Northern and Southern blotting

Product Number: T9285
 Batch Number: SLCJ3277
 Brand: SIGMA
 Quality Release Date: 19 FEB 2021
 Recommended Retest Date: FEB 2024



Test	Specification	Result
Appearance (Color)	Colorless	Colorless
Appearance (Form)	Liquid	Liquid
Appearance (Turbidity)	Clear	Clear
pH 1x, H2O	7.8 - 8.2	7.9
DNase, Exonuclease Detection	None Detected	None Detected
RNase Detection	None Detected	None Detected
Conductivity	≥ 30000	33090
Standard Plate Count	≤ 10 CFU/ml	0 CFU/ml

Brian Dulle, Supervisor
 Quality Assurance
 St. Louis, Missouri US

Sigma-Aldrich warrants, that at the time of the quality release or subsequent retest date this product conformed to the information contained in this publication. The current Specification sheet may be available at Sigma-Aldrich.com. For further inquiries, please contact Technical Service. Purchaser must determine the suitability of the product for its particular use. See reverse side of invoice or packing slip for additional terms and conditions of sale.

