





SV300 Ventilator

| Physical Specification | | |
|--|--|--|
| Dimensions | 354 mm*315 mm*249 mm (Excluding the trolley) | |
| Weight | Approximately 10 kg (Excluding the trolley) | |
| Screen | | |
| Display Size | 12.1 Color active matrix TFT touch | |
| Display Resolution (H) x (V) | 1280*800 pixels | |
| Brightness | Adjustable | |
| Ventilation Specifications | | |
| Patient Type | Adults, children, infants (body weight of at least 3 kg) | |
| Ventilation Mode | V-A/C (Volume assist/control) P-A/C (Pressure assist/control) V-SIMV (Volume - Synchronized Intermittent Mandatory Ventilation) P-SIMV (Pressure - Synchronized Intermittent Mandatory Ventilation) DuoLevel (Duo Level Ventilation) CPAP (Continuous Positive Airway Pressure) PSV (Pressure Support Ventilation) APRV (Airway Pressure Release Ventilation) PRVC (Pressure Regulated Volume Control) PRVC-SIMV (Pressure Regulated Volume Control-Synchronized Intermittent Mandatory Ventilation) AMV (Adaptive Minute Ventilation) | |
| | NIV (Non-invasive Ventilation) Apnea Ventilation | |
| Controlled Parameters | | |
| O ₂ % | 21 - 100% (increments of 1 %) | |
| TV(Tidal Volume) | Adult: 100 - 2000 mL (increments of 10 mL) Pediatric: 20 - 300 mL (increments of 1 mL) | |
| MV% | 25% to 350% | |
| f (Ventilation frequency) | 1 - 100 bpm (increments of 1 bpm) | |
| fSIMV (Ventilation frequency in SIMV m | node) 1 - 60 bpm (increments of 1 bpm) | |
| l:E range | 4:1 - 1:10 (increments of 0.5) | |
| Tinsp (Inspiratory time) | 0.20 - 10 s (increments of 0.05 s) | |
| Tslope (Time of Pressure Rising) | 0 - 2.00 s (increments of 0.05 s) | |
| Thigh | 0.2 - 30 s (increments of 0.1 s) | |
| Tlow | 0.2 - 30 s (increments of 0.1 s) | |
| Tpause | 5 % - 60 % (increments of 5 %), Off | |
| ΔPinsp | 5 - 80 cmH ₂ O (increments of 1 cmH ₂ O) | |
| ΔPsupp | $0-80~\text{cmH}_2\text{O}$ (increments of 1 cmH $_2\text{O}$) | |
| Phigh | $0-80~\text{cmH}_2\text{O}$ (increments of 1 cmH ₂ O) | |
| Plow | $0-45 \text{ cmH}_2\text{O}$ (increments of 1 cmH ₂ O) | |
| PEEP | 1 - 45 cmH₂O (increments of 1 cmH₂O), Off | |
| Flow trigger | 0.5 -15 L/min (increments of 0.1 L/min), Off | |
| Pressure trigger | -10 to - 0.5 cmH $_2$ O (increments of 0.5 cmH $_2$ O), Off | |
| Exp % (Expiration termination level) | 10 - 85% (increments of 5%), Auto | |
| Apnea Ventila tion | | |
| Tvapnea | Adult: 100 - 2000 mL (increments of 10 mL) Pediatric: 20 - 300 mL (increments of 1 mL) | |
| ΔPapnea | 5 - 80 cmH ₂ O (increments of 1 cmH ₂ O) | |

| Sigh | | |
|--|--|--|
| Sigh Switch | ON, Off | |
| Interval | 20 s - 180 min (increments of 1 s from 20 to 59 s, increments of 1 min from 1 to 180 min) | |
| Cycles Sigh | 1 - 20 (increments of 1) | |
| Δint.PEEP | 1 - 45 cmH₂O (increments of 1 cmH₂O), Off | |
| Automatic Tube Resistance Compensation | | |
| Tube Type | ET Tube, Trach Tube, Disable ATRC | |
| Tube I.D. | Adult: 5.0 - 12.0 mm (increments of 0.5 mm) Pediatric: 2.5 - 8.0 mm (increments of 0.5 mm) | |
| Compensate | 0 -100 % (increments of 1 %) | |
| Expiration Compensation Switch | ON, Off | |
| Monitored parameters | | |
| Airway pressure range | Ppeak, Pplat, Pmean, PEEP (Range 0 - 120 cmH₂O) | |
| Tidal volume range | TVi, TVe, TVe spn (Range 0 - 4000 mL) | |
| Frequency range | ftotal, fmand, fspn (Range 0 - 200 bpm) | |
| Minute volume range | MV, MVspn, MVleak (Range 0 - 100 L/min) | |
| Resistance | Rinsp, Rexp (0 - 600 cmH₂O/L/s) | |
| Compliance | Cstat, Cdyn (0 - 300 mL/cmH₂O) | |
| Inspired Oxygen(FiO ₂) | 15 - 100 % | |
| RSBI | 0 - 999 1/(L•min) | |
| WOB | 0 - 100 J/min | |
| P0.1 | -20 - 0 cmH ₂ O | |
| NIF | -45 - 0 cmH ₂ O | |
| PEEPi | 0 - 80 cmH ₂ O | |
| RCexp | 0 - 80 cmH₂O 0 - 10 s | |
| TVe/IBW | 0 - 50 ml/kg | |
| I:E | 100:1 -1:150 | |
| Tinsp | 0.00 - 60.00s | |
| Waveforms | Airway pressure - time, Flow - time, Volume - time | |
| Loops | Paw - Volume, Flow - Volume, Paw - Flow | |
| Ventilator Accuracy | - an ionality from totality, an incom- | |
| Control Accuracy | | |
| O ₂ % | ± (3 vol.% +1 % of setting) | |
| TV | ± (10 mL + 10 % of setting) (BTPS) | |
| Tinsp | \pm 0.1 s or \pm 10 % of setting, whichever is greater | |
| I: E | 2: 1 to 1: 4: ± 10 % of setting, other range: ± 15% of setting | |
| f | ± 1 bpm | |
| fSIMV | ± 1 bpm | |
| Tslope | ± (0.2 s + 20 % of setting) | |
| PEEP | ± (2.0 cmH₂O + 5 % of setting) | |
| ΔPinsp | ± (2.0 cmH₂O + 5 % of setting) | |
| ΔΡѕирр | \pm (2.0 cmH ₂ O + 5 % of setting) | |
| Phigh | \pm (2.0 cmH ₂ O + 5 % of setting) | |
| Plow | \pm (2.0 cmH ₂ O + 5 % of setting) | |
| Thigh | \pm 0.2 s or \pm 10 % of setting, whichever is greater | |
| Tlow | \pm 0.2 s or \pm 10 % of setting, whichever is greater | |
| Pressure Trigger | \pm (1.0 cmH ₂ O + 10 % of setting) | |
| Flow Trigger | ± (1.0 L/min + 10 % of setting) | |
| Δint.PEEP | \pm (2.0 cmH ₂ O + 5% of setting) | |
| Exp % | ± 10 % | |
| fapnea | ± 1 bpm | |
| ΔPapnea | \pm (2.0 cmH ₂ O + 5 % of setting) | |
| TVapnea | ± (10 mL + 10 % of setting) (BTPS) | |
| Apnea Tinsp | ± 0.1 s or ± 10% of setting, whichever is greater | |
| pca | = 5 5 of ± 10/6 of setting, whichever is greater | |

| Monitoring Accuracy | | | | |
|--|----------|---|--|--|
| Airway pressure (Ppeak, Pplat, Pmean, PEEP) | | \pm (2 cmH ₂ O + 4 % of the actual reading) | | |
| Tidal Volume (TVi, TVe, TVe/IBW, TVe spn) | | 0 ml - 100 ml: \pm (10 ml + 3 % of the actual reading) (BTPS); 100 ml - 4000 ml: \pm (3 ml + 10 % of the actual reading) (BTPS) | | |
| Minute Volume (MV, MVspn, MVleak) | | \pm (0.2 L/min + 10 % of the actual reading) (BTPS) | | |
| Frequency (ftotal, fmand, fsp | on) | $\pm5\%$ of reading or $\pm1\text{bpm}$, whichever is greater | | |
| Inspired Oxygen (FiO₂) | | ± (2.5 vol.% + 2.5 % of the actual reading) | | |
| Resistance | | 0 to 20: ± 10 cmH₂O/L/s Other range: 50 % of actual reading | | |
| Compliance | | \pm (2 ml/cmH ₂ O + 20 % of the actual reading) | | |
| RSBI | | ± (3 1/(L•min)+15 % of the actual reading) | | |
| WOB | | ± (1 J/min+15 % of the actual reading) | | |
| NIF | | \pm (2 cmH ₂ O + 4 % of the actual reading) | | |
| P0.1 | | \pm (2 cmH ₂ O + 4 % of the actual reading) | | |
| PEEPi | | No declaration | | |
| RCexp | | \pm (0.2 s + 20 % of the actual reading) | | |
| Alarm settings | | | | |
| Tidal Volume | High | Adult: 110 - 4000 mL, Off Pediatric: 25 - 600 mL, Off | | |
| | Low | Adult: 50 - 4000 ml, Off Pediatric: 10 - 600 mL, Off | | |
| Minute Volume | High | Adult: 0.2 - 100.0 L/min Pediatric: 0.2 - 60.0 L/min | | |
| | Low | Adult: 0.1 - 50.0 L/min Pediatric: 0.1 - 30.0 L/min | | |
| Air pressure | High | 10 - 85 cmH₂O | | |
| Frequency | High | 1 - 150 bpm, Off | | |
| Inspired oxygen (FiO ₂) | High | Auto, FiO ₂ exceeds the alarm limit for at least 30 s, internal alarm limit: min (set value+max (7 % or set value*10 %), 100 %) | | |
| | Low | Auto, FiO_2 lower than the alarm limit for at least 30 s, internal alarm limit: max (set value-max (7 % or set value*10 %),18 %), Absolute FiO_2 low limit: 18 % | | |
| Apnea alarm time | | 5 - 60 s | | |
| SideStream CO₂ Module (o | ptional) | | | |
| Displayed numerics | | EtCO ₂ | | |
| Measurement Range | | 0 - 99 mmHg | | |
| Measurement accuracy | | 0 to 40 mmHg \pm 2 mmHg 41 to 76 mmHg \pm 5% of reading 77 to 99 mmHg \pm 10% of reading | | |
| Waveforms | | EtCO ₂ - time | | |
| Resolution | | 1 mmHg | | |
| Sampling rate | | Adult: 70 ml/min, 100 ml/min, 120 ml/min, 150 ml/min | | |
| | | Pediatric: 70 ml/min, 100 ml/min | | |
| Sampling rate Accuracy | | \pm 15% of the set value or \pm 15 mL/min, whichever is greater | | |
| System response time | | Using Adult water trap, Adult sampling line: < 7.5 s @ 150 ml/min < 8.0 s @ 120 ml/min < 8.5 s @ 100 ml/min < 9.5 s @ 70 ml/min Using Pediatric water trap, Pediatric sampling line: < 7.5 s @ 100 ml/min < 8.0 s @ 70 ml/min | | |
| Rise time | | Adult water trap: < 400 ms @70 ml/min < 330 ms @100 ml/min < 300 ms @120 ml/min < 240 ms @150 ml/min Pediatric water trap: < 400 ms @70 ml/min < 330 ms @100 ml/min | | |
| Water trap cleaning time | | Adult water trap: ≥24 h @150 ml/min ≥48 h @70 ml/min Pediatric water trap: ≥24 h @100 ml/min ≥48 h @70 ml/min | | |

| EtCO ₂ | High | 2 - 99 mmHg | | |
|--|--------------|--|--|--|
| LICO ₂ | Low | 0 - 97 mmHg | | |
| MainStream CO ₂ Module(o | | | | |
| Displayed numerics | ptional | EtCO ₂ | | |
| EtCO₂ Measurement range | | 0 -150 mmHg | | |
| EtCO ₂ Measurement Accurac | у | 0 to 40 mmHg ± 2 m 41 to 70 mmHg ± 5% | mHg of reading of reading of reading % of reading | |
| Resolution | | 1 mmHg | | |
| Waveforms | | EtCO ₂ - time, Volume - EtCO ₂ | | |
| Other Parameters | | | | |
| SlopeCO₂ (slope of the alveo | lar plateau) | Range: 0 - 9.99 %/L | Resolution: 0.01 %/L | |
| Vtalv (Alveolar tidal ventilati | | Range: 0 - 9999 ml | Resolution: 1 ml | |
| V'alv (Alveolar minute ventil | ation) | Range: 0 - 20 L/min | Resolution: 0.01 L/min for < 1 L/min, 0.1 L/min for ≥ 1 L | |
| V'CO ₂ (CO ₂ elimination) | | Range: 0 - 9999 mL/min | Resolution: 1 ml/min | |
| VDaw (Airway death space) | | Range: 0 - 999 mL | Resolution: 1 ml | |
| VDaw/TVe (Physiological dea fraction at the airway openin | | Range: 0 - 100 % | Resolution: 1 % | |
| VeCO ₂ (exha l ed CO ₂ volume) | | Range: 0 - 999 mL | Resolution: 1 ml | |
| ViCO ₂ (inspired CO ₂ volume) | | Range: 0 - 999 mL | Resolution: 1 ml | |
| System response time | | < 2.0 s | | |
| CO ₂ alarm limits | | | | |
| EtCO ₂ | High | 2 - 150 mmHg | | |
| | Low | 0 - 148 mmHg | | |
| SpO ₂ module(optional) | | | | |
| Measurement Range and R | esolution | | | |
| SpO ₂ | | Range: 0 - 100 % | Resolution: 1 % | |
| PR | | Range: 20 - 254 1/min | Resolution: 1 1/min | |
| PI | | Range: 0.05 - 20 % | | |
| Measurement Accuracy | | 9 | | |
| | | 70 to 100 %: ±2 % | | |
| SpO ₂ | | 0 % to 69 %: Not specified | d. | |
| PR | | ± 3 1/min | | |
| SpO₂ alarm limits | | | | |
| SpO ₂ | High | 2 -100 % | | |
| | Low | 0 - 98 % | | |
| | Desat | 0 - 98 % | | |
| PR | High | 17 - 300 1/min | | |
| | Low | 15 - 298 1/min | | |
| Trend | | | | |
| Туре | | Tabular, Graphic | | |
| Length | | 72 hours | | |
| Content | | Monitor Parameters, Setti | ing Parameters (Setting Ventilation mode and Parameters | |
| O₂ Therapy | | | | |
| Controlled Parameters | | | | |
| O ₂ % | | 21 - 100 % (increments of | f 1 %) | |
| Flow | | 2 - 60 L/min | | |
| Controlled Accuracy | | | | |
| O ₂ % | | ± (3 vol.% +1 % of setting | ı) | |
| Flow | | ± (2 L/min +10 % of settir | ng) (BTPS) | |
| Monitor Accuracy | | | | |

| Flow | ± (2 L/min + 10 % of the actual reading) (BTPS) | |
|---|--|--|
| Log | | |
| Туре | Alarm, Operation | |
| Max number | 5000 | |
| O ₂ Sensor | | |
| Туре | Galvanic fuel cell | |
| Response time | < 15 s | |
| Communication interface | | |
| Communication interface | Rs232, Ethernet, VGA, USB port, Nurse call | |
| Gas supply | | |
| Gas type | 02 | |
| Pipe Connector | NIST or DISS | |
| Gas supply pressure | 280 - 600 kPa | |
| Peak flow in case of single supply gas(air) | ≥ 210 L/min (BTPS)* ¹ | |
| Operation Data | | |
| Environmental specifications | | |
| Temperature | 5 - 40 °C (operating); -20 to 60 °C (storage and transport, O_2 sensor: -20 to 50 °C) | |
| Relative Humidity | 10 - 95 % (operating); 10 - 95 % (storage and transport) | |
| Barometric Pressure | 62 - 106 kPa (operating); 50 -106 kPa (storage and transport) | |
| Power and Battery Backup | | |
| External AC power supply | | |
| Input voltage | 100 - 240 V | |
| Input frequency | 50/60 Hz | |
| Input current | 2.7 - 1.1 A | |
| Fuse | T3.15 AH/250 V | |
| External DC power supply | | |
| Input voltage | 12 V | |
| Input current | 15 A | |
| Internal battery | | |
| Number of batteries | One or Two | |
| Battery type | Build-in Lithium-ion battery, 14.8 VDC, 5800 mAh | |
| Battery run time | 180 min (Powered by one new fully - charged battery in standard working condition 360 min (Powered by two new fully - charged battery in standard working condition) | |
| Trolley | | |
| Dimensions | 1039 mm*528 mm*544 mm | |
| Weight | Approximately 20 kg | |
| Special Functions and procedures | | |
| Sigh | | |
| 100% O ₂ | | |
| Suction | | |
| Manual breath | | |
| Expiratory hold | | |
| Inspiratory hold | | |
| P0.1 | | |
| NIF | | |
| PV - Tool | | |
| PEEPi | | |
| Nebulizer | | |
| | | |

^{*}¹ BTPS =Body Temperature and Pressure Saturated
*² The standard work condition is:Ventilation mode:P-A/C; ΔPinsp:10 cmH₂O; f:10 bpm; Tinsp:2 s; Tslope:0.2 s; O₂%:21 Vol.%; PEEP:5 cmH₂O; R:20 cmH₂O/L/s; C:20 ml/cmH₂O; Gas supply nominal work pressure:400±100 kPa.