



中国认可
国际互认
检测
TESTING
CNAS L0227



No.TJA20221109-0489

INSPECTION REPORT

Sample Name: 95 kPa Class A biosafety transport tank (UN2814)

Inspected Company: Shanghai SQBQ Biotechnology Co., Ltd

Test Type: Commission

China Packaging Research & Test Center
National Package Product Quality Inspection & Test Center(Tianjin)
No.2 Haichuan Avenue Huanghai Rd. TEDA Tianjin China
TEL:86.22.66231623/1625 FAX:86.22.66231624
<http://www.packagetest.net>
Email:cprtclab@packagetest.net



DECLARE

- 1 .This report is void either without “the special inspection and test seal” or the seal of the test Organization.
2. Copy of the report is void without re-stamp of “the special inspection and test seal” or the seal of the test Organization. Without written approval, this report cannot be reproduced except in full.
3. This report is void without the signatures of the approver, auditor and editor.
4. Any erase and rewrite to the report will make the report void.
5. The consigner is responsible for the authenticity and representative of the delivered samples and any related information. The test organization do not undertake any relevant responsibilities.
6. Generally, delivered sample tests are only responsible for the incoming samples.
7. Objection for the report should be raised to the test organization within 15 (Fifteen) days from the date receiving the report.

China Packaging Research & Test Center

National Package Product Quality Inspection & Test Center (Tianjin)

INSPECTION REPORT

No.TJA20221109-0489

Sample Name	95 kPa Class A biosafety transport tank (UN2814)			Model/ Type	/	
Date of Production	/	Brand	/			
Production Lot Number	/					
Applicant	Shanghai SQBQ Biotechnology Co., Ltd			Grade	/	
Address	Room 202, Room 205, Room 303-304, Building 2, No. 152, Lane 468, Beiheng Shahe Road, Minhang District, Shanghai					
Inspected Company	Shanghai SQBQ Biotechnology Co., Ltd					
Manufacturer	/					
Test Type	Commission	Applicant Rep.	Chenyu Yao	Date of Application	2022/06/23	
Sample (s) Quantity	5 units	Receipt Date	2022/06/23	Date Tested	2022/06/29	
Sample Description	95 kPa Class A biosafety transport tank, the appearances of the packaged-products are normal.					
Test Standard	IATA Dangerous Goods Regulations 63th Ed.;ICAO regulations of dangerous goods Doc 9284 AN/905 2021-2022.					
Judgement Standard	IATA Dangerous Goods Regulations 63th Ed.;ICAO regulations of dangerous goods Doc 9284 AN/905 2021-2022.					
Test Item(s)	Leakage test					
Test Conclusion	Test results as follows.			Inspection seal Issue Date: 2022/06/30	检验检测专用章	
Note	/					

Edited by: 

Checked by: 

Approved by: 

INSPECTION REPORT

No.TJA20221109-0489

No	TEST ITEM	UNIT	TEST STANDARD	TEST RESULT	JUDGE
1	Leakage test	pressure differential 95kPa, -40°C, 30min	/	No leakage from the secondary receptacles.	No leakage from the secondary receptacles. PASS
		pressure differential 95kPa, 55°C, 30min	/	No leakage from the secondary receptacles.	
NOTE	/				



-----THE END-----