

COMPLIANCE SHEET

To whom may concern:

Center for Centralized Public Procurement in Health

We **MEDCAPTAIN MEDICAL TECHNOLOGY CO., LTD**, 12th Floor, Baiwang Research Building, No.5158 Shahe West Road, Xili, Nanshan, 518055, Shenzhen, Guangdong, PEOPLE'S REPUBLIC OF CHINA

Do confirm, that the offered infusion pumps

HP-30Pro

Offered in cooperation with **Aelo-Grup SRL** for the public procurement

21040135 Lot 31
95 Units

Complies with the following technical specification:

REQUIRED:

Infusion pump (infuser) with syringe (advanced features)

Description A syringe infuser is a device that allows the intravenous administration of medical fluids with high accuracy.

Parameter Specification

General features Type With 1 channel

LCD display

the possibility of adjusting the brightness

Possibility of placement in a column with fixing yes

Display Data displayed Infusion rate yes

The infusion volume yes

Infusion time yes

Time to the end of the infusion yes

Syringe size yes

The syringe manufacturer does

Occlusion level yes

Rate KVO da

Date, time is

Battery charge level yes

Events yes

Alarm is

Pump properties Flow rate range 0.1 - 999 ml / h

0.01 mL / h increment step

Bolus dosage Increase 0.1 - 1200 ml / h

KVO rate 0.1 - 3 ml / h, with incremental step 0.1 ml / h

Accuracy ≤2%

Compatible with all required syringe sizes produced by different manufacturers min. 8 producers, with the presentation of the list of producers

Possibility to add new types of syringes in the library yes

Accepted syringes 2, 3, 5, 10, 20, 30, 50, 60

OFFERED:

Infusion pump (infuser) with syringe (advanced features)

Description A syringe infuser is a device that allows the intravenous administration of medical fluids with high accuracy.

Parameter Specification

General features Type With 1 channel

LCD display

the possibility of adjusting the brightness

Possibility of placement in a column with fixing yes

Display Data displayed Infusion rate yes

The infusion volume yes

Infusion time yes

Time to the end of the infusion yes

Syringe size yes

The syringe manufacturer does

Occlusion level yes

Rate KVO da

Date, time is

Battery charge level yes

Events yes

Alarm is

Pump properties Flow rate range 0.1 - 9999.99 ml / h

0.01 mL / h increment step(Page10)

Bolus dosage Increase 0.1 - 2200 ml / h(Page10)

KVO rate 0.1 - 5 ml / h, with incremental step 0.01 ml / h(Page10)

Accuracy ≤2%(Page10)

Compatible with all required syringe sizes produced by different manufacturers min. 8 producers, with the presentation of the list of producers

Possibility to add new types of syringes in the library yes

Accepted syringes 2, 3, 5, 10, 20, 30, 50, 60

Automatic syringe detection yes

<p>Automatic syringe detection yes Maximum occlusion pressure level 75 - 880 mmHg min. 9 levels Possibility to add other names to the library in the library yes Protection against the ingress of liquids into the device minimum IP23 Infusion modes Infusion rate yes Dosage in time yes The ascending and descending mode yes The dosage is micro-infusion Alarms and indicators Occlusion yes System error yes Empty syringe yes Infusion at last yes Discharged battery yes Audible alarm Adjustable volume Yes Silent mode button yes Number of stored events yes Logbook Stored events Buttons pressed yes Error codes yes Alarm is Medically injected by The amount infused yes Settings yes Bolus is requested Power supply 220 V, 50 Hz yes Internal Rechargeable Battery yes Operating time ≥ 10 h at 5 mL / h Accessories Support the system for fixing the pump to the infusion stand yes Syringes 50 ml 100 pcs. Syringe interconnect circuit 100 pcs.</p>	<p>Maximum occlusion pressure level 75 - 975 mmHg min. 13 levels Possibility to add other names to the library in the library yes Protection against the ingress of liquids into the device minimum IP23 Infusion modes Infusion rate yes Dosage in time yes The ascending and descending mode yes The dosage is micro-infusion Alarms and indicators Occlusion yes System error yes Empty syringe yes Infusion at last yes Discharged battery yes Audible alarm Adjustable volume Yes Silent mode button yes Number of stored events yes Logbook Stored events Buttons pressed yes Error codes yes Alarm is Medically injected by The amount infused yes Settings yes Bolus is requested Power supply 220 V, 50 Hz yes Internal Rechargeable Battery yes Operating time ≥ 10 h at 5 mL / h(Page9) Accessories Support the system for fixing the pump to the infusion stand yes Syringes 50 ml 100 pcs. Syringe interconnect circuit 100 pcs.</p>
<p>Manufacturer's representative</p>	<p>stamp</p> <div data-bbox="1082 1352 1426 1585" data-label="Image"> </div>