

Adjustment and calibration certificate #0000211741

Object	Instrument:	MAS-100 NT / NT Ex Microbial AirSampler		
	Serial number:	0000121475		
	FW version:	1.7.18		
	HW version:	8		
	Name:	Head 1		
	Department:	Location 1		
	Perforated lid:	400 × 0.7 mm		
	Calibration procedure:	MBV SOP / Version 1.6.31		
Adjustment equipment	Serial number:	18007		
	Adjustment date:	31.7.2021		
	Deviation:	0.0 [l/min]		
Adjustment data	Adjustment date:	13.04.2022		
	Adjustment conditions:	Temp.: 24.0 [°C], Amb. Pressure: 975 [mbar]		
	ADC at 0 flow:	794 [mV]		
	ADC at target flow:	3019 [mV]		
	Gain:	78		
Calibration equipment	Serial number:	18007		
	Adjustment date:	31.7.2021		
	Deviation:	0.0 [l/min]		
Calibration data	Calibration date:	13.04.2022		
	Calibration conditions:	Temp.: 24.0 [°C], Amb. Pressure: 975 [mbar]		
	Run time check:	Passed		
Calibration results:				

		Sequence			Result
		1	2	3	
Target air flow	SLPM	100.0	100.0	100.0	100.0
Air flow on anemometer measurement 1	SLPM	99.8	99.8	99.8	NA
Air flow on anemometer measurement 2	SLPM	99.8	100.0	99.9	NA
Air flow on anemometer measurement 3	SLPM	100.0	99.6	99.8	NA
Air flow	SLPM	99.9	99.8	99.8	NA
Flow deviation of the anemometer ¹	l/min	0.0	0.0	0.0	NA
Calibration result (corr. air flow of anemometer)	SLPM	99.9	99.8	99.8	99.8
Deviation from target air flow Maximum allowable deviation: ± 2.5%	%	-0.1	-0.2	-0.2	-0.2
Calibration check status:					Accepted

¹⁾ For the DA-100 NT the deviation mentioned on the calibration certificate is not normalized to SLPM conditions (20°C; 1013 mbar) as it is not significant: The correction would be 0.34% per °C (1/293) and 0.09% (1/1013) per mbar difference to the SLPM standard conditions. E.g. a 0.2 l/min deviation measured at 25°C and 940 mbar leads to an error of 8.8% or 0.018 l/min.

Calibrated by:	J. Guevara	_ Signature:	Noverala
This report has been produced	by an electronic system an	d is valid with a singl	le signature of the calibrator.
Date of activation by user: _		_ Signature: _	
Due date: ²			
²) 12 months after the first activ	ation		