## **EU DECLARATION OF CONFORMITY**

Manufacturer:

**MERCATOR MEDICAL S.A.** 

UL. H.MODRZEJEWSKIEJ 30 31-327 KRAKÓW, POLSKA

Declares under its sole responsibility that non-sterile examination and protective gloves:

| Brand                         | Туре                                          | Sizes                | Reference Numbers                                                   |
|-------------------------------|-----------------------------------------------|----------------------|---------------------------------------------------------------------|
| nitrylex <sup>®</sup> classic | nitrile, powder-free, blue, for single<br>use | XS (5-6) - XL (9-10) | a'100: RD30019001-05<br>a'200: RD30096001-05                        |
|                               | nitrile, powder-free, white, for single use   | XS (5-6) - XL (9-10) | a'50: RD30174001-05<br>a'100: RD30143001-05<br>a'200: RD30097001-05 |
|                               | nitrile, powder-free, violet, for single use  | XS (5-6) - XL (9-10) | a'100: RD30169001-05<br>a'200: RD30168001-05                        |
|                               | <b>Basic UDI-DI:</b> 5906615                  | RD NS N PF 9C        |                                                                     |

meet the provisions of the Regulation (EU) 2017/745 of the European Parliament and the Council of 5 April 2017 on medical devices, are classified as medical device class I according to Annex VIII of the Regulation (EU) 2017/745 and comply with European harmonized standards: EN 455-1:2000, EN 455-2:2009+A2:2013, EN 455-3:2006, EN 455-4:2009, EN ISO 15223-1:2016, EN 1041:2008.

The products described above are also classified as Personal Protective Equipment Category III and comply with Regulation (EU) 2016/425 of the European Parliament and the Council of 9 March 2016 on Personal Protective Equipment and resolution of the Council Directive 89/686/EEC and European standards: EN 420:2003+A1:2009, EN ISO 374-1:2016, EN 374-2:2014, EN 16523-1:2015, EN 374-4:2013, EN ISO 374-5:2016.

The products described above are identical to the Personal Protective Equipment, which is the subject to the EU Type Examination (Module B) under certificate No. 2777/10015-03/E17-01 issued by notified body:

## Satra Technology Europe Limited (2777)

Bracetown Business Park, Clonee, Dublin 15, Dublin, Ireland

and are subject to the conformity to type procedure based on the internal production control plus supervised product checks at random intervals (Module C2) under surveillance of the notified body:

## Satra Technology Europe Limited (2777)

Bracetown Business Park, Clonee, Dublin 15, Dublin, Ireland

Date and place of issue: 05.05.2020, Kraków

MERCATOR MEDICAL S.A.
ul. Heleny Modrzejewskiej 30, 31-327 Kraków
tel. 12 66 55 400, fax 12 66 55 415
Rejestracja: Sąd Rejonowy dla Krakowa - Śródmieścia w Krakowie,
XI Wydział Gospodarczy (RS, KRS: 0000036244
Kapitał zakładowy (w całości wpłacony): 10.589.100 PLN
NIP: 677-10-36-424, REGON: 350967107
Numer BDO: 000056063

Signed on the behalf of the Manufacturer:

Wojciech Hercka

**Product Documentation Manager**