

# Certificate of Analysis



## Cathode Buffer Container CE-SeqStudio

Product No.                      **A33401**  
Lot No.                              **2409142**  
Date of Manufacture              **24SEP2024**

TEST	SPECIFICATION	RESULT
<b>PHYSICAL INSPECTION</b> Material is clear, colorless and free of particulates		
Appearance - Conformance		Pass
<b>pH TEST</b> Measured at 25.0°C ± 0.3°C		
pH Value	7.6 - 8.2	7.920
<b>CONDUCTIVITY TEST</b> Measured at 25.0°C ± 0.1°C		
Conductivity	1500 - 2000 µS/cm	1790 µS/cm

*For Research Use Only.*  
*Not for use in diagnostic procedures.*

**Manufactured in compliance with our ISO 13485 certified quality management system.**  
**Site: Bedford, MA, USA**

Corporate Legal Entity  
Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

*Janelle Him*  
**Quality Assurance**  
**Issued 25SEP2024**