




CARESCAPE B850/B650/B450 Modular Monitors CARESCAPE ONE

Clinical Reference Guide
Software Version 3.1





Notice

The materials contained in this document are intended for educational purposes only. This document does not establish specifications, operating procedures or maintenance methods for any of the products referenced. Always refer to the official written materials (labeling) provided with the product for specifications, operating procedures and maintenance requirements.

Contents

01 Monitoring Basics.....	5	Changing an Alarm Limit	37
Acquisition Modules	5	Critical Alarms for HR and SpO ₂	38
Main Screen Layout for B850/B650/B450	6	Breakthrough Alarms	39
Admitting and Discharging a Patient Overview.....	7	08 Trends	40
Accessing the Admit/Discharge Menu.....	7	Viewing Numeric Trends.....	40
Loading Patient Information from the CARESCAPE Network	8	Viewing Graphic Trends.....	41
Combination Monitoring Mode	9	Managing Snapshots	42
Roving Functionality.....	10	Managing Events	43
Discharging a Patient	11	09 Troubleshooting	44
Starting Standby	11	ECG Troubleshooting	44
Selecting Pause Monitor & Central.....	12	Respiration Troubleshooting.....	47
02 ECG.....	13	SpO ₂ Troubleshooting	48
ECG Setup Menu:.....	13	NIBP Troubleshooting.....	49
Skin Preparation.....	14	10 CARESCAPE ONE	50
Lead Placement	14	Overview	50
12 Lead Analysis.....	19	CARESCAPE ONE System Components	51
Impedance Respiration	20	CARESCAPE ONE Front, Back, and Side Views	52
Neonatal Impedance Respiration	21	Docking	53
03 NIBP.....	23	Turning the CARESCAPE ONE On/Off	54
Manual NIBP Measurement from the Main Menu	23	Inserting and Removing a Battery.....	54
The NIBP Setup Menu	24	CARESCAPE ONE Monitoring Basics	55
04 Pulse Oximetry	25	Profiles.....	58
Overview.....	25	Starting Monitoring.....	59
05 Invasive Pressure Monitoring	28	Entering Patient Data.....	60
Invasive Pressure Connections	28	Resetting a Case / Discharging a Patient.....	61
Zeroing a Transducer	29	Standby	62
The Invasive Pressure Setup Menu.....	30	Alarms	63
06 Airway Gases	31	Pausing Alarms	63
Airway Gas Modules and Connectors.....	31	Adjusting the Alarm Volume	63
CARESCAPE CO ₂ Measurement	32	Changing an Alarm Limit.....	64
Airway Gas Setup	33	Trends.....	65
Starting and Stopping the Pump.....	34	Transport Use	66
07 Alarms.....	35		
Alarm Priority Levels	35		
Pausing Alarms	35		
Adjusting the Alarm Volume	36		



01 Monitoring Basics

Acquisition Modules

There are two multi-parameter acquisition modules that process the patient data signals for the main patient parameters such as ECG, Impedance Respiration, Temperature, SpO₂, Invasive Pressure, and Non-Invasive Blood Pressure (NIBP).

1. The PDM (Patient Data Module)
2. The CARESCAPE ONE


Only one of these modules may be used at a time. The main difference between the two is the PDM does not have any type of display, it must be connected to a standalone monitor such as the CARESCAPE B850, B650 or B450 (CARESCAPE B Series Monitor). The CARESCAPE ONE has its own display and can be used as a standalone monitor or as an acquisition module.



Note! If both modules are connected simultaneously, the CARESCAPE ONE will have priority.

PDM

1. ECG/Imp. Respiration
2. T1-T2/CO
3. P1-P4
4. Tab for removing the module
5. SpO₂
6. NIBP
7. Defib/Sync
8. Dual function Power On and Zero All button
9. Power indicator
10. Communication Indicator




CARESCAPE ONE



Note! The CARESCAPE ONE is discussed in greater detail later in this material.

1. Parameter Connectors
2. Latch to disconnect from the dock
3. NIBP Connector
4. Defib/Sync
5. Screen Lock
6. Power on/standby button status
7. Battery status indicator
8. Alarm Pause/off status
9. Alarm Light




Main Screen Layout for B850/B650/B450

The main screen displays alarms, information, trends, waveforms, parameter windows, and the main menu in pre-defined areas.

When the information area of the screen is selected, it opens the Admit/Discharge or Case Setup menu and provides access to settings related to patient and administrative information, standby and profile selections.

1. Alarm area
2. Split screen area (option).
You may also select Minitrend as an option for this area.
3. Waveform area
4. Information area
5. Upper parameter windows
6. Lower parameter windows (option)
7. Main menu area



Admitting and Discharging a Patient Overview

A patient is automatically admitted when the monitor detects any of the following vital signs: ECG, Impedance Respiration, Art, Fem, UAC, NIBP, SpO₂, CO₂, EEG, BIS, or Entropy. (EEG, BIS, and Entropy are not used with the CS ONE.)

A patient is also automatically admitted when any patient data is entered or loaded. Patient data can be entered locally using the monitor, scanner using a barcode reader, loaded from an Admit- Discharge-Transfer (ADT) server over the CARESCAPE Network, or entered remotely using a central station.

Accessing the Admit/Discharge Menu

The Admit/Discharge menu is used to add or edit patient information. There are two ways to access the Admit/Discharge menu:

1. Select the patient information area on the screen.
2. Select **Data & Pages** from the main menu area at the bottom of the display and then select **Admit/ Discharge** or **Case Setup**.

Selecting a Profile

Overview

A profile is a group of unique settings suited to a particular care unit or patient demographic and can be customized. Settings for a profile could include alarm limits, screen layouts, trends and snapshot settings. When you start monitoring a patient, you can use the startup profile or select another profile. According to the configuration, your monitor software may have up to eight profiles to choose from.

You can select another profile while monitoring a patient without losing any patient data.

To Select a Profile:

1. Select the patient information area on the screen.
2. Select the **Patient** tab.
3. Select a profile from the **Profile** list.
4. You can return to the previous profile by selecting **Return to Previous Profile**.

The screenshot shows the 'Admit/Discharge' menu with a dark background. At the top, there are five tabs: 'Patient' (highlighted with a blue bar and a blue circle with the number 2), 'Load Patient', 'Administr. Information', 'Standby', and 'Care Unit & Bed'. Below the tabs, the 'Patient' section contains fields for 'MRN:', 'Second ID:', and 'Name:'. To the right of these fields are buttons for 'Edit Name & MRN' and 'Scan from Barcode'. Below these fields are sections for 'Date of Birth:', 'Age:', 'Height:', 'Weight:', 'BSA:', 'Gender:', and 'Ethnicity: Unknown'. To the right of these sections are buttons for 'Height cm', 'Weight kg', and 'Edit All Demographics'. Below these sections is a 'Profile' dropdown menu showing 'Medical' (highlighted with a blue circle with the number 3). Below the 'Profile' dropdown is a button labeled 'Return to Previous Profile' (highlighted with a blue circle with the number 4). At the bottom right, there is a 'Discharge Patient' button with a right arrow. At the very bottom, there is a 'Close' button.

Loading Patient Information from the CARESCAPE Network

In the CARESCAPE Network, patient information can be loaded from the ADT server. You cannot merge data between the monitor and the CARESCAPE Gateway Server. This can only be done on the host or Central Station. You cannot do barcode scanning and ADT on the CS ONE.

1. Select the patient information area on the screen.
2. Select the **Load Patient** tab.
3. Select **Find Patients**.
4. Select the **Medical Record Number** and/or **Last Name** field and enter the information you have available.
5. Select **Find**.
6. When the patient list appears, select the patient.
7. Select **Load Patient Information** to load the data from the CARESCAPE Gateway server.



Note! You can also add the First Name information but the search does not function with this information only.

The screenshot shows the 'Admit/Discharge' screen with the 'Load Patient' tab selected. A yellow box highlights the 'Find Patients' button. A blue circle with the number '2' points to the 'Load Patient' tab, and a blue circle with the number '3' points to the 'Find Patients' button. Below the button is a table with columns for Patient Name, Medical Record Number, Date of Birth, and Gender. A large empty area below the table is for the patient list. At the bottom right, there is a 'Load Patient Information' button and a 'Close' button.

Patient Name	Medical Record Number	Date of Birth	Gender
--------------	-----------------------	---------------	--------

Load Patient Information

Close

Combination Monitoring Mode

Combination monitoring is a licensed feature where ECG is acquired from a telemetry receiver system. This ECG data acquisition capability enhances basic telemetry monitoring by providing access to all of the available parameters from bedside monitors, while acquiring the ECG data from telemetry. All data is viewed at the central station and the bedside monitor.

Selecting the ECG Source



Note! Not available with the NICU software package.

1. Select the HR parameter window and then **ECG**.
2. Select **Setup**.
3. Select a source from the **ECG Source** list.
4. Select the correct patient transmitter/transceiver from the drop-down list.
5. Confirm the source selection with **Confirm**.

Exiting Combo/Discharging from Combo

1. Select the HR Parameter window and then **ECG**.
2. Select **Setup**.
3. Select a source from the **ECG Source** list.
4. Select **Monitor** from the drop-down list.
5. Confirm the source selection with **Confirm**.

The screenshot shows the ECG Setup screen. At the top, there are tabs for ECG, ST, QT, and Arrhythmia. Below these are tabs for Setup, Advanced, and Alarms. The Setup tab is selected. On the left, there are three ECG Lead dropdown menus: ECG 1 Lead (set to II), ECG 2 Lead (set to V1), and ECG 3 Lead (set to aVL). Below these is a Va Lead Position dropdown set to V1. On the right, there are controls for Beat Volume (set to 0), Beat Source (set to Primary HR), Size (set to 1x), and Hemo Sweep Speed (set to 25 mm/s). At the bottom, there are buttons for Update Lead Set, ECG Source (with a right arrow), 12 Lead Analysis (with a right arrow), All ECG Waveforms (with a right arrow), and a Close button.

1. ECG

2. Setup

3. ECG Source

Roving Functionality

Roving functionality allows you to move, or rove, the monitor to fit the patient's acuity needs, rather than moving the patient to a higher acuity monitored room. When you move the monitor to a new location on the CARESCAPE Network, you can update the unit and/or bed names from drop-down lists or add new names manually. Available selections depend on what has been allowed in configuration.

Roving between units

If roving between units is allowed, you can update the unit name when moving the monitor to a new location.

1. Select the patient information area on screen.
2. Select the **Care Unit & Bed** tab.
3. Select the **Care Unit Name** from the drop-down list.

Changing the Care Unit Name will also update the contents of the Bed Name list.

You can also change the name manually through **New Unit & Bed**. This selection is available in the **Care Unit & Bed** menu if it has been allowed in the Roving settings.

Roving between beds

If roving between beds is allowed, you can update the bed name when needed.

1. Select the patient information area.
2. Select the **Care Unit & Bed** tab.
3. Select the **Bed Name** from the drop-down list.

The new name appears in the upper right corner of the display. The unit name is given first, then a dash and the bed name (for instance, UNIT1–BED1).

You can also change the name manually through New Unit & Bed, or through New Bed. These selections are available in the Care Unit & Bed menu according to what has been allowed in the Roving settings. Monitors being roved to a different room need to be discharged before leaving the patient room and admitting another patient to the monitor.

Adding new units and beds (manual roving)

If manual roving between beds and/or units is allowed, you can also enter their names manually.

1. Select the patient information area.
2. Select the **Care Unit & Bed** tab. If the Roving settings do not allow roving between units, the **New Unit & Bed** is not available. In this case, select New Bed to enter a new bed name.
3. Select **New Unit & Bed**.
4. Select the **Care Unit Name** or the **Bed Name** field and type the new name with the on-screen keyboard. The maximum number of characters for the **Care Unit Name** is seven, and for the **Bed Name** it is five.

The screenshot shows the 'Admit/Discharge' screen with the 'Care Unit & Bed' tab selected. The 'Care Unit Name' is set to 'ICU' and the 'Bed Name' is 'Bed 12'. A 'Location ID' field shows '1'. A numeric keypad is visible for manual entry, and buttons for 'New Unit & Bed', 'Printing', and 'Close' are present.

Discharging a Patient

1. Disconnect patient cables.
2. Print necessary data and wait until the printing is completed.
3. Select the patient information area on the screen.
4. Then select **Admit/ Discharge**.
5. Select the **Patient** tab.
6. Select **Discharge Patient** and then **Confirm**.
(In PACU and Anesthesia software, rather than Discharge Patient, you will select **Reset Case**.)



Note! Monitor settings, including alarm limits, return to their default settings. All patient data and trend data is removed from both the monitor and a connected acquisition device.

5

The screenshot shows the 'Admit/Discharge' screen with the 'Patient' tab selected. The 'Patient' tab is highlighted with a blue circle and the number 5. The 'Patient' information area is highlighted with a yellow box. The 'Discharge Patient' button is highlighted with a blue box and the number 6. The 'Close' button is at the bottom right.

Starting Standby

When you remove the patient temporarily from the monitor, you can use the standby option.

1. Select the patient information area on the screen. Standby can also be accessed via **Data** and **Pages**.
2. Select the **Standby** tab.
3. Select the radio button for an appropriate standby location.
4. Select **Prepare for Standby**. If patient cables are still connected and the monitor receives vital signs, a text indicating that audio alarms have been paused appears.
5. Remove the acquisition device or disconnect patient cables to start the standby. If you do not disconnect the cables and vital signs are still present after the audio pause time expires, the standby is canceled. Check that the **NIBP Auto** is turned off.
6. The screen goes blank and the GE logo along with a message such as **Patient temporarily in MRI** (according to the location selected) appears.

2

The screenshot shows the 'Admit/Discharge' screen with the 'Standby' tab selected. The 'Standby' tab is highlighted with a blue circle and the number 2. The 'Operating Room' radio button is highlighted with a blue circle and the number 3. The 'Prepare for Standby' button is highlighted with a yellow box and the number 4. The 'Close' button is at the bottom right.

Selecting Pause Monitor & Central

If Pause Monitoring has been allowed in the Care Unit Settings (password protected), you can select to pause audible alarms at both the monitor and central. This selection also pauses all visual alarms and alarm printouts at the monitor. The system stops sending the digital output signal to the CARESCAPE Remote Alarm Device (RAD) when Pause Monitor & Central is active.

For more information regarding the Pause Monitoring setting, see the supplemental information provided. Pause Monitoring is not available in the CS ONE at this time.

1. Select **Alarms Setup** from the main menu.
2. Select the **Pause Monitoring** tab. If there is no network connection, it is indicated with the text **Monitor is not connected to the network** on an orange background. Pausing monitoring in this case would result in a complete loss of monitoring.
3. Select **Pause Monitor & Central**.

The screen saver that follows replaces the display of patient data, and alarms will be off at the monitor. In case there is no admitted patient/active patient case, it is indicated by the text **Monitoring paused - Patient discharged or Monitoring paused - Case ended**, respectively. Touch the screen or press any key to continue.

Automatic admit and discharge are not possible during **Pause Monitor & Central**. If remote discharge from a central station has been enabled, it can also be done during **Pause Monitor & Central**.

02 ECG

ECG Setup Menu:

To access the ECG setup menu:

1. Select the **ECG** parameter window.
2. Select the **Setup** tab.

The screenshot displays the ECG Setup Menu with a dark grey background. At the top, a header bar contains the title 'ECG' and four tabs: 'ECG' (highlighted in blue), 'ST', 'QT', and 'Arrhythmia'. Below this, a sub-header bar features three tabs: 'Setup' (highlighted in blue), 'Advanced', and 'Alarms'. The main content area is organized into two columns. The left column contains four dropdown menus for 'ECG 1 Lead' (set to II), 'ECG 2 Lead' (set to V1), 'ECG 3 Lead' (set to aVL), and 'Va Lead Position' (set to V1). The right column includes a 'Beat Volume' control with a speaker icon and a value of 0, a 'Beat Source' dropdown set to 'Primary HR' with a heart icon, a 'Size' dropdown set to '1x', and a 'Hemo Sweep Speed' dropdown set to '25 mm/s'. At the bottom of the main area are four buttons: 'Update Lead Set', 'ECG Source' with a right arrow, '12 Lead Analysis' with a right arrow, and 'All ECG Waveforms' with a right arrow. A 'Close' button is located in the bottom right corner.

ECG			
ECG	ST	QT	Arrhythmia
Setup Advanced Alarms ECG 1 Lead: II ECG 2 Lead: V1 ECG 3 Lead: aVL Va Lead Position: V1 Beat Volume: 0 Beat Source: Primary HR Size: 1x Hemo Sweep Speed: 25 mm/s Update Lead Set ECG Source > 12 Lead Analysis > All ECG Waveforms > Close			


Skin Preparation

Excessive body hair or skin oil reduces electrode contact with the skin and decreases the quality of electrode signal. When preparing the electrode sites, avoid bones close to skin, obvious layers of adipose tissue and major muscles.

1. Shave any hair from the electrode site.
2. Gently rub the surface of the skin to increase capillary blood flow.
3. Clean the skin with a mild soap and water solution to remove skin oil and dead or abraded skin cells.
4. Dry the skin completely before applying the electrodes.


Lead Placement

3-lead or 5-lead ECG Electrode Placement




IEC Lead	AAMI/AHA Lead	Electrode Placement
R - Red	RA - White	Just below the right clavicle
L - Yellow	LA - Black	Just below the left clavicle
User defined	User defined	For the 5-lead placement, place the precordial electrode according to the physicians preference
N - Black	RL - Green	Lower right edge of the rib cage
F - Green	LL - Red	Lower left edge of the rib cage

3-lead or 5-lead ECG Electrode Placement for the Neonatal Patient



3-lead Placement




5-lead Placement

IEC Lead	AAMI/AHA Lead	Electrode Placement
R - Red	RA - White	Just below the right clavicle
L - Yellow	LA - Black	Just below the left clavicle
F - Green	LL - Red	Lower left edge of the rib cage

IEC Lead	AAMI/AHA Lead	Electrode Placement
R - Red	RA - White	Just below the right clavicle
L - Yellow	LA - Black	Just below the left clavicle
User defined	User defined	For the 5-lead placement, place the precordial electrode according to the physicians preference
N - Black	RL - Green	Lower right edge of the rib cage
F - Green	LL - Red	Lower left edge of the rib cage


6-lead ECG Electrode Placement



6-lead Placement


IEC Lead	AAMI/AHA Lead	Electrode Placement
R - Red	RA - White	Just below the right clavicle
L - Yellow	LA - Black	Just below the left clavicle
Ca/C1 - White	Va/V1 - Brown	4th intercostal space, right sternal border
Cb/C5 - White	Vb/V5 - Brown	Left anterior axillary line at C4/V4 level
N - Black	RL - Green	Lower right edge of the rib cage
F - Green	LL - Red	Lower left edge of the rib cage

10-lead ECG Electrode Placement for Cardiac Monitoring



IEC Lead	AAMI/AHA Lead	Electrode Placement
R - Red	RA - White	Just below the right clavicle
L - Yellow	LA - Black	Just below the left clavicle
N - Black	RL - Green	Lower right dege of the rib cage
F - Green	LL - Red	Lower left of the rib cage
C/C1 - White/Yellow	V/V1 - Brown	4th intercostal space, right sternal border
C2 - White/Yellow	V2 - Brown/Yellow	4th intercostal space, left sternal border
C3 - White/Green	V3 - Brown/Green	Midway between C2/V2 and C4 /4
C4 - White/Brown	V4 - Brown/Blue	5th intercostal space, mid-clavicular line
C5 - White	V5 - Brown/Orange	Left anterior axillary line at C4/V4 level
C6 - White/Purple	V6 - White/Purple	Mid axillary line at C4/V4 and C5/V5 levels

Standard Resting 10-lead ECG Electrode Placement



IEC Lead	AAMI/AHA Lead	Electrode Placement
R - Red	RA - White	Right deltoid or wrist
L - Yellow	LA - Black	Left deltoid or wrist
N - Black	RL - Green	Right thigh or ankle
F - Green	LL - Red	Left thigh or ankle
C/C1 - White	V/V1 - Brown	4th intercostal space, right border of sternum
C2 - White/Yellow	V2 - Brown/Yellow	4th intercostal space, left border of sternum
C3 - White/Green	V3 - Brown/Green	Midway between C2/V2 and C4 /4
C4 - White/Brown	V4 - Brown/Blue	5th intercostal space, mid-clavicular line
C5 - White/Black	V5 - Brown/Orange	Left anterior axillary line at C4/V4 level.
C6 - White/Purple	V6 - Brown/Purple	Mid axillary line at C4/V4 and C5/V5 levels

12 Lead Analysis

To access the 12 Lead Analysis Menu:

1. Select the **HR** parameter window > **ECG > Setup**.
2. Select **12 Lead Analysis**.

To access the 12 Lead Analysis Settings Menu:

1. Select the **HR** parameter window > **ECG > Setup**.
2. Select **12 Lead Analysis**.
3. Select **Settings**.

Performing a 12 Lead Analysis

1. Select the **HR parameter** window.
2. Select **12 Lead Analysis**.
3. Enable **Stabilize Baseline**.
4. Select **12 Lead ECG Now**.


All the waveforms in the 12 Lead Analysis View freeze during the analysis except for the ECG I waveform. Analysis takes less than one second to complete. At that time, the monitor generates a 12 Lead report, saves the report locally, and displays the report on the screen.

The monitor can store up to fifteen 12 lead reports locally.

The 12 Lead Analysis Program


The 12 Lead ECG Analysis program assists the physician in interpreting and measuring the resting ten seconds of ECG data. This program generates a diagnostic textual report on patient's cardiovascular condition. This report can be routed to the MUSE Cardiology Information System via the CARESCAPE Network.

- a. Patient information, including patient Name, MRN, Date, and Time: the report was generated. Under Settings you can find additional details and options, such as: Tech ID, reason for 12 lead, Order number, Auto 12 leads, Diagnostic tools.
- b. Available values including Ventricular Rate, PR Interval, QRS Duration, QT/QTc, and P-R-T Axis.
- c. Diagnostic statements and/ or error messages.
- d. Waveform area.




Note! Complete analysis requires a 10-leadwire cable.

Impedance Respiration




1

ECG **Lead I** for
upper chest breather



2

ECG **Lead II** for
upper abdominal breather



3

RL-LL for
upper abdominal breather

Respiration Lead Placement

- 1. Lead I:** Provides good thoracic (upper chest) breath detection. However, lead I is more susceptible to cardiogenic artifact than the RL-LL vector.
- 2. Lead II:** Provides good thoracic breath detection and upper abdominal (lower chest) breath detection. However, lead II is more susceptible to both cardiogenic and motion (head, neck, or arm) artifact than the RL-LL vector.
- 3. RL-LL:** Vector provides good abdominal breath detection and is not as susceptible to cardiogenic artifact or motion artifact. When monitoring respiration through the RL-LL vector, use a standard 5-leadwire electrode placement, except place the RL electrode on the fifth intercostal space on the right side of the chest. This is also available for the 6- and 10-leadwire sets. It is not available with a 3-lead cable.




Note! *RL-LL vector only applies to PDM module. It is also available with the CS ONE.*

Neonatal Impedance Respiration


Overview

Impedance respiration is measured across the thorax between ECG electrodes. The respiration signal is made by supplying current between the electrodes and by measuring the differential current from the electrodes. The signal measured is the impedance change caused by breathing.


Respiration Lead Placement



ECG **Lead I** for upper chest breather



ECG **Lead II** for upper abdominal breather



RL-LL for upper abdominal breather (for 5-Lead placement)

- 1. Lead I:** Provides good thoracic (upper chest) breath detection. However, lead I is more susceptible to cardiogenic artifact than the RL-LL vector.
- 2. Lead II:** Provides good thoracic breath detection and upper abdominal (lower chest) breath detection. However, lead II is more susceptible to both cardiogenic and motion (head, neck, or arm) artifact than the RL-LL vector.
- 3. RL-LL:** Vector provides good abdominal breath detection and is not as susceptible to cardiogenic artifact or motion artifact. When monitoring respiration through the RL-LL vector, use a standard 5-leadwire electrode placement, except place the RL electrode on the fifth intercostal space on the right side of the chest. This is also available for the 6- and 10-leadwire sets. It is not available with a 3-lead cable.



Note! RL-LL vector only applies to PDM module. It is also available with the CS ONE.

Using Impedance Respiration

If the respiration measurement has not been setup as a configuration setting, it can be turned on.

1. Select the **Impedance Respiration** parameter window.
2. Select the **Setup** tab.
3. Select **Respiration Measurement > On**.



Note! When the respiration measurement is on, the message **Imped. on** is displayed in the ECG1 waveform field. You can also perform the following tasks from the **Impedance Respiration Setup** menu:

4. Enable **Autosize Waveform** to automatically size the current waveform to fit the available space.
5. Adjust the respiration waveform **Size** (the greater the value, the larger the waveform size).
6. Adjust the respiration waveform **Sensitivity** (the lower the value, the greater the sensitivity).
7. Select the **Imped. Resp Smoothing** check box to enable smoothing (deselect it to disable smoothing). When respiration smoothing is enabled, 10 seconds of respiration rate values are averaged.
8. Select the **Insp./Exp. Markers** check box to turn the inspiration and expiration waveform markers on (deselect it to turn them off).
9. Adjust the respiration waveform **Resp Sweep Speed** by selecting a value from the list (lower the value the slower the sweep speed).
10. Select the respiration **I, II or RL-LL** from the respiration lead list. Lead selections are presented as graphical icons.
11. Select the **Relearn Respiration** to relearn the respiration pattern (if the patient's breathing pattern changes after the initial learning process has taken place).

Impedance Respiration

Setup **Alarms**

Resp Measurement: **On** ▼

Autosize Waveform

Size: 40% ▼

Sensitivity: 40% ▼

☒ Imped. Resp Smoothing

Resp Sweep Speed: 6.25 mm/s ▼

☒ Insp./Exp. Markers

Relearn Respiration

Caution: This measurement may affect Rate Responsive Pacemakers.

Close

03 NIBP



Manual NIBP Measurement from the Main Menu

1. Start the measurement by selecting **NIBP Start** from the Main Menu area. (After **NIBP Start** is selected the menu tab becomes **NIBP Stop**.)
2. Stop the measurement by selecting **NIBP Stop** from the Main Menu area.

Automatic NIBP Measurement from the Main Menu

The NIBP Auto mode initiates repeated measurements for the selected Cycle Time. There will be at least a 30 second delay between two consecutive NIBP measurements during auto cycling.

3. Select **NIBP Auto Start**. (After **NIBP Auto Start** is selected the menu tab becomes **NIBP Auto Stop**.)
4. Stop the measurement by selecting **NIBP Auto Stop**.

The NIBP Setup Menu

To Access the NIBP Setup Menu:

1. Select the **NIBP** parameter window.
2. Select the **Setup** tab.

You can perform the following tasks from the NIBP Setup Menu:

3. Start automatic, manual or Stat NIBP measurements.



Note! STAT NIBP Mode

initiates a continuous cycle of measurement for five minutes.

4. Change the cycle time for automatic NIBP measurements.
5. Adjust the tone volume for a completed NIBP measurement.
6. Select the display format (large Systolic/Diastolic or large Mean values)
7. Select the cuff size.
8. Select the initial NIBP cuff inflation pressure automatically based on the cuff size.
9. Select the target NIBP inflation pressure for the first NIBP measurement.

The screenshot shows the 'Non-Invasive Blood Pressure' setup menu with two tabs: 'Setup' (active) and 'Alarms'. The 'Setup' tab contains the following controls:

- NIBP Auto**: A toggle switch.
- Start Cycling**: A large button with a horizontal line through it.
- Start Manual NIBP**: A button.
- Start STAT**: A button.
- Cycle Time**: A dropdown menu set to '15 min'.
- Completed NIBP Volume**: A volume control slider set to '3'.
- Cuff Size**: A dropdown menu set to 'Adult'.
- Display Format**: A dropdown menu set to 'Sys/Dia (Mean)'.
- Auto Initial Inflate**: A checkbox that is checked.
- Init. Pressure mmHg**: A numeric input field set to '135'.
- Previous Menu**: A button at the bottom right.

04 Pulse Oximetry


Overview

There are three supported pulse oximetry technologies:

1. Masimo SET
2. Nellcor OxiMax
3. GE TruSignal (CARESCAPE ONE)

SpO₂ monitoring starts automatically when the monitor detects an SpO₂ signal.

It is possible to measure SpO₂ from two different measurement sources simultaneously. The primary SpO₂ source is labeled SpO₂ and the secondary SpO₂ source is labeled SpO₂(2).



GE oximeter

The SpO₂ Setup Menu

To access the SpO₂ Setup Menu:

1. Select the **SpO₂** parameter window.
2. Select the **SpO₂** or **SpO₂(2)** tab.
3. Select **Setup**.

You may be able to perform the following tasks from the SpO₂ Setup Menu:



Note! The selections available from the SpO₂ menu will vary depending on the pulse oximetry technology and acquisition module.

1. Adjust the SpO₂ size from the **Scale** list.
2. Change the SpO₂ waveform sweep speed by selecting a numeric value from the Hemo Sweep Speed list. (The smaller the value, the slower the sweep speed.) Note that changing one parameter sweep speed will change all parameter sweep speeds, except for impedance respiration.
3. Select the heart rate source from the **Primary HR Source list** (ECG, SpO₂ or Invasive Pressure).
4. Show the SpO₂ pulse rate by checking **Show Pulse Rate**.
5. Select the SpO₂ **Averaging** time.
6. Adjust the SpO₂ **Beat Volume** by using the arrows.
7. Select the beat source from the **Beat Source** list.

The screenshot shows the 'SpO2' menu with two tabs: 'SpO2' and 'SpO2(2)'. The 'SpO2' tab is active. Inside, there are two sub-tabs: 'Setup' and 'Alarms'. The 'Setup' sub-tab is active. The 'Setup' menu contains the following settings:

- Scale:** A dropdown menu with 'AUTO' selected.
- Hemo Sweep Speed:** A dropdown menu with '25 mm/s' selected.
- Primary HR Source:** A dropdown menu with 'AUTO' selected.
- Show Pulse Rate:** A checkbox that is currently unchecked.
- Averaging:** A dropdown menu with '12 s' selected.
- Beat Volume:** A control with a speaker icon, a volume indicator set to '0', and a mute/unmute icon.
- Beat Source:** A dropdown menu with 'Primary HR' selected, accompanied by a heart icon.
- Variable Beat Tone Source:** A dropdown menu with 'Off' selected.

A 'Close' button is located at the bottom right of the menu.

GE TruSignal technology and sensors

You may be able to perform the following tasks from the SpO₂ Setup Menu for Masimo or Nellcor OxiMax technology and sensors.

Masimo technology and sensors

1. Adjust the SpO₂ size from the **Size** list.
2. Change the SpO₂ waveform sweep speed by selecting a numeric value from the Hemo Sweep Speed list. (The smaller the value, the slower the sweep speed.) Note that changing one parameter sweep speed will change all parameter sweep speeds, except for impedance respiration.
3. Select the heart rate source from the **Primary HR Source list** (ECG, SpO₂ or Invasive Pressure).
4. Show the SpO₂ pulse rate by checking **Show Pulse Rate**.
5. Select the SpO₂ **Averaging** time.
6. Select the SpO₂ sensor **Sensitivity** level.
7. Adjust the SpO₂ **Beat Volume** by using the arrows.
8. Select the beat source from the **Beat Source** list.

Nellcor OxiMax technology and sensors


1. Adjust the SpO₂ size from the Size list.
2. Change the SpO₂ waveform sweep speed by selecting a numeric value from the Hemo Sweep Speed list. (The smaller the value, the slower the sweep speed.) Note that changing one parameter sweep speed will change all parameter sweep speeds, except for impedance respiration.
3. Select the heart rate source from the **Primary HR Source list** (ECG, SpO₂ or Invasive Pressure).
4. Show the SpO₂ pulse rate by checking **Show Pulse Rate**.
5. Show the Saturation seconds in the SpO₂ parameter window by checking **Show Sat. Seconds**.
6. Set the **Saturations Seconds** threshold.
7. Select the Response time to Normal or Fast.
8. Adjust the SpO₂ **Beat Volume** by using the arrows.
9. Select the beat source from the **Beat Source** list.

05 Invasive Pressure Monitoring

Invasive Pressure Connections


Invasive Pressure equipment to patient connection with PDM

1. Module with invasive pressure measurement capability.
2. Fluid bag with pressure infusor.
3. Transducer setup.
4. Invasive blood pressure adapter cable; single or dual cable (optional).



Invasive Pressure equipment to patient connection with CARESCAPE ONE

1. CARESCAPE ONE
2. Fluid bag with pressure infusor.
3. Transducer setup.
4. CARESCAPE IP Micro parameter module.



Zeroing a Transducer



1. Level the transducer following your care unit's policy (usually level of the phlebostatic axis).
2. Close the transducer stopcock to the patient and open the venting stopcock to air.
3. If the pressure line you are trying to zero does not have the transducer open to air, the message "Pressure Sensed" displays.
4. You can zero all connected pressure transducers simultaneously by selecting **Zero All Pressures** from the monitor's main menu or from the remote control, or you can zero a single active pressure transducer by selecting the invasive pressure parameter window and then selecting **Setup > Zero**.
5. Check that a zero reference has been established. Watch the pressure parameter window for messages.
6. Close the venting stopcock to air and open the transducer stopcock to the patient.
7. Check that pressure numerics display on screen.

The Invasive Pressure Setup Menu

1. To access the Invasive Pressure Setup menu, select the **Invasive Pressure** parameter window. Select a slot or port (1-8) that you will be monitoring, and then select the **Setup** tab.
2. Select a channel label from the **Label** list.
3. Set the waveform scale with the **Scale** arrows. The larger the scale value, the smaller the waveform size.
4. Select **Optimize Scale**. The Scale selection will now show the automatic limit range.
5. Select the heart rate source from the **Primary HR Source** list (ECG, SpO₂ or Invasive Pressure).

The screenshot displays the 'Invasive Pressures' configuration interface. At the top, a row of tabs allows selection of different pressure sources: Art 1, CVP 2, Fem 3, PA 4, P5, P6, P7, and Art 8. Below this, a secondary row of tabs includes 'Setup' (the active tab), 'Advanced', 'Art 1 Alarms', and 'HR Alarms'. The 'Setup' tab contains several configuration options. On the left, there are three callout boxes: box 2 points to the 'Label' dropdown menu which is set to 'Art'; box 3 points to the 'Scale' control, which shows 'mmHg' and a range of '0-200'; box 4 points to the 'Optimize Scale' button. On the right side of the 'Setup' tab, there are three callout boxes: box 5 points to the 'Primary HR Source' dropdown menu set to 'AUTO'; another box points to the 'Show Pulse Rate' checkbox, which is currently unchecked; and a third box points to the 'SPV Source' dropdown menu set to 'Art 1'. Below these, there are buttons for 'Measure SPV', 'Zero', 'Zeroed', and 'Standby Art 1'. At the bottom right of the screen is a 'Previous Menu' button.

06 Airway Gases

Airway Gas Modules and Connectors

CARESCAPE Respiratory Module




Note! E-sCOV is not suitable for use with patients weighing less than 5 kg (11 lbs).

1. Patient Spirometry keys.
2. Water trap release/locking latch.
3. Gas sample, sampling line connector on the water trap.
4. Water trap container.

Make sure you use a water trap that is compatible with the module:

- **CARESCAPE Respiratory Modules:** D-Fend Pro or D-Fend Pro+
- **E-miniC:** Mini D-Fend
- **In Anesthesia:** Replace the D-Fend Pro or Mini D-Fend water trap when the message **Replace Water Trap** appears. The maximum lifetime of a water trap is two months.
- **In Critical Care:** It is recommended to replace the D-Fend Pro + or Mini D-Fend water trap for each new patient, when the message **Replace Water Trap** appears, or every 24 hours.

5. Connectors for Patient Spirometry tubes.
6. Gas exhaust, connector for the gas exhaust line.




E-miniC



Note! E-miniC is not suitable for use with patients weighing less than 5 kg (11 lbs).

1. Water trap latch.
2. Sampling line connector.
3. Mini D-fend water trap with a washable container.
4. Sample gas outlet.




CARESCAPE CO₂ Measurement

Equipment Connection with CARESCAPE CO₂

For intubated and non-intubated patients.

- Connect the LoFlo Sidestream Module to CARESCAPE CO₂.
- Connect CARESCAPE CO₂ to the CS ONE.
- Connect the sample line to the LoFlo Sidestream Module.
- When the LoFlo Sidestream module is connected it will zero and go through a warm-up procedure automatically.


1. CARESCAPE ONE
2. CARESCAPE CO₂
3. Exhaust port
4. LoFlo Sidestream Module
5. Accessories, i.e. sample lines, nasal cannulas, etc.




About Sidestream

Sidestream kits are intended for monitoring CO₂ of intubated and non-intubated patients. Following are things to consider when selecting a sidestream kit for a non-intubated patient.

- Select the kit that has the best fit and most comfort for the patient. For example, the kit labeled as Pediatric may have the best fit and be more comfortable on a small-statured adult than the adult cannula kit. The cannula kit are equally functional regardless of which is used.
- Verify that the cannula kit is clean, dry and undamaged. Replace the cannula kit, if necessary.
- Insert the sample cell into the sample cell receptacle. You will hear a click when the sample cell is properly inserted.
- When the LoFlo Sidestream module is connected it will zero and go through a warm-up procedure automatically.
- Place the nasal cannula kit onto the patient.



Airway Gas Setup



Note! Check that the sample line is connected to the water trap before connecting the module to the monitor or turning on the monitor.

1. Make sure that the water trap container is empty and properly attached.
2. Connect the gas sampling line to the sampling line connector on the water trap.
3. Connect the sample gas outlet to gas scavenging if N₂O or volatile agents are used. This is done only when anesthetic agents are used, typically in the OR. This not something that is done in the ICU.
4. Turn on the monitor or connect the module to the monitor. The monitor performs a self-check for the module when the module is connected. Automatic agent identification is activated in those modules that have this feature.
5. Wait until the message **Calibrating** disappears.
6. Connect the sampling line to the airway adapter or the airway adapter to the ventilator circuit. Position the adapter with the sampling port upwards to minimize the amount of condensed water possibly entering the sampling line.
7. Check that the airway adapter connections are tight and that the adapter is operating properly.



Note! To minimize the amount of dust drawn into the gas sampling system, always keep the water trap connected to the module. See the **Starting and Stopping the Pump** section that follows.

Starting and Stopping the Pump



Note! *Pertains to CARESCAPE respiratory modules only.*

1. Select a gas related parameter window.
2. Select the **CO₂ tab > Setup > Measurement Source > Gas Module (Mod)**.
3. Start the pump by selecting **Start Gas Module pump**. The pump starts and the selection changes to **Stop Gas Module pump**.
4. You can stop the pump by selecting **Stop Gas Module pump**.

The text Turning off Gas Module pump appears below the selection. Once the module has stopped the pump, the selection changes back to **Start Gas Module pump**.


07 Alarms

Alarm Priority Levels

Alarm Priority Levels

Physiological and technical alarms are categorized by priority level:

- 1. **High priority** alarms require an immediate response.
- 2. **Medium priority** alarms require a prompt response.
- 3. **Low priority** alarms require you to be aware of the condition.
- 4. **Informational priority** messages provide information you should know.



Pausing Alarms



Selecting the Pause Audio key results in different alarm behaviors depending on whether the alarms are active and/or latched or not. Acknowledging or pausing audio alarms does not affect other alarm indicators. They will still continue indicating alarms.

When the monitor is on the network, alarms can also be paused and acknowledged at the central station. The ability to pause alarms from the central station is defined by the unit and is a password protected function.

Adjusting the Alarm Volume

The selections in the Alarm Setup menu vary according to what has been configured in the Care Unit Settings (password protected).

1. Select **Alarm Setup** from the main menu.
2. Select the **Audible & Visual** tab.
3. Adjust the volume according to what is available in the menu:
 - a. Adjust the **Alarm Volume** value. This is the volume for all alarms.
 - b. Adjust the **Alarm Volume for**: separately for **High & Medium Priority** and **Low Priority**.

The lower the number, the quieter the alarm volume.



Note! That the minimum allowed volume levels are set in the **Care Unit Settings**.

Note! Pause All Audio for 5 min: The Pause All Audio for 5 min selection includes Alarm Breakthrough alarms. This is intended to be used when there are caregivers present at the bedside for emergency situations like cardiac arrests.

The screenshot shows the 'Alarms Setup' interface. At the top, there are five tabs: 'Alarm Limits', 'Alarm Priorities', 'Arrhythmia', 'Audible & Visual' (which is selected and highlighted in blue), and 'Pause Monitoring'. A blue circle with the number '2' points to the 'Audible & Visual' tab. Below the tabs, there are several controls. On the left, there are two buttons: 'Activate All Audible Alarms' (highlighted with a yellow border) and 'Pause All Audio for 5 min'. Below these is the text 'Set audible alarms permanently Off'. Further down are five radio button options: 'None' (selected), 'Apnea Audio Off', 'ECG Audio Off', 'Apnea & ECG Audio Off', and 'All Alarms Audio Off'. On the right side, there is a section titled 'Alarm Volume for:'. It contains three sub-sections: 'High & Medium Priority' with a bell icon and a volume slider set to 7; 'Low Priority' with a bell icon and a volume slider set to 7; and 'Alarm Light %' with a dropdown arrow and a value of 100. A blue box labeled '3a' encompasses the 'High & Medium Priority' and 'Low Priority' sections. A blue box labeled '3b' encompasses the 'Low Priority' section. A 'Close' button is located at the bottom right of the screen.

Changing an Alarm Limit

Setting Parameter Alarm Limits

Parameter alarm limits may be set in the Alarms Setup menu, or in the parameter menu's own Alarm tab. Alarm limits should not be set beyond reasonable physiological boundaries for maintaining patient safety. Setting limits outside of reasonable boundaries would cause the alarms to be ineffective.

1. Select **Alarm Setup** from the main menu.
2. Select the **Alarm Limits** table.
3. Select a parameter label.
4. That parameter menu's Alarm tab opens, where you can select the alarms on or off, and set their limits.


The screenshot displays the 'Alarms Setup' interface with a dark theme. At the top, there are five tabs: 'Alarm Limits' (highlighted in blue), 'Alarm Priorities', 'Arrhythmia', 'Audible & Visual', and 'Pause Monitoring'. Below the tabs is a grid of parameter settings. The first column, 'HR ECG', is highlighted with a yellow border. Each parameter cell shows a current value, a red upper limit line, and a green lower limit line. At the bottom of the grid are four buttons: 'Auto Limits', 'Default Limits', 'Undo Settings', and a right-pointing arrow. A 'Close' button is located at the bottom right of the screen.

HR ECG	HR ECG Crit.	ST Lat	ST Inf	ST Ant	PVC	SpO2
150	180	2.0	2.0	2.0	10	100
50	40	-2.0	-2.0	-2.0	0	90
SpO2 Crit.	SpO2 (2)	NIBP Sys	NIBP Mean	NIBP Dia	Art 1 Sys	CVP 2 Mean
100	100	180	140	100	180	15
88	90	80	60	40	80	0


Auto Limits Default Limits Undo Settings ►

Close

Critical Alarms for HR and SpO2



The SpO2 Alarms configuration screen shows three main sections: SpO2, SpO2 Critical, and PR SpO2. Each section has a status indicator (OFF), a numerical value (92, 88, 50), and a range (30-100%). The SpO2 section is highlighted with a yellow box around the 92 value. Below the sections are buttons for 'Alarm On', 'Critical Alarm On', and 'Alarm Off'. At the bottom are 'Delays Info' and 'Close' buttons.



The ECG HR/PR Alarms configuration screen shows a table of alarm settings. The 'Critical Tachy/Brady' row is highlighted in blue. The 'Alarm on' column has checkboxes for each row. The 'Primary HR Source' is set to ECG.

Source	Value /min	High	Low	Alarm on
HR(ECG) Tachy/Brady	60	150	50	☒
Critical Tachy/Brady	60	180	40	☒
PR(SpO2)	60	150	50	☐
PR(SpO2(2))	60	150	50	☐
PR(Art 1)	60	150	50	☐
PR(Fem 3)	60	150	50	☐

Primary HR Source: ECG

You can configure two different levels of alarms for HR and SpO₂/SpO₂(2) alarms: basic and critical. This enables an alarm to start with the basic alarm value and priority level and it changes to a high alarm priority once the set critical alarm value is breached. The critical alarm is always indicated as a high priority while the basic alarm priority can be adjusted.

Always check that you are configuring the correct alarm. Parameter Limit alarm settings are available for the following settings: HR Tachy/Brady, Critical Tachy/Brady, SpO₂ High/Low, Critical SpO₂ High/Low.

You must always set the critical alarm level limits so that they are outside the basic alarm limits. If you try to set them inside the basic alarm limits, the basic alarm limits will also change at the same time.

Breakthrough Alarms

Breakthrough alarms will break through (interrupt) an **All Alarms Audio Off** or a 2 or 5 minute audio alarm pause condition.

The **FiO₂ low**, **EtO₂ low**, **FiN₂O high**, and **Ppeak high** alarms will always break through when escalated to or activated at high priority alarm conditions regardless of the **All Alarms Audio Off** selection or any alarm pausing.

The following alarms will breakthrough when escalated to or activated at high priority alarm condition regardless of the 2 to 5 minute audible alarm pause: **Asystole**, **V Fib/V Tach**, **V Tach** (in all software packages), and **Brady** (in the NICU software packages only).

Setting Arrhythmia Alarm Priorities

1. Select **Alarms Setup** from the main menu.
2. Select the **Arrhythmia** tab.
3. Select **Lethal**, **Ventricular** or **Atrial**.
4. Select the **Alarm Priority**, **Create Snapshot** and/or Print on alarm options per arrhythmia.
5. V Tach criteria can be configured if enabled in the password protected profile settings.
 - a. Minimum HR: Enables you to set the minimum heart rate required for the set run duration of the V Tach.
 - b. Event Duration: Enables you to set the minimum run duration/length of time (seconds) required for a V Tach event to alarm.



Note! Although the informational, low and medium priority alarm settings may be allowed for V Tach, the alarm will always be a high priority if the V Tach duration is more than 30 seconds, the HR is higher than the set HR high limit and the HR exceeds 150 beats/minute. In the NICU software the high priority is enforced at 180 beats/minute.

V Tach Criteria

V Tach alarms when the following criteria are met

Minimum HR/min: 100

Event Duration: OFF

AND

Minimum estimated PVC count for V Tach alarm: 6.

Previous Menu

08 Trends


Viewing Numeric Trends

1. Select **Trends** from the main menu area at the bottom of the display.
2. Select **Numeric** from the View list.
 - To see other parameters, select their tabs in the **Numeric** trend view
 - To see more numeric trend data, use the cursor to scroll the data in horizontal direction
 - To change the time interval, select a new value from the **Time Interval** list



Viewing Graphic Trends

1. Select **Trends** from the main menu area at the bottom of the display.
2. Select **Graphic** from the View list.
 - To see more parameters, select tabs 1 to 4
 - To see numeric values of a certain time, move the cursor to that point of time. The numeric values are displayed next to the cursor
 - To change the time scale, select a new value from the **Time Scales** list
 - To change the trend scale, select **Trend Scales** and make the appropriate adjustment



Managing Snapshots

A snapshot is a set of measured data saved from a certain moment of time. Snapshots can contain waveform clips and graphic trends. You can take up to 400 snapshots.

Creating Manual Snapshots

You can create a snapshot manually by selecting **Freeze/ Snapshot** from the main menu area at the bottom of the display.

The monitor saves the image of pre-configured waveforms or trends at that moment in time. When a snapshot is taken manually, it is automatically numbered.


Creating Automatic Snapshots

You can select alarms that will automatically create a snapshot independent of their alarm priority.

1. Select **Trends** from the main menu area at the bottom of the display.
2. Select **Snapshot** from the View list.
3. Select **Snapshot Setup**.
4. Select which alarms will automatically create a snapshot.

Viewing Snapshots

1. Select **Trends** from the main menu area at the bottom of the display.
2. Select **Snapshot** from the View list.
 - To change the Snapshot time scale, select a new value from the **Time Scale** list
 - To change the trend scale, select **Trend Scales** and make the appropriate adjustment



Managing Events

Events are timestamps that are shown in their own list. An event is created automatically upon an alarm. An event records the time of and reason for its creation. Some events may also record a snapshot. Manually created events contain only the time and a manually added reason for the event. You cannot configure the Event trend pages.

An event is created automatically from:

- Medium and high priority physiological or technical alarms
- Low priority alarms that have a snapshot
- Manually created snapshots or ST snapshots

Viewing Events

The **Event** trend view shows event data on horizontal axis and time on vertical axis. The top of the view shows the highest priority realtime waveform and the bottom of the view shows a sample waveform if an event has a snapshot.

1. Select **Trends** from the main menu area at the bottom of the display.
2. Select **Event** from the View list.
 - The **Priority** column shows an alarm priority symbol for events created automatically from an alarm.
 - The **Event** column shows the reason the event was created for. If the event was created automatically, the alarm message is shown. If the event was created manually, a possible manually added text is shown.
 - If there is a manual annotation added to the event, this text is shown in quotation marks and also includes the prefix **NOTE**.
 - The Snapshot column shows a snapshot symbol if there is a snapshot attached to an event.

The screenshot displays the 'Trends' interface with the 'Event' view selected. The interface includes a sidebar with options like 'View', 'Sort by', 'Create Event', 'Annotate Event', 'Delete Event', 'Show Deleted', 'Print Page', and 'Close'. The main area shows a table of events with columns for Time, Priority, Event, and Snapshot. The table lists several events, including 'Brady' and 'Tachy' events with associated timestamps and priority symbols. A waveform is visible at the top and bottom of the screen.

Time	Priority	Event	Snapshot
13:19:23 09 Feb 2018	High	Brady	Camera icon
13:18:51 09 Feb 2018	High	Tachy	Camera icon
18:56:42 08 Feb 2018	Low	Brady	Camera icon
18:55:57 08 Feb 2018	Low	Tachy	Camera icon
17:52:46 08 Feb 2018	Low	SpO2 low	Camera icon
15:40:33 08 Feb 2018	Low	Brady	Camera icon
12:36:07 08 Feb 2018	Low	SpO2 low	Camera icon

09 Troubleshooting

ECG Troubleshooting

ECG

Problem	Solution
ECG signal is noisy or no QRS is detected	• Ensure that the patient is not shivering. • Select the correct filter by selecting the HR parameter window > Advanced > Waveform Filter. • Check the electrode quality and positioning. Do not place electrodes on body hair, bones close to skin, layers of fat and major muscles. Pre-gelled electrodes are recommended. • Change the lead in ECG1 to the best available signal and consider using the Single lead mode. • Consider using ECG Size > 2x. • Try an alternative location for the Va lead to improve signal quality. In some cases, like if the patient has a significant heart failure, changing for example from V5 to V1 can result in a considerable difference in the signal amplitude. • Remove the ECG cable from the module and reinsert it.



Pacemaker

Problem	Solution
How does activating pacemaker detection impact monitoring?	- Beats that would otherwise be classified as ventricular are instead classified as V-paced if a ventricular pacemaker event is detected. - Residual pacemaker energy that might otherwise appear in the ECG is removed, and a pacemaker enhanced spike is placed in the ECG. - On the ECG waveform, pacemaker detection is indicated by uniform, upright pacemaker enhancement spikes in the ECG data, both displayed and graphed.
How can pacemaker detection be improved?	Use additional alarms, including: <i>Pause</i> and <i>HR low, Brady</i> Possible problems include: - Heart rate double counting. - Inaccurate alarms for low heart rate or asystole. - Pacemaker spikes not recognized by the software. - False PVC detections and arrhythmia alarms. Possible solutions include: - Relearn arrhythmia. - Try an alternate electrode placement. - Try single-lead analysis, if available. - E-modules, CARESCAPE B Series Monitor, CARESCAPE ONE, and Combination monitoring: Switch to another pacemaker detection mode.
Why is the monitor double-counting the heart rate, alarming for a low heart rate, or not detecting pacemaker spikes?	The monitor is not detecting pacemaker activity. Causes may include: - PDM, CARESCAPE B Series Monitor, CARESCAPE ONE, and Combination monitoring: The pacemaker detection program is turned off. Turn it on, reprep the skin and reposition the electrodes if necessary. Relearn ECG. - The pacemaker signal is too weak for the monitor to detect. - The ECG signal is too weak for the monitor to detect. - The monitor is detecting atrial pacemaker artifact or non-QRS features as beats. If the monitor is alarming for low heart rate or asystole, assess the QRS amplitude: - View all ECG leads to assess the amplitude of the QRS complexes. To ensure correct HR readings, a 0.5 mV QRS amplitude is recommended for a normal ECG signal. If the QRS amplitude drops below 0.5 mV or an abnormal QRS width occurs (more than 120 ms), QRS detection may be reduced, leading to false asystole alarms. - If necessary, reprep the skin and reposition the electrodes. - Relearn ECG.

Arrhythmia

Problem	Solution
Why is the monitor alarming for asystole, bradycardia, pause, or inaccurate heart rate when a visible QRS waveform is present?	The monitor may not be detecting sufficient QRS amplitude in all analyzed leads. Multiple leads are used for arrhythmia processing. 1. Assess the patient. 2. Check the ECG signal acquired from the patient. 3. View all ECG leads to assess the amplitude of the QRS complexes. To ensure correct HR readings, a 0.5 mV QRS amplitude is recommended for a normal ECG signal. If the QRS amplitude drops below 0.5 mV or an abnormal QRS width occurs (more than 120 ms), QRS detection may be reduced, leading to false Asystole alarms. 4. Relearn arrhythmia. It is important to relearn the patient's ECG pattern any time the electrode configuration is adjusted. If the problem continues, switch to the ECG lead with the greatest amplitude, display that lead, then switch to single lead analysis so all arrhythmia interpretations are based on this single ECG lead.
How does the IntelliRate algorithm impact an Asystole alarm with a QRS waveform?	IntelliRate will report Asystole when the following conditions are met: • The ECG HR has been valid and has changed by 1/min or less during the previous 30 seconds . • The invasive pressure pulse rate has been valid for the previous 60 seconds, and it has been 0/min during the previous 30 seconds. • The mean arterial blood pressure of the invasive pressure pulse rate source is below the userselected limit. • The SpO₂ parameter if available does not indicate beat detections in the previous 30 seconds.
Why is the monitor calling V Tach when the patient is not in V Tach?	The monitoring system may be detecting a wider QRS complex or artifact in some of the analyzed ECG waveforms. In addition, the V leads may be exhibiting polarity changes, which may occasionally cause an inaccurate call. 1. Assess the patient. 2. Check the ECG signal acquired from the patient. • View all ECG leads to assess the width of the QRS complexes in the analyzed leads. • If artifact exists in any of the analyzed leads, reprep the patient's skin, replace electrodes, and adjust the electrode placement. • It may be beneficial to move V lead electrodes chest lead) to alternate precordial electrode placements to improve detection. 3. Relearn arrhythmia. It is important to relearn the patient's ECG pattern any time the electrode configuration is adjusted. If the problem continues, determine the lead with the narrowest QRS complex, display that lead, then switch to single lead analysis so all arrhythmia interpretations are based on this single ECG lead.

Respiration Troubleshooting

Problem	Solution
What can I do if the respiration measurement fails?	• Check electrode quality and positioning. • Adjust the breath detection sensitivity. During ventilator-supported breathing, the respiration calculation may count only ventilator-produced inspirations and expirations. • Other electrical devices may interfere with the measurement.
Why does the waveform have a combination of shallow and deep breaths, but the monitor is not detecting the shallow breaths?	If the detection sensitivity threshold is set too high, shallow breaths will not be detected, as shown in the following example of incorrect detection (1 = breath).  • Decrease the detection sensitivity percentage until the markers correctly identify each inspiration and expiration or set to AUTO (PSM). If the detection mode is AUTO, the grid lines represent the minimum limits. The limits in use may be a larger range. The following is an example of correct detection.  Respiration detection is not dependent on the size of the waveform. Size is for visual purposes only.
Why is the monitor detecting cardiac artifact as breaths?	The breath detection threshold is too low (1 = breath, 2 = artifact). The following is an example of incorrect detection.  • Increase the detection sensitivity percentage until the markers correctly identify each inspiration and expiration. The following is an example of correct detection. 

SpO₂ Troubleshooting


Problem	Solution
SpO₂ signal is poor	• Check the sensor and sensor position. • Make sure the patient is not shivering, moving, or does not have tremors. • The patient's pulse may be too low to measure.
Unable to adjust alarm limits	The alarm limits are not adjustable when the measurement source is from an external device connected to the Unity Network ID connectivity device.
Deactivated SpO₂ probe off alarm keeps alarming when the sensor is disconnected from the patient.	• Ensure that the sensor is protected from ambient light.
Why does the pulse oximeter sometimes read differently than a blood gas analyzer?	Blood gas analyzers calculate the O ₂ saturation based on normal values for pH, PaCO ₂ , Hb, temperature, etc. (i.e., a normal oxyhemoglobin dissociation curve). Depending on the patient's physiologic and metabolic status, this curve and all values may be shifted away from normal. Thus the oximeter, which measures O ₂ saturation, may not agree with the blood gas.
What effect can ambient light have on pulse oximetry monitoring?	Light sources such as surgical lamps, biliubin lamps, fluorescent lights, infrared heating lamps, and sunlight can cause poor waveform quality and inaccurate readings. Error messages are possible. Shielding the sensor with opaque tape, the posey wrap, or other dark or opaque material can increase oximetry accuracy, verified by good waveforms and signal strength.
What does motion artifact look like, what problems can it cause and how can it be corrected?	Electrosurgical interference is most obvious on the displayed waveform. It is a very spiky, erratic looking waveform caused by the electrosurgical unit's overwhelming interference. It can result in grossly inaccurate pulse oximeter results. Electrosurgical interference can be minimized by: • Making sure the pulse oximeter sensor is as far away from the return pad and operating site as possible. • Making sure the sensor is not between the return pad and operating site. • Keeping the power cord and sensor cable away from the power cord of the electrosurgical unit. • Plugging the electrosurgery unit into a separate set of outlets from the monitor.

NIBP Troubleshooting

Problem	Solution
NIBP measurement does not work or the values seem unstable.	• Check that the cuff tubing is not bent, stretched, compressed, or loose. • Check the cuff position and cuff tube connection. • Prevent motion artifact. • Use NIBP cuffs of correct size.
NIBP measurement does not start.	• Ensure that the cuff size has been selected.
Why is the monitor re-inflating the cuff automatically?	The cuff target pressure must be higher than the patient's systolic pressure to obtain an accurate systolic and diastolic measurement. If a systolic blood pressure cannot be found, the monitor searches for a systolic reading by re-inflating the cuff at a higher pressure. This systolic search may occur once per NIBP determination cycle. During a systolic search, the maximum cuff inflation pressure will not exceed the normal pressure range of the cuff. For more information, refer to the technical specifications. A control measurement may be taking place. If the new value (the control measurement) also exceeds the alarm limits the alarm priority escalates to medium. In Manual mode and STAT mode there are at least four seconds between the first measurement and the control measurement for Adult and Child cuffs, eight seconds for Infant cuffs. In Auto mode there are at least 30 seconds between the first measurement and the control measurement.

10 CARESCAPE ONE

Overview




CARESCAPE ONE • Front View

The CARESCAPE ONE (CS ONE) can be used as an accessory to the B850/B650/B450 monitor with version 3.1 software, or as a standalone monitor. As a standalone monitor, the CS ONE provides ECG, ST segment, arrhythmia detection, impedance respiration, SpO₂, non-invasive blood pressure, invasive pressure, temperature, and CO₂ airway gas parameter acquisition and monitoring.

The CS ONE includes an integrated display, touchscreen, alarm light, user input buttons, and an audio subsystem. Realtime physiological parameter measurements and waveforms are displayed on the integrated display. Visual alarms are conveyed using both the integrated display and alarm light. Audible alarms are conveyed using the integrated audio subsystem. User input for configuration control and interaction is provided via the touchscreen. In addition, both a power button and touchscreen lock button are provided.

CARESCAPE ONE System Components

Your system may not include all these components. Consult your local representative for the available components.



CARESCAPE ONE Front, Back, and Side Views

Front and Back Views

1. Alarm light area (blue, yellow, or red).
2. Audio alarm paused/off indicator (blue).
3. Battery status indicator (yellow or green).
4. Power on/standby button.



On/standby (yellow): The yellow on/standby key should be lit when one of the following occurs:

- The device is connected to an external power source and the device is turned off
- The device is booting up or is in standby





On/standby (green): The green on/standby key indicates the device is ready for operation.




On/standby (unlit): The unlit on/standby key indicates there is no power applied to the device.

5. Screen lock/unlock button.
6. Pull tab for removing the acquisition platform from the docking mount.
7. Battery door.
8. Docking interface connector to the CARESCAPE Dock F0 or Mini Dock.



Side View

1. CARESCAPE Parameter connectors.
2. Analog out / Defibrillator synchronization connector. Connect IEC 60601-1 compliant medical devices to this port only.
3. NIBP hose connector.




Docking

Connecting the CARESCAPE ONE to a Dock

To connect the CS ONE to a dock:


1. Align the CS ONE with the dock rails.
2. Push the CS ONE into the dock until it stops.
You will hear a click when it is fully docked.



Disconnecting the CARESCAPE ONE from a Dock

To disconnect the CS ONE from a dock:

1. Pull the tab out and slide the CS ONE out of the dock rails.
2. Hold onto the CS ONE to make sure it does not drop when it comes out.



Turning the CARESCAPE ONE On/Off

Turning On the CARESCAPE ONE

CS ONE can operate via battery power using its own battery, or via DC power when connected to a powered docking station.

1. Ensure all cables are properly connected.
2. Press the **on/standby** button on the handle side of the CS ONE. The welcome screen will appear.

Turning Off the CARESCAPE ONE

1. Ensure that the patient has been discharged/the case has ended.
2. Press the **on/standby** button for approximately 1 second.

The **Powering Down** window appears to confirm you want to shut down the CS ONE.

3. Press the **on/standby** button again within 10 seconds to shut down the CS ONE.

If you do not press the on/standby button again within 10 seconds, the CS ONE remains on and the **Powering Down** window is removed from the display.



Inserting and Removing a Battery


Caution: To prevent loss of monitoring, or the CS ONE from powering down, change the CS ONE battery when the CS ONE is connected to a powered CS ONE dock F0.

Inserting a Battery

1. Open the battery door by gently pulling on the battery door pull tab.
2. Position the battery with the connector end facing towards the battery slot and insert the battery all the way into the battery slot.

Removing a Battery

1. Open the battery door by gently pulling on the battery door pull tab.
2. Pull on the battery cord to remove the battery from the battery slot.




CARESCAPE ONE Monitoring Basics

CARESCAPE ONE Main Screen Layout

The main screen displays alarms, information, waveforms, parameter windows, and the main menu in pre-defined areas.

1. Alarm area
2. Information area
3. Waveform area.
4. Upper parameter windows
5. Lower parameter windows (optional).
6. Main menu area




Information Area

When the **Information area** of the screen is selected via touch, it opens the **Admit/Discharge** or **Case Setup** menu and provides access to settings related to patient information, standby and profile selections.

The Information area of the screen displays the following information:


- Patient name (if entered).
The patient name may be covered by an alarm message if there are three or more active alarms.
- Profile name that is being used for patient monitoring. The profile name may be covered by an alarm message if there are three or more active alarms.
- Patient icon . You can access the **Admit/Discharge** or **Case Setup** menu by selecting this icon.
- Battery icon . You can access the **Battery Status** menu by selecting this icon.
- Current time of day.



Main Menu Keys

Various functions of the monitor can be accessed through the main menu keys.



Key	Function
Home 	Close all menus/applications displayed on the monitor.
Alarms Setup 	Allows access to alarm limits and priorities, arrhythmia alarm settings, audible and visual alarm indicators.
Monitor Setup 	The Main Setup tab allows access to screen setup, colors, sound volumes, parameter setup, battery status, brightness settings, admit/discharge (ICU, ED, NICU software packages) or case setup (OR, PACU software packages). The Defaults & Service tab allows access to default setup, service calibrations, and service, which are password protected. The Monitor Info tab presents monitor hardware and software information.
Pt. Data & Trends 	Allows access to numeric trends, trends time interval, and admit/discharge (ICU, ED, NICU software packages) or case setup (OR, PACU software packages).
NIBP Start or NIBP Cancel 	Start or stop a non-invasive blood pressure measurement. After selection, the key toggles to NIBP Cancel. Measurement does not start unless the cuff size is defined.
NIBP Auto Start of NIBP Auto Stop 	Start or stop automatic non-invasive blood pressure measurements at timed intervals. After selection, the button toggles to NIBP Auto Stop. Measurement does not start unless the cuff size is defined.
Zero All Pressures 	Zero all invasive pressure channels. This does not apply to ICP.
	Pause audio alarm. Pauses active audio alarms or pre-pause audio for incoming active alarms.

Menu Example

1. Menu title (for example, **ECG**).
2. Submenu tabs (for example, **Setup, ST**).
3. Selection lists: when selecting the arrow, a list of options appear.
4. Arrow selector for increasing/decreasing a value.
5. Radio button for selecting or deselecting a feature from the available options.
6. Help text area.
7. Further menu selections
8. Selector for increasing or decreasing the volume.
9. Check box for selecting or deselecting a feature.
10. Tabs to access additional pages (for example, **Page 1, Page 2**).




Note! These tabs may also be labeled with feature names (for example, **HR, PVC/SVC Arrhythmia**).

11. Exit key (for example, **Previous Menu, Close**).



Note! Not all menus have these same components.



Selecting Menu Options with a Touchscreen

In this guide, the term select means using the touchscreen display to select an item on the screen.

1. Touch the menu option with your finger.
2. The highlight on screen moves to this option.
3. Lift your finger off the screen, and the selected function is performed (for example, a list opens).


Profiles

When you start monitoring a patient, you can use the startup profile or select another profile. According to the configuration, your software may have up to eight profiles to choose from. Profiles control many settings, including parameter defaults, alarm detection limits, and alarm functionality.

1. Select the patient information area on the screen.
2. Select the **Patient** tab.
3. Select a profile from the **Profile** list.

If you make changes to a profile while using it and need to return to its previous settings, first select another profile and then re-select the one you were using. If only one profile has been enabled during configuration, return to saved profile settings through patient discharge/ case end or device reset.

If a profile other than the startup menu has been selected during the monitoring session, upon discharge the CS ONE will return to the startup profile for the next monitoring session.



Starting Monitoring

A case automatically starts/a patient is admitted when the monitor detects any of the following vital signs: ECG, impedance respiration, Art, Fem, UAC, NIBP, SpO₂, or CO₂. Each vital sign has activation criteria that must be met before the vital sign is considered active. For more information about activation criteria, see the User Manual.

A case manually starts/a patient is admitted when any patient data is entered, such as the patient name or medical record number.

Always observe the monitor and the patient carefully during startup periods and when connecting CARESCAPE Parameters.

Here are the basic steps for starting monitoring. Parameter-specific instructions are more detailed and should always be followed.

1. Connect the patient to the CS ONE according to the measurement setup requirements in the parameter-specific chapters of the User Manual. The alarms and parameter settings become active.
2. If the startup profile is not suitable, select another profile.
3. Enter patient demographics. Alarms will be activated based on the criteria, even if demographic information is not entered. Demographic information can be entered at a later point, if needed.
4. Start the measurement.
5. Zero invasive pressure lines. Refer to the invasive pressures chapter in the User Manual.
6. If required, change the parameters displayed on the screen.
7. Check alarm limits and adjust, if necessary.

Entering Patient Data

The monitor automatically displays an on-screen keyboard for you to use for data entry.

1. Select the patient information area on the screen.
2. Select the **Patient** tab.
3. Edit or enter patient data by selecting the characters using the touchscreen.
 - a. Select **MRN: Second ID:**, select the field to be edited, and enter the data. Entering the **Second ID:** in addition to the **Medical Record Number** allows a flexible use of local patient identification methods.
 - b. Select **Name:**, select the field to be edited, and enter the data.
 - c. Select **Date of Birth: Age:** to enter or edit the patient's date of birth or age. If **Day, Month,** and **Year** are entered, **Age** is automatically calculated.
 - d. Select the **Height: Weight: BSA: Gender: Ethnicity:** field, then select values for different types of data. If height and weight are entered, the patient's BSA (body surface area) is automatically calculated.



Resetting a Case / Discharging a Patient

Resetting a case/discharging a patient deletes all patient information and returns the monitor to the defined startup profile.

All settings, including alarm limits return to their default values. All patient data is removed from the monitor.

The CS ONE may be configured in care unit settings with an automatic case reset/patient discharge timer. Care unit settings are password protected. If an automatic case reset/patient discharge is configured and there are no active parameters and no use of the touchscreen, monitoring will end automatically after the configured time has elapsed.

1. Disconnect patient cables.
2. Select the patient information area on the screen.
3. Select the **Discharge / Reset Case** tab.
4. Select **Confirm**.



Standby

About Standby

When you remove the patient temporarily from the monitor, you can use the standby option to put the monitor on standby until the patient is reconnected to the monitor. You can also use the standby option before a patient is admitted to the monitor/patient case is started. Configure the current patient settings for the incoming patient, then put the monitor in standby to prevent nuisance alarms or automatic discharge due to monitor inactivity.

Starting Standby


1. Select the patient information area on the screen.
2. Select the **Standby** tab.
3. If a patient is connected, select **Prepare for Standby** to pause audible alarms for two minutes. This allows time to disconnect patient cables.

If no patient is connected (no CARESCAPE Parameters are connected to the CS ONE), select **Standby** from the **Standby** tab, and the monitor will immediately enter standby.

4. Disconnect the patient cables and check that NIBP Auto is turned off.

If patient cables are still connected and the monitor receives vital signs, a message indicating that audio alarms have been paused appears.

If you do not disconnect the cables and vital signs are still present after the audio pause time expires, the standby is cancelled. You can also cancel the standby before the timer expires by selecting **Cancel Standby**. The monitor will automatically enter standby when all cables are disconnected and **NIBP Auto** is turned off. The screen goes blank and the GE logo with the message **Standby** appears.



End of Standby

The monitor ends the standby automatically when any of the following conditions occur:


- The touchscreen is pressed.
- At least one CARESCAPE Parameter is connected and active parameter data is detected.

Alarms


Alarm Priority Levels

Physiological and technical alarms are categorized by priority level:

1. **High priority** alarms require an immediate response.
2. **Medium priority** alarms require a prompt response.
3. **Low priority** alarms require you to be aware of the condition.
4. **Informational priority** messages provide information you should know.



Pausing Alarms




Selecting the **Pause Audio** key results in different alarm behaviors depending on whether the alarms are active and/or latched or not. The length of audio pause (2 min or 5 min) has been configured in Care Unit Settings > Alarms (password protected). Acknowledging or pausing audio alarms does not affect the visual alarm indicators. They will **still continue** indicating alarms.

Adjusting the Alarm Volume

The selections in the Alarms Setup menu vary according to what has been configured in the Care Unit Settings > Alarms (password protected).

1. Select **Alarms Setup** from the main menu.
2. Select the **Audible & Visual** tab.
3. Select the **Volume & Light** tab.
 - Adjust the **Alarm Volume** value. This is the volume for all alarms.
 - Adjust the **Alarm Volume** separately for **High & Medium Priority** and **Low Priority**.

The lower the number, the quieter the alarm volume. Note that the minimum allowed alarm volume levels are set in the **Care Unit Settings**.




Changing an Alarm Limit

Setting Parameter Alarm Limits

Parameter alarm limits may be set in the **Alarms Setup** menu, or in the parameter menu's Alarm tab. Alarm limits should not be set beyond reasonable physiological boundaries in to maintain patient safety. Setting outside of reasonable boundaries would cause the alarms to be ineffective.

1. Select **Alarms Setup** from the main menu.
2. Select the **Alarm Limits** tab.
3. Select a parameter label.
That parameter menu's **Alarms** tab opens, where you can select the alarms on or off, and set their limits.




Trends

Viewing Numeric Trends

Numeric trends contain three pages with 24 hours of trend data. The lowest row shows the time indication and the top of the view shows the highest priority realtime waveform. You cannot configure the layout of the **Numeric Trend** view.

1. Select **Pt. Data & Trends**.
2. Select **Numeric Trends**.
3. Select a parameter group tab to view the trends.
 - **Vital:** HR, **SpO₂**, invasive pressure 1, **NIBP S/D(M)**, **RR(Imp)**, **RR(CO₂)**
 - **Pressures:** HR, invasive pressure 1, invasive pressure 2, **NIBP S/D(M)**
 - **Temp & Gases:** **T1**, **T2**, **CO₂ ET/FI**
4. To see more numeric data, use the scroll bar to scroll the data in horizontal directions.



2

1

3

About Snapshots and Graphic Trends

Snapshots: The CS ONE records snapshots when it is operating as a standalone monitor. However, snapshots are only viewable on the host monitor when the CS ONE is connected to a host monitor as an acquisition device.

Graphic Trends: Graphic trends are only viewable on the host monitor when the CS ONE is connected to a host monitor as an acquisition device. Only numeric trends are available on the CS ONE when it is operating as a standalone monitor.



4

Transport Use

You can use the CS ONE for uninterrupted monitoring when transferring a patient from one location to another. When the CS ONE is used as an acquisition module, with a host monitor, you can disconnect it from the host monitor, retain the patient data in the CS ONE, and continue to monitor as you move the patient. After transport, you can return to the CS ONE to the host monitor, reconnect it as an acquisition module, and resume monitoring the patient on the host monitor.

Dual Monitoring and Transport

During normal monitoring when the CS ONE is connected to the host monitor as an acquisition device, the CS ONE display is not active. A screen saver is displayed, and all parameter data and alarm signals are presented on the host monitor.

You can activate the CS ONE display screen by touching it. When activated, the message **Acquisition mode – view only** is displayed, indicating that only view is allowed.

Upon return from transport when the CS ONE is reconnected to the host monitor, the CS ONE display remains illuminated for two minutes so you can ensure the host monitor is displaying waveforms and parameter data.

Preparing to Transport a Patient

1. Place the host monitor into standby mode. Refer to the host monitor user manual for instructions.
2. Add parameters needed for transport to the CS ONE, or remove those not needed for transport.
3. Ensure that the parameters you need to monitor during transport are selected to display on the CS ONE screen.
4. Ensure that there is a charged battery installed in the CS ONE. Refer to the User Manual if the battery needs to be charged or changed.
5. Disconnect the CS ONE from the CARESCAPE Dock F0.
6. Place the CS ONE in the transport location. Use a Mini Dock and/or a Parameter Dock 1 or Parameter Dock 5 to secure the CS ONE and CARESCAPE Parameters during transport.
7. Transport the patient.

Returning a Patient from Transport

Follow these steps to return the patient to the same host monitor after transport, retaining the patient data.

1. Return the patient to the room.
2. Remove the CS ONE from the Mini Dock.
3. Connect the CS ONE to the CARESCAPE Dock F0 at the host monitor.

Patient monitoring will automatically resume on the host monitor if the patient case matches. Patient data acquired on the CS ONE during transport will be available on the host monitor.

If the patient case does not match between the host monitor and the CS ONE, then the Select **Patient and Data** window will open on the host monitor. There are four possible options.

You can:

- Select the patient in the CS ONE to continue the CS ONE patient case. This option is shown as CS ONE, along with the patient name and MRN number.
 - Select the patient in the host monitor to continue the host monitor patient case. This option will only appear if a patient is admitted to the host monitor. It is shown as the host monitor bed name plus the patient name and MRN number.
 - Merge the CS ONE patient case with the host monitor patient case. This option is shown as the host monitor bed name + CS ONE.
 - Select **New Patient** to discharge the patient from the CS ONE, deleting all patient data and history, and start monitoring a new patient/case at the host monitor.
4. If needed, reconnect any parameters that were disconnected during transport.



www.gehealthcare.com

GE Healthcare provides transformational medical technologies and services that are shaping a new age of patient care.

Our broad expertise in medical imaging and information technologies, medical diagnostics, patient monitoring systems, drug discovery, biopharmaceutical manufacturing technologies, performance improvement and performance solutions services help our customers to deliver better care to more people around the world at a lower cost.

In addition, we partner with healthcare leaders, striving to leverage the global policy change necessary to implement a successful shift to sustainable healthcare systems.

imagination at work

Product may not be available in all countries and regions. Full product technical specification is available upon request. Contact a GE Healthcare Representative for more information. Please visit www.gehealthcare.com/promotional-locations.

© 2017 General Electric Company – All rights reserved.

General Electric Company reserves the right to make changes in specifications and features shown herein, or discontinue the product described at any time without notice or obligation.

GE, GE Monogram, and CARESCAPE are trademarks of General Electric Company.

Masimo SET is a trademark of Masimo Corporation.

Nellcor, and OxiMax are trademarks of a Medtronic company.

LoFlo is a trademark of Koninklijke Philips Electronics N.V.

Notice: The materials contained in this document are intended for educational purposes only. This document does not establish specifications, operating procedures or maintenance methods for any of the products referenced. Always refer to the official written materials (labeling) provided with the product for specifications, operating procedures and maintenance requirements. Specifications are subject to change.

All patient names or other protected health information or data contained in any image within this material is fictitious. Any similarity to actual persons is coincidental.