



Manufacturer's Name:	Agilent Technologies Australia (M) Pty Ltd.
Manufacturer's Address:	679 Springvale Rd, Mulgrave, Victoria, 3170, Australia

# Declares under sole responsibility that the product as originally delivered

Product Name:	Atomic Absorption Spectrometers 240FS, 240FS with GTA camera, 280FS
Product Number:	G8432A, G8433A, G8434A
Product Options:	This declaration covers all options of the above products

# complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

Low Voltage Directive	2014/35/EU
EMC Directive	2014/30/EU
RoHS Directive	2011/65/EU as amended by (EU) 2015/863

## and conforms with the following product standards:

	Standard
EMC:	IEC 61326-1:2012 / EN 61326-1:2013 Canada: ICES-001 Issue 4 Australia/New Zealand: IEC 61326-1:2012
Safety:	IEC 61010-1:2010 / EN 61010-1:2010 IEC 61010-2-061:2015 / EN 61010-2-061:2015 Canada: CAN/CSA C22.2 No. 61010-1-12 Canada: CAN/CSA C22.2 No. 61010-2-061-15 USA: UL 61010-1 (3 <sup>rd</sup> Edition) USA: UL 61010-2-061:2015
Other:	EN IEC 63000:2018

# **Supplementary Information:**

The product was tested in a typical configuration with Agilent Technologies test systems.

# This DoC applies to above-listed product placed on the EU market after:

30 June 2021

Date

Cale Hunkin, Quality Manager

For further information, please contact your local Agilent Technologies sales office, agent or distributor, or Agilent Technologies Deutschland GmbH, Hewlett-Packard-Strasse 8, 76337 Waldbronn, Germany.