


# Test Report

## – Eppendorf PCR clean –

### PCR Tube 0.2 mL



1000 Tubes 0.2 mL, thin walled  
PCR clean  
US-PAT.NO. 5,863,791

LOT K198325M

2026-07-28

Order no.: 0030 124.332  
Eppendorf AG  
22331 Hamburg · Germany  
Email: eppendorf@eppendorf.com

Cat. no.: 951010006  
Distributed in North America by  
Eppendorf North America, Inc. · USA  
Email: info@eppendorf.com

#### • Human DNA

Method: real-time PCR acc. to SOP 99880/V2  
Limit: < 2 pg; less than one human cell

#### • DNase

Method: DNA Digestion acc. to SOP 100570/V1  
Limit: Not detectable (LOD:  $1.0 \times 10^{-6}$  Kunitz units)

#### • PCR inhibition

Method: real-time PCR acc. to SOP 100559/V1  
Limit: Less than 10 targets amplifiable

#### • RNase

Method: RNA Digestion acc. to SOP 100571/V1  
Limit: Not detectable (LOD:  $1.0 \times 10^{-9}$  Kunitz units)


Customer: Eppendorf AG, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 23. 09. 21

Testing of the above described lot showed conformity within the limits of detection.

28.09.21  
Date of release

Epp 21 09 9508  
Registration number

  
Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.