

Certificate of CE-Notification

This is to certify that, in accordance with the *In Vitro* Diagnostic Medical Device Directive 98/79/EC, **CEpartner4U BV** agrees to perform all duties and responsibilities as the Authorized Representative for

Monocent Inc.
9237 Eton Ave.,
Chatsworth, CA 91311
United States

as stipulated and demanded by the aforementioned Directive. The Dutch Competent Authorities have accepted the manufacturer's medical device registrations by CEpartner4U as listed on the product list attached to the manufacturer's Declaration of Conformity:

IVD devices were registered with the Dutch Competent Authority with registration number:


IVD Devices groups:	Registration number:
CLIA Test Kits	NL-CA002-2020-50897
ELISA Test Kits	NL-CA002-2020-50898
IFA Test Kits	NL-CA002-2020-50899
Instruments	NL-CA002-2020-50900
PCR Test Kits	NL-CA002-2020-50901
Rapid Tests	NL-CA002-2020-50902
Serology Test Kits	NL-CA002-2020-50903

see appendix

The manufacturer has provided CEpartner4U with all necessary documentation, together with an appropriate Declaration of Conformity that the IVD medical devices fulfil the essential requirements of Directive 98/79/EC.

Issue date: 2022-10-31

This Certificate of CE-Notification is valid until May 26, 2025


R. Nusselder
Sr. consultant CEpartner4U BV

C e p a r t n e r 4 U
Esdoornlaan13
3951 DB Maarn NL
tel: +31 (0)343 442 524
www.cepartner4u.nl

Appendix

List of devices.

CLIA Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
Allergy Assays					
IgE	CL3-5055	Low Risk	C	30275	2020-04-14
Thyroid Assays					
T3	CL3-5028	Low Risk	C	30312	2020-04-14
T4	CL3-5029	Low Risk	C	30314	2020-04-14
TSH	CL2-5030	Low Risk	C	30318	2020-04-14
T3 Uptake	CL3-5072	Low Risk	C	30313	2020-04-14
FT3	CL3-5026	Low Risk	C	30309	2020-04-14
FT4	CL3-5027	Low Risk	C	30308	2020-04-14
Tg (Thyroglobulin)	CL3-5073	Low Risk	C	30490	2020-04-14
TBG	CL3-5074	Low Risk	C	30316	2020-04-14
Anti-Tg	CL3-5075	Low Risk	C	30490	2020-04-14
Anti-TPO	CL3-5076	Low Risk	C	30317	2020-04-14
Ultra-Sensitive TSH	CL2-5077	Low Risk	C	30318	2020-04-14
Fertility Assays					
LH	CL3-5006	Low Risk	C	38965	2020-04-14
FSH	CL3-5004	Low Risk	C	30322	2020-04-14
Prolactin	CL3-5008	Low Risk	C	30325	2020-04-14
hCG	CL2-5005	Low Risk	B	30513	2020-04-14
AMH	CL3-5069	Low Risk	C	43148	2020-04-14
Beta hCG	CL2-5055	Low Risk	B	30332	2020-04-14
HGH	CL3-5007	Low Risk	C	30333	2020-04-14
PAPP-A	CL3-5068	Low Risk	C	31533	2020-04-14
Diabetes Assays					
Insulin	CL2-5003	Low Risk	C	30338	2020-04-14
C-peptide	CL2-5002	Low Risk	C	30336	2020-04-14
Tumor Markers Assays					
AFP	CL3-5031	Low Risk	C	30295	2020-04-14
CEA	CL3-5036	Low Risk	C	30288	2020-04-14
Free Beta hCG	CL2-5037	Low Risk	C	30333	2020-04-14
Beta 2 Microglobulin	CL2-5032	Low Risk	C	30296	2020-04-14
NSE	CL2-5039	Low Risk	C	30301	2020-04-14
CA-12-5	CL3-5034	Low Risk	C	30283	2020-04-14
CA-19-9	CL2-5035	Low Risk	C	30280	2020-04-14
CA-15-3	CL2-5033	Low Risk	C	30279	2020-04-14
Ferritin	CL3-5001	Low Risk	C	30377	2020-04-14
Cyfra21-1	CL2-5079	Low Risk	C	44431	2020-04-14
Pro-GRP	CL2-5080	Low Risk	C	44438	2020-04-14
PAP	CL2-5081	Low Risk	C	34226	2020-04-14
Steroid Assays					
Progesterone	CL3-5021	Low Risk	C	30294	2020-04-14
Estradiol	CL3-5016	Low Risk	C	30321	2020-04-14
Testosterone	CL3-5022	Low Risk	C	30327	2020-04-14

CLIA Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
Free Testosterone	CL9-5023	Low Risk	C	30327	2020-04-14
Testosterone (Saliva)	CL9-5025	Low Risk	C	30327	2020-04-14
5a-Androstane-3a, 17b-diol Glucuronide (3a- Diol G)	CL9-5009	Low Risk	C	31533	2020-04-14
17 OH Progesterone	CL3-5010	Low Risk	C	30324	2020-04-14
Androstenedione	CL3-5070	Low Risk	C	30319	2020-04-14
Aldosterone	CL3-5011	Low Risk	C	31428	2020-04-14
Cortisol	CL3-5012	Low Risk	C	31394	2020-04-14
DHEA	CL3-5013	Low Risk	C	39894	2020-04-14
DHEA-S	CL3-5014	Low Risk	C	39894	2020-04-14
uE3	CL3-5041	Low Risk	C	30330	2020-04-14
Estriol (Saliva)	CL9-5018	Low Risk	C	30329	2020-04-14
Estrone (Saliva)	CL9-5019	Low Risk	C	33293	2020-04-14
Estrone	CL3-5020	Low Risk	C	33293	2020-04-14
Plasma Renin Activity (PRA)	CL9-5024	Low Risk	C	43444	2020-04-14
SHBG	CL3-5071	Low Risk	C	30326	2020-04-14
Procalcitonin	CL3-5067	Low Risk	C	12069016	2020-04-14
Infectious Disease Assays					
Digoxin	CL3-5059	Low Risk	C	30386	2020-04-14
hs-CRP	CL2-5060	Low Risk	C	30499	2020-04-14
CK-MB	CL3-5061	Low Risk	C	30499	2020-04-14
Myoglobin	CL3-5062	Low Risk	C	30264	2020-04-14
cTn I	CL2-5063	Low Risk	C	30266	2020-04-14
Bone Metabolism					
ACTH	CL3-5017	Low Risk	C	39005	2020-04-14
Calcitonin	CL3-5064	Low Risk	C	30342	2020-04-14
PTH	CL3-5065	Low Risk	C	30353	2020-04-14
Vitamin D	CL3-5066	Low Risk	C	30350	2020-04-14
Autoimmune Disease					
Cardiolipin IgA	CL2-5051	Low Risk	C	30475	2020-04-14
Cardiolipin IgG	CL2-5052	Low Risk	C	30475	2020-04-14
Cardiolipin IgM	CL2-5053	Low Risk	C	30475	2020-04-14
ds-DNA	CL2-5054	Low Risk	C	30458	2020-04-14
RF IgM	CL2-5114	Low Risk	C	30500	2020-04-14
B2GP1 IgA	CL2-5115	Low Risk	C	30478	2020-04-14
B2GP1 IgG	CL2-5116	Low Risk	C	30478	2020-04-14
B2GP1 IgM	CL2-5117	Low Risk	C	30478	2020-04-14
Thyroglobulin IgG	CL2-5118	Low Risk	C	30315	2020-04-14
Anti-CCP	CL2-5119	Low Risk	C	44202	2020-04-14
Anemia Assays					
Folate	CL3-5056	Low Risk	C	30378	2020-04-14
Vitamin B12	CL3-5057	Low Risk	C	30384	2020-04-14
Transferrin Soluble Receptor (sTfR)	CL3-5058	Low Risk	C	30253	2020-04-14
NeoNatal Assays					
Neonatal TSH	CL2-5078	Low Risk	C	30310	2020-04-14

CLIA Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
Infectious Disease Assays					
H. pylori IgA	CL2-5048	Low Risk	B	30691	2020-04-14
H. pylori IgG	CL2-5049	Low Risk	B	30691	2020-04-14
H. pylori IgM	CL2-5050	Low Risk	B	30691	2020-04-14
H. pylori IgG (Quantitative)	CL2-5082	Low Risk	B	30691	2020-04-14
H. pylori Antigen	CL2-5083	Low Risk	B	30691	2020-04-14
EBV VCA IgA	CL2-5084	Low Risk	D	30809	2020-04-14
EBV VCA IgG	CL2-5085	Low Risk	D	30809	2020-04-14
EBV VCA IgM	CL2-5086	Low Risk	D	30809	2020-04-14
EBV EA-D IgA	CL2-5087	Low Risk	D	30809	2020-04-14
EBV EA-D IgG	CL2-5088	Low Risk	D	30809	2020-04-14
EBV EA-D IgM	CL2-5089	Low Risk	D	30809	2020-04-14
EBNA IgA	CL2-5090	Low Risk	D	30808	2020-04-14
EBNA IgG	CL2-5091	Low Risk	D	30808	2020-04-14
EBNA IgM	CL2-5092	Low Risk	D	30808	2020-04-14
Measles IgG	CL2-5093	Low Risk	C	44019	2020-04-14
Measles IgM	CL2-5094	Low Risk	C	44019	2020-04-14
VZV IgG	CL2-5095	Low Risk	C	44027	2020-04-14
VZV IgM	CL2-5096	Low Risk	C	44027	2020-04-14
Mumps IgG	CL2-5097	Low Risk	C	33908	2020-04-14
Mumps IgM	CL2-5098	Low Risk	C	33908	2020-04-14
Dengue IgG	CL2-5099	Low Risk	C	32481	2020-04-14
Dengue IgM	CL2-5100	Low Risk	C	32481	2020-04-14
HSV 1/2 IgG	CL2-5101	Low Risk	C	40176	2020-04-14
HSV 1/2 IgM	CL2-5102	Low Risk	C	40176	2020-04-14
HSV 1 IgA	CL2-5103	Low Risk	C	38870	2020-04-14
HSV 1 IgG	CL2-5104	Low Risk	C	38870	2020-04-14
HSV 1 IgM	CL2-5105	Low Risk	C	38870	2020-04-14
HSV 2 IgA	CL2-5106	Low Risk	C	38875	2020-04-14
HSV 2 IgG	CL2-5107	Low Risk	C	38875	2020-04-14
HSV 2 IgM	CL2-5108	Low Risk	C	38875	2020-04-14

ELISA Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
Allergy					
Total Human IgE	EL1-1000, EL2-1000	Low Risk	B	30275	2020-04-14
Human Specific IgG	EL15-1001	Low Risk	C	44211	2020-04-14
Human Specific IgG4	EL15-1002	Low Risk	C	44211	2020-04-14
Histamine	EL30-1003	Low Risk	C	30274	2020-04-14
Anemia					
Vitamin B12	EL1-1007	Low Risk	B	30384	2020-04-14
Folate	EL1-1005	Low Risk	B	30378	2020-04-14
sTfR-Transferrin Soluble Receptor	EL3-1006	Low Risk	B	30253	2020-04-14
Ferritin	EL1-1004	Low Risk	B	30377	2020-04-14

ELISA Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
Hepcidin	EL1-1008	Low Risk	B	12070190	2020-04-14
Autoimmune Disease					
Anti-CCP	EL2-1011	Low Risk	B	44202	2020-04-14
Anti-CP IgG	EL20-1288	Low Risk	B	44202	2020-04-14
Beta 2 Glycoprotein 1 IgA	EL2-1017	Low Risk	B	30478	2020-04-14
Beta 2 Glycoprotein 1 IgG	EL2-1018	Low Risk	B	30478	2020-04-14
Beta 2 Glycoprotein 1 IgM	EL2-1019	Low Risk	B	30478	2020-04-14
Anti-Tissue Transglutaminase IgG	EL20-1015	Low Risk	C	44385	2020-04-14
Anti-Tissue Transglutaminase IgA	EL20-1014	Low Risk	C	44385	2020-04-14
ANA Screen IgG	EL1-1009	Low Risk	B	30454	2020-04-14
ENA IgG Profile-6	EL10-1024	Low Risk	B	30455	2020-04-14
ENA Screen IgG	EL20-1025	Low Risk	B	30455	2020-04-14
Rheumatoid Factor (RF) IgA	EL15-1034	Low Risk	B	30500	2020-04-14
Rheumatoid Factor (RF) IgG	EL15-1035	Low Risk	B	30500	2020-04-14
Rheumatoid Factor (RF) IgM	EL2-1038	Low Risk	B	30500	2020-04-14
Sm/RNP IgG	EL1-1040	Low Risk	B	30464	2020-04-14
Sm IgG	EL1-1041	Low Risk	B	17276	2020-04-14
Jo-1 IgG	EL21-1029	Low Risk	C	30461	2020-04-14
Scl-70 IgG	EL1-1039	Low Risk	B	30463	2020-04-14
SS-A (Ro)	EL1-1042	Low Risk	B	44202	2020-04-14
SS-B (La)	EL1-1043	Low Risk	B	44202	2020-04-14
dsDNA	EL1-1023	Low Risk	B	30458	2020-04-14
Cardiolipin IgG	EL1-1021	Low Risk	C	30475	2020-04-14
Cardiolipin IgM	EL1-1022	Low Risk	C	30475	2020-04-14
Cardiolipin IgA	EL1-1020	Low Risk	C	30475	2020-04-14
Cardiolipin Total Ab	EL1-1044	Low Risk	C	30475	2020-04-14
Mitochondrial Antibody (MA)	EL1-1031	Low Risk	C	30476	2020-04-14
Thyroglobulin Antigen (Anti-Tg)	EL3-1016	Low Risk	C	30315	2020-04-14
PR3 (c-ANCA)	EL20-1033	Low Risk	B	30484	2020-04-14
ANCA screen IgG	EL10-1010	Low Risk	B	30483	2020-04-14
MPO, Myeloperoxidase (p-ANCA)	EL20-1032	Low Risk	B	30483	2020-04-14
Gliadin IgG	EL36-1026	Low Risk	C	30480	2020-04-14
Gliadin IgA	EL36-1027	Low Risk	C	30480	2020-04-14
TPO	EL1-1012	Low Risk	C	30317	2020-04-14
Anti-Phospholipids Screen	EL20-1013	Low Risk	B	30582	2020-04-14
ASMA	EL29-1302	Low Risk	B	30274	2020-04-14
Beta-2-Glycoprotein IgA	EL2-1017	Low Risk	B	30478	2020-04-14
Beta-2-Glycoprotein IgG	EL2-1018	Low Risk	B	30478	2020-04-14
Beta-2-Glycoprotein IgM	EL2-1019	Low Risk	B	30478	2020-04-14
Tumor markers					
Prostatic Acid Phosphatase (PAP)	EL2-1289	Low Risk	C	34226	2020-04-14
Beta-2-Microglobulin	EL2-1277	Low Risk	C	30296	2020-04-14
AFP (Alpha Fetoprotein)	EL1-1276	Low Risk	C	43480	2020-04-14
CEA	EL1-1283	Low Risk	C	30288	2020-04-14
CA-15-3	EL1-1279	Low Risk	C	30279	2020-04-14
CA-12-5	EL1-1278	Low Risk	C	30283	2020-04-14

ELISA Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
CA-19-9	EL1-1280	Low Risk	C	30280	2020-04-14
NSE	EL2-1286	Low Risk	C	30301	2020-04-14
Free Beta HCG	EL1-1284	Low Risk	C	30333	2020-04-14
Pro-GRP (Gastrin-Releasing Peptide)	EL2-1290	Low Risk	C	44438	2020-04-14
Chromogranin A	EL1-1281	Low Risk	C	30289	2020-04-14
HE4	EL1-1306	Low Risk	C	30289	2020-04-14
Cyfra21-1	EL2-1034	Low Risk	C	30289	2020-04-14
Bone Metabolism					
Intact PTH	EL3-1048	Low Risk	C	30353	2020-04-14
25-OH Vitamin D	EL1-1045	Low Risk	B	30350	2020-04-14
ACTH	EL3-1046	Low Risk	C	39005	2020-04-14
Cardiac					
Digoxin	EL3-1051	Low Risk	C	30386	2020-04-14
CK-MB	EL3-1050	Low Risk	C	30499	2020-04-14
Troponin I	EL1-1054	Low Risk	C	30266	2020-04-14
Myoglobin	EL6-1053	Low Risk	C	30264	2020-04-14
C-Reactive Protein (CRP)	EL1-1049	Low Risk	C	30499	2020-04-14
Diabetes					
Insulin	EL1-1058	Low Risk	C	30338	2020-04-14
C-peptide	EL1-1055	Low Risk	C	30336	2020-04-14
Leptin	EL9-1059	Low Risk	B	12069017	2020-04-14
Adiponectin	EL9-1056	Low Risk	B	12069017	2020-04-14
(IGFBP-1) Insulin-Like Growth Factor Binding Protein-1	EL9-1057	Low Risk	B	42852	2020-04-14
Anti-GAD	EL8-1060	Low Risk	B	30340	2020-04-14
IAA	EL8-1061	Low Risk	B	30339	2020-04-14
IGF-1	EL8-1062	Low Risk	B	30361	2020-04-14
Pro-Insulin	EL1-1063	Low Risk	C	42852	2020-04-14
Fertility					
Human Growth Hormone (HGH)	EL1-1083	Low Risk	B	30333	2020-04-14
hCG Visual	EL6-1082	Low Risk	B	30513	2020-04-14
Beta hCG (Total)	EL2-1078	Low Risk	B	30332	2020-04-14
FSH	EL1-1080	Low Risk	B	31533	2020-04-14
LH	EL1-1084	Low Risk	B	38246	2020-04-14
Prolactin	EL1-1086	Low Risk	B	30325	2020-04-14
PAPP-A	EL3-1085	Low Risk	B	31533	2020-04-14
SHBG	EL3-1261	Low Risk	B	30326	2020-04-14
AMH	EL3-1079	Low Risk	B	43148	2020-04-14
hCG	EL1-1081	Low Risk	B	30332	2020-04-14
Sperm Ab	EL8-1087	Low Risk	B	30486	2020-04-14
Infectious Diseases					
Adenovirus IgG	EL15-1102	Low Risk	C	39468	2020-04-14
Adenovirus IgA	EL15-1101	Low Risk	C	39468	2020-04-14
Adenovirus IgM	EL15-1103	Low Risk	C	39468	2020-04-14
Influenza A IgA	EL15-1365	Low Risk	B	39463	2020-04-14
Influenza A IgG	EL15-1366	Low Risk	B	39463	2020-04-14

ELISA Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
Influenza A IgM	EL15-1367	Low Risk	B	39463	2020-04-14
Influenza B IgA	EL15-1368	Low Risk	B	39463	2020-04-14
Influenza B IgG	EL15-1369	Low Risk	B	39463	2020-04-14
Influenza B IgM	EL15-1370	Low Risk	B	39463	2020-04-14
Chikungunya IgG	EL4-1114	Low Risk	D	32481	2020-04-14
Chikungunya IgM	EL4-1113	Low Risk	D	32481	2020-04-14
COVID-19 IgA	EL45-1373	Low Risk	D	42994	2020-04-14
COVID-19 IgG	EL1-1360	Low Risk	D	42994	2020-04-14
COVID-19 IgM	EL1-1361	Low Risk	D	42994	2020-04-14
COVID-19 IgG	EL36-1360R	Low Risk	D	42994	2020-04-14
COVID-19 IgM	EL36-1361R	Low Risk	D	42994	2020-04-14
COVID-19 IgG	EL45-1360	Low Risk	D	42994	2020-04-14
COVID-19 IgM	EL45-1361	Low Risk	D	42994	2020-04-14
COVID-19 Total Ab	EL45-1379	Low Risk	D	42994	2020-12-06
Mycobacterium Tuberculosis (TB) IgA	EL15-1317	Low Risk	C	30635	2020-04-14
Mycobacterium Tuberculosis (TB) IgG	EL15-1201	Low Risk	C	30635	2020-04-14
Mycobacterium Tuberculosis (TB) IgM	EL15-1202	Low Risk	C	30635	2020-04-14
Herpes Simplex 1 IgG (HSV1 IgA)	EL2-1162	Low Risk	C	38870	2020-04-14
Herpes Simplex 1 IgG (HSV1 IgG)	EL1-1163	Low Risk	C	38870	2020-04-14
Herpes Simplex 1 IgM (HSV1 IgM)	EL1-1164	Low Risk	C	38870	2020-04-14
Herpes Simplex 2 IgG (HSV2 IgG)	EL1-1165	Low Risk	C	38875	2020-04-14
Herpes Simplex 2 IgM (HSV2 IgM)	EL1-1166	Low Risk	C	38875	2020-04-14
Herpes Simplex 1,2 IgG (HSV1,2 IgG)	EL1-1167	Low Risk	C	40176	2020-04-14
Herpes Simplex 1,2 IgM (HSV1,2 IgM)	EL1-1168	Low Risk	C	40176	2020-04-14
Epstein Barr Virus VCA IgA (EBV, VCA IgA)	EL2-1135	Low Risk	D	30809	2020-04-14
Epstein Barr Virus VCA IgG (EBV, VCA IgG)	EL1-1136	Low Risk	D	30809	2020-04-14
Epstein Barr Virus VCA IgM (EBV, VCA IgM)	EL1-1137	Low Risk	D	30809	2020-04-14
Epstein Barr Virus Early Antigen (EA) IgM	EL2-1134	Low Risk	D	30809	2020-04-14
Epstein Barr Virus Early Antigen (EA) IgG	EL2-1133	Low Risk	D	30809	2020-04-14
Epstein Barr Virus Early Antigen (EA) IgA	EL2-1132	Low Risk	D	30809	2020-04-14
Epstein Barr Virus Nuclear Antigen (EBNA) IgG	EL2-1130	Low Risk	D	30808	2020-04-14
Epstein Barr Virus Nuclear Antigen (EBNA) IgM	EL2-1131	Low Risk	D	30808	2020-04-14
Epstein Barr Virus Nuclear Antigen (EBNA) IgA	EL2-1129	Low Risk	D	30808	2020-04-14
Measles IgG	EL1-1177	Low Risk	C	44019	2020-04-14
Measles IgM	EL1-1178	Low Risk	C	44019	2020-04-14
Mumps IgG	EL1-1179	Low Risk	C	33908	2020-04-14
Mumps IgM	EL1-1180	Low Risk	C	33908	2020-04-14
Mycoplasma pneumonia IgG	EL1-1181	Low Risk	C	30657	2020-04-14
Mycoplasma pneumonia IgM	EL1-1182	Low Risk	C	30657	2020-04-14
Syphilis (TPA) IgG	EL1-1195	Low Risk	C	30685	2020-04-14
Syphilis (TPA) IgM	EL1-1197	Low Risk	C	30685	2020-04-14
Legionella urine Ag detection	EL16-1175	Low Risk	C	30692	2020-04-14
H. pylori IgG	EL1-1140	Low Risk	B	30691	2020-04-14
H. pylori IgA	EL1-1139	Low Risk	B	30691	2020-04-14
H-Pylori IgM	EL1-1141	Low Risk	B	30691	2020-04-14
H. pylori Antigen	EL2-1138	Low Risk	B	30691	2020-04-14

ELISA Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
	EL32-1138				
Varicella-Zoster IgG	EL1-1209	Low Risk	C	44027	2020-04-14
Varicella-Zoster IgM	EL1-1210	Low Risk	C	44027	2020-04-14
HEV IgG	EL13-1156	Low Risk	C	30757	2020-04-14
HEV IgM	EL13-1161	Low Risk	C	30758	2020-04-14
HAV Ab	EL7-1142	Low Risk	C	30721	2020-04-14
HAV IgM	EL7-1143	Low Risk	C	30722	2020-04-14
HDV IgG	EL7-1153	Low Risk	D	30750	2020-04-14
HDV IgM	EL7-1155	Low Risk	D	30752	2020-04-14
HDV Ab	EL13-1315	Low Risk	D	30750	2020-04-14
HDV Ag	EL13-1316, EL7-1154	Low Risk	D	30747	2020-04-14
HTLV 1 + 2 Ab	EL7-1160	Low Risk	C	30789	2020-04-14
Lyme Disease IgG	EL10-1171	Low Risk	C	30697	2020-04-14
Lyme Disease IgM	EL10-1172	Low Risk	C	30697	2020-04-14
Lyme Disease IgG, M	EL10-1173	Low Risk	C	30697	2020-04-14
Bordetella Pertussis IgA	EL15-1110	Low Risk	C	37723	2020-04-14
Bordetella Pertussis IgG	EL15-1111	Low Risk	C	37723	2020-04-14
Bordetella Pertussis IgM	EL15-1112	Low Risk	C	37723	2020-04-14
RSV IgA	EL15-1186	Low Risk	B	30814	2020-04-14
RSV IgG	EL15-1187	Low Risk	B	30814	2020-04-14
RSV IgM	EL15-1188	Low Risk	B	30814	2020-04-14
Tetanus	EL5-1205	Low Risk	C	38876	2020-04-14
Diphtheria IgG	EL5-1124	Low Risk	D	33499	2020-04-14
Salmonella typhi IgG	EL1-1193	Low Risk	C	30709	2020-04-14
Salmonella typhi IgM	EL1-1194	Low Risk	C	30709	2020-04-14
Salmonella Antigen detection	EL4-1192	Low Risk	C	30709	2020-04-14
Anthrax IgG	EL1-1105	Low Risk	C	32481	2020-04-14
Babesia IgG	EL4-1109	Low Risk	C	32481	2020-04-14
Dengue IgM	EL5-1127	Low Risk	C	32481	2020-04-14
Dengue IgG/IgM	EL5-1125	Low Risk	C	32481	2020-04-14
Dengue IgG	EL5-1126	Low Risk	C	32481	2020-04-14
Dengue NS1 Antigen	EL4-1128	Low Risk	C	32481	2020-04-14
Japanese Encephalitis IgG	EL4-1169	Low Risk	C	44321	2020-04-14
Japanese Encephalitis IgM	EL4-1170	Low Risk	C	44321	2020-04-14
Leprosy IgG/IgM	EL4-1176	Low Risk	C	32481	2020-04-14
Parvovirus B19 IgG	EL30-1183	Low Risk	C	40443	2020-04-14
Parvovirus B19 IgM	EL30-1184	Low Risk	C	40444	2020-04-14
Rotavirus (fecal)	EL16-1185	Low Risk	C	30815	2020-04-14
Scrub Typhus IgG	EL4-1199	Low Risk	C	44028	2020-04-14
Scrub Typhus IgM	EL4-1200	Low Risk	C	44028	2020-04-14
TB IgA	EL15-1317	Low Risk	C	30635	2020-04-14
TB IgG	EL15-1201	Low Risk	C	30635	2020-04-14
TB IgM	EL15-1202	Low Risk	C	30635	2020-04-14
Zika Virus IgG	EL1-1203	Low Risk	C	32481	2020-04-14
Zika Virus IgM	EL1-1204	Low Risk	C	32481	2020-04-14

ELISA Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
West Nile IgG	EL4-1211	Low Risk	C	42926	2020-04-14
West Nile IgM	EL4-1212	Low Risk	C	42926	2020-04-14
Parasitology					
Schistosoma IgG	EL5-1227	Low Risk	C	30824	2020-04-14
Chagas	EL5-1213	Low Risk	D	30820	2020-04-14
Cysticercosis IgG (T. solium)	EL5-1220	Low Risk	B	39979	2020-04-14
Campylobacter	EL16-1229	Low Risk	B	33948	2020-04-14
E. coli 0157 Ag detection	EL16-1232	Low Risk	B	37727	2020-04-14
E. histolytica IgG (Amebiasis)	EL5-1221	Low Risk	B	39979	2020-04-14
E. histolytica Dispar	EL16-1233	Low Risk	B	39979	2020-04-14
Echinococcus IgG	EL5-1222	Low Risk	B	30822	2020-04-14
Fasciola IgG	EL5-1216	Low Risk	B	34068	2020-04-14
Fasciola gigantica	EL5-1217	Low Risk	B	34068	2020-04-14
Filaria IgG4	EL4-1218	Low Risk	B	34068	2020-04-14
Leishmania	EL5-1223	Low Risk	C	30823	2020-04-14
Leptospira IgG	EL5-1224	Low Risk	C	30716	2020-04-14
Leptospira IgM	EL5-1226	Low Risk	C	30716	2020-04-14
Leptospira IgG/IgM	EL5-1225	Low Risk	C	30716	2020-04-14
Toxocara IgG	EL5-1228	Low Risk	C	34068	2020-04-14
Trichinella IgG	EL5-1215	Low Risk	C	33379	2020-04-14
Ascaris IgG	EL5-1219	Low Risk	B	39979	2020-04-14
Strongyloides IgG	EL5-1214	Low Risk	C	34068	2020-04-14
Crypto/Giardia Ag detection	EL16-1230	Low Risk	B	30675	2020-04-14
Cryptosporidium Ag detection	EL16-1231	Low Risk	B	30675	2020-04-14
Giardia antigen	EL16-1235	Low Risk	B	36173	2020-04-14
Giardia coprpantigen in stool	EL5-1361	Low Risk	B	36173	2020-04-14
Anti-Giardia IgA ELISA in saliva	EL5-1362	Low Risk	B	36173	2020-04-14
Entamoeba histolytica coproantigen in stool	EL5-1363	Low Risk	B	39979	2020-04-14
Adenovirus Antigen	EL16-1104	Low Risk	C	41274	2020-04-14
Steroid					
Aldosterone	EL3-1247	Low Risk	C	31428	2020-04-14
Cortisol	EL1-1249	Low Risk	C	31394	2020-04-14
Aldosterone	EL3-1247	Low Risk	B	31428	2020-04-14
Cortisol	EL1-1249	Low Risk	C	31394	2020-04-14
Cortisol Saliva	EL9-1250	Low Risk	C	31394	2020-04-14
Estradiol	EL1-1254	Low Risk	B	30321	2020-04-14
DHEA-S	EL1-1251	Low Risk	C	30320	2020-04-14
DHEA	EL3-1252	Low Risk	C	39894	2020-04-14
Progesterone	EL1-1259	Low Risk	C	30323	2020-04-14
Progesterone Saliva	EL9-1260	Low Risk	C	30294	2020-04-14
Testosterone	EL1-1263	Low Risk	B	30327	2020-04-14
Testosterone Saliva	EL9-1265	Low Risk	B	30327	2020-04-14
Free Testosterone	EL1-1264	Low Risk	B	30327	2020-04-14
Androstenedione	EL1-1248	Low Risk	C	30321	2020-04-14
Free Estriol	EL1-1257	Low Risk	B	30330	2020-04-14
Dihydrotestosterones (DHT)	EL9-1253	Low Risk	C	30327	2020-04-14

ELISA Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
17-OH Progesterone	EL1-1245	Low Risk	C	30324	2020-04-14
5a-Androstane-3a, 17b-diol Glucuronide (3a- Diol G)	EL9-1246	Low Risk	C	31533	2020-04-14
Total Estrogen	EL9-1255	Low Risk	B	38858	2020-04-14
Estrone	EL3-1256	Low Risk	B	33293	2020-04-14
Pregnenolone	EL9-1258	Low Risk	B	33301	2020-04-14
Total Estriol	EL8-1266	Low Risk	B	30330	2020-04-14
Thyroid					
T3	EL1-1270	Low Risk	C	30314	2020-04-14
T4	EL1-1271	Low Risk	C	30312	2020-04-14
TSH	EL1-1273	Low Risk	C	30489	2020-04-14
U-TSH	EL6-1275	Low Risk	C	30489	2020-04-14
Free T4	EL1-1268	Low Risk	C	30308	2020-04-14
Free T3	EL1-1267	Low Risk	C	30309	2020-04-14
Reverse T3	EL9-1274	Low Risk	C	30311	2020-04-14
T Uptake	EL3-1269	Low Risk	C	30313	2020-04-14
Tg (Thyroglobulin)	EL1-1272	Low Risk	C	30490	2020-04-14
TBG (Thyroxine-Binding Globulin)	EL3-1262	Low Risk	C	30316	2020-04-14
Neo-Natal Panel					
Neo-Natal T4	EL1-1240	Low Risk	C	30273	2020-04-14
Neo-Natal TSH	EL1-1239	Low Risk	C	30310	2020-04-14
Neo-Natal TBG	EL3-1242	Low Risk	C	30316	2020-04-14
Neo-Natal 17-OH Progesterone	EL1-1236	Low Risk	C	30324	2020-04-14
Neo-Natal MSUD	EL1-1237	Low Risk	C	30273	2020-04-14
Neo-Natal PKU	EL1-1238	Low Risk	C	30273	2020-04-14
Neo-Natal IRT	EL1-1241	Low Risk	C	30273	2020-04-14
Neo-Natal Total Galactose	EL1-1243	Low Risk	C	30273	2020-04-14
G6PD	EL1-1303	Low Risk	C	30273	2020-04-14
Neo-Natal Biotinidase	EL1-1244	Low Risk	C	30273	2020-04-14
Others					
Procalcitonin	EL3-1309	Low Risk	C	12069016	2020-04-14
Calcitonin	EL3-1292	Low Risk	C	30342	2020-04-14
Renin	EL9-1300	Low Risk	B	43444	2020-04-14

IFA Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
Autoimmune Diseases and others					
ANA Rat Liver IFA Kit	IF17-4002, IF17-4019	Low Risk	C	41420	2020-04-14
ANA Mouse Kidney IFA Kit	IF17-4003	Low Risk	C	41420	2020-04-14
ANA Hep-2 IFA Kit	IF17-4004, IF17-4005, IF17-4018	Low Risk	C	17269	2020-04-14
AMA IFA Kit	IF17-4022, IF17-4023	Low Risk	C	17267	2020-04-14
AAS Rat Kidney Stomach Liver Tissue	IF17-4000	Low Risk	C	30274	2020-04-14

IFA Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
ASMA IFA Kit	IF17-4006, IF17-4015	Low Risk	C	30274	2020-04-14
ATA IFA Kit	IF17-4030, IF174031	Low Risk	C	30274	2020-04-14
ASA IFA Kit	IF17-4008, IF17-4034	Low Risk	C	30274	2020-04-14
nDNA IFA Kit	IF17-4007, IF17-4051, IF17-4052	Low Risk	C	30274	2020-04-14
Endomysial (Primate Endomysial)	IF17-4032, IF17-4033	Low Risk	C	12109016	2020-04-14
Anti-Reticulin IgA	IF17-4041, IF17-4042	Low Risk	C	30526	2020-04-14
Anti-Reticulin IgG	IF17-4043, IF17-4044	Low Risk	C	30526	2020-04-14
C-ANCA	IF17-4059	Low Risk	C	30484	2020-04-14
P-ANCA	IF17-4060	Low Risk	C	30483	2020-04-14
Bacterial Diseases					
Legionella pneumophila 1-6 IFA Poly (HT)	IF17-4063, IF17-4064	Low Risk	C	30694	2020-04-14
Legionella pneumophila 1-6/bdglmj/C Specimen	IF17-4061	Low Risk	C	30694	2020-04-14
Legionella pneumophila 1-6/bdglmj DFA Screen	IF17-4062	Low Risk	C	30694	2020-04-14
FTA-ABS Double Stain (Syphilis) IFA Kit	IF17-4013, IF17-4066	Low Risk	C	32455	2020-04-14
FTA-ABS (T. pallidum)	IF17-4012, IF17-4067	Low Risk	C	32455	2020-04-14
FTA-ABS (Syphilis) Titrable IFA Kit	IF17-4014	Low Risk	C	32455	2020-04-14
Viral diseases					
HSV-1 IgG IFA Kit	IF17-4016	Low Risk	C	39502	2020-04-14
HSV-2 IgG IFA Kit	IF17-4080	Low Risk	C	39502	2020-04-14
HSV-1 IgM IFA Kit	IF17-4017	Low Risk	C	39502	2020-04-14
HSV-2 IgM IFA Kit	IF17-4081	Low Risk	C	39502	2020-04-14
HSV 1&2 IgG	IF17-4078	Low Risk	C	39502	2020-04-14
HSV 1&2 IgM	IF17-4079	Low Risk	C	39502	2020-04-14
EBV-VCA IgG IFA Kit	IF17-4074	Low Risk	C	33971	2020-04-14
EBV-VCA IgM IFA Kit	IF17-4075	Low Risk	C	33971	2020-04-14
EBV-EA IFA Kit	IF17-4077	Low Risk	C	33971	2020-04-14
EBNA IFA Kit	IF17-4076	Low Risk	C	33971	2020-04-14
RMSF Rocky Mountain Spotted Fever (R. rickettsii)	IF17-4065	Low Risk	C	32473	2020-04-14
Measles IgG IFA Kit	IF17-4092	Low Risk	C	44019	2020-04-14
Measles IgM IFA Kit	IF17-4093	Low Risk	C	44019	2020-04-14
Mumps IgG IFA Kit	IF17-4094	Low Risk	C	33908	2020-04-14
Mumps IgM IFA Kit	IF17-4095	Low Risk	C	33908	2020-04-14
RSV IgG (Respiratory Syncytial Virus)	IF17-4096	Low Risk	C	30814	2020-04-14

IFA Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
RSV IgM (Respiratory Syncytial Virus)	IF17-4097	Low Risk	C	30814	2020-04-14
Varicella-Zoster Virus IgG IFA Kit	IF17-4098	Low Risk	C	44027	2020-04-14
Varicella-Zoster Virus IgM IFA Kit	IF17-4099	Low Risk	C	44027	2020-04-14
West Nile Virus IgG	IF17-4100	Low Risk	C	42926	2020-04-14
West Nile Virus IgG	IF17-4101	Low Risk	C	42926	2020-04-14

RT-PCR	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
SARS-CoV-2	PR31-8000	Low Risk	D	42994	2020-04-14
SARS-CoV-2	PR4-8000	Low Risk	D	42994	2020-04-14
SARS-CoV-2 pap-PCR	PR45-8000	Low Risk	D	42994	2020-12-06
SARS-CoV-2/Flu/RSV RT-PCR	PR31-8001	Low Risk	D	42994	2020-12-06

Rapid Tests Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
Tumor Markers Tests					
FOB Cassette	RT27-2182	Low Risk	C	38217	2020-04-14
FOB Strip	RT27-2181	Low Risk	C	38217	2020-04-14
CEA	RT27-2180	Low Risk	C	30288	2020-04-14
AFP	RT27-2179	Low Risk	C	30295	2020-04-14
Cardiac markers					
CK-MB Cassette (Serum/Plasma/Whole Blood)	RT27-2001	Low Risk	C	30499	2020-04-14
C-Reactive Protein (CRP) Cassette (Serum/Plasma/Whole Blood)	RT27-2003	Low Risk	C	30507	2020-04-14
C-Reactive Protein (CRP) Strip (Serum/Plasma/Whole Blood)	RT27-2002	Low Risk	C	30507	2020-04-14
D-Dimer Cassette (Plasma/Whole Blood)	RT27-2004	Low Risk	C	30576	2020-04-14
Myoglobin Cassette (Serum/Plasma/Whole Blood)	RT27-2005	Low Risk	C	30264	2020-04-14
Troponin I Cassette (Serum/Plasma/Whole Blood)	RT27-2007	Low Risk	C	30509	2020-04-14
3 in 1 Troponin I/Myoglobin/CKMB Cassette (Serum/Plasma/Whole Blood)	RT27-2006	Low Risk	C	42649	2020-04-14
Drug Test					
Alcohol Urine Strip	RT27-2010	Low Risk	B	30443	2020-04-14
Alcohol Saliva Strip	RT27-2009	Low Risk	B	30443	2020-04-14
Amphetamine Urine Cassette	RT27-2012	Low Risk	C	30516	2020-04-14
Amphetamine Urine Strip	RT27-2011	Low Risk	C	30516	2020-04-14
Barbiturates Urine Cassette	RT27-2014	Low Risk	C	30517	2020-04-14
Barbiturates Urine Strip	RT27-2013	Low Risk	C	30517	2020-04-14
Buprenorphine Urine Cassette	RT27-2016	Low Risk	C	31584	2020-04-14
Buprenorphine Urine Strip	RT27-2015	Low Risk	C	31584	2020-04-14
Benzodiazepine Urine Cassette	RT27-2018	Low Risk	C	30518	2020-04-14
Benzodiazepine Urine Strip	RT27-2017	Low Risk	C	30518	2020-04-14
Cocaine Urine Cassette	RT27-2022	Low Risk	C	30520	2020-04-14
Cocaine Urine Strip	RT27-2021	Low Risk	C	30520	2020-04-14

Rapid Tests Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
Cotinine Cassette	RT27-2024	Low Risk	C	37270	2020-04-14
Cotinine Strip	RT27-2023	Low Risk	C	37270	2020-04-14
EDDP Urine Cassette	RT27-2028	Low Risk	C	30521	2020-04-14
EDDP Urine Strip	RT27-2027	Low Risk	C	30521	2020-04-14
Fentanyl Urine Cassette	RT27-2030	Low Risk	C	31582	2020-04-14
Fentanyl Urine Strip	RT27-2029	Low Risk	C	31582	2020-04-14
Ketamine Urine Cassette	RT27-2032	Low Risk	C	31582	2020-04-14
Ketamine Urine Strip	RT27-2031	Low Risk	C	31582	2020-04-14
MDMA(Ecstasy) Cassette	RT27-2038	Low Risk	C	30423	2020-04-14
MDMA(Ecstasy) Strip	RT27-2037	Low Risk	C	30423	2020-04-14
Methadone (MTD) Urine Urine Cassette	RT27-2040	Low Risk	C	30521	2020-04-14
Methadone (MTD) Urine Urine Strip	RT27-2039	Low Risk	C	30521	2020-04-14
Methamphetamine Urine Cassette	RT27-2042	Low Risk	C	30423	2020-04-14
Methamphetamine Urine Strip	RT27-2041	Low Risk	C	30423	2020-04-14
Marijuana (THC) Urine Cassette	RT27-2057	Low Risk	C	30519	2020-04-14
Marijuana (THC) Urine Strip	RT27-2056	Low Risk	C	30519	2020-04-14
Opiates Urine Cassette	RT27-2044	Low Risk	C	30522	2020-04-14
Opiates Urine Strip	RT27-2043	Low Risk	C	30522	2020-04-14
Oxycodone Urine Cassette	RT27-2047	Low Risk	C	31584	2020-04-14
Oxycodone Urine Strip	RT27-2046	Low Risk	C	31584	2020-04-14
Phencyclidine (PCP) Urine Cassette	RT27-2049	Low Risk	C	30523	2020-04-14
Phencyclidine (PCP) Urine Strip	RT27-2048	Low Risk	C	30435	2020-04-14
Tricyclic Antidepressants (TCA) Cassette	RT27-2055	Low Risk	C	30524	2020-04-14
Tricyclic Antidepressants (TCA) Strip	RT27-2054	Low Risk	C	30523	2020-04-14
Tramadol Urine Cassette	RT27-2059	Low Risk	C	31582	2020-04-14
Tramadol Urine Strip	RT27-2058	Low Risk	C	31582	2020-04-14
2-Drug Cassette (Any Combination)	RT27-2060	Low Risk	C	30261	2020-04-14
3-Drug Cassette (Any Combination)	RT27-2061	Low Risk	C	30261	2020-04-14
4-Drug Cassette (Any Combination)	RT27-2062	Low Risk	C	30261	2020-04-14
5-Drug Cassette (Any Combination)	RT27-2063	Low Risk	C	30261	2020-04-14
6-Drug Cassette (Any Combination)	RT27-2064	Low Risk	C	30261	2020-04-14
7-Drug Cassette (Any Combination)	RT27-2065	Low Risk	C	30261	2020-04-14
8-Drug Cassette (Any Combination)	RT27-2066	Low Risk	C	30261	2020-04-14
9-Drug Cassette (Any Combination)	RT27-2067	Low Risk	C	30261	2020-04-14
10-Drug Cassette (Any Combination)	RT27-2068	Low Risk	C	30261	2020-04-14
11-Drug Cassette (Any Combination)	RT27-2069	Low Risk	C	30261	2020-04-14
12-Drug Cassette (Any Combination)	RT27-2070	Low Risk	C	30261	2020-04-14
2-Drug Strip (Any Combination)	RT27-2071	Low Risk	C	30261	2020-04-14
3-Drug Strip (Any Combination)	RT27-2072	Low Risk	C	30261	2020-04-14
4-Drug Strip (Any Combination)	RT27-2073	Low Risk	C	30261	2020-04-14
5-Drug Strip (Any Combination)	RT27-2074	Low Risk	C	30261	2020-04-14
6-Drug Strip (Any Combination)	RT27-2075	Low Risk	C	30261	2020-04-14
7-Drug Strip (Any Combination)	RT27-2076	Low Risk	C	30261	2020-04-14
8-Drug Strip (Any Combination)	RT27-2077	Low Risk	C	30261	2020-04-14
9-Drug Strip (Any Combination)	RT27-2078	Low Risk	C	30261	2020-04-14
10-Drug Strip (Any Combination)	RT27-2079	Low Risk	C	30261	2020-04-14

Rapid Tests Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
11-Drug Strip (Any Combination)	RT27-2080	Low Risk	C	30261	2020-04-14
12-Drug Strip (Any Combination)	RT27-2081	Low Risk	C	30261	2020-04-14
Drug Test/Cup					
2-Drug Cup (Any Combination)	RT27-2082	Low Risk	C	30261	2020-04-14
3-Drug Cup (Any Combination)	RT27-2083	Low Risk	C	30261	2020-04-14
4-Drug Cup (Any Combination)	RT27-2084	Low Risk	C	30261	2020-04-14
5-Drug Cup (Any Combination)	RT27-2085	Low Risk	C	30261	2020-04-14
6-Drug Cup (Any Combination)	RT27-2086	Low Risk	C	30261	2020-04-14
7-Drug Cup (Any Combination)	RT27-2087	Low Risk	C	30261	2020-04-14
8-Drug Cup (Any Combination)	RT27-2088	Low Risk	C	30261	2020-04-14
9-Drug Cup (Any Combination)	RT27-2089	Low Risk	C	30261	2020-04-14
10-Drug Cup (Any Combination)	RT27-2090	Low Risk	C	30261	2020-04-14
11-Drug Cup (Any Combination)	RT27-2091	Low Risk	C	30261	2020-04-14
12-Drug Cup (Any Combination)	RT27-2092	Low Risk	C	30261	2020-04-14
Infectious Diseases and others					
Legionella Urinary Antigen Cassette	RT27-2147	Low Risk	C	30692	2020-04-14
Legionella Urinary Antigen Strip	RT27-2146	Low Risk	C	30692	2020-04-14
Adeno/Rotavirus Antigen Cassette	RT27-2131	Low Risk	C	42994	2020-04-14
Adeno Antigen Cassette	RT27-2132	Low Risk	C	42994	2020-04-14
Rotavirus Antigen Cassette	RT27-2161	Low Risk	C	30815	2020-04-14
Chagas Cassette	RT27-2133	Low Risk	C	30820	2020-04-14
Chikungunya IgG/IgM Cassette	RT27-2135	Low Risk	C	42994	2020-04-14
Gonorrhoea Cassette	RT27-2140	Low Risk	C	38851	2020-04-14
Influenza A&B Cassette	RT27-2145	Low Risk	C	39466	2020-04-14
Leishmania IgG/IgM Cassette	RT27-2149	Low Risk	C	30823	2020-04-14
Leishmania Cutaneous Strip	RT27-2148	Low Risk	C	30823	2020-04-14
Leptospira IgG/IgM	RT27-2150	Low Risk	C	30716	2020-04-14
Syphilis Cassette	RT27-2172	Low Risk	C	30687	2020-04-14
Syphilis Strip	RT27-2173, RT24-2173	Low Risk	C	30687	2020-04-14
Mononucleosis Cassette (Mono) (S/P)	RT27-2177	Low Risk	C	30826	2020-04-14
Strep A Cassette	RT27-2169	Low Risk	C	30826	2020-04-14
Strep A Strip	RT27-2168	Low Risk	C	30826	2020-04-14
Strep B Cassette	RT27-2171	Low Risk	C	30827	2020-04-14
Strep B Strip	RT27-2170	Low Risk	C	30827	2020-04-14
H1N1 Strip	RT40-2209	Low Risk	C	39461	2020-04-14
H. Pylori Ab Cassette (Serum/Plasma)	RT27-2141	Low Risk	B	30825	2020-04-14
H. Pylori Ab Cassette (Serum/Plasma/Whole Blood)	RT27-2142, RT24-2142	Low Risk	B	30825	2020-04-14
H. Pylori Antigen Cassette	RT27-2143, RT24-2203	Low Risk	B	30689	2020-04-14
HAV IgM	RT27-2108	Low Risk	C	30720	2020-04-14
Dengue IgG&IgM	RT27-2138, RT24-2197	Low Risk	C	42994	2020-04-14
Dengue NS1	RT24-2139	Low Risk	C	42994	2020-04-14
Dengue IgG/IgM/NS1	RT24-2208	Low Risk	C	42994	2020-04-14

Rapid Tests Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
Malaria P.f./Pv	RT24-2204	Low Risk	C	30674	2020-04-14
Malaria Pan	RT24-2206	Low Risk	C	30674	2020-04-14
Malaria P.f./Pan	RT24-2205, RT27-2154	Low Risk	C	30674	2020-04-14
Malaria P.f. Cassette	RT24-2207, RT27-2151	Low Risk	C	30674	2020-04-14
Malaria P.f. Strip	RT27-2152	Low Risk	C	30674	2020-04-14
Malaria P.f./vivax	RT27-2153	Low Risk	C	30674	2020-04-14
Norovirus	RT27-2156	Low Risk	C	32459	2020-04-14
Salmonella typhi Antigen Cassette	RT27-2163	Low Risk	C	30709	2020-04-14
Salmonella typhi IgG/IgM Cassette	RT27-2164	Low Risk	C	30709	2020-04-14
Salmonella typhi/paratyphi antigen	RT27-2165	Low Risk	C	30709	2020-04-14
Scrub typhus IgG Strip	RT4-2166	Low Risk	C	30717	2020-04-14
Scrub typhus IgM Strip	RT4-2167	Low Risk	C	30717	2020-04-14
Zika Virus IgG/IgM Cassette	RT27-2178	Low Risk	C	42994	2020-04-14
COVID-19 IgG/IgM	RT24-2198, RT28-2198, RT45-2198	Low Risk	D	44022	2020-04-14
SARS-CoV2 Antigen Rapid Test	RT45-2214	Low Risk	D	44022	2020-08-24
Tuberculosis (TB) Cassette	RT27-2175	Low Risk	C	44020	2020-04-14
Tuberculosis (TB) Strip	RT27-2174	Low Risk	C	44020	2020-04-14
HEV IgG/IgM	RT27-2119	Low Risk	D	30756	2020-04-14
Cryptococcus Ag	RT27-2137	Low Risk	C	37746	2020-04-14
Hantavirus IgG/IgM	RT27-2144	Low Risk	C	15048014	2020-04-14
Mycoplasma pneumoniae Ag	RT27-2155	Low Risk	C	17311	2020-04-14
Rickettsia IgG/IgM	RT24-2160	Low Risk	C	30717	2020-04-14
RSV	RT27-2162	Low Risk	C	30814	2020-04-14
Tetanus	RT27-2176	Low Risk	C	38876	2020-04-14
Fertility					
FSH Urine Cassette	RT27-2094	Low Risk	B	30512	2020-04-14
FSH Urine Strip	RT27-2093	Low Risk	B	30512	2020-04-14
Ovulation					
LH Urine Cassette	RT27-2106	Low Risk	B	30515	2020-04-14
LH Urine Strip	RT27-2105	Low Risk	B	30515	2020-04-14
Pregnancy					
hCG 10 mIU/ml Midstream	RT27-2099	Low Risk	B	30513	2020-04-14
hCG 20 mIU/ml Midstream	RT27-2102	Low Risk	B	30513	2020-04-14
hCG 10mIU/ml urine Cassette	RT27-2095	Low Risk	B	30513	2020-04-14
hCG 10mIU/ml urine Strip	RT27-2097	Low Risk	B	30513	2020-04-14
hCG 10mIU/ml urine/serum	RT27-2098	Low Risk	B	30513	2020-04-14
hCG 20 mIU/ml urine Cassette	RT27-2101	Low Risk	B	30513	2020-04-14
hCG 20 mIU/ml urine Strip	RT27-2100	Low Risk	B	30513	2020-04-14
hCG 10mIU/ml urine/serum/p	RT27-2096	Low Risk	B	30513	2020-04-14
hCG 20 mIU/ml urine/serum/p Cassette	RT27-2104	Low Risk	B	30513	2020-04-14
hCG 20 mIU/ml urine/serum/p Strip	RT27-2103	Low Risk	B	30513	2020-04-14
Others					

Rapid Tests Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
Micro-Albumin (HAS) Strip	RT27-2197	Low Risk	C	30246	2020-04-14
Ferritin	RT27-2196	Low Risk	C	30377	2020-04-14
H-FABP	RT27-2107	Low Risk	C	1230190	2020-04-14
Nt-proBNP	RT27-1157	Low Risk	C	12130190	2020-04-14
Procalcitonin (S/P/WB)	RT27-2158	Low Risk	C	12069016	2020-04-14
Procalcitonin (S/P)	RT27-2159	Low Risk	C	12069016	2020-04-14
Urine Reagent Strips					
URS-1G	RT27-2185	Low Risk	B	17419	2020-04-14
URS-2PK	RT27-2186	Low Risk	B	30226	2020-04-14
URS-3 GKpH	RT27-2187	Low Risk	B	30226	2020-04-14
URS-4 GKpHB	RT27-2188	Low Risk	B	30226	2020-04-14
URS-5GKpHBP	RT27-2189	Low Risk	B	30226	2020-04-14
URS-6GKpHBPBili	RT27-2190	Low Risk	B	30226	2020-04-14
URS-7GKpHBPBiliU	RT27-2191	Low Risk	B	30226	2020-04-14
URS-8GKpHBPBiliUN	RT27-2192	Low Risk	B	30226	2020-04-14
URS-9GKpHBPBiliUNS	RT27-2193	Low Risk	B	30226	2020-04-14
URS-10GKpHBPBiliUNSL	RT27-2194	Low Risk	B	30226	2020-04-14
URS-11	RT27-2195	Low Risk	B	30226	2020-04-14

Serology Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
C- Reactive Protein (CRP)	SL25-3002, SL25-3003	Low Risk	C	30499	2020-04-14
RF	SL25-3008, SL25-3009	Low Risk	C	30500	2020-04-14
Anti- Streptolysin O(ASO)	SL25-3000, SL25-3001	Low Risk	C	30495	2020-04-14
Infectious Mononucleosis Screening (Mono)	SL25-3004, SL25-3005	Low Risk	C	30810	2020-04-14
RPR	SL25-3011, SL25-3012	Low Risk	C	17393	2020-04-14
Lupus Erythematosus (SLE)	SL25-3007	Low Risk	C	30487	2020-04-14
TPHA	SL25-3016	Low Risk	C	32453	2020-04-14
Rotavirus	SL25-3010	Low Risk	C	17381	2020-04-14
S. Aureus	SL25-3013	Low Risk	C	33887	2020-04-14
Streptococci Lancefield grouping	SL25-3015	Low Risk	C	17389	2020-04-14
VDRL Antigen	SL25-3017	Low Risk	C	17395	2020-04-14
PARATYPHOID A (Salmonella, flagellar a antigen)	SL25-3022	Low Risk	C	39453	2020-04-14
PARATYPHOID B (Salmonella, flagellar b antigen)	SL25-3023	Low Risk	C	39453	2020-04-14
PARATYPHOID C (Salmonella typhi, flagellar c antigen)	SL25-3024	Low Risk	C	39453	2020-04-14
SALMONELLA Group A Antigen (somatic antigen)	SL25-3028	Low Risk	C	39453	2020-04-14
SALMONELLA Group B Antigen (somatic antigen)	SL25-3029	Low Risk	C	39453	2020-04-14

Serology Device Group	Ref. No.	IVDD Risk class	IVDR Risk class	GMDN code	First CE-marking
SALMONELLA Group C Antigen (somatic antigen)	SL25-3030	Low Risk	C	39453	2020-04-14
TYPHOID H (Salmonella typhi, flagellar d antigen)	SL25-3031	Low Risk	C	39453	2020-04-14
TYPHOID O (Salmonella typhi, somatic Group D antigen)	SL25-3032	Low Risk	C	39453	2020-04-14
Brucella Melitensis	SL25-3018	Low Risk	C	39536	2020-04-14
Brucella Abortus	SL25-3019	Low Risk	C	39536	2020-04-14
PROTEUS OX2 (somatic antigen)	SL25-3026	Low Risk	C	39543	2020-04-14
PROTEUS OX19 (somatic antigen)	SL25-3025	Low Risk	C	39543	2020-04-14
PROTEUS OXK (somatic antigen)	SL25-3027	Low Risk	C	39543	2020-04-14



Certificate of Registration

This is to certify the Quality Management System of:

MONOCENT, INC.
9237 Eton Avenue
Chatsworth, CA 91311

has been assessed and found to be in compliance with the requirements of

ISO 13485:2016

for the following scope:

**Manufacturing and Distribution of IVD Products
(Serology, Rapid, ELISA, CLIA, IFA Test Systems and Instrumentation)**

Medical Device Code: In Vitro Diagnostics (IVD) & Non-active Medical Device

Certificate Number: **SARA-2019-CA-0253-02-A**

Originally Registered:
January 10, 2020

Latest Issue:
December 20, 2022

Certification Cycle:
January 10, 2023 – January 9, 2026

Expiration Date:
January 9, 2026

President, SARA Registrar



MSCB-194

This registration is subject to the company maintaining its system to the required standard which will be monitored annually by SARA Registrar.
This certificate remains the property of Standards American Registrations Authority (SARA Registrar) and shall be returned immediately upon request.
SARA Registrar Headquarter Mailing: 1807H Santa Rita Road, #175, Pleasanton, CA 94566



Certificate of Registration

This is to certify the Quality Management System of:

MONOCENT, INC.
9237 Eton Avenue
Chatsworth, CA 91311

has been assessed and found to be in compliance with the requirements of

ISO 9001:2015

for the following scope:

**Manufacturing and Distribution of IVD Products
(Serology, Rapid, ELISA, CLIA, IFA Test Systems and Instrumentation)**

IAF Code: 31 & 35

Certificate Number: **SARA-2019-CA-0253-01-A**

Originally Registered:
January 10, 2020

Latest Issue:
December 20, 2022

Certification Cycle:
January 10, 2023 – January 9, 2026

Expiration Date:
January 9, 2026

President, SARA Registrar



MSCB-194

This registration is subject to the company maintaining its system to the required standard which will be monitored annually by SARA Registrar. This certificate remains the property of Standards American Registrations Authority (SARA Registrar) and shall be returned immediately upon request.
SARA Registrar Headquarter Mailing: 1807H Santa Rita Road, #175, Pleasanton, CA 94566



ENGLISH

ANA SCREEN IgG ELISA TEST SYSTEM



EL1-1009



96 TESTS



INTENDED USE

The Monocent, Inc.'s ANA IgG ELISA Test System is an enzyme-linked immunosorbent assay (ELISA) for the detection of IgG class antibodies to ANA in human serum or plasma.

SUMMARY AND EXPLANATION

Antinuclear antibodies (ANA) are frequently present in patients with systemic lupus erythematosus (SLE) and, less commonly, in other autoimmune diseases Rheumatoid arthritis, Collagen vascular diseases, chronic liver diseases and systemic sclerosis (scleroderma). ANA bind to several nuclear antigens including DsDNA, SSDNA, RNP, Sm, SSA and SSB. ANA frequency increases with age in apparently healthy people, especially women after the age of 45 years. ANA ELISA is widely used as a screening procedure for different autoimmune diseases.

PRINCIPLE OF THE TEST

Diluted patient serum is added to wells coated with purified nuclear antigens. ANA IgG specific antibody, if present, binds to the antigen. All unbound materials are washed away and the enzyme conjugate is added to bind to the antibody-antigen complex, if present. Excess enzyme conjugate is washed off and substrate is added. The plate is incubated to allow the hydrolysis of the substrate by the enzyme. The intensity of the color generated is proportional to the amount of IgG specific antibody in the sample.

MATERIALS AND COMPONENTS

• Microwells coated with nuclear antigens	12x8x1
• Sample Diluent: 1 bottle (ready to use)	22ml
• Calibrator 1 Vial (ready to use)	1ml
• Positive Control 1 vial (ready to use)	1ml
• Negative Control 1 vial (ready to use)	1ml
• Enzyme conjugate: 1 bottle (ready to use)	12ml
• TMB Substrate: 1 bottle (ready to use)	12ml
• Stop Solution: 1 bottle (ready to use)	12ml
• Wash concentrate 20X: 1 bottle	25ml

MATERIALS REQUIRED BUT NOT PROVIDED

- Distilled or deionized water.
- Precision pipettes.
- Disposable pipette tips.
- ELISA reader capable of reading absorbance at 450 nm.
- Absorbance paper or paper towel.

STORAGE CONDITIONS

- Store the kit at 2-8°C.
- Keep microwells sealed in a dry bag with desiccants.
- The reagents are stable until expiration of the kit.
- Do not expose test reagents to heat, sun or strong light.

PRECAUTIONS

1. For Research Use Only. Not for use in diagnostic procedures.
2. For Laboratory use.
3. Not for Internal or External Use in Humans or Animals.
4. There should be no eating or drinking within work area.
5. Always wear gloves and a protective lab coat.
6. No pipetting should be done by mouth. Handle all specimens and reagents as potentially infectious and biohazardous.
7. Do not add sodium azide to samples as preservative.
8. Do not use external controls containing sodium azide.
9. Use disposable pipette tips to avoid contaminating chromogenic substrate reagent. Discard reagent if it turns blue.
10. Do not pour chromogenic substrate back into container after use.
11. Do not freeze reagents.
12. Do not mix reagents from different kit lot numbers.
13. Keep reagents out of direct sunlight.
14. Handle stop reagent with care, since it is corrosive.
15. Bring all reagents to room temperature.
16. Viscous forensic samples should always be diluted in phosphate buffered saline or distilled water prior to pipetting.
17. Ensure the bag containing the micro-plate strips and desiccant is sealed well, if only a few strips are used.

SPECIMEN COLLECTION

1. Collect blood specimens and separate the serum.
2. Specimens may be refrigerated at 2-8°C for up to seven days or frozen for up to six months. Avoid repetitive freezing and thawing.

REAGENT PREPARATION

Prepare 1x wash buffer by adding the contents of the bottle (25 ml, 20x) to 475 ml of distilled or deionized water. Store at room temperature (20-25°C).

PREPARATION FOR TEST

Bring all specimens and kit reagents to room temperature (20-25°C) and gently mix.

TEST PROCEDURE

1. Place the desired number of coated strips into the holder.
2. Negative control, positive control, and calibrator are ready to use. Prepare 1:21 dilution of test samples, by adding 10 µl of the sample to 200 µl of sample diluent. Mix well.
3. Dispense 100 µl of diluted sera, calibrator and controls into the appropriate wells. For the reagent blank, dispense 100 µl sample diluent in 1a well position. Tap the holder to remove air bubbles from the liquid and mix well. Incubate for 20 minutes at room temperature.
4. Remove liquid from all wells. Wash wells three times with 300 µl of 1x wash buffer. Blot on absorbance paper or paper towel.
5. Dispense 100 µl of enzyme conjugate to each well and incubate for 20

minutes at room temperature.

6. Remove enzyme conjugate from all wells. Wash wells three times with 300 µl of 1x wash buffer. Blot on absorbance paper or paper towel
7. Dispense 100 µl of TMB substrate and incubate for 10 minutes at room temperature.
8. Add 100 µl of stop solution.
9. Read OD at 450 nm using ELISA reader within 15 min. A dual wavelength is recommended with reference filter of 600-650 nm.

CALCULATION OF RESULTS

1. Check Calibrator Factor (CF) value on the calibrator bottle. This value might vary from lot to lot. Make sure you check the value on every kit.
2. Calculate the cut-off value: Calibrator OD x Calibrator Factor (CF).
3. Calculate the Ab (Antibody) Index of each determination by dividing the O.D. value of each sample by cut-off value.

REFERENCES

1. Emlen w; o'neill I clinical significance of antinuclear antibodies: comparison of detection with immunofluorescence and enzyme-linked immunosorbent assays. Arthritis rheum 1997;40(9):1612-8.
2. Gonz'alez c; martin t; arroyo t; garc'ia-isidoro m; navajo ja; gonz'alez-buitrago jm. Comparison and variation of different methodologies for the detection of autoantibodies to nuclear antigens (ana). J clin lab anal 1997;11(6):388-92.
3. Parveen s; morshed sa; nishioka m. High prevalence of antibodies to recombinant cenp-b in primary biliary cirrhosis: nuclear immunofluorescence patterns and elisa reactivities. J gastroenterol hepatol 1995;10(4):438-45.
4. Welin henriksson e; hansson h; karlsson-parra a; pettersson i. Autoantibody profiles in canine ana-positive sera investigated by immunoblot and elisa. Vet immunol immunopathol 1998;61(2-4):157-70.
5. Koh wh; dunphy j; whyte j; dixey j; mchugh nj. Characterisation of anticytoplasmic antibodies and their clinical associations [see comments]. Ann rheum dis 1995;54(4):269-73.
6. Spronk pe; bootsma h; horst g; huitema mg; limburg pc; cohen tervaert jw; kallenberg cg. Antineutrophil cytoplasmic antibodies in systemic lupus erythematosus. Br j rheumatol 1996;35(7):625-31.




**Manufactured by
Monocent, Inc**

9237 Eton Ave. Chatsworth, CA 91311, USA

Info@monocent.com | Tel: 424-310-0777

www.monocent.com



CEpartner4U

ESDOORNLAAN 13, 3951DB MAARN, THE NETHERLANDS.

www.cepartner4u.com

Revision Date: 2017-04-13



ENGLISH

ScL-70 IgG ELISA TEST SYSTEM



REF

EL1-1039



96 TESTS

IVD

INTENDED USE

The Monocent, Inc.'s ScL-70 ELISA Kit is intended for the detection of IgG antibody to ScL-70 in human serum or plasma.

SUMMARY AND EXPLANATION

Systemic autoimmune disease is characterized by the presence of circulating auto-antibodies directed to a wide variety of cellular antigens. Systemic lupus erythematosus (SLE), commonly referred to as Lupus is the best known of these diseases. Other possible connective tissue diseases include mixed connective tissue disease (MCTD), Sjogren syndrome, scleroderma, and polymyositis/dermatomyositis. The majority can be diagnosed by clinical presentation and their antibody profiles to the various antigens involved, which include dsDNA, SM, RNP, SSA, SSB ScL-70, Jo1 and Histones. Therefore, immunoassays for autoantibodies are useful for diagnostic and prognostic evaluations of autoimmune disease. ScL-70 IgG antibodies react with human topoisomerase I of 100 kd molecular weight as well as its 70 kd fragment. ScL-70 antibodies are present in 20-40% of diffuse scleroderma patients and in about 20% of patients with limited scleroderma. ScL-70 antibodies are sometimes reported in classical SLE without features of scleroderma, which may explain the unexpected co-existence of marker autoantibodies for SLE and scleroderma. Some patients with silica-associated systemic sclerosis (SSc) have ScL-70 antibodies.

PRINCIPLE OF THE TEST

Diluted patient serum is added to wells coated with purified antigen. IgG specific antibody, if present, binds to the antigen. All unbound materials are washed away and the enzyme conjugate is added to bind to the antibody-antigen complex, if present. Excess enzyme conjugate is washed off and substrate is added. The plate is incubated to allow the hydrolysis of the substrate by the enzyme. The intensity of the color generated is proportional to the amount of IgG specific antibody in the sample.

MATERIALS AND COMPONENTS

- Microwells coated with ScL-70 antigen - 12x8x1
- Sample Diluent: 1 bottle (ready to use) - 22ml
- Calibrator: 1 Vial (ready to use) - 1ml

- Positive Control: 1 vial (ready to use) - 1ml
- Negative Control: 1 vial (ready to use) - 1ml
- Enzyme conjugate: 1 bottle (ready to use) - 12ml
- TMB Substrate: 1 bottle (ready to use) - 12ml
- Stop Solution: 1 bottle (ready to use) - 12ml
- Wash concentrate 20X: 1 bottle - 25ml

MATERIALS REQUIRED BUT NOT PROVIDED

- Distilled or deionized water
- Precision pipettes
- Disposable pipette tips
- ELISA reader capable of reading absorbance at 450nm
- Absorbance paper or paper towel
- Graph paper

STORAGE CONDITIONS

- Store the kit at 2-8°C.
- Keep microwells sealed in a dry bag with desiccants.
- The reagents are stable until expiration of the kit.
- Do not expose test reagents to heat, sun or strong light.

PRECAUTIONS

1. For Research Use Only. Not for use in diagnostic procedures.
2. For Laboratory Use.
3. Potential biohazardous materials:
The calibrator and controls contain human source components which have been tested and found non-reactive for hepatitis B surface antigen as well as HIV antibody with FDA licensed reagents. However, there is no test method that can offer complete assurance that HIV, Hepatitis B virus or other infectious agents are absent. These reagents should be handled at the Biosafety Level 2, as recommended in the Centers for Disease Control/National Institutes of Health manual, "Biosafety in Microbiological and Biomedical Laboratories" 1984.
4. Optimal results will be obtained by strict adherence to the test protocol. Precise pipetting as well as following the exact time and temperature requirements is essential.
5. Do not pipette by mouth. Do not smoke, eat, or drink in the areas in which specimens or kit reagents are handled.
6. The components in this kit are intended for use as an integral unit. The components of different lots should not be mixed.
7. This product contains components preserved with sodium azide. Sodium azide may react with lead and copper plumbing to form explosive metal azide. On disposal, flush with a large volume of water.

SPECIMEN COLLECTION

1. Collect blood specimens and separate the serum.
2. Specimens may be refrigerated at 2-8°C for up to seven days or frozen for up to six months. Avoid repetitive freezing and thawing.

REAGENT PREPARATION

Prepare 1X Wash buffer by adding the contents of the bottle (25 ml, 20X) to 475 ml of distilled or deionized water. Store at room temperature (20-25°C).

TEST PROCEDURE

Bring all specimens and kit reagents to room temperature (20-25°C) and gently mix.

1. Place the desired number of coated strips into the holder.
2. Negative control, positive control, and calibrator are ready to use. Prepare 1:21 dilution of test samples, by adding 10 µl of the sample to 200 µl of sample diluent. Mix well.
3. Dispense 100 µl of diluted sera, calibrator and controls into the appropriate wells. For the reagent blank, dispense 100µl sample diluent in 1A well position. Tap the holder to remove air bubbles from the liquid and mix well. Incubate for 20 minutes at room temperature.
4. Remove liquid from all wells. Wash wells three times with 300 µl of 1X wash

buffer. Blot on absorbance paper or paper towel.

5. Dispense 100 µl of enzyme conjugate to each well and incubate for 20 minutes at room temperature.
6. Remove enzyme conjugate from all wells. Wash wells three times with 300 µl of 1X wash buffer. Blot on absorbance paper or paper towel.
7. Dispense 100 µl of TMB substrate and incubate for 10 minutes at room temperature.
8. Add 100 µl of stop solution.
9. Read O.D. at 450 nm using ELISA reader within 15 min. A dual wavelength is recommended with reference filter of 600-650 nm.

CALCULATION OF RESULTS

1. Check Calibrator Factor (CF) value on the calibrator bottle. This value might vary from lot to lot. Make sure you check the value on every kit.
2. Calculate the cut-off value: Calibrator OD x Calibrator Factor (CF).
3. Calculate the Ab (Antibody) Index of each determination by dividing the O.D. value of each sample by cut-off value.

LIMITATIONS OF THE TEST

Lipemic or hemolyzed samples may cause erroneous results.

REFERENCES

1. Hildebrandt S, Weiner ES, Senecal J-L, Noell GS, Earnshaw WC, Rothfield NF. Autoantibodies to topoisomerase I (ScL-70): analysis by gel diffusion, immunoblot, and enzyme-linked immunosorbent assay. Clin Immunol Immunopathol 1990;57:399-410.
2. Hildebrandt S, Weiner E, Senécal J, et al. The IgG, IgM, and IgA isotypes of anti- topoisomerase I and anticentromere autoantibodies. Arthritis Rheum 1990;33:724-7.
3. Steen VD, Powell DL, Medsger TA Jr. Clinical correlations and prognosis based on serum autoantibodies in patients with systemic sclerosis. Arthritis Rheum 1988;31:196-203.
4. Hildebrandt S, Jäckh G, Weber S, Peter H-H. A long-term longitudinal isotypic study of anti- topoisomerase I autoantibodies. Rheumatol Int 1993;12:321-4.
5. Briggs DC, Vaughan RW, Welsh KI, Myers A, DuBois RM, Black CM. Immunogenetic prediction of pulmonary fibrosis in systemic sclerosis. Lancet 1991;338:661-2.
6. Mukai M, Sagawa A, Atsumi T, et al. 3 cases of anti-ScL-70 (topoisomerase I) antibody associated with central nervous system lupus without renal disorder. J Rheumatol 1993;20:91594-7.
7. Stojanov L, Satoh M, Dooley MA, Kuwana M, Jennette JC, Reeves WH. Autoantibodies to topoisomerase I in a patient with systemic lupus erythematosus without features of scleroderma. Lupus 1995;4:314-7.
8. Juarez C, Vila JL, Gelpi C, et al. Characterization of the antigen reactive with anti-ScL-70 antibodies and its application in an enzyme-linked immunosorbent assay. Arthritis Rheum 1988;31:108-15.



**Manufactured by
Monocent, Inc.**

9237 Eton Ave. Chatsworth, CA 91311, USA

Info@monocent.com | Tel: 424-310-0777

www.monocent.com

EC REP

CEpartner4U

ESDOORNLAAN 13, 3951DB MAARN, THE NETHERLANDS.

www.cepartner4u.com