

ACON BIOTECH (HANGZHOU) CO.,LTD

CERTIFICATE OF ANALYSIS

PRODUCT NAME: HBcAb EIA Test Kit

PURCHASE NO.: 37740068A

BATCH NO.: 2008762

EXPIRATION DATE: 2021/11/08

QUANTITY: 11kits

SPECIFICATION	STANDARD	CONCLUSION
Appearance	No visible deposit or floccules in liquid component	Complies
Negative Control	OD > 1.000	Complies
Positive Control	OD < 0.080	Complies
QCO reference	$0.500 \leq \text{QCO/Cutoff value} \leq 1.500$	Complies
Compliance rate of negative reference	15/15	Complies
Compliance rate of positive references	15/15	Complies
Precision criteria	CV < 20%	Complies

Quality Department: Hedy Zhang

Date: Sep. 23, 2020

