



Sysmex Europe SE · Bornbarch 1 · 22848 Norderstedt · Germany

To whom it may concern

Sysmex Europe SE
Bornbarch 1
22848 Norderstedt, Germany
Phone +49 40 527 26-0
Fax +49 40 527 26-100
info@sysmex-europe.com

LETTER OF AUTHORIZATION

Whereas Sysmex Europe SE ("Sysmex"), who are established, reputable and authorised representative in Europe, Africa and Middle East (EMEA region), officially announced by the manufacturer Sysmex Corporation, having its principal place of business at 1-5-1 Wakinohama-Kaigandoori, Chuo-ku, Kobe 651-0073, Japan, and having the power to grant authorizations to local representatives within the above mentioned markets,

do hereby declare that the company

ECHIPAMED Plus SRL
Valea Trandafirilor 24 "B", off. 80
MD-2001 Chisinau, Moldova (the "COMPANY")

is our distributor and local representative for the following Sysmex products:

Sysmex Haematology- and Urine- Analysers
with Reagents, Accessories, Software and Spare Parts
(the "**Products**")

In the territory of Moldova (the "**TERRITORY**")

The **COMPANY** is therefore authorized to carry out all commercial and support activities for the **PRODUCTS** including sales, marketing, application, registration and field service support in the **TERRITORY**.

The **COMPANY** is aware that this special authorisation is limited to the above listed **PRODUCTS** and does not create any further rights for the **COMPANY**.

We hereby grant our warranty following our general conditions of sale for the **PRODUCTS** delivered, consisting of and limited to:

Company Location Norderstedt
Registered AG Kiel
HRB 24262 KI
VAT-ID DE 118 687 842
WEEE/ElektroG Reg. Nr. DE 159 56 453

Chairman of the
Supervisory Board
Iwane Matsui

Management Board
Alain Baverel (CEO)
Yuki Hyogu
Dr Sanjeev Kumar
Stefanie Schaal
Matthias Voelkel

COMMERZBANK AG
Hamburg
IBAN DE20 2004 0000 0287 1879 00
SWIFT/BIC COBADEFFXXX

www.sysmex-europe.com



Sysmex Europe GmbH · Bornbarch 1 · 22848 Norderstedt · Germany

Sysmex Europe GmbH
Bornbarch 1
22848 Norderstedt, Germany
Phone +49 40 52726-0
Fax +49 40 52726-100
info@sysmex-europe.com

To whom it may concern

DECLARATION

We, Sysmex Europe GmbH, located at Bornbarch 1, 22848 Norderstedt, Germany, who are established, reputable and authorised representative in Europe (EC REP), Africa and Middle East (EMEA region), officially announced by the manufacturer Sysmex Corporation, Japan hereby confirm that our Haematology Analysers

XN-1000, XN-2000, XP-300 and UX-2000

are 'closed systems' and only to be used together with Sysmex Reagents, Sysmex Controls and Sysmex Calibrators. Every change of this closed system by the user is regarded as 'non-specified use' by Sysmex.

The technology of all Sysmex IVD analysers is fine-tuned together with the corresponding reagents used on each single analyser. Thereby, using Sysmex reagents maintains optimum performance as well as optimal and enhanced accuracy of the system. There is a high interdependency between research and using/finding optimal reagents for any new parameter(s). As Sysmex is actively doing research, it is thereby ensured that Sysmex reagents fulfil best practice requirements for any research parameter(s), which later will become diagnostic parameter(s) after the legally required procedures under Annex VIII-IVD-Directive 98/79/EC.


Therefore Sysmex Reagents offer best performance on Sysmex Analysers.

The Reagents, Controls and Calibrators listed on the following page are allowed to be used on Sysmex Haematology Analysers.

On behalf of Sysmex Europe GmbH

Date: January 14th, 2016

Place: 22848 Norderstedt, Germany



Sysmex Europe GmbH

i.A. Katharina Paucke
Manager Regulatory Affairs

„Design and specifications may be subject to changes due to further product development. Changes are confirmed by their appearance on a newer document and verification according to its date of issue.“

Company Location Norderstedt
Registered AG Kiel
HRB 4179
VAT-ID DE 118 687 842
WEEE/ElektroG Reg. Nr. DE 159 56 453

Managing Directors
Fernando Andreu
Kensuke Iizuka
Takeshi Kubota
Kazuya Obe
Dr. Michael Schaefer
Dr. Jürgen Schulze
Matthias Völkel

The Bank of Tokyo-Mitsubishi UFJ, Ltd. Hamburg
Bank ID-Code 300 107 00
Account Nr. 03 77 13
IBAN DE03 3001 0700 0000 0377 13
SWIFT/BIC Code BOTKDE3X



Reagents, Controls and Calibrators that are allowed to be used on Sysmex Haematology Analysers:

XN-1000	XN-2000	XP-300	UX-2000
CELLPACK DCL	CELLPACK DCL	CELLPACK	UX II PACK-BAC
CELLPACK DST	CELLPACK DST	STROMATOLYSER-WH	UX II PACK-SED
CELLPACK DFL	CELLPACK DFL	CELLCLEAN	UX II SEARCH -BAC
Lysercell WDF	Lysercell WDF	EIGHTCHECK-3WP	UX II SEARCH -SED
Lysercell WNR	Lysercell WNR	SCS-1000	UX II SHEATH
Lysercell WPC	Lysercell WPC		UX CLEAN -C
SULFOLYSER	SULFOLYSER		MEDITAPE II 10U
Fluorocell PLT	Fluorocell PLT		MEDITAPE II 9U
Fluorocell RET	Fluorocell RET		MEDITAPE II 10K
Fluorocell WDF	Fluorocell WDF		UF II CONTROL
Fluorocell WNR	Fluorocell WNR		MEDITAPE CHECK 1
Fluorocell WPC	Fluorocell WPC		MEDITAPE CHECK 2
CELLCLEAN	CELLCLEAN		UF II Calibrator
CELLCLEAN AUTO	CELLCLEAN AUTO		
XN CHECK	XN CHECK		
XN CHECK BF	XN CHECK BF		
XN CAL	XN CAL		
XN CAL PF	XN CAL PF		

End of list



To whom it may concern

Composition of Sysmex Reagents

The composition of Sysmex Reagents is highly confidential! Therefore only active components and those classified as dangerous must be declared on the product labelling.

The below listed table gives an overview of these components in Sysmex Reagents:

Cellpack	Sodium chloride 6.38 g/L Boric acid 1.0 g/L Sodium tetraborate 0.2 g/L EDTA-2K 0.2 g/L
CELLPACK DCL	Sodium chloride 0.7% Tris buffer 0.2% EDTA-2K 0.02%
CELLPACK DST	Sodium chloride 15.7% Tris buffer 4.3% EDTA-2K 0.4%
CELLPACK DFL	Tricine buffer 0.17%
CELLCLEAN	Sodium Hypochlorite (available chlorine concentration 5.0%)
CELLCLEAN AUTO	Sodium Hypochlorite (available chlorine concentration 5.0%)
Stromatolyser-WH	Organic quaternary ammonium salt 8.5 g/L Sodium chloride 0.6 g/L
Lysercell WDF	Organic quaternary ammonium salts 0.07% Nonionic surfactant 0.17%
Lysercell WNR	Organic quaternary ammonium salts 0.20% Nonionic surfactant 0.10%
Lysercell WPC	Anionic surfactant 0.03% Nonionic surfactant 0.12%
Sulfolyser	Sodium lauryl sulfate 1.7 g/L
Fluorocell PLT	Oxazine 0.003% Ethylene glycol 99.9%
Fluorocell RET	Polymethine 0.03% Methanol 7.9% Ethylene glycol 92.0%
Fluorocell WDF	Polymethine 0.002% Methanol 3.0% Ethylene glycol 96.9%
Fluorocell WNR	Polymethine 0.005% Ethylene glycol 99.9%

Fluorocell WPC	Polymethine 0.004% Ethanol 15.1% Ethylene glycol 84.8%
XN CHECK	quality control material; includes stabilized human red blood cells, human white blood cells, a platelet and nucleated red blood cell component in a preservative medium.
XN CHECK BF	quality control material; includes stabilized human red blood cells and white blood cells in a preservative medium.
XN CAL	calibrator; includes stabilized human red blood cells, human white blood cells, a platelet and nucleated red blood cell component in a preservative medium.
XN CAL PF	calibrator; includes stabilized human red blood cells and a platelet component in a preservative medium.
Eightcheck-3WP	quality control material; includes stabilized human red blood cells, fixed mammalian white blood cells and a platelet component in a preservative medium
SCS-1000	quality control material; contains stabilised human red blood cells, fixed mammalian white bloodcells, and a platelet component in a medium containing preservatives.
UX II PACK-BAC	Buffer 1.9% Cation surfactant 0.1%
UX II PACK-SED	Buffer 2.1%
UX II SEARCH -BAC	Polymethine Dye 0.01% (w / w) Ethylene glycol 99.9% (w / w)
UX II SEARCH -SED	Polymethine Dye 0.03% (w / w) Ethylene glycol 99.9% (w / w)
UX II SHEATH	Tris Buffer 0.14%
UX CLEAN -C	t-Octylphenoxypolyethoxyethanol < 1.0 % Sodium azide < 0.1 % Sodium phosphate tribasic dodecahydrate < 1.0 %
MEDITAPE II 10U	Reactive ingredients (per 100 test strips) [Glucose] Glucose oxidase: 700 I.U., Peroxidase: 175 P.U., 4-Aminoantipyrine: 14.0 mg, 1-Naphthol-3,6-disulfonic acid, disodium salt: 14 mg [Protein] Tetrabromophenol blue: 0.35 mg [Bilirubin] 2-Methyl-5-nitroaniline: 1.9 mg, Sodium nitrite: 1.0 mg [Urobilinogen] 3,3'-Dimethoxy-4,4'-biphenylbis (diazonium tetrafluoroborate): 0.16 mg [Creatinine] 2,6-Dichloro-4'-hydroxy-3',3''-dimethyl-3-sulfofuchsone-5',5''-dicarboxylic acid, trisodium salt: 0.34 mg, Palladium (II) chloride: 0.10 mg [pH] Bromocresol green: 0.07 mg, Bromoxyleneol blue: 0.72 mg [Blood] Cumene hydroperoxide: 30.0 mg, 3,3',5,5'-Tetramethylbenzidine: 15.0 mg [Ketones] Sodium nitroprusside: 12.0 mg [Nitrite] Sulfanilamide: 3.9 mg, N-1-Naphthylethylenediamine dihydrochloride: 0.3 mg [Leukocytes] 3-(N-Toluenesulfonyl-L-alanyloxy) indole: 0.69 mg, 2-Methoxy-4-(N-morpholino)benzenediazonium: 0.38 mg

MEDITAPE II 9U	Reactive ingredients (per 100 test strips) [Glucose] Glucose oxidase: 700 I.U., Peroxidase: 175 P.U., 4-Aminoantipyrine: 14.0 mg, 1-Naphthol-3,6-disufonic acid, disodium salt: 14 mg [Protein] Tetrabromophenol blue: 0.35 mg [Bilirubin] 2-Methyl-5-nitroaniline: 1.9 mg, Sodium nitrite: 1.0 mg [Urobilinogen] 3,3'-Dimethoxy-4,4'-biphenylbis (diazonium tetrafluoroborate): 0.16 mg [pH] Bromocresol green: 0.07 mg, Bromoxyleneol blue: 0.72 mg [Blood] Cumene hydroperoxide: 30.0 mg, 3,3',5,5'-Tetramethylbenzidine: 15.0 mg [Ketones] Sodium nitroprusside: 12.0 mg [Nitrite] Sulfanilamide: 3.9 mg, N-1-Naphthylethylenediamine dihydrochloride: 0.3 mg [Leukocytes] 3-(N-Toluenesulfonyl-L-alanyloxy) indole: 0.69 mg, 2-Methoxy-4-(N-morpholino)benzenediazonium: 0.38 mg	
MEDITAPE II 10K	Reactive ingredients (per 100 test strips) [Glucose] Glucose oxidase: 700 I.U., Peroxidase: 175 P.U., 4-Aminoantipyrine: 14.0 mg [Protein] Tetrabromophenol blue: 0.35 mg [Albumin] 4,5,6,7-Tetrachloro-2',4',5',7'-tetraiodofluorescein disodium salt: 0.14 mg [Bilirubin] 2-Methyl-5-nitroaniline: 1.9 mg, Sodium nitrite: 1.0 mg [Creatinine] 2,6-Dichloro-4'-hydroxy-3',3''-dimethyl-3-sulfofuchsone-5',5''-dicarboxylic acid, trisodium salt: 0.34 mg, Palladium (II) chloride: 0.10 mg [pH] Bromocresol green: 0.07 mg, Bromoxyleneol blue: 0.72 mg [Blood] Cumene hydroperoxide: 30.0 mg, 3,3',5,5'-Tetramethylbenzidine: 15.0 mg [Ketones] Sodium nitroprusside: 12.0 mg [Nitrite] Sulfanilamide: 3.9 mg, N-1-Naphthylethylenediamine dihydrochloride: 0.3 mg [Leukocytes] 3-(N-Toluenesulfonyl-L-alanyloxy) indole: 0.69 mg, 2-Methoxy-4-(N-morpholino)benzenediazonium: 0.38 mg	
UF II CONTROL	UF II CONTROL -H Control particles 0.4% (w / w) NOTE : This product contain Latex particle.	UF II CONTROL -L Control particles 0.1% (w / w) NOTE : This product contain Latex particle.
MEDITAPE CHECK 1	prepared from human urine; Chemical and biochemical substances as well as constituents of human origin are contained.	
MEDITAPE CHECK 2	prepared from human urine; Chemical and biochemical substances as well as constituents of human origin are contained.	
UF II Calibrator	Control particles 0.4% (w / w) NOTE : This product contain Latex particle.	

On behalf of Sysmex Europe GmbH



 i.A. Katharina Paucke
 Manager Regulatory Affairs

sysmex

 Sysmex Europe GmbH
 Bornbarch 1
 22848 Norderstedt

 Date: January 19th, 2016

Place: 22848 Norderstedt, Germany

„Design and specifications may be subject to changes due to further product development. Changes are confirmed by their appearance on a newer document and verification according to its date of issue.“



Certificate

Standard **ISO 9001:2015**

Certificate Registr. No. **09 100 89004**


Certificate Holder: **SYSMEX CORPORATION**
1-5-1 Wakino-hama-Kaigandori, Chuo-ku, Kobe 651-0073 Japan
including the locations according to annex

Scope: Development, design, production, sales and servicing of in-vitro diagnostic medical devices, laboratory equipment, reagents, laboratory information system and gene variants analysis set (for cancer genome profiling)

Proof has been furnished by means of an audit that the requirements of ISO 9001:2015 are met.

Validity: The certificate is valid from 2024-08-01 until 2027-07-31.
First certification 1998

2024-07-19



TÜV Rheinland Cert GmbH
Am Grauen Stein · 51105 Köln

www.tuv.com



Certificate

Quality Management System EN ISO 13485:2016

Registration No.: SX 1254782-1

Certificate Holder: SYSMEX CORPORATION
1-5-1 Wakinohama-Kaigandori,
Chuo-ku, Kobe
651-0073 Japan


Scope: Design and development, manufacture, distribution, installation and service of blood analyzer, urine analyzer, related reagents and accessories and gene variants analysis set (for cancer genome profiling)
Product categories: Analyzers and reagents for hematological test, blood coagulation test, immune serum test, biochemical test, genetic test, bacteriological test and urine test

Installation is not applicable for reagents and gene variants analysis set (for cancer genome profiling)

In accordance with EN ISO 13485:2016 Medical devices - Quality management systems - Requirements for regulatory purposes (ISO 13485:2016)

The Certification Body of TÜV Rheinland LGA Products GmbH certifies that the organization has established and applies a quality management system for medical devices. Proof has been furnished that the requirements specified in the abovementioned standard are fulfilled. The quality management system is subject to yearly surveillance.

Report No.: 150287325-301
Effective date: 2024-08-01
Expiry date: 2027-07-31
Issue date: 2024-07-04
Replaces certificate SX 1254782-1 issued 2023-08-28



Atsushi Kato
TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg · Germany

This certificate can be validated on <https://www.certipedia.com>

© TÜV, TÜEV and TÜV are registered trademarks. Utilisation and application requires prior approval.



EU Declaration of Conformity

Product identification:

Product name: CELLPACK DCL
 Model name: N/A
 REF code: CT-661-628, CU-228-496
 BUDI-DI: 4987562CELLPACKDCLX9
 Intended Purpose: See attachment

Manufacturer:

Name: SYSMEX CORPORATION
 Single Registration Number: JP-MF-000014037
 Address: 1-5-1 Wakinohama-Kaigandori, Chuo-ku, Kobe 651-0073 Japan

Authorised representative:

Name: SYSMEX EUROPE SE
 Single Registration Number: DE-AR-000022333
 Address: Bornbarch 1, 22848 Norderstedt, Germany

SYSMEX CORPORATION, as the manufacturer of the device, take sole responsibility for and hereby declare that the above mentioned device meets the provisions of the following Regulation:

- Regulation EU 2017/746 on *In vitro* Diagnostic Medical Devices
 Other Regulation(s)/Directive(s) as applicable for the device(s):

Risk class:

- A B C D

Conformity route:

Annex I+II+III according to Article 48 (10) of EU 2017/746

Common Specification:

N/A

<u>Takashi Demachi</u>	<u>Kobe Japan</u>	<u>25/05/2022</u>
Takashi Demachi	Name	Date
Executive Vice President	Function	(DD.MM.YYYY)



EU Declaration of Conformity

Product identification:

Product name:	SULFOLYSER
Model name:	N/A
REF code:	054-3351-4, 904-1131-7, AS788212, 904-1141-4
BUDI-DI:	4987562SULFOLYSERBV
Intended Purpose	See attachment

Manufacturer:

Name:	SYSMEX CORPORATION
Single Registration Number:	JP-MF-000014037
Address:	1-5-1 Wakinohama-Kaigandori, Chuo-ku, Kobe 651-0073 Japan

Authorised representative:

Name:	SYSMEX EUROPE SE
Single Registration Number:	DE-AR-000022333
Address:	Bornbarch 1, 22848 Norderstedt, Germany

SYSMEX CORPORATION, as the manufacturer of the device, take sole responsibility for and hereby declare that the above mentioned device meets the provisions of the following Regulation:

- Regulation EU 2017/746 on *In vitro* Diagnostic Medical Devices
 Other Regulation(s)/Directive(s) as applicable for the device(s):

Risk class:


- A B C D

Conformity route:

Annex I+II+III according to Article 48 (10) of EU 2017/746

Common Specification:

N/A

	<i>Kobe Japan</i>	<i>25/05/2022</i>
Takashi Demachi	Name	Date
Executive Vice President	Function	(DD.MM.YYYY)

Sysmex Corporation

1-5-1 Wakinohama-Kaigandori, Chuo-ku, Kobe 651-0073, Japan
Tel 81-78-265-0500 Fax 81-78-265-0524



www.sysmex.co.jp

EU Declaration of Conformity

Product identification:

Product name:	Lysercell WNR
Model name:	N/A
REF code:	BL-121-531, AN-577-063
BUDI-DI:	4987562LysercellWNRZ8
Intended Purpose	See attachment

Manufacturer:

Name:	SYSMEX CORPORATION
Single Registration Number:	JP-MF-000014037
Address:	1-5-1 Wakinohama-Kaigandori, Chuo-ku, Kobe 651-0073 Japan

Authorised representative:

Name:	SYSMEX EUROPE SE
Single Registration Number:	DE-AR-000022333
Address:	Bornbarch 1, 22848 Norderstedt, Germany

SYSMEX CORPORATION, as the manufacturer of the device, take sole responsibility for and hereby declare that the above mentioned device meets the provisions of the following Regulation:

- Regulation EU 2017/746 on *In vitro* Diagnostic Medical Devices
 Other Regulation(s)/Directive(s) as applicable for the device(s):

Risk class:

- A B C D

Conformity route:

Annex I+II+III according to Article 48 (10) of EU 2017/746

Common Specification:

N/A

		
Takashi Demachi	Kobe Japan	25/05/2022
Executive Vice President	Place	Date
	Function	(DD.MM.YYYY)

System Corporation



EU Declaration of Conformity

Product identification:

Product name: Lysercell WDF
 Model name: N/A
 REF code: AL-337-564, BG-689-680, AZ-124-801, AW-993-605
 BUDI-DI: 4987562LysercellWDFXJ
 Intended Purpose: See attachment

Manufacturer:

Name: SYSMEX CORPORATION
 Single Registration Number: JP-MF-000014037
 Address: 1-5-1 Wakinohama-Kaigandori, Chuo-ku, Kobe 651-0073 Japan

Authorised representative:

Name: SYSMEX EUROPE SE
 Single Registration Number: DE-AR-000022333
 Address: Bornbarch 1, 22848 Norderstedt, Germany

SYSMEX CORPORATION, as the manufacturer of the device, take sole responsibility for and hereby declare that the above mentioned device meets the provisions of the following Regulation:

- Regulation EU 2017/746 on *In vitro* Diagnostic Medical Devices
 Other Regulation(s)/Directive(s) as applicable for the device(s):

Risk class:

- A B C D

Conformity route:

Annex I+II+III according to Article 48 (10) of EU 2017/746

Common Specification:

N/A

<u>Takashi Demachi</u>	<u>Kobe Japan</u>	<u>25/05/2022</u>
Takashi Demachi	Name	Date
Executive Vice President	Function	(DD.MM.YYYY)

Sysmex Corporation



EU Declaration of Conformity

Product identification:

Product name:	Fluorocell WNR
Model name:	N/A
REF code:	BG128712, CP-066-715
BUDI-DI:	4987562FLUOROCELLWNR4
Intended Purpose	See attachment

Manufacturer:

Name:	SYSMEX CORPORATION
Single Registration Number:	JP-MF-000014037
Address:	1-5-1 Wakinohama-Kaigandori, Chuo-ku, Kobe 651-0073 Japan

Authorised representative:

Name:	SYSMEX EUROPE SE
Single Registration Number:	DE-AR-000022333
Address:	Bornbarch 1, 22848 Norderstedt, Germany

SYSMEX CORPORATION, as the manufacturer of the device, take sole responsibility for and hereby declare that the above mentioned device meets the provisions of the following Regulation:

- Regulation EU 2017/746 on *In vitro* Diagnostic Medical Devices
 Other Regulation(s)/Directive(s) as applicable for the device(s):

Risk class:



- A B C D

Conformity route:

Annex I+II+III according to Article 48 (10) of EU 2017/746

Common Specification:

N/A

		
Takashi Demachi	Kobe Japan	27/05/2022
Executive Vice President	Name	Date
	Function	(DD.MM.YYYY)



EU Declaration of Conformity

Product identification:

Product name: Fluorocell WDF
 Model name: N/A
 REF code: AE687941, BY458697, BJ284784, CV-377-552, AA-325-279
 BUDI-DI: 4987562FLUOROCELLWDFWE
 Intended Purpose: See attachment

Manufacturer:

Name: SYSMEX CORPORATION
 Single Registration Number: JP-MF-000014037
 Address: 1-5-1 Wakino-hama-Kaigandori, Chuo-ku, Kobe 651-0073 Japan

Authorised representative:

Name: SYSMEX EUROPE SE
 Single Registration Number: DE-AR-000022333
 Address: Bornbarch 1, 22848 Norderstedt, Germany

SYSMEX CORPORATION, as the manufacturer of the device, take sole responsibility for and hereby declare that the above mentioned device meets the provisions of the following Regulation:

- Regulation EU 2017/746 on *In vitro* Diagnostic Medical Devices
 Other Regulation(s)/Directive(s) as applicable for the device(s):

Risk class:


- A B C D

Conformity route:

Annex I+II+III according to Article 48 (10) of EU 2017/746

Common Specification:

N/A

	<u>Kobe Japan</u>	<u>25/05/2022</u>
Takashi Demachi	Name	Date
Executive Vice President	Function	(DD.MM.YYYY)



EU Declaration of Conformity

Product identification:

Product name:	CELLCLEAN
Model name:	N/A
REF code:	834-0162-1, BU037001
BUDI-DI:	4987562CELLCLEANP7
Intended Purpose	See attachment

Manufacturer:

Name:	SYSMEX CORPORATION
Single Registration Number:	JP-MF-000014037
Address:	1-5-1 Wakinohama-Kaigandori, Chuo-ku, Kobe 651-0073 Japan

Authorised representative:

Name:	SYSMEX EUROPE SE
Single Registration Number:	DE-AR-000022333
Address:	Bornbarch 1, 22848 Norderstedt, Germany

SYSMEX CORPORATION, as the manufacturer of the device, take sole responsibility for and hereby declare that the above mentioned device meets the provisions of the following Regulation:

- Regulation EU 2017/746 on *In vitro* Diagnostic Medical Devices
 Other Regulation(s)/Directive(s) as applicable for the device(s):

Risk class:

- A B C D

Conformity route:

Annex I+II+III according to Article 48 (10) of EU 2017/746

Common Specification:

N/A

	Kobe Japan	25/05/2022
Takashi Demachi	Name	Date
Executive Vice President	Function	(DD.MM.YYYY)

System Corporation



EC Declaration of Conformity

Application of Council Directive:

- 98/79/EC of 27 October 1998 on In Vitro Diagnostic Medical Devices.

Means of conformity:

The following products are in conformity with

- Directive 98/79/EC based on the conformity assessment procedures in accordance with Annex III.

Product identification:

Product name: XN CHECK

Classification: Other device (except Annex II and self-testing devices)

List of Applied Standards:

- Harmonised Standards used for conformity assessment are listed in the technical documentation.

Legal Manufacturer:

Name: SYSMEX CORPORATION

Address: 1-5-1 Wakinohama-Kaigandori, Chuo-ku, Kobe 651-0073 Japan

Authorised officer: Takashi Demachi Date: Apr. 28, 2022
Takashi Demachi, Executive Vice President

Authorised representative:

Name: SYSMEX EUROPE SE

Address: Bombarch 1, 22848 Norderstedt, Germany

Authorised officer: Sinem Yaman Date: 28.04.2022
Sinem Yaman, Vice President, Head of Regulatory Affairs,
Quality Assurance, Quality Control

This declaration of conformity is issued under the sole responsibility of the manufacturer and is valid until the end of transitional period stipulated in REGULATION (EU) 2017/746 & its related regulations. The Classification of this product under REGULATION (EU) 2017/746 is Class B.



EC Declaration of Conformity

Application of Council Directive:

- 98/79/EC of 27 October 1998 on In Vitro Diagnostic Medical Devices.

Means of conformity:

The following products are in conformity with

- Directive 98/79/EC based on the conformity assessment procedures in accordance with Annex III.

Product identification:

Product name: XN CAL

Classification: Other device (except Annex II and self-testing devices)

List of Applied Standards:

- Harmonised Standards used for conformity assessment are listed in the technical documentation.

Legal Manufacturer:

Name: SYSMEX CORPORATION

Address: 1-5-1 Wakinohama-Kaigandori, Chuo-ku, Kobe 651-0073 Japan

Authorised officer: Takashi Demachi Date: Apr. 28, 2022
Takashi Demachi, Executive Vice President

Authorised representative:

Name: SYSMEX EUROPE SE

Address: Bornbarch 1, 22848 Norderstedt, Germany

Authorised officer: Sinem Yaman Date: 28.04.2022
Sinem Yaman, Vice President, Head of Regulatory Affairs,
Quality Assurance, Quality Control

This declaration of conformity is issued under the sole responsibility of the manufacturer and is valid until the end of transitional period stipulated in REGULATION (EU) 2017/746 & its related regulations. The Classification of this product under REGULATION (EU) 2017/746 is Class B.

