

Certificate of Analysis

Issue date 2025-05-28
Product SVANOVIR® PRVgB-Ab
Cat. Number(s) SV-104905
Lot number F202500113
Expiry date 2027-05-31 (yyyy-mm-dd)
Storage temperature 2 °C to 8 °C

Quality Control

Plate incubation temperature: 37 ± 1 °C

Release Analysis Data-Serum

Component	OD Virus Antigen 450 nm	ODp
Negative Control	1.41	100
Positive Control	0.15	11
Weak Positive Serum	0.89	63

Release Analysis Data-Plate

Well type	C.V. Intra plate %	C.V. Inter plate %
Virus Antigen	4	3

Manufactured and tested at INDICAL BIOSCIENCE GmbH, Leipzig, Germany, and in compliance with our ISO 9001 certified quality management system. The product fulfills all INDICAL quality requirements.

Kristin Wappler, Quality Assurance

This certificate is a computer printout and therefore valid without signature.