

Declaration of Conformity

Manufacturer's Name: SHIMADZU CORPORATION
Analytical & Measuring Instruments Division
Address : 1, Nishinokyo-Kuwabara-cho, Nakagyo-ku,
Kyoto 604-8511, Japan

declares in sole responsibility that the following product

Product Name Atomic Absorption Spectrophotometer
Model Name AA-7000F, AA-7000G, AA-7000F/AAC
P/N 206-77000-58, 206-77100-58, 206-77500-58

referred to in this declaration conforms with following directives and standards

EMC Directive 2014/30/EU
EN 61326-1:2013
Low Voltage Directive 2014/35/EU
EN 61010-1:2010
EN 61010-2-061:2003
RoHS Directive 2011/65/EU

- Note 1) This declaration becomes invalid if technical or operational modifications are introduced without manufacturer's consent.
Note 2) This declaration is valid if this product is used alone or in combination with the accessories of this product which are mentioned in attached Appendix 1 or other instruments which fulfill with the requirement of mentioned directive.
Note 3) Authorised Representative in EU is as follows:
SHIMADZU EUROPA GmbH
Address :Albert-Hahn-Strasse 6-10, 47269 Duisburg, F.R. Germany

..... Kyoto, JAPAN 16 June, 2015

Place and date of issue

..... *K. Tanimizu*

Signature

..... Koji Tanimizu

Name

General Manager of Quality Assurance Department
Analytical & Measuring Instruments Division
Title



SHIMADZU





Appendix 1

Options	P/N
GFA-7000A	206-77777-58
ASC-7000	206-77600-58
ASK-7000	206-77550-41
MICROSAMPLING KIT	206-77540-91
AAC-7000	206-77701-41




SHIMADZU

改定内容一覧表

改訂記号	改訂理由	改訂内容	改訂日付	作成	検討	承認
制定	新規作成		2014-03-17	杉原		
A	規格変更	EMC指令と整合規格, 低電圧指令 を改定	2015.6.16			

検討 (他部門の了解および承認が必要な場合)

配布先

作成・制定 DATE	2014.03.17	作成 PREP	検討 CHK	承認 APP	名称 TITLE	番号 No.	SHEET 3 / 3	改定 REV
担当場所 DEPT	ZE	SPBU 杉原			DECLARATION OF CONFORMITY, AA-7000 RoHS			
 島津製作所 SHIMADZU CORPORATION		類別 CLASSIF			ZEAE-1082			
改定 REVISED								