

Certificate of Conformity

Product: Nunc SlideFlask Chamber, 9 cm²
Cell Culture Treated

Date: 2018-09-01
Page: 1

Item number: 170920

Lot number: 157913

Date of manufacture: June 2018

Date of expiry: June 2023

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

| Test Description | Requirements | Result |
|---|----------------------------|------------|
| Visual | According to specification | Complies |
| Test of function | According to specification | Complies |
| Cloning efficiency with V-79 cell line | According to specification | Complies* |
| USP <85> Bacterial endotoxins test | <0.5 endotoxins unit/ml | Complies* |
| USP <88> "Biological Reactivity test In Vivo, USP Class VI Test" | USP current edition | Complies** |
| Cell Culture Test L 929 Human Lung Primary chicken embryo | According to specification | Complies* |

*Performed as monitoring test

**Only performed once

Sterility is obtained through irradiation according to ISO 11137 "Sterilization of health care products - Radiation", with SAL 10E-6.

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2008 "Quality management system - Requirements" and ISO 13485:2003 "Medical devices - Quality management systems - System requirements for regulatory purposes".

Roskilde Site



Emil O'Hanlon Cohrt, PhD
Quality Systems, Support and Monitoring, Team Leader


