

**EKOTEKS LABORATUVAR ve GÖZETİM
HİZMETLERİ A.Ş.**

AB-0583-T

21003762-
İNG

02-21

TEST RESULT

WATER PERMEABILITY; ISO 811:2018

Hydrostatic Head Tester, Textest marka Fx 3000 model
Temperature of water 20°C. Pressure increase ratio 10 mbar/min.
Performed in the conditioned room (20±2°C-65%±4)
Less critical area

Sample 1
Sample 2
Sample 3
Sample 4
Sample 5

RESULT

25,3 cmH₂O
24,3 cmH₂O
30,8 cmH₂O
29,2 cmH₂O
25,9 cmH₂O

Average

27,1 cmH₂O

REQUIREMENT

≥ 10 cmH₂O

MICROBIAL CLEANLINESS (Bioburden)

Test Metod: Ref: EN ISO 11737-1:2018

The sample is put in extraciton liquid after shaking well, inoculated on the agar.
After incubation at 30 ± 1 ° C for 72 hours, growth microorganisms are counted on the agar.

Critical, Less Critical area

	RESULTS	REQUIREMENT
Microbial cleanliness (cfu/ 100 cm ²)	280 cfu/100 cm ²	≤300 cfu/100 cm ²

*cfu= Colony forming unit.

**ZENMED TIBBİ CİHAZLAR
VE SAĞLIK ÜRÜNLERİ SAN. VE TİC. LTD.ŞTİ.**
Mahmutbey Mah. 2567.Sk. No:1 İş Yeri No:11,21
Bağcılar / İSTANBUL / TÜRKİYE
Tel: 0212 445 45 25 Tic.Sic.No: 294704
Güneşli V.D.: 997 067 1942