

Advanced pH/ORP Meter

pH/ORP and Temperature



HI6221 is a streamlined benchtop meter with a large touch screen display, comprised of a housing and an integrated pH / ORP measurement module.

Compact and easy to operate, the benchtop meter is delivered with Hanna Instruments HI1131B double junction combination pH electrode, together with HI7662-TW temperature probe.

HI1131B is a glass body, double junction, refillable pH electrode with an indicating sensor made of High Temperature (HT) glass. The double junction reference and HT glass design allow the HI1131B to be used in a wide variety of applications including samples with metals and elevated temperatures.

Probe connection to the unit is secured through a galvanically isolated BNC connection.

HI7662-TW stainless steel temperature probe allows the meter to automatically temperature compensate (ATC) pH measurements.

This system responds to a complex range of measurement and monitoring requirements, providing accuracy, reproducibility, and reliability.

HI6221 is supplied with an electrode holder that has a flexible arm. The holder can be mounted quickly and provides secure support for electrodes while taking measurements in sample containers.

User interface

- 7-inch capacitive touch screen with multi-touch support
- Capacitive touch back, home and system menu keys
- User-friendly icons and symbols allow users to easily navigate and interpret the instrument functions.
- The user can select between five different views:
 - Basic measurement configuration
 - Simple GLP with calibration information
 - Full GLP with electrode status and calibration point details
 - Live updated, interactive graph
 - Tabulated data with date, time, and notes

Measurement

- Measure pH/mV (pH) or mV/Rel. mV (ORP) with temperature
- Application-specific profiles allow quick and direct measurement without the need to update the sensor and system settings

- Active log during measurement
- Measurement stability indicator (using the Stability Criteria setting)
- Reading modes: direct and direct/autohold
- Temperature compensation can be Automatic (using temperature probe) or set manually
- Audible and/or alarm messages for measurements outside of predefined limits
- Galvanic isolation for pH/ORP measurement

Calibration

- 5-point pH calibration with automatic recognition for standard buffers (Hanna and NIST buffers)
- Choice of standard or custom buffers for calibration
- Non-volatile memory saves data and settings

Logging

- Data log collection of at least 1,000,000 data points (with time and date stamp)
- Logging types: manual, automatic, autohold
- Sample ID for manual and Autohold data

Connectivity features & services

- Transfer logged data to a USB thumb drive
- Log files that include measurements and calibration data (as .csv file)
- FTP and email for log export via Ethernet and Wi-Fi connection
- USB type A for USB stick, keyboard, and printer
- USB type C for USB stick and PC connection

Help section for meter guidance

- Video support presentation of main functionalities



This system responds to a complex range of measurement and monitoring requirements, providing accuracy, reproducibility, and reliability.



1. Capacitive touch screen with multi-touch support

The benchtop unit has a 7-inch color display with 800 x 480p resolution. The capacitive, multi-touch screen supports video playback and data plotting.

2. Back key

3. Home key

4. System Menu key

This key will enter the system menu where User accounts, System Settings, and Logging can be configured. The Help menu is also accessed on the system menu screen.

5. Stability indicator

6. Current date

7. Current time

8. pH reading

9. mV reading

10. pH electrode icon

11. Calibration information: Electrode condition, Offset, Slope, Date and Time

12. Buffer trays

13. Temp. reading and Temp. compensation status

14. Measurement setup menu

Opens sensor setup parameters.

15. User name (default shown)

16. Direct/Autohold Readings

When Direct/Autohold is selected, measurement reading is held on display when measurement stability is reached. This option removes the subjective nature of stability as a measurement that has not reached equilibrium will not be used.

When not selected, sample measurements are displayed continuously.


17. Logging space availability

18. Logging start

19. USB connection status

20. Peripheral connection status

21. Wireless network connection status



System menu overview

Through the System Menu (☰), control user accessibility, system and connectivity configuration, access logged data and video-supported help.

- Add and delete user accounts through Users (👤).
- Access Network Connectivity, System and Info tabs through Settings (⚙️).
- Log Recall (▶️) recalls stored log sessions (Automatic continuous logging, Manual, or Autohold).
- Help (❓) guides users with video-supported help.




pH

benchtop




Users



Custom Users

New administrator or standard user accounts can be created. Standard accounts can be configured for specific accessibility.




User Account Management

Administrators can create and manage accounts from the Account Management Screen.




Settings



Network Screen


Determine how measurement logs are shared through network settings. Users can select network to be connected via Ethernet or Wi-Fi, or Disabled.



Connect and Print Screen

Activate connectivity options to allow the meter to connect to other devices.

- FTP access to meter, permits log file transfer to a FTP site and to connect the meter FTP server to a client for log download.
- Meter web server, permits log file download to a web client.
- Sending emails, permits log files to be transferred by email.



System Screen

The system screen enables users to configure options such as: Time, Date, Language, Meter ID, Decimal Separator, Backlight Saver, Audible signals, Startup Tutorial, and Factory Settings restore.



Info Screen

Displays information on meter, channel serial number, and Wi-Fi firmware version.



Log Recall


| Name | Parameter | Start/Stop | #Samples |
|--------------------------|-----------|--|----------|
| mV_20230101070237.csv | mV | 10:02:37 03/03/2022 10:03:21 03/03/2022 | 45 |
| pH_20230101070155.csv | pH | 10:01:58 03/03/2022 10:02:27 03/03/2022 | 30 |
| pH_20230101070403.csv | pH | 10:04:03 03/03/2022 10:04:32 03/03/2022 | 10 |
| relmV_20230101070334.csv | Rel. mV | 10:03:34 03/03/2022 10:03:53 03/03/2022 | 20 |

Log History and Sharing

The item allows users access and management (selection, deletion, and sharing) of measurement data. Only the user who generated the data has access to the logs created by that user.

Data can be viewed tabulated (complete with date, time, and notes), or plotted (as graph).

Log files can be shared via USB, FTP, web server and email.



Graph View

| pH | mV | T(°C) | Date | Time | Notes |
|-------|------|-------|------------|----------|-------|
| 7.044 | -2.4 | 25.0 | 03/03/2022 | 10:01:58 | "H" |
| 7.044 | -2.4 | 25.0 | 03/03/2022 | 10:01:59 | "H" |
| 7.044 | -2.4 | 25.0 | 03/03/2022 | 10:02:00 | "H" |
| 7.044 | -2.4 | 25.0 | 03/03/2022 | 10:02:01 | "H" |
| 7.044 | -2.4 | 25.0 | 03/03/2022 | 10:02:02 | "H" |
| 7.044 | -2.4 | 25.0 | 03/03/2022 | 10:02:03 | "H" |
| 7.044 | -2.4 | 25.0 | 03/03/2022 | 10:02:04 | "H" |
| 7.044 | -2.4 | 25.0 | 03/03/2022 | 10:02:05 | "H" |
| 7.044 | -2.4 | 25.0 | 03/03/2022 | 10:02:06 | "H" |
| 7.044 | -2.4 | 25.0 | 03/03/2022 | 10:02:07 | "H" |

Table View

10:07:53 01/01/2023 HI6221

pH_20230101070155.csv

GENERAL INFORMATION
 Username: HI6221
 Profile: default_pH

INSTRUMENT
 Instrument Name: HI6221-01
 Serial Number: 129456789LMN
 Firmware Version: 01.2.20825

CHANNEL INFO
 Channel Number: 1
 Module: General

Log Detail

Tapping the information icon displays log details such as user and profile name, instrument name and serial number, channel, lot information, as well as GLP data.



Help

Hanna Tutorial System

- 1. HI6221 First Look
- 2. Getting Started
 - 2.1. Get familiar with functionalities
 - 2.2. Users
 - 2.3. Meter settings
 - 2.4. Setting measurement
- 3. General Operations
- 4. General Operations
- 5. Troubleshooting guide
- 6. Accessories and Warranty

2.1. Get familiar with functionalities
 Screens explained
 Main View - This screen shows the current measurement according to the measurement settings and give access to the user calibration and measurement settings options.

Main Menu - This screen gives the user access to the 5 main icons

On-board Help

The HELP menu supports users with a brief overview of the system's main functionalities through text and video tutorials.

Hanna Tutorial System

- 1. HI6221 First Look
- 2. Getting Started
 - 2.1. Get familiar with functionalities
 - 2.2. Users
 - 2.3. Meter settings
 - 2.4. Setting measurement
- 3. General Operations
- 4. General Operations
- 5. Troubleshooting guide
- 6. Accessories and Warranty


2.4. Setting up measurement
 Your measurement screen can be configured by pressing **Q**.

READING

Mode
 pH mV Rel. mV



Measurement Setup Configuration




Calibration

Customize calibration options such as Last Calibration, Automatic, semi-automatic or manual calibration, First Calibration Point, daily or periodic Calibration Reminder, and buffer Groups.



Buffer groups

This option allows the user to select Buffers in Use for calibrating a pH electrode when using the Automatic calibration type.



Custom Buffers

Custom buffers can be created.




Reading

Customize measurement options such as Parameter, Resolution, Stability Criteria, Reading Mode



Temperature

Customize temperature options such as Automatic or manual temperature Source, °C, °F, or K temperature Unit, Manual Temperature input, Isopotential Point.



Alarm configuration

Alarm configuration allows users to set the high and low threshold limits for the measured parameters. When the parameter is enabled and the measurement exceeds the high-limit value or drops below the low-limit value, the alarm is triggered and will appear on the message banner along with an audible alarm (if Alarm Beepers is enabled).



Logging


Logging Type (Automatic, Manual or Autohold), Sampling Period (Automatic), File Name (Manual and Autohold), and Sample ID (Manual and Autohold) can be configured under this option menu.



Profiles

A profile is a sensor setup complete with required measurement unit, temperature unit, display preference, and alarm threshold options.

Once saved the profile can be loaded for applications that require similar configurations.



View Configuration

This screen allows users to select the preferred display configuration.

pH options: Basic, Simple GLP, Full GLP, Graph, Table
 mV options: Basic, Graph, Table
 Rel. mV options: Simple GLP, Basic, Graph, Table

Basic View

Basic screen displays the measured value, measurement unit as well as temperature source.




Simple GLP View

In addition to data displayed when Basic option is selected, Simple GLP screen also displays: last calibration date and time, Offset value, average slope (Avg. Slope), and electrode condition (Condition).



Full GLP View

In addition to data displayed when Simple GLP option is selected, Full GLP screen also displays: electrode symbol, used buffers trays together with calibration date, time, and temperature probe status.



Graph View

When Graph is selected, the measured value is plotted as a graph.



Table


When Table is selected, the measured values are displayed tabulated (complete with date, time, and notes made during logging). The newest data is displayed on the top of the table.

Electrode Holder

HI6221 is supplied with the HI764060 electrode holder featuring a flexible arm. The holder can be mounted on either side quickly and provides secure support for electrodes while taking measurements in sample containers.



Rear Ports



| Specifications | | HI6221 |
|--------------------------|-----------------------|--|
| pH | Range | -2.0 to 20.0 pH; -2.00 to 20.00 pH; -2.000 to 20.000 pH |
| | Resolution | 0.1 pH; 0.01 pH; 0.001 pH |
| | Accuracy | ±0.1 pH; ±0.01 pH; ±0.002 pH (±1 last significant digit) |
| mV | Range | ±2000.0 mV |
| | Resolution | 1 mV; 0.1 mV |
| | Accuracy | ±0.2 mV ±1 last significant digit |
| Temperature | Range | -20.0 to 120.0 °C; -4.0 to 248.0 °F; 253.0 to 393.0 K |
| | Resolution | 0.1 °C / 0.1 °F / 0.1 K |
| | Accuracy | ±0.2 °C / ±0.4 °F / ±0.2 K |
| Relative mV offset range | | ±2000.0 mV |
| pH Calibration | Calibration points | Up to 5 |
| | Type | Automatic; Semiautomatic; Manual |
| | Standard buffers | Hanna and NIST pH 1.68, 3.00, 4.01, 6.86, 7.01, 9.18, 10.01, 12.45 |
| | Custom buffers | Up to 5 |
| | Custom group | Up to 5 |
| | 1st calibration point | Offset or Points (user setting) |
| | Reminder | Disabled Daily: 0 min. to 23 hours and 59 min. Periodic: 1 min. to 500 days, 23 hours and 59 min. |
| Temperature Compensation | | Automatic or Manual |
| Reading | Modes | Direct; Direct/Autohold |
| | Stability criteria | Accurate; Medium; Fast |
| | Isopotential | 7.000 or 4.010 |
| | Sampling rate | 1000 ms |
| pH Views | Basic | Measurement (pH, mV, Rel.mV, Abs.mV) Temperature, Stability status |
| | Simple GLP | Basic view information Last calibration date, electrode offset, average slope, and electrode condition |
| | Full GLP | Simple GLP information and calibration point details |
| | Table | Measurements updated every second are displayed in table |
| | Graph (Plot) | pH (or mV) and temperature versus time graph can be panned or zoomed (pinch-to-zoom technology) |
| Logging | Type | Automatic, Manual, Autohold |
| | Number of records | 50 000 maximum per file Stores at least 1 000 000 data points per user |
| | Automatic interval | 1, 2, 5, 10, 30 seconds 1, 2, 5, 10, 15, 30, 60, 120, 150, 180 minutes |
| | Sample ID | Incremental mode |
| | Export option | .csv file format |
| Users | | Up to 9 users and admin. account (default) |
| Connectivity | USB-A | 2 ports for keyboard input or USB thumb drive |
| | USB-C | 1 port for PC connectivity and USB-C type thumb drive |
| | Wi-Fi & Ethernet | FTP Web server Log transfer and download Email |
| | RS232 | Connecting peripherals |
| Power supply | | DC adapter 100-240AC to 24VDC 2.5A |
| Environment | | 0 - 50 °C / 32 - 122 °F / 273 - 323 K maximum 95% RH non-condensing |
| Dimensions | | 205 x 160 x 77 mm (8.0 x 6.2 x 3.0") |
| Weight | | Approximately 1.2 kg (26.5 lbs.) |
| Ordering Information | | HI6221 is supplied with HI1131B pH electrode; HI7662-TW temperature probe; pH calibration starter kit consisting of: pH 4.01 buffer solution sachet (2 pcs.); pH 7.01 buffer solution sachet (4 pcs.); pH 10.01 buffer solution sachet (2 pcs.); HI700601 electrode cleaning solution sachet (2 pcs.); HI7082 3.5M KCl electrolyte solution (30 mL); HI764060 electrode holder; capillary pipette; 24 VDC power adapter; USB-C to USB-A cable; quick reference guide with instrument quality certificate. |