



Certificate of Analysis

22/09/2025

DONPREP®	
Product code	P50, P50B
Product size	10 and 50 columns
Batch number	NJ 254
Date of manufacture	09 September 2025
Expiry	09 September 2027
Storage recommendations	2 - 8 °C. Do not freeze. Alternatively this product can be stored at room temperature (21 - 25 °C) for 12 months from the date of manufacture.

This document is electronically created and therefore valid without a signature

QC/CERTIFICATES/TEMPLATES/V9/11.2023





Certificate of Analysis

Tested to Specification: Quality Control Product Release Inspection and Test Specification			
Recovery testing method	500 ng of DON is applied in 2 ml of water.		
Recovery specifications	≥85 % to ≤110 % of DON, according to CEN specification (CEN Method No. EN-15891).		
Recovery results (n = 6)	DON: 89 %	% RSD (n = 6)	DON: 3.6 %
Capacity testing method	2500 ng of DON is applied in 2 ml of water.		
Capacity specification	Contains antibody against DON with a capacity ≥1000 ng for DON, according to CEN specification (CEN Method No. EN-15891).		
Capacity results	Recovery ≥70 % when challenged with 2500 ng of DON. For optimum results, samples with greater amounts should be diluted.		

This document is electronically created and therefore valid without a signature

This is to certify that DONPREP® batch NJ 254 passes the required specifications and is released for sale.

R-Biopharm Rhône Ltd.
Quality Control Department



Emma Robinson
Principal Quality Scientist



QC/CERTIFICATES/TEMPLATES/V9/11.2023