



# Surgicon (Pvt) Ltd.

P.O. Box No. 244, Khadim Ali Road Sialkot-51310, Pakistan

Phone: +92 52 4262730 , Fax: +92 52 426273

Email: [info@surgicon.com.pk](mailto:info@surgicon.com.pk)

## HARDNESS TEST REPORT

Item Code: J-22-1230

Item Description: Wertheim Scissors 18cm

Material: AISI 420 J2

Hardness Value: ☐ 42 – 46

☒ 48 – 52

| SR. #                      | ACTUAL HARDNESS | HARDNESS AFTER ANNEALING |
|----------------------------|-----------------|--------------------------|
| 1                          | 50              |                          |
| 2                          |                 |                          |
| 3                          |                 |                          |
| 4                          |                 |                          |
| 5                          |                 |                          |
| Range (Samples): 48-52 HRC |                 | Mean Value: HRC          |

Approved By:

Name:

Muhammad Ijaz



Signature:

Date: 25-08-2022