



INMATEC GaseTechnologie
GmbH & Co. KG
Gewerbestraße 72
D-82211 Herrsching
Fon + 49 8152 9097 0
Fax + 49 8152 9097 10
info@inmatec.com
www.inmatec.com

INMATEC GaseTechnologie GmbH & Co. KG Gewerbestraße 72, D-82211 Herrsching

ROMEDCOM SRL

Mr. Stefan Turcanu
Str. Sciusev 16
MD 2001 Chisinau

Moldova

28-Dec-21

To whom may be concerned

On behalf of the Company INMATEC I hereby declare that for the generators in the MED range, in order to ensure the quality of oxygen according to the ISO 13485 standard, it is necessary to respect the following requirements for air :

Air quality according to ISO 8573.1 class 1
Inlet air temperature +5 ... + 50 degrees C
Guaranteed dew point + 3 degrees
Ambient temperature +5 ... + 40 degrees C

Also for the O2 generator:

IMT POC 8100 MED requires 11 Nm³ / h for 93% purity
IMT POC 8300 MED requires 16.8 Nm³ / h for 93% purity
IMT POC 8400 MED requires 26.6 Nm³ / h for 93% purity
IMT POC 8500 MED requires 37.3 Nm³ / h for 93% purity
Best regards

INMATEC GaseTechnologie GmbH & Co. KG

Dominik Fernandez

International Sales Manger



INMATEC GaseTechnologie
GmbH & Co. KG
Sitz Herrsching
HPA 79616

Geschäftsführer
Dr. Peter Bierenkopf
UST-IdNr.: DE22/388245

XEBEC Komplementär
GmbH
Sitz Herrsching
HPB 266556

Geschäftsführer
Nathalie Théberge
Marinus van Driel

Bankverbindung:
Hypovereinsbank Herrsching
IBAN DE85 7002 0270 3140 1567 19
SWIFT (BIC) HYVEDE33XXX

BTU Deutschland AG
IBAN DE94 7201 2300 0717 5957 00
SWIFT (BIC) BTVADE61XXX



Zertifiziert nach
ISO 9001



Zertifiziert nach
EN ISO 13485