

## Specificaţii tehnice (F4.1)

Numărul procedurii de achiziție: ocds-b3wdp1-MD-1601018361612 din 25.09.2020

Denumirea procedurii de achiziție: Consumabile pentru hemodializă pentru IMSP Spitalul Clinic Municipal Sfanta Treime

Cod CPV	Denumirea bunurilor	Modelul articolului	Ţara de origine	Producătorul	Specificarea tehnică deplină solicitată de către autoritatea contractantă	Specificarea tehnică deplină propusă de către ofertant	Standarde de referinţă
1	2	3	4	5	6	7	8
33100000-1	Lot 1 Dializator suprafaţa 1,5 – 1,6 m-2. Coeficientul de ultrafiltrare ≥ 10,0 ml/oră mm Hg	Hollow Fiber Dialyzer F16	China	Weigao Blood Purification Products Co.,		Hollow Fiber Dialyzer F16  1.International Certification:EC and ISO13485 - Yes, refer to the CE and ISO13485;  2.Compatibility with devices in use - Yes, refer to the page 2 of the "INSTRUCTIONS-Low Flux Series Hollow fiber dialyzer";  3.Warranty term ≥ 2 years - Yes, refer to the page 2 of the "INSTRUCTIONS-Low Flux Series Hollow fiber dialyzer";  4.Vapor sterilization, ethylene oxide, γ, β and other irradiation methods -Yes, refer to the page 1 of the "INSTRUCTIONS-Low Flux Series Hollow fiber dialyzer";  5.Fully synthetic membrane - Yes, refer to the page 1 of the "INSTRUCTIONS-Low Flux Series Hollow fiber dialyzer";  6.UF coefficient ≥ 10 ml / hour - Yes, refer to the page 1 of the "INSTRUCTIONS-Low Flux Series Hollow fiber dialyzer";  7.Filling volume ≤ 100 ml - Yes, refer to the page 1 of the "INSTRUCTIONS-Low Flux Series Hollow fiber dialyzer";  8.Specific requirements № Clearence Unit , ml / min - Yes, refer to the page 1 of the "INSTRUCTIONS-Low Flux Series Hollow fiber dialyzer";  urea ≥ 182  creatinine ≥ 168  phosphates ≥ 130  Vitamin B-12 ≥ 67	ISO;CE
33100000-1	Lot 2 Dializator suprafaţa 1,7 – 1,8 m-2. Coeficient de ultrafiltrare ≥ 12 ml/oră mm Hg	Hollow Fiber Dialyzer F18	China	Purification Products Co.,	2.Dializator suprafaţa 1,7 – 1,8 m-2. Coeficient de ultrafiltrare ≥ 12 ml/oră mm Hg	Hollow Fiber Dialyzer F18  1.International Certification:EC and ISO13485 - Yes, refer to the CE and ISO13485;  2.Compatibility with devices in use - Yes, refer to the page 2 of the "INSTRUCTIONS-Low Flux Series Hollow fiber dialyzer";  3.Warranty term ≥ 2 years - Yes, refer to the second page of the "INSTRUCTIONS-Low Flux Series Hollow fiber dialyzer";  4.Vapor sterilization, ethylene oxide, γ, β and other irradiation methods - Yes, refer to the page 1 of the "INSTRUCTIONS-Low Flux Series Hollow fiber dialyzer";  5.Fully synthetic membrane - Yes, refer to the page 1 of the "INSTRUCTIONS-Low Flux Series Hollow fiber dialyzer";  6.UF coefficient ≥ 12 ml / hour - Yes, refer to the page 1 of the "INSTRUCTIONS-Low Flux Series Hollow fiber dialyzer";  7.Filling volume ≤ 110 ml - Yes, refer to the page 1 of the "INSTRUCTIONS-Low Flux Series Hollow fiber dialyzer";  8.Specific requirements № Clearence Unit , ml / min - Yes, refer to the page 1 of the "INSTRUCTIONS-Low Flux Series Hollow fiber dialyzer";  urea ≥ 186  creatinine ≥ 173  phosphates ≥ 154  Vitamin B-12 ≥ 95 111	ISO;CE



33100000-4	Lot 3 Linii pentru sînge (magistrale), artera/vena, set maturi (pentru aparatele de dializă Fresenius 4008B și 4008S)	JRHLL B05	China	Weihai Weigao Blood Purification Products Co., Ltd	Linii pentru sînge (magistrale), artera/vena, set maturi (pentru aparatele de dializă Fresenius 4008B și 4008S)	Extracorporeal Blood Circuit for Blood Purification Equipment JRHLL B05  1.International Certification (EC or FDA or EMEA, ISO13485) - Yes, refer to the CE and ISO13485;  2.Compatibility with devices in use - Yes, refer to the page 1 of the "INSTRUCTIONS Extracorporeal Blood Circuit for Blood Purification Equipment";  3.Warranty term ≥ 2 years - Yes, refer to the page 1 of the INSTRUCTIONS Extracorporeal Blood Circuit for Blood Purification Equipment;  4.Supplying in set of vein and artery + needle for natrium chloridum connectivity - Yes, refer to the page 1 of the INSTRUCTIONS Extracorporeal Blood Circuit for Blood Purification Equipment;  5. Arterial air captivity chamber - Yes, refer to the page 1 of the "INSTRUCTIONS Extracorporeal Blood Circuit for Blood Purification Equipment";  6. Venous air captivity chamber - Yes, refer to the page 1 of the "INSTRUCTIONS Extracorporeal Blood Circuit for Blood Purification Equipment";  7. Clamps on arterial and venous segments - Yes, refer to the page 1 of the "INSTRUCTIONS Extracorporeal Blood Circuit for Blood Purification Equipment";  8. Segment for connectivity/monitoring of arterial/venous tension with double membrane with specific color marking - Yes, refer to the page 1 of the INSTRUCTIONS Extracorporeal Blood Circuit for Blood Purification Equipment:  9. Line for automatic heparinization - Yes, refer to the page 1 of the "INSTRUCTIONS Extracorporeal Blood Circuit for Blood Purification Equipment";  10. Atraumatic construction of the injection port with filling membrane - Yes, refer to the page 2 of the CATALOGUE: 5 Extracorporeal blood circuit for blood purification equipment-EN;  11. Specific requirements:  1 Pump segment (tube) diameter 8 mm - Yes, refer to the "Design of Blood Line";  2 Pump segment (tube) diameter 8 mm - Yes, refer to the page 2 of the CATALOGUE: 5 Extracorporeal blood circuit for blood purification equipment-EN;  3 Diameter of venous blood collection chamber 22 mm - Yes, refer to the "Design of Blood Line";  5 Total length of arterial lin	
33100000-	Ac pentru puncţie, artera (fistuline), 16 G	A.V. Fistula Needle Sets 1.6x25XD	China		Ac pentru puncţie, artera (fistuline), 16 G	A.V. Fistula Needle Sets 1.6x25XD  1. International Certification (EC or FDA or EMEA, ISO13485) - Yes, refer to the CE and ISO13485;  2. Compatibility with devices in use - Yes, refer to the page 2 of the "INSTRUCTIONS-A.V. Fistula Needle Sets";  3. Warranty term ≥ 3 years - Yes, refer to the page 3 of the "INSTRUCTIONS-A.V. Fistula Needle Sets";  4. Color / artery / vein color matching - Yes, refer to the page 2 of the CATALOGUE: 6 Fistula Needle Sets -RU  5. Specific requirements № Requirement Unit of measurement - Yes, refer to the page 2 of the CATALOGUE: 6 Fistula Needle Sets -RU:  1 Will have a lateral orifice (arterial)  2 Will contain integrated clamps.  3 They will contain "flyers" for 360° rotation.  4 They will have the following parameters:  -1.5 mm needle diameter  -25 mm needle length  -150 mm tube length	ISO;CE
33100000-	Ac pentru puncţie, vena (fistuline), 16G	A.V. Fistula Needle Sets 1.6x25XD	China	Weihai Weigao Blood Purification Products Co., Ltd	Ac pentru puncţie, vena (fistuline), 16G	A.V. Fistula Needle Sets 1.6x25XD  1.International Certification (EC or FDA or EMEA, ISO13485) - yes, refer to the CE and ISO13485;  2.Compatibility with devices in use - Yes, refer to the page 2 of the "INSTRUCTIONS-A.V. Fistula Needle Sets";  3.Warranty term ≥ 3 years - Yes, refer to the page 3 of the "INSTRUCTIONS-A.V. Fistula Needle Sets";  4.Color / artery / vein color matching - Yes, refer to the page 2 of the catalougue:6 Fistula Needle Sets -RU;  5.Specific requirements Ne Requirement Unit of measurement - Yes, refer to the page 2 of the catalougue:6 Fistula Needle Sets -RU  1 Will have a lateral orifice (arterial)  2 Will contain integrated clamps.  3 They will contain "flyers" for 360° rotation.  4 They will have the following parameters:  - 1.5 mm needle diameter  - 25 mm needle length  - 150 mm tube length	ISO;CE

Semnat:\_\_\_\_\_\_ Numele, Prenumele: <u>Tudosan Reghina</u> În calitate de: <u>Farmacist-diriginte</u>

Ofertantul: DITA ESTFARM SRL Adresa: mun. Chisinau, str. Burebista, 23